

Certification for Trace Elements and Methyl Mercury Mass Fractions in IAEA-452 Scallop (*Pecten maximus*) Sample



IAEA

International Atomic Energy Agency

CERTIFICATION FOR TRACE ELEMENTS
AND METHYL MERCURY MASS
FRACTIONS IN IAEA-452 SCALLOP
(*Pecten maximus*) SAMPLE

The following States are Members of the International Atomic Energy Agency:

AFGHANISTAN	GUATEMALA	PANAMA
ALBANIA	HAITI	PAPUA NEW GUINEA
ALGERIA	HOLY SEE	PARAGUAY
ANGOLA	HONDURAS	PERU
ARGENTINA	HUNGARY	PHILIPPINES
ARMENIA	ICELAND	POLAND
AUSTRALIA	INDIA	PORTUGAL
AUSTRIA	INDONESIA	QATAR
AZERBAIJAN	IRAN, ISLAMIC REPUBLIC OF	REPUBLIC OF MOLDOVA
BAHRAIN	IRAQ	ROMANIA
BANGLADESH	IRELAND	RUSSIAN FEDERATION
BELARUS	ISRAEL	RWANDA
BELGIUM	ITALY	SAUDI ARABIA
BELIZE	JAMAICA	SENEGAL
BENIN	JAPAN	SERBIA
BOLIVIA	JORDAN	SEYCHELLES
BOSNIA AND HERZEGOVINA	KAZAKHSTAN	SIERRA LEONE
BOTSWANA	KENYA	SINGAPORE
BRAZIL	KOREA, REPUBLIC OF	SLOVAKIA
BULGARIA	KUWAIT	SLOVENIA
BURKINA FASO	KYRGYZSTAN	SOUTH AFRICA
BURUNDI	LAO PEOPLE'S DEMOCRATIC REPUBLIC	SPAIN
CAMBODIA	LATVIA	SRI LANKA
CAMEROON	LEBANON	SUDAN
CANADA	LESOTHO	SWAZILAND
CENTRAL AFRICAN REPUBLIC	LIBERIA	SWEDEN
CHAD	LIBYA	SWITZERLAND
CHILE	LIECHTENSTEIN	SYRIAN ARAB REPUBLIC
CHINA	LITHUANIA	TAJIKISTAN
COLOMBIA	LUXEMBOURG	THAILAND
CONGO	MADAGASCAR	THE FORMER YUGOSLAV REPUBLIC OF MACEDONIA
COSTA RICA	MALAWI	TOGO
CÔTE D'IVOIRE	MALAYSIA	TRINIDAD AND TOBAGO
CROATIA	MALI	TUNISIA
CUBA	MALTA	TURKEY
CYPRUS	MARSHALL ISLANDS	UGANDA
CZECH REPUBLIC	MAURITANIA	UKRAINE
DEMOCRATIC REPUBLIC OF THE CONGO	MAURITIUS	UNITED ARAB EMIRATES
DENMARK	MEXICO	UNITED KINGDOM OF GREAT BRITAIN AND NORTHERN IRELAND
DOMINICA	MONACO	UNITED REPUBLIC OF TANZANIA
DOMINICAN REPUBLIC	MONGOLIA	UNITED STATES OF AMERICA
ECUADOR	MONTENEGRO	URUGUAY
EGYPT	MOROCCO	UZBEKISTAN
EL SALVADOR	MOZAMBIQUE	VENEZUELA
ERITREA	MYANMAR	VIETNAM
ESTONIA	NAMIBIA	YEMEN
ETHIOPIA	NEPAL	ZAMBIA
FIJI	NETHERLANDS	ZIMBABWE
FINLAND	NEW ZEALAND	
FRANCE	NICARAGUA	
GABON	NIGER	
GEORGIA	NIGERIA	
GERMANY	NORWAY	
GHANA	OMAN	
GREECE	PAKISTAN	
	PALAU	

The Agency's Statute was approved on 23 October 1956 by the Conference on the Statute of the IAEA held at United Nations Headquarters, New York; it entered into force on 29 July 1957. The Headquarters of the Agency are situated in Vienna. Its principal objective is "to accelerate and enlarge the contribution of atomic energy to peace, health and prosperity throughout the world".

IAEA Analytical Quality in Nuclear Applications No. IAEA/AQ/24

CERTIFICATION FOR TRACE ELEMENTS
AND METHYL MERCURY MASS
FRACTIONS IN IAEA-452 SCALLOP
(*Pecten maximus*) SAMPLE

COPYRIGHT NOTICE

All IAEA scientific and technical publications are protected by the terms of the Universal Copyright Convention as adopted in 1952 (Berne) and as revised in 1972 (Paris). The copyright has since been extended by the World Intellectual Property Organization (Geneva) to include electronic and virtual intellectual property. Permission to use whole or parts of texts contained in IAEA publications in printed or electronic form must be obtained and is usually subject to royalty agreements. Proposals for non-commercial reproductions and translations are welcomed and considered on a case-by-case basis. Enquiries should be addressed to the IAEA Publishing Section at:

Marketing and Sales Unit, Publishing Section
International Atomic Energy Agency
Vienna International Centre
PO Box 100
1400 Vienna, Austria
fax: +43 1 2600 29302
tel.: +43 1 2600 22417
email: sales.publications@iaea.org
<http://www.iaea.org/books>

For further information on this publication, please contact:

IAEA Environment Laboratories, Monaco
Marine Environmental Studies Laboratory
International Atomic Energy Agency
4a Quai Antoine 1er, MC 98000
Principality of Monaco

© IAEA, 2013
Printed by the IAEA in Austria
October 2013

CERTIFICATION FOR TRACE ELEMENTS AND METHYL MERCURY MASS FRACTIONS
IN IAEA-452 SCALLOP (*Pecten maximus*) SAMPLE

IAEA, VIENNA, 2013
IAEA/AQ/24
ISSN 2074-7659
© IAEA, 2013
Printed by the IAEA in Austria
October 2013

FOREWORD

The primary goal of the IAEA Environment Laboratories (NAEL) is to help Member States understand, monitor and protect the marine environment. The major impact exerted by large coastal cities on marine ecosystems is therefore of great concern to the IAEA, particularly to its Environment Laboratories. The marine pollution assessments needed to understand such impacts depend on accurate knowledge of contaminant concentrations in various environmental compartments. Two fundamental requirements to ensure the reliability of analytical results are quality control (QC) and quality assurance (QA). Since the early 1970s, NAEL has been assisting national laboratories and regional laboratory networks through its reference material programme for the analysis of radionuclides, trace elements and organic compounds in marine samples. Relevant activities include global interlaboratory comparison exercises and regional proficiency tests, the production of marine reference materials, and the development of reference methods for analysis of trace elements and organic pollutants in marine samples.

QA, QC and associated good laboratory practice should be essential components of all marine environmental monitoring. QC procedures are commonly based on the analysis of reference materials to assess reproducibility and measurement bias. QA can be realized by participation in externally organized laboratory performance studies, also known as interlaboratory comparison exercises, which compare and evaluate the analytical performance and measurement capabilities of participating laboratories.

The need for good QA/QC in the chemical analysis of marine environmental samples is widely recognized and has been tested in a number of international QA exercises. Such diligence also needs to be applied to other components of the monitoring exercise, since these may represent a greater source of error in many instances. Data that are not based on adequate QA/QC can be erroneous, and their misuse can lead to poor environmental management decisions. The IAEA has a long history of organizing interlaboratory studies, which have evolved to include an increasing array of potential contaminants in the marine environment.

A marine certified reference material (CRM), IAEA-452, prepared with a scallop (*Pecten maximus*) sample, was recently produced by the IAEA and certified for trace elements and methylmercury (MeHg). This species of scallop is a common, widely consumed seafood that is also used as a bioindicator for trace metal contamination in marine pollution studies.

This publication presents the sample preparation methodology, material homogeneity and stability studies, evaluation of certification campaign results, and assignment of property values and their associated uncertainty. The reference values and associated expanded uncertainty for nine trace elements (As, Cd, Cr, Cu, Fe, Hg, Mn, Pb and Zn) and MeHg in the scallop sample are established. The informative value for one more element (Ni) is also given. The new CRM can be used for the development and validation of analytical methods in the determination of trace elements and MeHg in seafood *Pecten maximus* as well as for QA/QC purposes.

The IAEA officers responsible for this publication were E. Vasileva and S. Azemard of the IAEA Environment Laboratories.

EDITORIAL NOTE

This publication (including the figures, tables and references) has undergone only the minimum copy editing considered necessary for the reader's assistance.

The views expressed do not necessarily reflect those of the IAEA, the governments of the nominating Member States or the nominating organizations.

The use of particular designations of countries or territories does not imply any judgement by the publisher, the IAEA, as to the legal status of such countries or territories, of their authorities and institutions or of the delimitation of their boundaries.

The mention of names of specific companies or products (whether or not indicated as registered) does not imply any intention to infringe proprietary rights, nor should it be construed as an endorsement or recommendation on the part of the IAEA.

The authors are responsible for having obtained the necessary permission for the IAEA to reproduce, translate or use material from sources already protected by copyrights.

The IAEA has no responsibility for the persistence or accuracy of URLs for external or third party Internet web sites referred to in this book and does not guarantee that any content on such web sites is, or will remain, accurate or appropriate..

CONTENTS

1.	INTRODUCTION.....	1
2.	METHODOLOGY	2
2.1.	Collection and preparation of the material.....	2
2.2.	Selection of laboratories for the certification campaign	2
2.3.	Homogeneity testing	3
2.4.	Stability study	4
2.5.	Characterization	4
2.6.	Moisture determination	5
3.	RESULTS AND DISCUSSION	5
3.1.	Results of the homogeneity study	5
3.2.	Results from the stability study.....	7
3.3.	Determination of certified values and uncertainties	8
3.4.	Methyl mercury (MeHg) certification.....	16
4.	METROLOGICAL TRACEABILITY AND COMMUTABILITY	17
5.	CONCLUSIONS	18
6.	ACKNOWLEDGEMENTS	18
	APPENDIX I: REFERENCE SHEET CERTIFIED REFERENCE MATERIAL.....	19
	APPENDIX II: DATA REPORT OF RESULTS SORTED BY ELEMENTS	23
	REFERENCES	27
	List of participants	29

1. INTRODUCTION

The society's growing interest in environmental issues requires the production of reliable information for policy makers, stakeholders and public in general. This information must be based on accurate and comparable results produced by qualified laboratories. National and international marine monitoring programmes have been initiated worldwide to assess the quality of the marine environment. In monitoring programme it is considered essential to ensure that the data produced from different laboratories over a number of years can be compared. If results are to be comparable, it is essential that they are based on reliable measurement standards whose values are linked to a stated reference. The difficulties inherent to traceability are generally recognised, but this task can be considerably simplified by the use of a suitable certified reference material (CRM). CRMs are key tools to establish traceability of measurement results. Indeed the best reference material should be of a matrix similar to the analysed sample and should contain comparable mass fraction of analytes of interest. Regular use of CRM by laboratories is essential also for the performance evaluation of analytical techniques and for the validation of analytical methods used in environmental monitoring programmes.

The Marine Environmental Studies Laboratory (MESL) of the International Atomic Energy Agency's Environment Laboratories (IAEA-NAEL) has the programmatic responsibility to provide assistance to its Member States' laboratories, in maintaining and improving the reliability of analytical measurement results, both in trace elements and organic pollutants. This is accomplished through the provision of CRMs of marine origin, validated analytical procedures, training in the implementation of internal quality control, and through the evaluation of measurement performance by the organization of worldwide and regional interlaboratory comparison (ILC) exercises. IAEA's analytical quality control service (AQCS), now named reference products for environment and trades represents important benchmark in upgrading the quality of laboratory performances and assessing the validity of the analytical methods used for marine monitoring studies in the Member States.

One of the outcomes from the IAEA's coordinated research project on seafood safety was the identification of the need of producing seafood matrix CRM with elevated level of toxic elements. The species selected was scallop (*Pecten maximus*), which is popular seafood in many countries and is used to assess the potential transfer of toxic elements through the food chain.

The CRM IAEA-452 will assist laboratories in validating their analytical methods and controlling the quality of produced analytical results for the determination of trace elements and methyl mercury in marine biota samples.

2. METHODOLOGY

2.1. COLLECTION AND PREPARATION OF THE MATERIAL

Two hundred kilograms of scallop (*Pecten maximus*) was collected in December 2007 and January 2008 by scuba diving in Pertuis Breton, western France. Organisms were immediately dissected. Soft tissues (the gills, mantle and digestive glands) were thawed overnight in a refrigerator, and lyophilised in a Labconco Freeze Dry System 4.5 (Labconco, Kansas City, USA). The freeze dried matter was grinded to a powder in a grinder Retsch SM 200 (Retsch, Haan, Germany). The powder was then sieved through a 250 µm sieve (Fritsch, Idar Oberstein, Germany). Biota particles retained in the sieve were collected, milled and sieved again. The sieved material with a particle size of less than 250 µm was further homogenized. The homogeneity was performed by mixing the material in a stainless steel rotating homogeniser Moritz ERM-BB124 (Moritz, Chatou, France) for 14 days in a clean atmosphere at a temperature of 20°C (+/-2) and relative humidity of 50%. After checking for the homogeneity of sample material, aliquots of about 8 g were packed into pre-cleaned brown borosilicate glass bottles with polyethylene screw caps and then sealed in plastic bags. The sample material was labelled as the IAEA-452 sample. The average moisture content of the sample after bottling was determined by drying to a constant weight at 85°C.

2.2. SELECTION OF LABORATORIES FOR THE CERTIFICATION CAMPAIGN

The selection of participants for this certification exercise was based on the measurement performances, demonstrated by laboratories in the previous IAEA's ILC. Each laboratory received one bottle of scallop (*Pecten maximus*) sample, accompanied by an information sheet and a report form. Participants were requested to analyse as many trace elements as possible, using a validated analytical method. They were asked to report the measurement results (six replicates and average value) along with the expanded uncertainty in addition to the information about the applied quality control procedure. The second request was to report results for the trace elements in CRM with similar to the candidate reference material matrix. Moisture determination method was prescribed.

Sixteen of the 23 invited laboratories sent the requested information/measurement results (value and expanded uncertainty), description of the method used, results from the analysis of CRM and results for moisture determination.

The reference value for MeHg was obtained from the results of the worldwide ILC run on the same matrix in 2009 [1].

List of laboratories participating in the certification exercise is presented in Table 1.

TABLE 1. LIST OF LABORATORIES PARTICIPATING IN THE IAEA-452 CERTIFICATION

LABORATORIES	COUNTRY
International Atomic Energy Agency, IAEA	
Brooks Rand Labs, BRL	USA
University of Plymouth, UP	UK
Finnish Customs Laboratory, FCL	FINLAND
Centro de Energia Nuclear na Agricultura, CENA	BRAZIL
Comisión Chilena de Energía Nuclear, CCEN	CHILE
Institute for Environmental Protection and Research, ISPRA	ITALY
Laboratory of Environmental Chemistry, NKUA	GREECE
Flett Research Ltd., FRLtd	CANADA
Ukrainian Scientific Center of the Ecology of the Sea, UkSCES	UKRAINE
Trent University, TU	CANADA
Laboratory of Public Health, Department of Health, LPH	MACAO
National Agency for New Technologies, Energy and Sustainable Economic Development, ENEA	ITALY
Institute of Nuclear Chemistry and Technology, INCT	POLAND
Israel Oceanographic & Limnological Research, IO-LR	ISRAEL
Institute of Public Health, IPH	CROATIA

2.3. HOMOGENEITY TESTING

Extensive homogeneity tests were carried out on this material in order to ensure its suitability as a candidate reference material and to estimate the uncertainty associated with homogeneity of the sample. Between bottle homogeneity was tested by the determination of the mass fraction of some typical elements (As, Cd, Cr, Cu, Fe, Hg, Mn, Ni, Pb, Zn and MeHg). In total, 10 bottles were selected using random stratified sampling of the whole batch. Care was taken to ensure that the order of measurements did not correspond to the filling sequence of the bottles, which enable the differentiation between potential trend in the filling sequence and analytical drift. Three subsamples from each bottle were analysed for their total element

mass fractions. The within bottle homogeneity was assessed by 15 replicate determinations of the content of investigated trace elements in one bottle. Subsamples of 0.2 g were mineralised with 5 ml conc. HNO₃ and digested in a microwave oven by adding 2 ml conc. HF according to the protocol described earlier [1]. The final measurements were performed by graphite furnace atomic absorption spectrometry under repeatability conditions and in a randomised way in order to be able to separate a potential analytical drift from a trend in the filling sequence. The determination of the total mercury was done in solid subsamples with solid mercury analyser. The method used for homogeneity study of MeHg was based on alkaline digestion and aqueous phase ethylation followed by gas chromatography separation, thermal desorption of Hg species and atomic fluorescence spectrometry (AFS) detection. All methods used for homogeneity studies were previously validated by MESL, IAEA.

2.4. STABILITY STUDY

Three sets of five bottles each were stored in the dark at different temperature, -20°C, +20°C and +60°C, just after bottling process and kept at described conditions over a period of two years. One isochronous study over 8 weeks was applied to evaluate stability of the materials during transport and one isochronous study over 12 months and 24 months, respectively, to evaluate stability during storage. Obtained results were compared with the results from samples kept at -20°C during this period (-20°C is considered the reference temperature). The stability investigation for the evaluation of long term stability is still ongoing.

2.5. CHARACTERIZATION

Characterization refers to the process of determining the reference values. The candidate reference material was initially analysed by MESL. The final characterization was based on the results delivered by selected laboratories with demonstrated measurement capabilities, based on criteria that comprised both technical and quality management aspects. Characterization of the trace element mass fraction in the IAEA-452 scallop (*Pecten maximus*) sample was based on the application of different analytical techniques as summarised in Figure 1. All participating laboratories have used validated methods for determination of trace elements in marine samples. In addition they provided results from the analysed CRM with similar matrix composition and the information on standard calibration solutions, used for every trace metal and MeHg. The results of laboratories that did not report any quality assurance information were excluded from the further evaluation.




FIG. 1. Analytical methods used for the certification of trace elements in the IAEA-452 biota sample.

Combined uncertainties were calculated in compliance with the Guide to the Expression of Uncertainty in Measurement (GUM) [3], including uncertainties due to possible heterogeneity and instability.

All participating laboratories claimed traceability of provided results to the International System of Units (SI) via standard calibration solutions and CRM applied as a part of their analytical procedures.

2.6. MOISTURE DETERMINATION

The determination of the moisture content of the samples is to some extent operationally defined. In view of the comparability of results, the protocol for the correction of the moisture was developed at NAEL and prescribed to other participants. The drying procedure at 85°C (+/-2) was established after experimental evaluation of sample stability. Correction for dry mass was obtained from separate portions of the material of minimum mass of 0.5 g (10 sub samples from 5 bottles). The weighing and repeated drying were performed until constant mass was attained. Moisture, determined at 85°C was found to be 6.5% ± 0.4% for bottles kept at 20°C.

3. RESULTS AND DISCUSSION

3.1. RESULTS OF THE HOMOGENEITY STUDY

For the homogeneity study, 10 samples (about 2% of the total batch) of the IAEA-452 were chosen using a random stratified sample picking scheme and analysed for their trace elements and MeHg contents in triplicate. The results were combined and evaluated to detect any trends regarding filling or analysis sequence and to estimate the uncertainty contribution from

possible heterogeneity. Grubbs tests were performed to identify potentially outlying individual results as well as outlying bottles means. Two individual results for Ni, Cr and Pb respectively were detected as outlier. These results were excluded as they were outliers at 95% but also at 99% confidence level.

The retained individual results and bottle means were checked whether they follow a normal distribution or are unimodally distributed. The series of results for investigated trace elements and MeHg were normally distributed. The distribution of Cr, Ni and Pb were skewed, but unimodal. One way analysis of variance ANOVA [2] was then applied to assess the between bottles and within bottles homogeneities. The ANOVA allowed the calculation of within unit standard deviation S_{wb} and also between bottles standard deviation S_{bb} :

$$s_{wb} = u_{wb} = \sqrt{MS_{wb}} \quad (1)$$

$$s_{bb} = u_{bb} = \sqrt{\frac{MS_{bb} - MS_{wb}}{n}} \quad (2)$$

For several elements (As, Fe, Zn, MeHg) MS_{bb} (ANOVA mean square between bottles) was smaller than MS_{wb} (ANOVA mean square within bottles) and S_{bb} could not be calculated. Instead u^*_{bb} , the heterogeneity that can be hidden by the method repeatability was calculated, as described by Linsinger et al. [4]:

$$u^*_{bb} = \frac{s_{wb}}{\sqrt{n}} 4 \sqrt{\frac{2}{v_{MS_{wb}}}} \quad (3)$$

where:

n is number of replicate subsamples per bottle;

$v_{MS_{wb}}$ is the degrees of freedom of MS_{wb} .

Heterogeneity could be quantified due to the good repeatability of the method used. The between bottles variations/heterogeneity were between 2.0 and 6.1%, small enough to ensure the homogeneity of the material. The uncertainty contributions due to the inhomogeneity were estimated according to ISO Guide 35 [2] as the maximum values obtained with Equation 2 or Equation 3. The results for sample size 0.2 g are presented in Table 2.

TABLE 2. THE ESTIMATE OF INHOMOGENEITY CONTRIBUTIONS TO THE TOTAL UNCERTAINTY FOR THE CERTIFIED TRACE ELEMENTS

Element	As	Cd	Cr	Cu	Fe	Hg	Mn	Ni	Pb	Zn	MeHg
$u_{bb}\%$		2.3	6.1	1.2		1.2	1.8	4.5	5.9		
$u^*_{bb}\%$	2.2				1.6					0.9	3.2

The conclusion from the presented results for the tested trace elements was that the homogeneity of the candidate reference material complied with the provisions given by the ISO 35 at the range of weights used. A minimum sample intake of 0.2 g was set, based on the smallest sample intake used in the characterization study.

3.2. RESULTS FROM THE STABILITY STUDY

Samples selected for stability study were analyzed and each of the elements was evaluated individually. No outliers were detected on 95% confidence level in any study. The evaluation of data was further carried out by performing a linear regression on the determined mass fractions vs. time.

The test material showed no significant trend to degradation over the time frame at different temperature -20°C , $+20^{\circ}\text{C}$ and $+60^{\circ}\text{C}$. Significant impact of storage conditions on the stability of the certified properties could be detected, neither of storage time nor of temperature (up to $+60^{\circ}\text{C}$). In all cases the slope of the linear regression did not significantly differ from zero. No significant slope at 95% level of confidence was detected for any of investigated analytes, neither in the short term study nor in the long term study. As no degradation could be observed under any conditions neither, in the short term nor in the long term study, it was concluded that no special precautions regarding temperature control during shipment were necessary. The uncertainty of the short term stability (u_{stis}) was assumed to be negligible since no degradation is expected to happen during this short time. Nevertheless, 20°C was chosen as the storage temperature.

Failure to detect degradation does, however, not prove stability. The uncertainty of stability u_{stab} describes the potential degradation which still can be reconciled with the data, even if the slope is not statistically significantly different from zero. Although under these conditions, an expansion of the total uncertainty of the certified values is generally not encouraged, but in this case the approach of ISO Guide 35 was followed, mainly due to the lack of sound alternatives. An uncertainty contribution related with stability of the candidate reference material was estimated as uncertainty of the regression line with a slope of 0 multiplied with the chosen shelf life, as described by Linsinger et al. [5]. Factor of 3 was selected, taking into account a minimum shelf life of 3 years. Stability during storage period was chosen as 1%,

which ensured the validity of the certificate for at least 3 years. Results for MeHg mass fractions in the IAEA-452 scallop (*Pecten maximus*) sample found in bottles kept at different storage conditions over 3 years period are presented in Table 3. Each result is obtained as a mean value for MeHg mass fractions for three different bottles (three replicates from each one).

TABLE 3. MASS FRACTIONS OF MeHg IN THE IAEA-452 SCALLOP (*Pecten maximus*) SAMPLE FOUND IN BOTTLES KEPT FOR DIFFERENT TIMES IN DIFFERENT STORAGE CONDITIONS

Temperature	-20°C	20°C	60°C
1 week	0.022±0.03	0.023±0.04	0.021±0.04
1 year	0.021±0.02	0.020±0.03	0.021±0.04
2 years	0.022±0.03	0.022±0.03	0.020±0.03
3 years	0.022±0.03	0.023±0.02	0.021±0.04

Obtained results from short term and long term studies provide evidence to a good stability of all analytes under consideration, including MeHg.

3.3. DETERMINATION OF CERTIFIED VALUES AND UNCERTAINTIES

The characterization campaign resulted in 7 to 14 results per element and 16 results for MeHg. The obtained data were first checked for compliance with the certification requirements and then for their validity based on technical reasons. All accepted set of results were submitted to the following statistical tests: Grubbs test to detect single and double outliers, Dixon's test to detect outlying lab means and Kolmogorov–Smirnov test for normal distribution.

Four outlying values (two for Zn, one for Cr and one for Cu) were detected in the received data sets. These results were outliers at 95%, but not at 99% confidence level. The outlying values were scrutinised but no technical reason for exclusion was found, therefore the results were accepted for further evaluation. The normal distribution at 95% level of confidence was found for the sample means of all studied elements except for Zn. At 99% level of confidence Zn means were also normally distributed. All laboratories reported combined standard and expanded uncertainties. As more important contributors to the combined uncertainty they stated recovery, within lab reproducibility and uncertainty of the reference value of the CRM applied.

The IAEA-452 reference values were calculated as a median of the accepted dataset, rounded off to the most significant number of the uncertainty, as described in ISO 13528 [5]. The robust mean and unweighted mean of the means were also calculated and compared with the

respective medians. As no differences were observed, the reference values obtained with the median approach were further used. These values are considered to be the most reliable estimates of the property values.

The uncertainties associated with the reference values were calculated according to the ISO standard 35 [2]. The relative combined uncertainty of the certified value of the CRM consists of uncertainty related to characterization u_{char} , between bottle heterogeneity (u_{bb}) and long term stability (u_{stab}). These different contributions were combined to estimate the expanded, relative uncertainty.

$$U^2_{CRM,rel} = 2*(u^2_{char} + u^2_{hom} + u^2_{stab}) \quad (4)$$

where

k : coverage factor equalling 2, representing a level of confidence of about 95%;

u_{hom} was estimated as a larger value of the standard deviation between bottles (u_{bb}) or the maximum heterogeneity potentially hidden by the method repeatability (u_{bb}^*) as previously shown in Table 2;

u_{stab} was the stability during storage period chosen as 1%, which as described before, ensured the validity of the certificate for at least 3 years;

u_{char} was estimated using an approach described by Pauwels, J., et al. [6]. In this approach, the uncertainty of characterization is separated into laboratory dependent uncertainty $u(I)$, uncertainty common to all laboratories $u(II)$ and uncertainty common to a groups of laboratories $u(III)$.

$$u(I) = \frac{\sqrt{\sum_1^n u_i^2}}{n} \quad (5)$$

where

u_i is the combined standard uncertainty for the mean value given by each laboratory;


n is the number of laboratories. For this study $u(II)$ and $u(III)$ were set to zero.

As it can be seen from the Figure 1, methods with different quantification steps (AAS, GF-AAS, AFS, ICP-OES, ICP-MS — see Table 4 for list of instrumental techniques) as well as methods without sample preparation step (NAA) were used for characterization of the material. The agreement between results confirms the absence of any significant method bias and demonstrates the identity of the analyte.

TABLE 4. INSTRUMENTAL TECHNIQUES


Method code	Instrumental technique
AAS	Atomic absorption spectrometry
ICP-MS	Inductively coupled plasma mass spectrometry
ID-MS	Isotope dilution ICP-MS
ICP-AES	Inductively coupled plasma atomic emission spectrometry
AFS	Atomic fluorescence spectrometry
CV	Cold vapour
GF	Graphite furnace

Provided by participants, results for trace elements and MeHg mass fractions grouped by methods are displayed in Figures 2 to 11. Detailed results as they were reported by participants are shown in Appendix II. In all figures the reported results are plotted versus the reference values, which are denoted by a bold line, while the dashed lines represent the expanded uncertainty ($k=2$) associated with the reference value. Error bars represent expanded uncertainty as reported by participants.




The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 2. Laboratory results for arsenic mass fraction (mg kg^{-1}).




The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 3. Laboratory results for cadmium (Cd) mass fraction (mg kg^{-1}).




The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 4. Laboratory results for chromium mass fraction (mg kg^{-1}).




The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 5. Laboratory results for copper mass fraction (mg kg^{-1}).




The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 6. Laboratory results for iron mass fraction (mg kg^{-1}).




The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 7. Laboratory results for mercury mass fraction (mg kg^{-1}).




The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG.8. Laboratory results for manganese mass fraction (mg kg^{-1}).




The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 9. Laboratory results for lead (*Pb*) mass fraction (mg kg^{-1}).



The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 10. Laboratory results for zinc mass fraction (mg kg^{-1}).



The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 11. Laboratory results for methyl mercury mass fraction (mg kg^{-1}).

A good agreement within the stated uncertainty was observed for results obtained with different methods. Therefore, all of them were considered in the deriving of reference values.


The reference values for As, Cd, Cr, Cu, Fe, Hg, Mn, Pb, Zn and MeHg are presented in Table 5, together with their expanded uncertainty ($k=2$). Due to the limited number of provided results, the information value was assigned to Ni – see Figure 12.

TABLE 5. REFERENCE VALUES FOR TRACE ELEMENTS AND MeHg MASS FRACTIONS AND THEIR EXPANDED UNCERTAINTY ($k=2$) IN THE IAEA-452 BIOTA SAMPLE

Element	Reference value ¹ mg kg^{-1}	Expanded uncertainty ² mg kg^{-1}	Number of accepted datasets
As	17.9	1.1	11
Cd	30.5	1.6	15
Cr	5.2	0.7	8
Cu	11.0	0.8	11
Fe	1068	47	9
Hg	0.16	0.02	10
Mn	279	14	7
Pb	2.4	0.3	11
Zn	167	10	13
MeHg	0.022	0.004	16

¹The value is the median of accepted sets of data, each set being obtained by different laboratory. The certified values are reported on dry mass basis and are traceable to the SI.

²Expanded uncertainty with a coverage factor $k=2$ according to the guide to the Expression of Uncertainty of Measurement (GUM), corresponding to the level of confidence of about 95%.



The median (solid line) and corresponding expanded uncertainty ($k=2$) (dashed line) are shown. The error bars correspond to the expanded uncertainty reported by each laboratory.

FIG. 12. Laboratory results for nickel mass fraction (mg kg^{-1}).

3.4. METHYL MERCURY (MeHg) CERTIFICATION

In total, 16 laboratories sent results for MeHg in the frame of the IAEA-452 ILC [1]. The sixteen results for MeHg were obtained using a variety of methodologies as described in Table 6. In the first step — release of MeHg from the binding sites, three techniques were mainly used: distillation, alkaline digestion and acid leaching. Further steps included separation of organic and inorganic mercury by ion exchange, solvent extraction or gas chromatography after appropriate derivatisation. The following methods were then applied for the detection of MeHg: atomic fluorescence spectrometry, gas chromatography with electron capture detector, inductively coupled plasma mass spectrometry and cold vapour atomic absorption spectrometry. Two laboratories used species specific isotope dilution inductively coupled plasma mass spectrometry (ID-ICP-MS).

ID-ICP-MS method is considered as potentially primary method of measurement and gives the results with highest metrological quality. As it can be seen in Figure 11, the results within uncertainty obtained with other methods are in good agreement with those obtained after application of species specific ID-ICP-MS. Therefore the reference value for MeHg in the IAEA-452 sample was calculated taking into account the results provided by all sixteen laboratories participating in the IAEA's ILC study.

TABLE 6. METHODS FOR DETERMINATION OF MeHg, USED BY THE LABORATORIES PARTICIPATING IN THE IAEA-452 WORLDWIDE ILC

Sample preparation	Detection	Number of Laboratories
Alkaline digestion aqueous phase ethylation (NaBEt ₄), GC separation, thermal desorption of Hg species at 600°C	AFS	7
Alkaline digestion, solvent extraction, derivatisation (NaBEt ₄), GC separation	ICP-MS	1
Alkaline digestion, solvent extraction, derivatisation (NaBEt ₄), GC separation	ID-ICP-MS	1
Distillation — aqueous phase ethylation (NaBEt ₄), GC separation	ID-ICP-MS	1
Alkaline digestion, separation by HPLC	AFS	1
Dithizone–toluene extraction, GC separation	GC-ECD	1
Distillation, BrCl oxidation	CV-AAS	1
Toluene extraction, L-cysteine back extraction	CV-AFS	1
Leaching 10N HCl	CV-AAS	1
H ₂ SO ₄ /CuSO ₄ /KBr leaching followed by toluene extraction, back extraction in aqueous thiosulfate solution	Solid AAS	1

Note: AFS: atomic fluorescence spectrometry; CV-AAS: cold vapour atomic absorption spectrometry; Solid AAS: solid Mercury analyser; ICP-AES: inductively coupled plasma atomic emission spectrometry; CV-AFS: cold vapour atomic fluorescence spectrometry; ICP-MS: inductively coupled plasma mass spectrometry; ID-ICP-MS: isotope dilution inductively coupled plasma mass spectrometry, GC-ECD: gas chromatography with electron capture detector.

4. METROLOGICAL TRACEABILITY AND COMMUTABILITY

If results obtained by different laboratories are to be comparable it is essential that all results are based on reliable measurement standards whose values are linked to a stated reference.

Only validated methods applied within stated scope were used by participating in the certification exercise laboratories. Matrix CRMs with stated SI traceability purchased from NIST, EC JRC IRMM, NRC-CNRC were used for validation of the applied in this study methods [1].

Pure metal standard solutions (CRM) with stated purity were employed for calibration from all laboratories participating in the certification campaign. As stated in the respective certificates of all CRM producers the mass fractions of the trace element and MeHg in the respective standard solutions were measured against another CRM (i.e. NIST, BAM or EMPA) with demonstrated SI traceability followed by gravimetric preparation using balances

calibrated with SI-traceable weights. Consequently the value calculated by this unbroken chain of comparison is traceable to the reference to which the starting material is compared.

In addition, the agreement between the results confirms absence of any significant method bias and demonstrates the identity of the analytes.

Commutable CRMs should exhibit a similar analytical behaviour for given method as a real laboratory sample. Characterization study has been selected such as to provide variety of analytical methods, regarding sample preparation, calibration and detection. The good agreement between the results obtained indicates commutability of the material. However CRMs might show behaviour different from that of real samples, in particular during digestion, due to their small particle size in contrast to the possible larger particle size for real laboratory samples. The commutability was also confirmed in the frame of the IAEA worldwide ILC for trace elements with the same sample material, in which 143 laboratories took part [1].

5. CONCLUSIONS

This exercise allows assignment of reference values for As, Cd, Cr, Cu Fe, Hg, Mn, Pb, Zn and MeHg with associated uncertainties following ISO guidelines. The certified values are derived from measurement results provided by the laboratories participating in the certification campaign. Only validated methods were applied in the certification of the IAEA-452 scallop (*Pecten maximus*) sample. As the certified values are the combination of SI traceable individual results, they are themselves traceable to SI. The obtained results are in excellent agreement with the results from the IAEA-452 worldwide ILC exercise.

The produced matrix CRM is suitable for QC purposes of environment and food laboratories, and has a high level Cd as requested by experts in food safety. As any certified reference material, it can be used for validation studies.

6. ACKNOWLEDGEMENTS

We highly acknowledge the participants and laboratories that took part in this interlaboratory comparison exercise and contributed their time and facilities to the present work.

The International Atomic Energy Agency is grateful to the Government of the Principality of Monaco for the support provided to its Environment Laboratories.

REFERENCE SHEET
CERTIFIED REFERENCE MATERIAL



International Atomic Energy Agency
Department of Nuclear Sciences and Applications
IAEA Environment Laboratories
Vienna International Centre, P.O. Box 100, 1400 Vienna, Austria

REFERENCE SHEET

CERTIFIED REFERENCE MATERIAL

IAEA-452

MASS FRACTIONS OF TRACE ELEMENTS AND METHYL MERCURY IN BIOTA SCALLOP (*Pecten maximus*)

Analyte	Reference value ¹ [mg kg ⁻¹]	Expanded uncertainty ² [mg kg ⁻¹]
As	17.9	1.1
Cd	30.5	1.6
Cr	5.2	0.7
Cu	11.0	0.8
Fe	1068	47.0
Hg	0.16	0.02
Mn	279	14.0
Pb	2.4	0.3
Zn	167	10.0
MeHg	0.022	0.004

⁽¹⁾ Median of the accepted data sets, each set being obtained by a different laboratory and/or a different method of determination.

⁽²⁾ Estimated expanded uncertainty with a coverage factor $k=2$, corresponding to a level of confidence of approximately 95%, as defined in the Evaluation of measurement data — Guide to the expression of uncertainty in measurement JCGM100:2008 [7].

Origin and preparation of the material

200 kg of Scallop (*Pecten maximus*) were collected in December 2007 and January 2008 by scuba diving in the “Pertuis Breton”, western France. Organisms were dissected and soft tissues (the gills, mantle and digestive glands) lyophilized. The freeze dried matter was milled to a powder in a grinder, then sieved through a 250 µm sieve. The sieved material with a particle size of less than 250 µm was further homogenized.

Homogeneity was achieved by mixing the material in a stainless steel rotating homogenizer for 14 days in a clean atmosphere at a temperature of (20 +/-2)° C and relative humidity of 50 %. After checking for the homogeneity of sample material, aliquots of about 8 g were packed into pre-cleaned brown borosilicate glass bottles and then sealed in plastic bags. The sample material was labelled IAEA-452.

Homogeneity of the material, dry mass determination and recommended minimum test portion

Homogeneity of the material

Between bottle homogeneity was tested by the determination of the mass fraction of all certified elements (As, Cd, Cr, Cu, Fe, Hg, Mn, Pb, Zn and MeHg). In total, 10 bottles were selected using random stratified sampling of the whole batch. Three subsamples from each bottle were analysed for their total element mass fractions. The within bottle homogeneity was assessed by 15 replicate determinations of mass fraction of trace elements in one bottle. The final measurements were performed by Graphite Furnace Atomic Absorption Spectrometry after microwave acid digestion. The determination of the total mercury was done in solid subsamples with solid mercury analyser. The method used for homogeneity study of MeHg was based on alkaline digestion and aqueous phase ethylation, followed by gas chromatography separation, thermal desorption of Hg species and Atomic Fluorescence Spectrometry detection. All analytical procedures used for homogeneity studies were previously validated in the IAEA’s Marine Environmental Studies Laboratory (MESL), Monaco.

The analysis of homogeneity study was performed under repeatability conditions and in a randomized way in order to separate a potential analytical drift from a trend in the filling sequence and minimize variations. The homogeneity test results provided experimental evidence that satisfactory levels of between and within bottles homogeneity were attained and the uncertainty due to between and within bottles heterogeneity was within acceptable limits.

Dry mass determination

The average moisture content of the material was determined by drying several test portions of 1 g in an oven at 85±2°C until constant weight and was found to be approximately 6.5±0.8% (k=2). Since the moisture content can vary with ambient humidity and temperature, it is recommended that it should be checked prior to analysis and that all results should be corrected on a dry mass basis.

Recommended minimum test portion

Based on results obtained for within and between bottles homogeneity, the minimum test portion was set to be 0.2 g.

Characterization study

Twenty laboratories from Austria, Brasil, Canada, Chile, Croatia, Finland, Greece, Italy, Israel, Macao, Poland, UK, USA, including the IAEA Monaco Laboratory, participated in the characterization study.

Laboratories were requested to determine as many trace elements as possible by their analytical method of choice. The following methods were used: Atomic absorption spectrometry (flame and graphite furnace), inductively coupled optical emission spectrometry, inductively coupled mass spectrometry, neutron activation analysis and atomic fluorescence spectrometry.

Assignment of values — certification procedure

The assigned property values were established on the basis of results reported by participating laboratories to the IAEA Environment Laboratories in Monaco. The median concentrations for the sets of individual data were chosen as the best estimate of the property values [8]. Certified values are assigned when at least 7 laboratory means were available, and at least 4 different analytical methods applied.

The details concerning all reported results as well as the criteria for certification may be found in "Certification for trace elements and methyl mercury mass fractions in IAEA-452 scallop (*Pecten maximus*) sample"[8]. The paper is available free of charge upon request.

Based on the evidence on calibrators used, quality control procedures applied by the participating laboratories and their generally high quality performance in the IAEA interlaboratory comparisons, the Certification Committee decided to accept these assigned values as certified.

Statement on metrological traceability and uncertainty of assigned values

Expanded uncertainties with a coverage factor of $k=2$, corresponding to a level of confidence of about 95%, were calculated according to the JCGM100:2008 Evaluation of measurement data — guide to the expression of uncertainty in measurement [7].

The quantity values assigned to the IAEA-452 certified reference material are mass fractions of specified trace elements, expressed in the derived SI unit mg kg^{-1} . Evidence on metrological traceability to the SI Units of reference materials and calibrators used in the characterization process was provided by all laboratories in their reports. More details can be found in the reference [8].

Intended use

This certified reference material (CRM) is intended to be used in the development, validation and quality control of analytical methods for the determination of trace elements in marine biota samples. Based on the above metrological traceability statement, this material is not to be used as calibrator.

Instructions for use

The reference material is supplied in 8 g units. The minimum sample volume (mass) laboratories should take when using the IAEA-452 is 0.2 g.

Handling and storage

The original unopened bottle should be stored securely at ambient temperature in a dark and dry place. Analysts are reminded to take appropriate precaution in order to avoid contamination of the material during handling.

Issue and expiry date

The issue date of this reference material is **March 2012**. The expiry date is **March 2017**. The IAEA is monitoring the long term stability of the material and customers will be informed in case of any observed change.

Legal disclaimer

The IAEA makes no warranties, expressed or implied, with respect to the data contained in this reference sheet and shall not be liable for any damage that may result from the use of such data.

Compliance with ISO Guide 31:2000

The content of this this IAEA Reference Sheet is in compliance with the ISO Guide 31:2000: Reference materials — content of certificates and labels [9].

Citation of this reference sheet

It is suggested to cite this reference sheet according to the following example, as appropriate to the citation format used: INTERNATIONAL ATOMIC ENERGY AGENCY, Reference Sheet for CRM IAEA-453, 'Trace elements and methyl mercury in biota'. IAEA, Vienna, 4 pp. (The latest version published applies; see "Note" below).

Note

Certified values as stated in this reference sheet may be updated if more information becomes available. Users of this material should ensure that the reference sheet in their possession is current. The current version can be found in the IAEA's Reference Materials online catalogue:

<http://nucleus.iaea.org/rpst/ReferenceProducts/ReferenceMaterials>

Further information:

For further information regarding this material, please contact:

Head, Marine Environmental Studies Laboratory
International Atomic Energy Agency
Environment Laboratories
4, Quai Antoine 1er
MC 98000 MONACO

Tel.: 377 97 97 72 72

Fax: 377 97 97 72 76

Email: mel@iaea.org

Appendix II

DATA REPORT OF RESULTS SORTED BY ELEMENTS

TABLE II.1. ARSENIC: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
8	18.04	0.37	0.74
9	17.9	0.86	1.7
18	20.8	0.98	2
26	17.2	2.7	5.3
54	17.7	0.36	0.73
65	19	0.8	1.5
72	17.9	0.86	2.29
74	15.6	0.34	0.68
89	16.2	1.135	2.27
94	19.5	1.1	2.2
121	18.1	0.39	0.78

TABLE II.2. CADMIUM: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
8	30.5	0.62	1.24
9	30.2	1.15	2.3
18	29.6	0.79	1.6
26	32	5.3	10.6
54	35.4	1	2
60	26.4	1.6	3.2
65	36.2	1.7	3.4
67	23.6	0.124	0.247
72	30	0.5	1.34
74	30.6	0.53	1.06
89	25.28	0.765	1.53
94	36	2.2	4.4
121	32.1	0.63	1.27
126	28.3	0.5	1
127	32.2	0.8	1.6

TABLE II.3. CHROMIUM: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
8	4.89	0.11	0.22
9	3.69	0.333	0.67
18	4.7	0.24	0.48
60	5.48	0.56	1.12
67	5.44	0.164	0.329
72	4.12	0.11	0.29
125	5.49	0.24	0.5
127	6.21	0.68	1.37

TABLE II.4. COPPER: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
8	9.6	0.21	0.42
9	11	0.632	1.26
18	12.8	0.46	0.91
60	9.04	0.19	0.37
67	9.25	0.07	0.141
72	9.29	0.11	0.29
74	11.1	0.26	0.52
94	11.9	0.55	1.1
121	11.37	0.18	0.35
127	11.43	0.34	0.68

TABLE II.5. IRON: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
8	1068	22	44
9	945	38	76
18	970	26	51
54	1058	16	32
60	1102	75	150
72	1033	31	83
74	1169	37	74
94	1100	34.8	70
125	1099	12	24

TABLE II.6. MERCURY: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
9	0.156	0.0077	0.015
26	0.167	0.013	0.025
60	0.132	0.021	0.042
70	0.165	0.0097	0.0195
72	0.159	0.003	0.007
74	0.156	0.001	0.002
89	0.1593	0.012	0.024
90	0.144	0.009	0.018
94	0.15	0.013	0.03

TABLE II.7. MANGANESE: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
9	274	24.4	49
18	283	5.9	12
60	283.8	6.5	13.1
67	266	0.04	0.09
72	273	5	14
74	279	6.6	1.32
121	283.76	9.07	18.14

TABLE II.8. LEAD: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
8	2.43	0.05	0.1
9	2.34	0.15	0.3
26	2.44	0.37	0.73
65	2.57	0.09	0.18
67	3.04	0.084	0.169
72	2.31	0.1	0.26
74	2.17	0.12	0.24
89	2.242	0.091	0.182
94	2.7	0.095	0.2
121	2.55	0.07	0.14
127	2.25	0.11	0.22

TABLE II.9. ZINC: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
8	156.2	3.4	6.8
9	165	7.5	15
18	167	5.1	10.2
54	167.8	2.2	4.4
60	170.1	3.9	7.8
65	198	9	18
67	166	0.077	0.155
72	165	3	7
74	166	3.8	7.6
94	176	5.2	11
121	165.8	3.42	6.84
125	170.7	3.1	6.2
127	176.7	4.5	9

TABLE II.10. NICKEL: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
8	2.46	0.06	0.12
9	2.25	0.135	0.27
72	2.5	0.05	0.44

TABLE II.11. METHYL MERCURY: RESULTS AS REPORTED BY PARTICIPANTS (in mg kg⁻¹)

Lab code	Mean	Combined uncertainty (u)	Expanded uncertainty (U)
1	0.0216	0.0009	0.0018
6	0.018	0.0017	0.0034
7	0.023	0.0022	0.0043
9	0.0203	0.0002	0.0004
24	0.020	0.0012	0.0024
37	0.0238	0.0008	0.0016
40	0.0234	0.0004	0.0008
43	0.0186	0.0015	0.0030
69	0.0219	0.0002	0.0005
70	0.0230	0.0016	0.0033
90	0.0221	0.0002	0.0004
101	0.0222	0.0008	0.0016
118	0.022	0.0009	0.0018
123	0.0216	0.0006	0.0012
134	0.0208	0.0007	0.0014
138	0.027	0.0010	0.0020

REFERENCES

- [1] INTERNATIONAL ATOMIC ENERGY AGENCY, Worldwide Laboratory Comparison on the Determination of Trace Elements in IAEA-452 Biota Sample, IAEA/AQ/23, IAEA, Vienna (2012).
- [2] INTERNATIONAL ORGANISATION FOR STANDARDISATION, Guide 35:2006 (E), Reference Materials — General and Statistical Principles for Certification, ISO, Geneva, Switzerland.
- [3] INTERNATIONAL ORGANISATION FOR STANDARDISATION, Guide to the Expression of Uncertainty of Measurements: (1995) ISO, Geneva, Switzerland.
- [4] LINSINGER, T., et al., Accreditation and quality assurance, **6** (2005) pp. 20–25.
- [5] INTERNATIONAL ORGANISATION FOR STANDARDISATION, Guide 13528:2005 (E), Statistics Methods for Use in Proficiency Testing by Inter-Laboratory Comparisons, ISO, Geneva, Switzerland.
- [6] PAUWELS, J., LAMBERTY, A., SCHIMMEL, H., Accreditation and quality assurance, **3** (1998) pp. 180–181.
- [7] JOINT COMMITTEE FOR GUIDES IN METROLOGY (JCGM), Evaluation of measurement data — Guide to the Expression of Uncertainty in Measurement, JCGM 100:2008 (GUM with minor corrections) (2008), http://www.bipm.org/utils/common/documents/jcgm/JCGM_100_2008_E.pdf
- [8] VASILEVA, E., AZEMARD, S., OH, J., BUSTAMANTE, P., BETTI, M., Certification for Trace Elements and Methyl Mercury Mass Fractions in IAEA-452 Scallop (*Pecten maximus*) Sample, Accreditation and Quality Assurance. **16** (2011) 439–447.
- [9] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION (ISO), Reference materials — content of certificates and labels, ISO Guide 31: 2000, ISO, Geneva (2000).

LIST OF PARTICIPANTS

AUSTRIA

Törvènyi, A. IAEA
A-2444 Seibersdorf

BRAZIL

De Nadai Fernandes, E. A. Laboratorio de Radioisotopos
Centro de Energia Nuclear na Agricultura
(CENA/USP)
Ave.Centenário 303, Caixa Postal 96
13400-970 Piracicaba - SP

CANADA

Flett, R. Flett Research Ltd
440 Desalaberry Ave
Winnipeg Manitoba R2L 0Y7

Hilderman, E. Flett Research Ltd
440 Desalaberry Ave
Winnipeg Manitoba R2L 0Y7

Hintelman, H. Department of Chemistry
1600 West Bank Drive
Peterbourg Ontario K9J 7B8

Wei, X. Flett Research Ltd
440 Desalaberry Ave
Winnipeg Manitoba R2L 0Y7

CHILE

Andonie, O. Comisión Chilena de Energía Nuclear
Nueva Bilbao 12501, Las Condes
7600713 Las Condes

CROATIA

Kras, M. Institute of Public Health
Kresimirova 52A
51000 Rijeka

Zivkovic, A. Institute of Public Health
Kresimirova 52A
51000 Rijeka

FINLAND

Niemi, E. Finnish Customs Lab.
Tekniikantie 13
FI-02150 Espoo

GREECE

Dassenakis, E. Laboratory of Environmental Chemistry
Dept. Chemistry NKUA
Panepistimiopolis Zogafrou
15771 Athens

Scoullou, M. Laboratory of Environmental Chemistry
Dept. Chemistry NKUA
Panepistimiopolis Zogafrou
15771 Athens

ISRAEL

Herut, B. Israel Oceanographic & Limnological Research
Tel Shikmona
P.O. Box 8030
31080 Haifa

Kress, N. Israel Oceanographic & Limnological Research
Tel Shikmona
P.O. Box 8030
31080 Haifa

Shefer, E. Israel Oceanographic & Limnological Research
Tel Shikmona
P.O. Box 8030
31080 Haifa

ITALY

Belli, M. Environmental Metrology Service Head
ISPRA
Via di Castel Romano, 100
00128 Roma

Cremisini, C. ENEA
via Anguillarese 301
S. Maria Di Galeria
00123 Roma

Pinto, V. ENEA
via Anguillarese 301
S. Maria Di Galeria
00123 Roma

MACAO

Sun, W. Laboratory of Public Health
Dept of Health
Estrada dos Parses,
Edif Laboratorio de Saude Publica
Macao

MONACO

Vasileva, E. IAEA
4 Quai Antoine 1er
98000 Monaco

POLAND

Chajank, E. Institute of Nuclear Chemistry and Technology
ul. Dorodna 16
03-195 Warsaw

Danko, B. Institute of Nuclear Chemistry and Technology
ul. Dorodna 16
03-195 Warsaw

Polkowska-Motrenko, H. Institute of Nuclear Chemistry and Technology
ul. Dorodna 16
03-195 Warsaw

Samczyski, Z. Institute of Nuclear Chemistry and Technology
ul. Dorodna 16
03-195 Warsaw

UNITED KINGDOM

Fisher, A. School of Earth, Ocean and Environmental Sciences
University of Plymouth
Drake Circus
Plymouth, Devon PL4 8AA

UNITED STATES OF AMERICA

Mc Farland, F. Brooks Rand Labs
3958 6th Ave. N.W. (U.S.A)
Seattle, Washington 98107

UKRAINE

Denga, Y.

Ukranian Scientific Center of the Ecology of the Sea
89 Frantsuzki Blvd
65009 Odessa

Hapchenko, L.

Ukranian Scientific Center of the Ecology of the Sea
89 Frantsuzki Blvd
65009 Odessa



IAEA

International Atomic Energy Agency

No. 23

ORDERING LOCALLY

In the following countries, IAEA priced publications may be purchased from the sources listed below, or from major local booksellers.

Orders for unpriced publications should be made directly to the IAEA. The contact details are given at the end of this list.

AUSTRALIA

DA Information Services

648 Whitehorse Road, Mitcham, VIC 3132, AUSTRALIA

Telephone: +61 3 9210 7777 • Fax: +61 3 9210 7788

Email: books@dadirect.com.au • Web site: <http://www.dadirect.com.au>

BELGIUM

Jean de Lannoy

Avenue du Roi 202, 1190 Brussels, BELGIUM

Telephone: +32 2 5384 308 • Fax: +61 2 5380 841

Email: jean.de.lannoy@euronet.be • Web site: <http://www.jean-de-lannoy.be>

CANADA

Renouf Publishing Co. Ltd.

Telephone: +1 613 745 2665 • Fax: +1 643 745 7660

5369 Canotek Road, Ottawa, ON K1J 9J3, CANADA

Email: order@renoufbooks.com • Web site: <http://www.renoufbooks.com>

Bernan Associates

4501 Forbes Blvd., Suite 200, Lanham, MD 20706-4391, USA

Telephone: +1 800 865 3457 • Fax: +1 800 865 3450

Email: orders@bernan.com • Web site: <http://www.bernan.com>

CZECH REPUBLIC

Suweco CZ, spol. S.r.o.

Klecakova 347, 180 21 Prague 9, CZECH REPUBLIC

Telephone: +420 242 459 202 • Fax: +420 242 459 203

Email: nakup@suweco.cz • Web site: <http://www.suweco.cz>

FINLAND

Akateeminen Kirjakauppa

PO Box 128 (Keskuskatu 1), 00101 Helsinki, FINLAND

Telephone: +358 9 121 41 • Fax: +358 9 121 4450

Email: akatilaus@akateeminen.com • Web site: <http://www.akateeminen.com>

FRANCE

Form-Edit

5, rue Janssen, PO Box 25, 75921 Paris CEDEX, FRANCE

Telephone: +33 1 42 01 49 49 • Fax: +33 1 42 01 90 90

Email: fabien.boucard@formedit.fr • Web site: <http://www.formedit.fr>

Lavoisier SAS

14, rue de Provigny, 94236 Cachan CEDEX, FRANCE

Telephone: +33 1 47 40 67 00 • Fax: +33 1 47 40 67 02

Email: livres@lavoisier.fr • Web site: <http://www.lavoisier.fr>

L'Appel du livre

99, rue de Charonne, 75011 Paris, FRANCE

Telephone: +33 1 43 07 50 80 • Fax: +33 1 43 07 50 80

Email: livres@appeldulivre.fr • Web site: <http://www.appeldulivre.fr>

GERMANY

Goethe Buchhandlung Teubig GmbH

Schweitzer Fachinformationen

Willstaetterstrasse 15, 40549 Duesseldorf, GERMANY

Telephone: +49 (0) 211 49 8740 • Fax: +49 (0) 211 49

Email: s.dehaan@schweitzer-online.de • Web site: <http://www.goethebuch.de/>

HUNGARY

Librotade Ltd., Book Import

PF 126, 1656 Budapest, HUNGARY

Telephone: +36 1 257 7777 • Fax: +36 1 257 7472

Email: books@librotade.hu • Web site: <http://www.librotade.hu>

INDIA

Allied Publishers

1st Floor, Dubash House, 15, J.N. Heredi Marg
Ballard Estate, Mumbai 400001, INDIA
Telephone: +91 22 2261 7926/27 • Fax: +91 22 2261 7928
Email: alliedpl@vsnl.com • Web site: <http://www.alliedpublishers.com>

Bookwell

3/79 Nirankari, Dehli 110009, INDIA
Tel.: +91 11 2760 1283 • +91 11 27604536
Email: bkwell@nde.vsnl.net.in • Web site: <http://www.bookwellindia.com/>

ITALY

Libreria Scientifica "AEIOU"

Via Vincenzo Maria Coronelli 6, 20146 Milan, ITALY
Tel.: +39 02 48 95 45 52 • Fax: +39 02 48 95 45 48
Email: info@libreriaaeiou.eu • Web site: <http://www.libreriaaeiou.eu/>

JAPAN

Maruzen Co., Ltd.

1-9-18 Kaigan, Minato-ku, Tokyo 105-0022, JAPAN
Tel.: +81 3 6367 6047 • Fax: +81 3 6367 6160
Email: journal@maruzen.co.jp • Web site: <http://maruzen.co.jp>

NETHERLANDS

Martinus Nijhoff International

Koraalrood 50, Postbus 1853, 2700 CZ Zoetermeer, NETHERLANDS
Tel.: +31 793 684 400 • Fax: +31 793 615 698
Email: info@nijhoff.nl • Web site: <http://www.nijhoff.nl>

Swets

PO Box 26, 2300 AA Leiden
Dellaertweg 9b, 2316 WZ Leiden, NETHERLANDS
Telephone: +31 88 4679 263 • Fax: +31 88 4679 388
Email: tbeydens@nl.swets.com • Web site: www.swets.com

SLOVENIA

Cankarjeva Založba dd

Kopitarjeva 2, 1515 Ljubljana, SLOVENIA
Tel.: +386 1 432 31 44 • Fax: +386 1 230 14 35
Email: import.books@cankarjeva-z.si • Web site: http://www.mladinska.com/cankarjeva_zalozba

SPAIN

Díaz de Santos, S.A.

Librerías Bookshop • Departamento de pedidos
Calle Albasanz 2, esquina Hermanos Garcia Noblejas 21, 28037 Madrid, SPAIN
Telephone: +34 917 43 48 90
Email: compras@diazdesantos.es • Web site: <http://www.diazdesantos.es/>

UNITED KINGDOM

The Stationery Office Ltd. (TSO)

PO Box 29, Norwich, Norfolk, NR3 1PD, UNITED KINGDOM
Telephone: +44 870 600 5552
Email (orders): books.orders@tso.co.uk • (enquiries): book.enquiries@tso.co.uk • Web site: <http://www.tso.co.uk>

On-line orders:

DELTA International Ltd.

39, Alexandra Road, Addlestone, Surrey, KT15 2PQ, UNITED KINGDOM
Email: info@profbooks.com • Web site: <http://www.profbooks.com>

United Nations (UN)

300 East 42nd Street, IN-919J, New York, NY 1001, USA
Telephone: +1 212 963 8302 • Fax: 1 212 963 3489
Email: publications@un.org • Web site: <http://www.unp.un.org>

UNITED STATES OF AMERICA

Bernan Associates

4501 Forbes Blvd., Suite 200, Lanham, MD 20706-4391, USA
Tel.: +1 800 865 3457 • Fax: +1 800 865 3450
Email: orders@bernan.com • Web site: <http://www.bernan.com>

Renouf Publishing Co. Ltd.

812 Proctor Avenue, Ogdensburg, NY 13669, USA
Tel.: +800 551 7470 (toll free) • +800 568 8546 (toll free)
Email: orders@renoufbooks.com • Web site: <http://www.renoufbooks.com>

Orders for both priced and unpriced publications may be addressed directly to:

IAEA Publishing Section, Marketing and Sales Unit, International Atomic Energy Agency
Vienna International Centre, PO Box 100, 1400 Vienna, Austria
Telephone: +43 1 2600 22529 or 22488 • Fax: +43 1 2600 29302
Email: sales.publications@iaea.org • Web site: <http://www.iaea.org/books>

INTERNATIONAL ATOMIC ENERGY AGENCY
VIENNA
ISSN 2074-7659