

IAEA Analytical Quality in Nuclear Applications Series No. 7

Worldwide Open Proficiency Test IAEA-CU-2007-09/A: Determination of Po-210 in Water



IAEA

International Atomic Energy Agency

Worldwide Open Proficiency Test
IAEA-CU-2007-09/A
Determination of Po-210 in Water

The following States are Members of the International Atomic Energy Agency:

AFGHANISTAN	GHANA	NIGERIA
ALBANIA	GREECE	NORWAY
ALGERIA	GUATEMALA	OMAN
ANGOLA	HAITI	PAKISTAN
ARGENTINA	HOLY SEE	PALAU
ARMENIA	HONDURAS	PANAMA
AUSTRALIA	HUNGARY	PARAGUAY
AUSTRIA	ICELAND	PERU
AZERBAIJAN	INDIA	PHILIPPINES
BAHRAIN	INDONESIA	POLAND
BANGLADESH	IRAN, ISLAMIC REPUBLIC OF	PORTUGAL
BELARUS	IRAQ	QATAR
BELGIUM	IRELAND	REPUBLIC OF MOLDOVA
BELIZE	ISRAEL	ROMANIA
BENIN	ITALY	RUSSIAN FEDERATION
BOLIVIA	JAMAICA	SAUDI ARABIA
BOSNIA AND HERZEGOVINA	JAPAN	SENEGAL
BOTSWANA	JORDAN	SERBIA
BRAZIL	KAZAKHSTAN	SEYCHELLES
BULGARIA	KENYA	SIERRA LEONE
BURKINA FASO	KOREA, REPUBLIC OF	SINGAPORE
BURUNDI	KUWAIT	SLOVAKIA
CAMEROON	KYRGYZSTAN	SLOVENIA
CANADA	LATVIA	SOUTH AFRICA
CENTRAL AFRICAN REPUBLIC	LEBANON	SPAIN
CHAD	LESOTHO	SRI LANKA
CHILE	LIBERIA	SUDAN
CHINA	LIBYAN ARAB JAMAHIRIYA	SWEDEN
COLOMBIA	LIECHTENSTEIN	SWITZERLAND
CONGO	LITHUANIA	SYRIAN ARAB REPUBLIC
COSTA RICA	LUXEMBOURG	TAJIKISTAN
CÔTE D'IVOIRE	MADAGASCAR	THAILAND
CROATIA	MALAWI	THE FORMER YUGOSLAV REPUBLIC OF MACEDONIA
CUBA	MALAYSIA	TUNISIA
CYPRUS	MALI	TURKEY
CZECH REPUBLIC	MALTA	UGANDA
DEMOCRATIC REPUBLIC OF THE CONGO	MARSHALL ISLANDS	UKRAINE
DENMARK	MAURITANIA	UNITED ARAB EMIRATES
DOMINICAN REPUBLIC	MAURITIUS	UNITED KINGDOM OF GREAT BRITAIN AND NORTHERN IRELAND
ECUADOR	MEXICO	UNITED REPUBLIC OF TANZANIA
EGYPT	MONACO	UNITED STATES OF AMERICA
EL SALVADOR	MONGOLIA	URUGUAY
ERITREA	MONTENEGRO	UZBEKISTAN
ESTONIA	MOROCCO	VENEZUELA
ETHIOPIA	MOZAMBIQUE	VIETNAM
FINLAND	MYANMAR	YEMEN
FRANCE	NAMIBIA	ZAMBIA
GABON	NEPAL	ZIMBABWE
GEORGIA	NETHERLANDS	
GERMANY	NEW ZEALAND	
	NICARAGUA	
	NIGER	

The Agency's Statute was approved on 23 October 1956 by the Conference on the Statute of the IAEA held at United Nations Headquarters, New York; it entered into force on 29 July 1957. The Headquarters of the Agency are situated in Vienna. Its principal objective is "to accelerate and enlarge the contribution of atomic energy to peace, health and prosperity throughout the world".

IAEA Analytical Quality in Nuclear Applications No. IAEA/AQ/7

Worldwide Open Proficiency Test IAEA-CU-2007-09/A: Determination of Po-210 in Water

IAEA-CU-2007-09/A

COPYRIGHT NOTICE

All IAEA scientific and technical publications are protected by the terms of the Universal Copyright Convention as adopted in 1952 (Berne) and as revised in 1972 (Paris). The copyright has since been extended by the World Intellectual Property Organization (Geneva) to include electronic and virtual intellectual property. Permission to use whole or parts of texts contained in IAEA publications in printed or electronic form must be obtained and is usually subject to royalty agreements. Proposals for non-commercial reproductions and translations are welcomed and considered on a case-by-case basis. Enquiries should be addressed to the IAEA Publishing Section at:

Sales and Promotion, Publishing Section
International Atomic Energy Agency
Vienna International Centre
PO Box 100
1400 Vienna, Austria
fax: +43 1 2600 29302
tel.: +43 1 2600 22417
email: sales.publications@iaea.org
<http://www.iaea.org/books>

For further information on this publication, please contact:

Chemistry Unit, Agency's Laboratories, Seibersdorf
International Atomic Energy Agency
2444 Seibersdorf
Austria

WORLDWIDE OPEN PROFICIENCY TEST IAEA-CU-2007-09/A
DETERMINATION OF Po-210 IN WATER

IAEA-CU-2007-09/A

IAEA, VIENNA, 2009

IAEA/AQ/7

ISSN 2074-7659

© IAEA, 2009

Printed by the IAEA in Austria

November 2009

FOREWORD

The IAEA helps the Member States' laboratories to maintain their readiness by coordination activities, by development of standardized methods for sample collection and analysis, and by conducting interlaboratory comparisons and proficiency tests as a tool for external quality control.

The Chemistry Unit of the Physics, Chemistry and Instrumentation Laboratory in the International Atomic Energy Agency's Seibersdorf Laboratories in Austria, has the programmatic responsibility to support global radionuclide measurement systems. To fulfil this obligation and ensure a reliable worldwide, rapid and consistent response, the Chemistry Unit organises interlaboratory studies and proficiency tests.

The Po-210 poisoning event which occurred in November 2006 brought into focus a number of issues, including the capacity of laboratories to rapidly and accurately determine this radionuclide in environmental samples. A number of requests were received from Member States to address this issue. Responding to these requests, the Chemistry Unit of the Physics, Chemistry and Instrumentation Laboratory in the Agency's Laboratories, conducted a world wide proficiency test on the determination of Po-210 in water. The aim was to gather information on the current state of practice for Po-210 measurements at various levels in aqueous samples. This report describes the methodology employed and the results obtained in this proficiency test.

The IAEA officer responsible for this publication is A. Shakhashiro of the Agency's Laboratories, Seibersdorf.

EDITORIAL NOTE

The use of particular designations of countries or territories does not imply any judgement by the publisher, the IAEA, as to the legal status of such countries or territories, of their authorities and institutions or of the delimitation of their boundaries.

The mention of names of specific companies or products (whether or not indicated as registered) does not imply any intention to infringe proprietary rights, nor should it be construed as an endorsement or recommendation on the part of the IAEA.

CONTENTS

1. INTRODUCTION	1
2. MATERIALS AND METHODS.....	2
2.1. Proficiency test objectives	2
2.2. Participants.....	2
2.3. Composition of the proficiency test materials	2
2.4. Preparation of the spiked samples.....	2
2.5. Homogeneity testing	5
3. PERFORMANCE CRITERIA.....	6
3.1. Relative bias	6
3.2. PT evaluation criteria	6
3.2.1. Trueness.....	7
3.2.2. Precision	7
3.2.3. Blank evaluation.....	7
3.3. The z-score value	8
3.4. The u-score value	8
4. RESULTS AND DISCUSSION	9
4.1. General	9
4.2. Technical information provided by the participants	9
4.3. False positive reporting	9
4.4. Measurement repeatability	30
4.5. Recommendations to the laboratories	30
5. CONCLUSIONS.....	30
APPENDIX I. DATA EVALUATION TABLES SORTED BY SAMPLE CODE	33
APPENDIX II. DATA EVALUATION TABLES SORTED BY LABORATORY CODE.....	57
APPENDIX III. LIST OF PARTICIPATING LABORATORIES	115
REFERENCES.....	131
CONTRIBUTERS TO DRAFTING AND REVIEW	133

1. INTRODUCTION

The Po-210 poisoning event which occurred in November 2006 [1] demonstrated the need for rapid and accurate determination of this radionuclide in environmental samples. Responding to many requests from Member States, it was decided to conduct a world wide proficiency test on the determination of Po-210 in water. The aim of this proficiency test was to gather information on the current state of practice for Po-210 measurements at various levels in aqueous samples and to assist the participating laboratories in improving the quality of the analytical results.

In the proficiency test described in this report, 635 test water samples were prepared and distributed to the participating laboratories during the last week of March 2007. Laboratories were sent five water samples containing known activities of Po-210 and were requested to return the results within one week of receipt of the samples.

The participating laboratories were requested to analyse the samples employing the methods used in their routine work, so that their performance on the test samples could be directly related to the real performance of the rapid reporting time.

114 laboratories from 127 initially registered laboratories, reported their results to the IAEA. The analytical results of the participating laboratories were compared with the reference values assigned to the reference materials, and a rating system was applied. The list of participants is reported in Appendix III.

The participants and laboratories who responded to this proficiency test and contributed their efforts to the present work are highly appreciated and acknowledged.




Fig.1. Summary evaluation of 570 measurement results of Po-210.

Although the matrix was pure acidified water and the activity concentrations were relatively high, 19% of the reported results failed to pass the proficiency test criteria. In few cases positive results were reported for the blank sample which suggests a possibility of false positive reporting.

The result of the overall summary evaluation of this proficiency test showed that 86 laboratories reported results which fit the purpose of rapid detection of Po-210 in water.

Figure 1 reports the summary of the analytical data evaluation of this proficiency test. 68% of all reported results were 'acceptable'.

2. MATERIALS AND METHODS

2.1. Proficiency test objectives

Rapid measurement of spiked water, with an unknown (to the participants) amount of Po-210 was aimed at:

- checking the preparedness of Member States laboratories for rapid determination of Po-210 in liquid matrix,
- evaluating the probability of reporting false positive and false negative
- evaluating the repeatability of the reported results, and
- encouraging the participating laboratories to implement remedial actions where shortcomings in analytical performance are detected.

2.2. Participants

In this proficiency test 127 laboratories from 56 countries all over the world were registered on-line. 114 participants from 46 countries reported their results back to the IAEA via the designated on-line application. The list of participants can be found in Appendix III. The participating laboratories geographical distribution is shown in Figures 2 and 3.

2.3. Composition of the proficiency test materials

The set of the proficiency test materials consisted of 5 samples each 50 mL. The following proficiency test design was applied:

- two spiked demineralised water samples (sample codes 01, 03) ~50 g each containing ~2.5 Bq Po-210,
- two spiked demineralised water samples (sample codes 02, 04) ~50 g each containing ~5 Bq Po-210,
- one blank demineralised water (sample code 05). This is the same water which was used as raw material to spike the test materials,

Figure 4 shows a set of the packed PT samples.

Table 1 lists the target values and the associated combined standard uncertainty of the PT materials and the PT performance criteria LAP and MAB (see Section 3).

2.4. Preparation of the spiked samples

The spiked water samples were gravimetrically prepared in two batches: one batch for samples 01 and 03 and one batch for samples 02 and 04. To prepare each batch 20 kg of acidified demineralised water was spiked with a certified single Po-210 solution traceable to the international standard of radioactivity. Then a pump with multiple outlets was used to homogenise the bulk water sample in a 50 L tank. The first batch was divided in two samples: 01 and 03, the second batch in samples 02 and 04.

Sample 05 was prepared from the same bulk water used in preparation of the spiked samples 01 to 04. This sample (blank) was used to check for the false positive reporting.




Fig.2. Participants' distribution by country.




Fig. 3. Geographical distribution of the participants.



Fig. 4. The PT materials set.

TABLE 1. TARGET VALUES AND THE ASSOCIATED COMBINED STANDARD UNCERTAINTIES OF THE PT MATERIALS AND THE PT PERFORMANCE CRITERIA (SECTION 3)

Sample code	Samples 01 and 03	Samples 02 and 04	Sample 05
Activity [Bq.kg ⁻¹]	52.8±1.4	101.6±2.8	<0.1
LAP [%]	15	15	-
MAB [%]	20	20	-

The final target activity concentration for Po-210 was calculated from the certified activity value assigned to the certified standard solution of Po-210, taking into account the successive dilution steps, the mass of spiking mixture and the amount of water being spiked as determined from weighing. The combined standard uncertainty includes two major components: uncertainty of the certified solution and weighing uncertainty. The initial activity concentration of the standard solution was 377±10 Bq.g⁻¹.

The reference date for results reporting was set to the 1st of April 2007.

2.5. Homogeneity testing

Three bottles from each batch were measured using liquid scintillation counter in the Agency's Seibersdorf Laboratories to verify the homogeneity and stability of the PT materials. The three bottles were stored at ambient temperature and measured four times in the period from 19 March to 7 May 2007. Measurement results are presented in Figure 5.

The variations of the obtained measurement results are comparable to the method reproducibility and therefore it can be concluded that there was not any significant uncertainty arising from between bottles heterogeneity or material instability.




Fig. 5. Homogeneity and stability test results, four sets of measurements, and one set every two weeks. Reference date: 1st of April 2007.

3. PERFORMANCE CRITERIA

Currently most laboratories produce test results accompanied, at best, with an indication of their repeatability only and provide no indication of their analytical uncertainty. However, testing laboratories intending to follow international best practice will need to quantify and report their measurement uncertainty. In particular, this is a requirement under international standard ISO/IEC 17025:2005.

Several rating systems have been developed for determining a laboratory's performance and the meaning of the results of the different scoring systems are not always comparable. Among various statistics, z-scores and u-scores are most often used. The drawback of z-scores is that the uncertainty of the participant's measurement result is not taken into account in the evaluation of performance. In the case of u-scores, the evaluation includes uncertainties of the participant measurements and the uncertainty of the assigned value. Laboratories performing well in classical proficiency testing (z-scores) will not necessarily exhibit the same level of performance when their analytical uncertainties are considered in the evaluation.

The proficiency testing scoring system applied by the Chemistry Unit in the Agency's laboratories takes into consideration the trueness and the precision of the reported data and it includes in the evaluation both the combined standard uncertainty associated with the target value of proficiency testing samples and the combined standard uncertainty reported by the participating laboratories. According to the newly adopted approach, the reported results are evaluated against the acceptance criteria for accuracy and precision and assigned the status 'acceptable' or 'not acceptable' accordingly. A result must pass both criteria to be assigned the final status of 'acceptable'. The advantage of this approach is that it checks the credibility of the uncertainty statement given by the participating laboratories. Results are no longer compared against fixed criteria but participants establish their individual acceptance range on the basis of the uncertainties assigned to the values. Such an approach highlights not only methodological problems affecting the accuracy of the reported data but also identifies shortcomings in uncertainty estimation.

In addition, three other statistical parameters namely: relative bias, z-score and IAEA/Laboratory result ratio are calculated as complementary information for the participating laboratories.

3.1. Relative bias

The first stage in producing a score for a result $Value_{Analyst}$ (a single measurement of analyte concentration in a test material) is obtaining the estimate of the bias. To evaluate the bias of the reported results, the relative bias between the Analyst's value and the IAEA value is calculated and expressed as a percentage:

$$Relative\ bias = \frac{Value_{Analyst} - Value_{IAEA}}{Value_{IAEA}} \times 100\%$$

3.2. PT evaluation criteria

The proficiency test results were evaluated against the acceptance criteria for trueness and precision and assigned the status 'acceptable', 'Warning' or 'not acceptable' accordingly [3].

3.2.1. *Trueness*

The participant result is assigned ‘acceptable’ status for trueness if:

$$A1 \leq A2$$

where:

$$A1 = \left| Value_{IAEA} - Value_{Analyst} \right|$$

$$A2 = 2.58 \times \sqrt{Unc_{IAEA}^2 + Unc_{Analyst}^2}$$

3.2.2. *Precision*

For evaluation of precision an estimator P is calculated for each participant, according to the following formula:

$$P = \sqrt{\left(\frac{Unc_{IAEA}}{Value_{IAEA}}\right)^2 + \left(\frac{Unc_{Analyst}}{Value_{Analyst}}\right)^2} \times 100\%$$

P directly depends on the measurement uncertainty claimed by the participant. The Limit of Acceptable Precision (LAP) for each analyte respectively is defined for the respective proficiency test in advance, including any adjustment due to the concentration or activity level of the analytes concerned and the complexity of the analytical problem. Participants’ results are scored as ‘acceptable’ for precision when $P \leq LAP$. The LAP value used in the evaluation of all radionuclides is listed in Table 1.

In the final evaluation, both scores for trueness and precision are combined. A result must obtain an ‘acceptable’ score in both criteria to be assigned the final score ‘acceptable’. Obviously, if a score of ‘not acceptable’ was obtained for both trueness and precision, the final score will also be ‘not acceptable’. In cases where either precision or trueness is ‘not acceptable’, a further check is applied. The reported result relative bias (R. Bias) is compared with the maximum acceptable bias (MAB). If $R. Bias > MAB$, the result will be ‘not acceptable’. However, if $R. Bias \leq MAB$, the final score will be ‘warning’. A ‘warning’ will reflect mainly two situations. The first situation will be a result with small measurement uncertainty; however its bias is still within MAB. The second situation will appear when results close to the assigned property value are reported, but the associated uncertainty is large. The MAB value used in the evaluation of all radionuclides is listed in Table 1.

3.2.3. *Blank evaluation*

The results of the blank (sample 05) were evaluated to check if a false positive was reported using the following rule: if the reported result fulfils the following criteria it was considered acceptable:

$$\left| Value_{analyst} - Unc_{analyst} \right| < 0.1$$

Also if the laboratory reported the MDL as a result (a value with a sign <) it was considered acceptable. Otherwise, the reported value was not acceptable.

3.3. The z-score value

The z-score is calculated from the laboratory results, the assigned value and a standard deviation in accordance with the following equation:

$$z_{Score} = \frac{Value_{Analyst} - Value_{IAEA}}{\sigma}$$

On the basis of the ‘fitness for purpose’ principle, the target value for the standard deviation (σ) is:

$$0.10 \times Value_{IAEA}$$

The laboratory performance is evaluated as satisfactory if $|z_{Score}| \leq 2$; questionable for $2 < |z_{Score}| < 3$, and unsatisfactory for $|z_{Score}| \geq 3$.

3.4. The u-score value

The value of the u_{test} was calculated according to the following equation [4]

$$u_{test} = \frac{|Value_{IAEA} - Value_{Analyst}|}{\sqrt{Unc_{IAEA}^2 + Unc_{Analyst}^2}}$$

This value is compared with the critical value listed in the t-statistic tables to determine if the reported result differs significantly from the expected value at a given level of probability. The advantage of the u_{test} is that it takes into consideration the propagation of measurement uncertainties when defining the normalised error. This is especially useful when evaluating results, which uncertainty may overlap with the reference interval.

It should be noted that the choice of the significance level is subjective. For this proficiency test we have set the limiting value for the u-test parameter to 2.58 for a level of probability at 99% to determine if a result passes the test ($u < 2.58$).

If the evaluation approach and/or acceptance criteria applied in this PT are not appropriate for the types of analyses and application performed in one of the participating laboratories, it is suggested to apply a self- scoring evaluation system which could fit specific requirements.

4. RESULTS AND DISCUSSION

4.1. General

570 measurement results were reported to the IAEA in this PT from 114 laboratories. The participants' data along with the statistical performance evaluation were compiled and presented in two tables which constitute an integral part of this report. Appendix I shows the data evaluation tables sorted by sample code. Performance evaluation tables sorted by laboratory code are reported in Appendix II.

The overall evaluation showed that 68% of all reported results fulfilled the PT criteria for both trueness and precision. Despite the fact that the matrix was easy and there was not any interference effect, 19% of all reported results were not acceptable against the PT criteria.

4.2. Technical information provided by the participants

The summary of the technical information provided by the participants on the analytical procedures used in their own laboratories is compiled in Tables 4 and 5. The information is coded with the same laboratory code used in data evaluation. The participants can benefit from the information exchange without revealing the laboratories' identity. The provided technical information was compiled in the same format as it was received, without any modification or editing.

From the reported information on the applied analytical procedure, most of the participants did not use any separation method due to the nature of the matrix. For source preparation 88 laboratories used auto deposition method on silver or stainless steel disk, only 8 participants used electro-deposition, one laboratory (42) used co-precipitation.

Different measurement techniques were used by the participating laboratories as illustrated in Table 2.

From the technical details of the analytical procedure provided by the participants who had low performance score, it was not possible to find any indication of a methodological error or problem. There was no substantial difference in the described procedures to which the root cause of discrepancy could be attributed.

4.3. False positive reporting

In this PT the method detection limit (MDL) was an important performance indicator. The participants were asked to report the MDL estimated through method validation.

From the technical information provided by the laboratories, it can be observed that there is no harmonised method for MDL estimation amongst Member States laboratories which could lead to inappropriate comparison of MDL estimated in different laboratories. The summary of the reported MDL and the used procedure to derive the MDL is shown in Table 6.

It can be noticed that 21 laboratories (group C, Table 3) reported false positive for the blank sample 05, this might indicate that the MDL was not appropriately estimated, or the method validation for such a matrix was not yet performed.

TABLE 2. AVERAGE PERFORMANCE SCORE AGAINST THE USED MEASUREMENT TECHNIQUE

Measurement technique	Number of laboratories	Average score [%]
Alpha spectrometry	88	76
Liquid scintillation	11	69
Proportional counter	5	56
Gross alpha counter	2	75
ZnS counter	2	10
Not reported	6	-

TABLE 3. LABORATORIES GROUPING RELATED TO THE CATEGORY OF RECOMMENDATIONS

Group	Laboratory code
Group A	1, 2, 3, 4, 5, 7, 8, 9, 12, 14, 15, 16, 18, 19, 21, 24, 25, 26, 27, 28, 30, 31, 33, 34, 35, 38, 39, 40, 42, 43, 46, 47, 50, 51, 55, 56, 57, 58, 59, 60, 63, 64, 65, 67, 69, 71, 72, 73, 74, 76, 78, 79, 81, 83, 84, 85, 87, 88, 89, 91, 92, 93, 94, 96, 97, 100, 102, 103, 104, 105, 106, 107, 108, 109, 110, 112, 114, 115, 117, 119, 121, 123, 124, 125, 126, 127.
Group B	2, 6, 17, 18, 19, 20, 23, 24, 26, 29, 37, 41, 44, 49, 51, 52, 56, 57, 62, 69, 70, 77, 82, 83, 86, 89, 90, 91, 93, 95, 98, 99, 101, 110, 125
Group C	3, 9, 13, 17, 20, 23, 30, 59, 62, 73, 80, 84, 85, 93, 95, 98, 99, 101, 110, 114, 118.
Group D	13, 20, 22, 23, 29, 36, 44, 49, 54, 70, 80, 82, 86, 95, 99, 101, 116.

TABLE 4. SUMMARY INFORMATION ON THE ANALYTICAL PROCEDURE AS REPORTED BY THE PARTICIPANTS IS PRESENTED AGAINST THE AVERAGE SCORE, WHERE NUMBER OF ‘A’ WAS MULTIPLIED BY 20, ‘W’ BY 10. NR: NOT REPORTED

Lab.	Sample	Source preparation	Measurement technique	Average score
1	Evaporation	Auto deposition	Alpha spec.	100
2	Evaporation	Auto deposition	Alpha spec.	90
3	Evaporation	Auto deposition	Alpha spec.	80
4	Evaporation	Auto deposition	Alpha spec.	70
5	Evaporation	Auto deposition	Alpha spec.	100
6	Chelation	Electro deposition	ZnS counter	20
7	Evaporation	Auto deposition	Alpha spec.	100
8	Evaporation	Auto deposition	Alpha spec.	90
9	Evaporation	Auto deposition	Alpha spec.	80
12	Evaporation	Auto deposition	Alpha spec.	100
13	Evaporation	Auto deposition	ZnS counter	0
14	Evaporation	Auto deposition	Alpha spec.	70
15	Evaporation	Auto deposition	Alpha spec.	90
16	Evaporation	Auto deposition	Alpha spec.	90
17	Deposition	Deposition	Liquid scintillation counter	60
18	Evaporation	Auto deposition	Alpha spec.	80
19	Evaporation	Auto deposition	NR	90
20	Evaporation	Auto deposition	Alpha spec.	0
21	Evaporation	Auto deposition	Alpha spec.	80
22	Evaporation	Auto deposition	Alpha spec.	30
23	NR	NR	NR	0
24	Evaporation	Auto deposition	Alpha spec.	90

Lab.	Sample	Source preparation	Measurement technique	Average score
25	Evaporation	Auto deposition	Alpha spec.	100
26	Evaporation	Auto deposition	Alpha spec.	100
27	Evaporation	Auto deposition	Alpha spec.	100
28	NR	NR	Liquid scintillation counter	60
29	Evaporation	Auto deposition	Alpha spec.	30
30	Evaporation	Auto deposition	Alpha spec.	50
31	Evaporation	Auto deposition	Alpha spec.	60
33	Evaporation	Auto deposition	Alpha spec.	100
34	Evaporation	Auto deposition	Alpha spec.	100
35	Evaporation	Auto deposition	NR	90
36	NR	NR	Liquid scintillation counter	0
37	Evaporation	Auto deposition	Alpha spec.	60
38	Evaporation	Auto deposition	Alpha spec.	100
39	Evaporation	Auto deposition	Alpha spec.	100
40	Evaporation	Auto deposition	Alpha spec.	80
41	Chelation	Electro deposition	Proportional counter	50
42	Coprecipitation	Coprecipitation	Proportional counter	90
43	NR	NR	NR	100
44	Scintillation cocktail	Scintillation cocktail	Liquid scintillation counter	10
46	Evaporation	Auto deposition	Alpha spec.	100
47	NR	NR	Liquid scintillation counter	100
49	None	None	Alpha spec.	20
50	Evaporation	Auto deposition	Alpha spec.	100
51	Evaporation	Auto deposition	Alpha spec.	100
52	Evaporation	Auto deposition	Alpha spec.	50
54	Evaporation	Auto deposition	Alpha spec.	20

Lab.	Sample	Source preparation	Measurement technique	Average score
55	NR	NR	Liquid scintillation counter	90
56	Evaporation	Auto deposition	Alpha spec.	90
57	Evaporation	Auto deposition	Gross Alpha counting	70
58	NR	NR	NR	80
59	Evaporation	Auto deposition	Gross Alpha counting	80
60	NR	NR	Liquid scintillation counter	60
62	Evaporation	Auto deposition	Alpha spec.	60
63	Evaporation	Auto deposition	Alpha spec.	70
64	Evaporation	Auto deposition	Alpha spec.	90
65	Evaporation	Auto deposition	Alpha spec.	100
67	Evaporation	Auto deposition	Alpha spec.	100
69	NR	NR	Alpha spec.	80
70	Evaporation	Auto deposition	Alpha spec.	20
71	Evaporation	Auto deposition	Alpha spec.	100
72	Evaporation	Auto deposition	Alpha spec.	90
73	Evaporation	Auto deposition	Alpha spec.	80
74	Evaporation	Auto deposition	Alpha spec.	100
76	Evaporation	Auto deposition	Alpha spec.	100
77	NR	NR	Alpha spec.	10
78	NR	NR	Liquid scintillation counter	100
79	Evaporation	Auto deposition	Alpha spec.	100
80	Chelation	Electro deposition	Alpha spec.	0
81	Evaporation	Auto deposition	Alpha spec.	100
82	Evaporation	Auto deposition	Alpha spec.	20
83	Evaporation	Auto deposition	Alpha spec.	100
84	Evaporation	Auto deposition	Alpha spec.	100

Lab.	Sample	Source preparation	Measurement technique	Average score
85	LSC	LSC	Liquid scintillation counter	100
86	Evaporation	Auto deposition	Proportional Counter	20
87	Evaporation	Auto deposition	Alpha spec.	60
88	Evaporation	Auto deposition	Alpha spec.	60
89	Evaporation	Auto deposition	Alpha spec.	90
90	NR	FALSE	Proportional Counter	40
91	Evaporation	Auto deposition	Alpha spec.	90
92	Evaporation	Auto deposition	Alpha spec.	100
93	Filtration	Filtration	Liquid scintillation counter	80
94	LSC	LSC	Liquid scintillation counter	100
95	Evaporation	Auto deposition	Alpha spec.	20
96	Evaporation	Auto deposition	Alpha spec.	100
97	Evaporation	Auto deposition	Alpha spec.	70
98	Evaporation	Auto deposition	Alpha spec.	60
99	Evaporation	Auto deposition	Alpha spec.	0
100	Chelation	Electro deposition	Alpha spec.	30
101	Evaporation	Auto deposition	Alpha spec.	40
102	Evaporation	Auto deposition	Alpha spec.	60
103	Evaporation	Auto deposition	Alpha spec.	100
104	Evaporation	Auto deposition	Alpha spec.	80
105	Evaporation	Auto deposition	Alpha spec.	100
106	Evaporation	Auto deposition	Alpha spec.	100
107	Chelation	Electro deposition	Alpha spec.	100
108	Evaporation	Auto deposition	Alpha spec.	100
109	Chelation	Electro deposition	Alpha spec.	100
110	Evaporation	Auto deposition	Alpha spec.	80

Lab.	Sample	Source preparation	Measurement technique	Average score
112	Evaporation	Auto deposition	Alpha spec.	100
114	Evaporation	Auto deposition	Proportional Counter	80
115	Chelation	Electro deposition	Alpha spec.	100
116	Evaporation	Auto deposition	Alpha spec.	40
117	Evaporation	Auto deposition	Alpha spec.	100
118	Evaporation	Auto deposition	Alpha spec.	40
119	Chelation	Electro deposition	Alpha spec.	100
120	Evaporation	Auto deposition	Alpha spec.	20
121	Evaporation	Auto deposition	Alpha spec.	90
123	Evaporation	Auto deposition	NR	100
124	NR	NR	Alpha spec.	80
125	Evaporation	Auto deposition	Alpha spec.	80
126	Deposition	Deposition	Alpha spec.	100
127	Evaporation	Auto deposition	Alpha spec.	80

TABLE 5. Po-210 ANALYTICAL PROCESS PARAMETERS AS REPORTED BY THE PARTICIPANTS. NR: NOT REPORTED

Lab code	Sample mass [g]	Counting time [s]	Gross counts	Counting efficiency [%]	Tracer added activity [Bq]	Unc. of added Tracer	Tracer recovery [%]
1	5.023	79711	6314	26	0.31	0.016	100
2	26.63	60000	16566	22.4	0.1738	0.0087	96.8
3	8.113	90000	4383	12.8	0.2	0.0009	87.3
4	74.3	81759	26519	9.6	0.0792	0.0034	93.1
5	30.35	11820	1618	18.8	1.48	0.01	45.8
6	60.15	7200	929	16	NR	NR	63
7	20.75	55402	10409	17.6	0.202	0.0051	98
8	10.1429	60000	6467	27.4	0.397	0.0012	78.6
9	1.0348	246228	2646	24	0.036	0.002	80
12	2	61000	1420	29	0.1	0.005	73
13	20.6448	30000	10760	52	0	0	93
14	25.95	60000	9044	16.58	0.6262	0.0188	69.25
15	15.347	60000	6448	25	0.96	0.01	60.9
16	5.0788	102477	6793	NR	0.193	0.0028	NR
17	26.77	7200	2553	50	NR	NR	NR
18	10	5000	683	0.32	NR	NR	NR
19	16.8	8000	1067	23	0.24	0.02	63
20	2.0526	64829	924	29.72	0.136	0.002	46
21	5.0812	30301	630	10.91	0.5717	0.0286	74.43
22	31.60977	10000	2994	25	0	0	0
23	3.1394	17220	355	28.63	0.134	0.002	45
24	2.725	92605	2618	19	0.039	0.001	95
25	57.5844	59653	40413	26	12.456	0.029	87

Lab code	Sample mass		Counting time		Gross counts	Counting efficiency [%]	Tracer added activity [Bq]	Unc. of added Tracer	Tracer recovery [%]
	[g]		[s]						
26	0.5041		84272		519	28.92	0.0105	0.00105	60.86
27	20		60000		NR	18.2	0.32	0.006	49.3
28	10		5400		NR	85	NR	NR	NR
29	21.0891		228910		40753		0.0353	0.00001	
30	19.68		15000		2191	34	1.858	0.013	36.5
31	10.392		50000		4811	21	0.861	0.0018	90.2
33	4.1578		200000			37.9	0.043	0.0004	
34	13.8		160000		11200	14.5	2.2	0.02	70.2
35	5.1858		86405		4914	27	0.08	0.001	83
36	5		6000		370	57	NR	NR	NR
37	3.089		24583		1715	36.67	0.0696	0.0003	83.31
38	4.1065		48540		2050	29.1	0.163	0.005	76.6
39	10.35		10800		1131	37.2	NR	NR	96
40	19.9193		10690		2044	19.89	NR	NR	95.15
41	55.41		20400		7625	36.6	0	0	73.7
42	29.74		216000		86962	31.33	0.0387	0.001	100
43	1.0459		398098		2191	20	0.089	0.0015	53
44	0.25		12000		4324	57	NR	NR	NR
46	3.0928		75161		2197	23.28	0.03149	0.00013	65.15
47	4.2		24900		5769	99.5	0	0	0
49	51.9497		1800		231	38.073	0	0	0
50	99.6218		13608		649	26.59	0.423	0.0145	95
51	10.381		107000			20.6	NR	NR	NR
52	26.64		8000		984	20	NR	NR	NR

Lab code	Sample mass		Counting time		Gross counts	Counting efficiency [%]	Tracer added activity [Bq]	Unc. of added Tracer	Tracer recovery [%]
	[g]		[s]						
53	50		78820		699	27.9	0.57	0.01	15.4
54	5.1885		21600		1190		0.4404	0.0048	54.8
55	15.5		28440		15642	71.1	NR	NR	NR
56	10		21200		2452	25	0.0511	0.00025	87
57	50.2562		500		494	32.94	0	NR	NR
58	15		74495		NR	NR	NR	NR	NR
59	10		400000		1195	17.3	NR	NR	NR
60	4		15000		3189	86.7	NR	NR	NR
62	20.07		80000		10382	17.52	6.03	0.12	76.1
63	21.06		55177		8813	25.42	3.66	0.05	75
64	53.3781		33000		10650	14	0.17	0.0017	83
65	47.1712		9154		5086	36.3	0.534	0.027	68.94
66	69.01		86404		12657	9	13.2865	0.0279	50.6
67	5.199		82800		3783	NR	7.39	0.22	91
69									
70	20.09		84417		11730	19.4	0.065766	0.002	62.1
71	2.6095		252812		9338	29.95	0.050289	0.001106	99.9
72	5.17		85000		1735	14.8	0.523	1.7	54.2
73	2.07		200000		3692	23	0.02	0.0006	75.3
74	8.31		57600		5395	25.5	0.0542	0.0011	91
76	19.4066		334800		47474	25	0.923	0.02	61.5
77	0.005		323300		361	2.48	0		
78	8.1709		29344		13860	100			
79	22.7		60000		15360	23.6	0.0813	0.002	

Lab code	Sample mass		Counting time		Gross counts	Counting efficiency [%]	Tracer added activity [Bq]	Unc. of added Tracer	Tracer recovery [%]
	[g]		[s]						
80	10.4181		246612		11610	11.92	0.437	0.004	51.27
81	31.509		170000		30740	20.6	0.2002	0.0005	57
82	32.47		18327		3560	20	0.047	0.001	76
83	10.37		36000		5566	36.8	0.1935	0.0097	79
84	1		75600		838	38.9386	0.0774	0.0031	56.227
85	3.63		18000		7449	98	NR	NR	NR
86						NR			
87	25.036		150000		30022	13.9	0.163	0.003	124
88	20		81315		12787	29.83	0.160333		59.8
89	8.1675		23000		1595	16.3	0.2599	0.002	16.3
90	0.94		3600		90	0.503	0	0	0
91	19.85		120000		36696	31.17	0.08758	0.00036786	96.67
92	5		70000		2139		0.169	0.003	
93	10.48		2500		1319	85.1		NR	
94						NR			
95	20.6842		7200		2036	41	0.473	0.015	75
96	5.2108		200000		9497	25.7	0.229	0.0033	68.3
97	5.305		60000		2734	29.9	6.901	0.015	61.38
98	14.98		5800		426	36	0	0	43
99	5		70023		2284	40	0.074	0.004	21.5
100	2.01633		144693		3366	23.33	0.0525	0.0016	80.51
101	20		78994		6127	20.6	0.236	0.0127	16.156
102	70.44		20000		26901	47.2	0.1	0.0013	88.7
103	15.56		21650		835	6.27	0.556	0.011	82.5

Lab code	Sample mass		Counting time		Gross counts	Counting efficiency [%]	Tracer added activity [Bq]	Unc. of added Tracer	Tracer recovery [%]
	[g]		[s]						
104	60.8		86475		18932	11.1	0.156	0.019	68.2
105	25		200000		56412	25	1.07	0.05	51
106	25.0004		86400		14347	16.85	0.071	0.002	70.4
107	10.0104		82		5665	NR	0.035	0.000035	
108	19.71318		84370		5891	24	2.61329	0.0261329	29
109	10.328		80007		9814	26.3	0.1964	0.0023	88.2
110	21		79847		19773	15.65	0.5	0.035	68.6
112	14.7		43200		5005	25	0.89	NR	62
114	50		900		665	27		NR	
115	2.5622		252844		2656	NR	0.0816	0.0016	
116	2		242322		4842	23	0.077	0.002	95
117	0.982		80000		1749	45	0.123	0.00074	101
118	5.177		251004		4453	NR	0.1129	0.0063	
119	30.5346		149048		NR	8.66	7.2	0.3	8.66
120	2.062		55000		1114	26	0.124	0.012	102
121	57.3374		6000		4653	40.6	0.36	0.004	68.88
123	5.2		43200		1182	NR	0.403	0.002	10.7
124	8.214		151200		9179	17.2	0.261	0.8	17.2
125	2.58		60000		1608	28	0.09	0.00018	85
126	3.0314		183605		6730	25	0.1205	0.0025	98
127	25.022		38447		12296	26.9	2.0265	0.009727	86.386

TABLE 6. INFORMATION AS REPORTED BY THE PARTICIPANTS RELATED TO THE METHOD DETECTION LIMIT (MDL) OF PO-210

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
1	no	NR	A
2	yes	Minimum detection limit (MDL) = 0.06Bq/kg, Repeatability = 0.64 at 10.3Bq/kg	A
3	yes	Reproducibility : analysis on SRM (f.e. IAEA 134 + 368), one intercomparison test, Repeatability on spiked water samples (variation coefficient of 5.3% with uncertainty of 15%)	N
4	yes	Detection Limit (MDL) calculated according to ISO-11929 standard, MDL= 0.013 Bq/Kg (99%) (Confidence level)	A
5	no	For a 1000 minute count the MDL= 0.01Bq/kg	A
6	yes	Documents and records were first set to meet trackability and traceability requirements, where internal quality control mechanisms have been adopted. Methods stability was checked by means of Z-score control charts. Internal method validation parameters including method detection limits, repeatability limits, reproducibility limits, recovery coefficient and relative error were estimated. External method validation has been achieved by participating in national inter-laboratory exercises and international intercomparison exercises and proficiency tests.	A
7	no	detection limit: 0.01 - 0.06 Bq/kg	A
8	no	NR	A
9	yes	Minimum detection limit: ranges between 0.2-0.4 mBq depending on the camera used.	N
12	yes	All 5 samples analysed first based on 20 mL aliquots with DL's below 0.1 Bq/kg. Analyses of samples 1-4 repeated based on 2-3 mL aliquots to match spike amounts.	A
13	no	NR	N
14	yes	MDL: 0.04 Bq/kg (Po-210)	A
15	no	We do not routinely analyse for Po-210. We have performed limited cross-checking with other laboratories.	A

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
16	no	NR	A
17	no	NR	N
18	yes	MDL is 0.001 Bq/sample	-
19	yes	Minimum Detection Limit : 0.04 Bq/kg, Repeatability: 5%, Reproducibility: 8%	A
20	no	NR	N
21	no	NR	-
22	yes	a) Spectrometric system is checked by counting tracer (Pu-242) for MDL level concentration. b) Background count rate reproducibility	A
23		NR	N
24	yes	MDL is about 1 mBq for about 24 h counting time	A
25	yes	Method was validated against SRM4337 (Pb-210 standard that is equilibrium with Po-210 standard)	A
26	no	NR	A
27	no	NR	A
28	yes	ROI 40-400 keV, statistics of single measurement $\pm 5\%$, statistics of 3 repetitive measurements for result	A
29	yes	NR	A
30	no	NR	N
31	yes	Minimum detection limit was calculated by L. A. Currie's equation. The MDL is 0.07 Bq/kg. Repeatability was not tested. Each sample was counted once by alpha spectrometer. Reproducibility was tested. The identical sample was measured three times. The RSD of results is less than 1.5%.	A

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
33	yes	MDL = < 0.05 Bq/kg	A
34	yes	MDL = < 0.05 Bq/kg	A
35	no	NR	A
36		MDL = 0.65 +/- 0.16 Bq/kg, Repeatability limit = 0.88 Bq/kg. No reproducibility limits are available at this time. Minimum detectable activity is determined by applying the average of eight replicate analysis of a blank sample to the Currie formula. Repeatability is determined using eight replicate analysis of a NIST traceable reference material. Similar reference materials are used to determine four point linearity.	A
37	yes	Minimum Detection Limit = 0.002 Bq/L, Standard deviation of the mean = 13%, Bias = 1,4 %	A
38	yes	MDL is dependant on sample mass. For low level work where 100g of sample is taken the MDL is 0.005Bq/kg. For these IAEA samples where the activity was very high 4 g of sample was used to achieve an MDL around 0.1Bq/kg. Repeatability at k=2 is 4.2%	A
39	yes	For this PT MDL=0.15Bq/kg, at Counting Time 25200s, sample Mass 10.4g, Counting Efficiency 37.2%, Recovery 96% (for low salinity waters)	A
40	no	Although we did not participate yet in a Po-210 intercomparison exercise, we have calculated our Minimum detection limit: for 20 grams of sample, 3 hours of measuring time, efficiency 0,4 and chemical yield 0,95= 0,08 Bq/Kg	-
41	yes	Minimum detection activity MDL: 0.0123 Bq on TENNELEC S5HP / Gas flow type proportional counter , 36000 sec. counting time ; Minimum detection activity: 0.0062 Bq on TENNELEC TC 257/IPA Silicon detector, SCA counting mode, 60000 sec Counting time .	A
42	no	NR	A
44	no	NR	A
46	yes	1-MDL=0.00481Bq/L, 2-Triplicate samples analysed, 3-Above 95%.	A
47	no	NR	A

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
49	yes	Minimum Detection Limit: 0.31 Bq/kg, Repeatability limit: 0.15%, Reproducibility limit: 0.25%	A
50	yes	Minimum detection limit: 0.1 Bq/kg	A
51	yes	Not especially for Po-210, Validation parameters.	A
52	no	NR	A
54	no	Method validation for polonium in solid, water and urine samples are in progress.	A
55	yes	Testing of reproducibility with certified standard	A
56	yes	The validation method was carried out for urine sample and is believed to be valid for water sample as well, Minimum detection limit: 1 mBq/L when 500 ml of urine is introduced in the analysis, repeatability: 10% at the 20 mBq/L level of activity.	A
57	yes	Minimum Detection Limit for Sample Code 1 to 4 - 0.2 Bq/Kg(3 sigma). Repeatability $\pm 3\%$, Reproducibility $\pm 4\%$	A
58	no	NR	A
59	no	NR	N
60	yes	<p>The detection limit for Po-210 detection in water samples was calculated by the following formula:</p> $DL = 3 \cdot \sqrt{Cb} / (T \cdot E \cdot M)$ <p>where DL is the detection limit, Cb is the background counts, T is the background counting time, E is the efficiency of detecting alpha particles and M is the sample mass. For Cb=3000 counts, T=30000sec, E=0.867 and M=4g the detection limit is DL=0.16 Bq/Kg. If a water sample of 7g is used the detection limit is reduced to 0.08 Bq/Kg.</p> <p>The repeatability limit was calculated by using the formula $R = 1.96 \cdot \sqrt{2} \cdot S$ where S is the standard deviation of activity A calculated by repeated measurements of each sample (4 measurements for each sample). Repeatability Limits in Bq/Kg were calculated for each sample and were found: sample#1: 3.5, sample#2: 5.4, sample#3: 8.0, sample#4: 8.8, sample#5: 1.2.</p>	A

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
62	yes	In order to determine any bias, the method validation procedure was carried out via the analysis of a standard Po-210 solution supplied in National Metrological Institute, with identical sample counting geometry and detector configuration. For the purposes of this exercise, the minimum detection limit achieved for this methodology was 0.34 mBq/kg (95% confidence level). The repeatability limit was 3%.	N
63	yes	NR	A
64	no	NR	A
65	yes	MDL=0.01 Bq/kg, repeatability = 6%	A
67	yes	MDC 0.1 Bq/kg for 5g sample and 23 hour counting time.	A
69	no	NR	A
70	no	NR	A
		Genie 2000 software validated by CANBERRA	
71	yes	The method validation was performed by analysing IAEA-326 Soil and IAEA-315 Sediment. The obtained data were all in good agreement with the recommended values. The obtained precision (relative standard deviations) is < 10% and the accuracy (relative bias) is < 2%. The minimum detection limit for 5 kg of water sample is 0.016 mBq/kg and the corresponding value for 2.5 g of water 32 mBq/kg.	A
72	yes	I could give the validation parameters for these matrices. Furthermore, the minimum detection limit highly depends on the processed sample volume. Again there is no sense in giving a detection limit as such. A water sample is much simpler than a silicon-containing sample with a strongly oxidizing chemical treatment.	A
73	yes	The method validation was performed on urine samples of 500 ml. Considering a counting time of 20000s the validation parameters were: Minimum detection limit: 5 mBq/l; Repeatability limit : 5%; Reproducibility limit : 9%.	N
74	no	NR	A

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
76	yes	Detection limit (MDL): 0.005 bq/kg, Repeatability limit: 1.4 bq/kg for sample of 50 bq/kg radioactivity, at 95% confidence level. Reproducibility limit: No data	A
77	no	NR	A
78	no	NR	A
79	yes	MDL 0.1 Bq/kg, reproducibility limit 10%	A
80	yes	Validation was performed as emergency method; MDL: 0.16 Bq/kg.	N
81	no	NR	A
82	no	NR	A
83	no	NR	A
84	yes	1 pCi/L MDL, 20% precision limit, 25% bias limit	N
85	no	NR	N
86	no	NR	A
87	yes	Limit of detection (MDL) 0.05 Bq/kg, Repeatability 0.6%	A
88	yes	MDL = 6.225E-04 Bq/sample, LCS = Found-to-Added 98%	A
89	no	NR	A
90	no	MDL = 0.0089 Bq	A
91	yes	Water samples were spiked with a NIST traceable standard at 4.92, 11.5, 26.3, 42.6, 82.2, 392, and 755 mBq/L. The %recoveries ranged from 91% to 112%. The relative percent difference ranged from 3.5% to 12%. The typical MDL values for samples analyzed in this study ranged from 0.00477 to 0.09648 Bq/kg.	A

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
92	no	NR	A
93	yes	MDL 0.3 Bq/l	N
94	yes	NR	A
95	no	NR	N
96	no	The Currie detection limits ranged over 0.09 to 0.17Bq/kg for these particular samples	A
97	no	NR	A
98	yes	The MDL for our procedure is less than 0,05 Bq/Kg. The reproducibility parameter is 0,95.	N
99	yes	Minimum detection limit: 0.08 Bq/kg, Repeatability limit: smaller than 8%, Reproducibility limit: smaller than 10%	N
100	yes	MDL: 0.02. Use two different method to check reproductively.	A
101	no	NR	N
102	no	NR	A
103	yes	MDL: 0.05 Bq/kg	A
104	no	NR	A
105	yes	MDL is 0.05 Bq/kg	A
106	yes	MDL=0.04 Bq/kg where efficiency=0.2; yield=0.6; sample mass=25g, counting time=86400s. Accuracy: 15%. Precision: 10% (12 samples on day of analyses). Reproducibility: 13% (more than 50 test samples analysed over a year as test control samples)	A
107	no	NR	A

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
108	no	Based on peer reviewed publication	A
109	no	Minimum detection limit (MDL): 12 counts/measuring sample	A
110	yes	Minimum detection limit: 0.5Bq/sample in about 24hours (95% confidence level) Repeatability: 5%. Reproducibility: 7%	N
112	no	NR	A
114	no	MDL: 0.004 Bq/ml. Reproducibility: 0.95	N
115	no	NR	A
116	yes	Minimum detection limit, Repeatability limit, Reproducibility limit.	A
117	yes	NR	A
118	no	NR	N
119	no	NR	A
120	yes	Reagent blank (triplicates) was spiked with the same amount of tracer (Po-209). MDA was determined to be 0.084 Bq/Kg.	A
121	yes	The detection limit (MDL) for a 1 gram solid sample is 0.005Bq/g. Note, the detection limit will vary depending on the sample weight. Historically, the method has yielded a precision (relative standard deviation) of +/- 10%. The method has yielded an accuracy [Average Recovery (%) ± standard deviation (%)] of 99.0 ± 9.1	A
122	NR	NR	A
123	yes	MDL=0.007Bq/kg. Repeatability limit: r=2.8*S, #1: 2.7; #2: 7.6; #3: 3.9; #4: 8.0; #5: 0.093 Bq/kg, S was used as Standard deviation of 4 analyses.	A
124	no	NR	A

Lab code	Method validated	Reported procedure for Method Detection Limit estimation in the participants laboratories	Blank sample score
125	no	NR	A
126	yes	Repeated analysis on all samples. Note! reported values denote arithmetic means	A
127	yes	MDL: 3.00E-02 Bq/g. Accuracy : u=2.41. Precision: 6.41%	-

4.4. Measurement repeatability

The PT samples contained duplicate samples 01, 03 and 02, 04. The difference between the results of the duplicate samples was checked. 45 laboratories of 114 had a variation between the duplicate samples more than 5%, which might indicate the need for improving the method stability. The method statistical control and repeatability should be controlled and monitored to insure the method capability to detect low activity concentrations with high reliability.

Figures 6 and 7 show the graphical presentation of the variations between the duplicate samples.

4.5. Recommendations to the laboratories

Based on the performance evaluation results the recommendations to the participants could be divided into four categories:

a) Eighty six laboratories (Table 3, group A) reported results with a quality which fits for the purpose of rapid responding in emergency situation to trigger an alarm for remediation or any other decision for an action to be taken. However more efforts should be invested on method validation to determine the method performance characteristics in the laboratory's local conditions and to demonstrate that the targeted quality criteria of the analytical procedure are attained.

b) Thirty five laboratories (Table 3, group B) should improve the repeatability and the reproducibility of their determinations and to find out the source of variations, is it from the plating or from inappropriate recovery correction or from other source. Replicate analysis of spiked samples should be used to optimise the method and to reduce the source of variations. Target repeatability and reproducibility standard deviations should be set up by the analyst and the analytical procedure should be optimised to attain these targets.

c) Twenty one laboratory (Table 3, group C) reported false positive or a value for the blank sample higher than the target value. These laboratories should evaluate the analytical procedure blank and to subtract it from the sample value. Eurachem Guide on method validation suggests some guidelines on MDL determination. Many participants reported in Table 6 the procedure they applied in the estimation of MDL.

d) Seventeen laboratories (Table 3, group D) could not report acceptable results due to either significant bias or in few cases due to unstable method. These laboratories should revise their method and look for the root cause of bias or instability and perform method validation to check the reliability of the reported results.

5. CONCLUSIONS

The IAEA-CU-2007-09 world wide proficiency test for the determination of Po-210 in water was successfully conducted, 127 participants received the PT samples, and 88% of the participants reported back their results to the IAEA which indicates a high rate of results reporting in this PT.

The PT results demonstrated that around 70% of the participants were able to report results which fit the purpose of rapid detection of Po-210 in water.




Fig. 6 and 7. Variations in % between the reported results for the duplicate samples 01, 03 and 02, 04.

However, although the matrix was a straightforward easy one and the activity concentrations were relatively high 19% of the reported results failed to pass the PT criteria. In few cases positive results were reported for the blank sample which suggests a possibility of false positive reporting.

The PT organiser proposed four categories of general recommendations to the participating laboratories to improve their analytical performance. However, if any participant needs any technical assistance to improve the analytical performance of Po-210 determination, Chemistry Unit at the Agency's Seibersdorf laboratories will be glad to respond to such requests.

The PT results revealed the need for a harmonised analytical procedure for Po-210 rapid determination in case of emergency for high and low levels of activities. The procedure should also contain a standardized quality control protocol to assist the analyst in the validation of the reported results.

APPENDIX I. DATA EVALUATION TABLES SORTED BY SAMPLE CODE

All results listed in this Appendix are expressed in Bq/kg units at a reference date set to 2007-April-01. The abbreviations and calculation formulas used in the individual evaluation tables are explained in Section 3 of this report.

The individual laboratory evaluation reports are presented in ascending order of the laboratory code.

On the S-shape charts the IAEA target value is represented by a red line, and the respective combined standard uncertainty [u] is represented by two green lines.

On the z score charts warning limits are represented by blue lines, action limits by red lines.

Data evaluation of sample 01




Fig. I-01: S-shape chart of sample 01




Fig. I-02: z-score chart of sample 01

Data evaluation of sample 01

Target Value: 52.8 ± 1.4 Bq/kg

TABLE I-01: DATA EVALUATION OF SAMPLE 01

Lab code	Rep. Value	Rep Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
1	48.10	2.60	5.41	4.70	7.62	A	6.02	A	A
2	47.00	4.10	8.72	5.80	11.18	A	9.12	A	A
3	52.80	4.10	7.77	0.00	11.18	A	8.21	A	A
4	48.80	2.00	4.10	4.00	6.30	A	4.88	A	A
5	51.90	3.60	6.94	0.90	9.97	A	7.43	A	A
6	20.50	2.50	12.20	32.30	7.39	N	12.48	A	N
7	52.60	1.90	3.61	0.20	6.09	A	4.48	A	A
8	48.30	1.00	2.07	4.50	4.44	N	3.36	A	W
9	54.00	2.00	3.70	1.20	6.30	A	4.55	A	A
12	50.50	5.60	11.09	2.30	14.89	A	11.40	A	A
13	36.00	3.00	8.33	16.80	8.54	N	8.74	A	N
14	51.00	1.80	3.53	1.80	5.88	A	4.41	A	A
15	45.60	2.35	5.15	7.20	7.06	N	5.80	A	W
16	58.92	1.42	2.41	6.12	5.14	N	3.58	A	W
17	43.00	5.40	12.56	9.80	14.39	A	12.84	A	A
18	59.80	4.70	7.86	7.00	12.65	A	8.29	A	A
19	54.02	3.76	6.96	1.22	10.35	A	7.45	A	A
20	99.00	5.00	5.05	46.20	13.40	N	5.70	A	N
21	49.41	3.01	6.09	3.39	8.56	A	6.64	A	A
22	40.98	2.98	7.27	11.82	8.49	N	7.74	A	N
23	102.00	10.00	9.80	49.20	26.05	N	10.16	A	N
24	53.76	3.09	5.75	0.96	8.75	A	6.33	A	A
25	52.00	1.10	2.12	0.80	4.59	A	3.39	A	A
26	48.02	6.66	13.87	4.78	17.56	A	14.12	A	A
27	53.10	2.70	5.08	0.30	7.85	A	5.73	A	A
28	65.00	10.00	15.38	12.20	26.05	A	15.61	N	N
29	41.64	1.03	2.47	11.16	4.48	N	3.63	A	N
30	58.15	1.65	2.84	5.35	5.58	A	3.88	A	A
31	48.12	0.88	1.83	4.68	4.27	N	3.22	A	W
33	58.00	3.00	5.17	5.20	8.54	A	5.81	A	A
34	50.70	2.30	4.54	2.10	6.95	A	5.25	A	A
35	47.90	1.70	3.55	4.90	5.68	A	4.43	A	A
36	20.00	2.00	10.00	32.80	6.30	N	10.35	A	N
37	54.10	2.90	5.36	1.30	8.31	A	5.98	A	A

Lab code	Rep. Value	Rep Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
38	45.90	3.20	6.97	6.90	9.01	A	7.46	A	A
39	55.00	3.00	5.45	2.20	8.54	A	6.06	A	A
40	50.00	3.40	6.80	2.80	9.49	A	7.30	A	A
41	48.36	1.33	2.75	4.44	4.98	A	3.82	A	A
42	45.10	3.20	7.10	7.70	9.01	A	7.57	A	A
43	49.90	1.60	3.21	2.90	5.49	A	4.16	A	A
44	36.00	38.00	105.56	16.80	98.11	A	105.5	N	N
46	62.35	3.57	5.73	9.55	9.90	A	6.31	A	A
47	51.10	2.70	5.28	1.70	7.85	A	5.91	A	A
49	6.43	0.86	13.32	46.37	4.24	N	13.58	A	N
50	56.57	4.08	7.21	3.77	11.13	A	7.68	A	A
51	48.30	5.02	10.39	4.50	13.45	A	10.73	A	A
52	34.60	1.42	4.10	18.20	5.14	N	4.89	A	N
54	71.20	4.90	6.88	18.40	13.15	N	7.38	A	N
55	46.00	3.00	6.52	6.80	8.54	A	7.04	A	A
56	53.00	4.00	7.55	0.20	10.93	A	8.00	A	A
57	65.60	3.00	4.57	12.80	8.54	N	5.29	A	N
58	51.70	0.40	0.77	1.10	3.76	A	2.76	A	A
59	71.00	7.00	9.86	18.20	18.42	A	10.21	A	A
60	58.40	1.20	2.05	5.60	4.76	N	3.35	A	W
62	49.00	2.00	4.08	3.80	6.30	A	4.87	A	A
63	39.50	4.42	11.19	13.30	11.96	N	11.50	A	N
64	49.97	2.12	4.24	2.83	6.55	A	5.00	A	A
65	46.20	3.00	6.49	6.60	8.54	A	7.01	A	A
67	50.00	1.70	3.40	2.80	5.68	A	4.31	A	A
69	46.00	1.40	3.04	6.80	5.11	N	4.04	A	W
70	59.00	6.60	11.19	6.20	17.41	A	11.50	A	A
71	47.37	3.32	7.01	5.43	9.30	A	7.49	A	A
72	47.50	1.90	4.00	5.30	6.09	A	4.80	A	A
73	50.60	2.70	5.34	2.20	7.85	A	5.96	A	A
74	48.00	3.00	6.25	4.80	8.54	A	6.79	A	A
76	50.23	2.26	4.50	2.57	6.86	A	5.22	A	A
77	35.00	3.10	8.86	17.80	8.78	N	9.25	A	N
78	49.30	6.60	13.39	3.50	17.41	A	13.65	A	A
79	46.70	2.00	4.28	6.10	6.30	A	5.04	A	A
80	73.88	2.17	2.94	21.08	6.66	N	3.96	A	N
81	48.81	4.14	8.48	3.99	11.28	A	8.89	A	A
82	40.00	3.00	7.50	12.80	8.54	N	7.95	A	N
83	50.80	2.90	5.71	2.00	8.31	A	6.29	A	A

Lab code	Rep. Value	Rep Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
84	50.69	3.13	6.17	2.11	8.85	A	6.72	A	A
85	54.00	2.00	3.70	1.20	6.30	A	4.55	A	A
86	7.12	0.17	2.39	45.68	3.64	N	3.57	A	N
87	44.80	2.20	4.91	8.00	6.73	N	5.58	A	W
88	43.03	1.71	3.97	9.77	5.70	N	4.78	A	W
89	50.09	2.00	4.00	2.71	6.30	A	4.80	A	A
90	45.31	5.51	12.16	7.49	14.67	A	12.45	A	A
91	50.18	2.36	4.70	2.62	7.07	A	5.39	A	A
92	49.30	3.80	7.71	3.50	10.45	A	8.15	A	A
93	56.20	2.10	3.74	3.40	6.51	A	4.58	A	A
94	51.30	3.60	7.02	1.50	9.97	A	7.50	A	A
95	42.90	2.10	4.90	9.90	6.51	N	5.57	A	W
96	52.10	1.50	2.88	0.70	5.29	A	3.91	A	A
97	45.60	2.40	5.26	7.20	7.17	N	5.89	A	W
98	32.00	3.00	9.38	20.80	8.54	N	9.74	A	N
99	74.82	5.80	7.75	22.02	15.39	N	8.19	A	N
100	61.60	1.80	2.92	8.80	5.88	N	3.95	A	W
101	115.42	9.54	8.26	62.62	24.87	N	8.68	A	N
102	45.70	1.35	2.95	7.10	5.02	N	3.97	A	W
103	48.60	2.92	6.01	4.20	8.35	A	6.57	A	A
104	48.30	1.70	3.52	4.50	5.68	A	4.41	A	A
105	48.30	2.60	5.38	4.50	7.62	A	6.00	A	A
106	57.70	6.40	11.09	4.90	16.90	A	11.40	A	A
107	52.40	2.39	4.56	0.40	7.15	A	5.28	A	A
108	51.23	0.94	1.83	1.57	4.35	A	3.22	A	A
109	52.20	1.30	2.49	0.60	4.93	A	3.64	A	A
110	63.50	4.40	6.93	10.70	11.91	A	7.42	A	A
112	53.50	0.70	1.31	0.70	4.04	A	2.96	A	A
114	60.00	6.00	10.00	7.20	15.90	A	10.35	A	A
115	51.50	1.90	3.69	1.30	6.09	A	4.54	A	A
116	42.70	7.10	16.63	10.10	18.67	A	16.84	N	W
117	49.10	1.70	3.46	3.70	5.68	A	4.36	A	A
118	47.62	2.94	6.17	5.18	8.40	A	6.72	A	A
119	49.00	2.00	4.08	3.80	6.30	A	4.87	A	A
120	38.70	4.30	11.11	14.10	11.67	N	11.42	A	N
121	52.50	3.30	6.29	0.30	9.25	A	6.82	A	A
123	50.00	2.60	5.20	2.80	7.62	A	5.84	A	A
124	49.36	1.05	2.13	3.44	4.52	A	3.40	A	A
125	46.70	0.20	0.43	6.10	3.65	N	2.69	A	W

Lab code	Rep. Value	Rep Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
126	50.60	1.30	2.57	2.20	4.93	A	3.69	A	A
127	54.48	0.79	1.44	1.68	4.14	A	3.02	A	A

Data evaluation of sample 02




Fig. I-03: S-shape chart of sample 02




Fig. I-04: z-score chart of sample 02

Data evaluation of sample 02

Target Value: 101.6 ± 2.8 Bq/kg

TABLE I-02: DATA EVALUATION OF SAMPLE 02

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
1	93.30	4.80	5.14	8.30	14.34	A	5.84	A	A
2	81.40	7.20	8.85	20.2	19.93	N	9.26	A	W
3	99.00	7.70	7.78	2.60	21.14	A	8.25	A	A
4	81.60	3.30	4.04	20.0	11.17	N	4.89	A	W
5	99.40	3.50	3.52	2.20	11.56	A	4.47	A	A
6	25.10	3.00	11.95	76.5	10.59	N	12.2	A	N
7	100.00	3.35	3.35	1.60	11.26	A	4.34	A	A
8	95.40	2.00	2.10	6.20	8.88	A	3.46	A	A
9	106.00	3.00	2.83	4.40	10.59	A	3.95	A	A
12	84.40	9.40	11.14	17.2	25.31	A	11.4	A	A
13	72.00	6.00	8.33	29.6	17.08	N	8.78	A	N
14	83.90	2.80	3.34	17.7	10.22	N	4.33	A	W
15	92.93	4.68	5.04	8.67	14.07	A	5.74	A	A
16	102.01	2.56	2.51	0.41	9.79	A	3.73	A	A
17	75.80	7.90	10.42	25.8	210.6	N	10.7	A	N
18	109.00	7.00	6.42	7.40	19.45	A	6.99	A	A
19	107.28	7.35	6.85	5.68	20.29	A	7.38	A	A
20	144.00	7.00	4.86	42.40	19.45	N	5.59	A	N
21	101.15	6.16	6.09	0.45	17.46	A	6.68	A	A
22	80.55	5.78	7.18	21.0	16.57	N	7.69	A	N
23	148.00	12.00	8.11	46.4	31.79	N	8.56	A	N
24	85.75	4.34	5.06	15.85	13.33	N	5.76	A	W
25	98.80	2.10	2.13	2.80	9.03	A	3.48	A	A
26	97.98	14.44	14.74	3.62	37.95	A	14.99	A	A
27	101.60	5.10	5.02	0.00	15.01	A	5.73	A	A
28	110.00	15.00	13.64	8.40	39.37	A	13.91	A	A
29	55.19	1.11	2.01	46.4	7.77	N	3.41	A	N
30	116.50	2.66	2.28	14.9	9.96	N	3.58	A	W
31	89.92	1.32	1.47	11.6	7.99	N	3.12	A	W
33	113.00	7.00	6.19	11.4	19.45	A	6.78	A	A
34	100.20	3.60	3.59	1.40	11.77	A	4.53	A	A
35	93.60	3.20	3.42	8.00	10.97	A	4.39	A	A
36	46.00	2.00	4.35	55.6	8.88	N	5.15	A	N
37	105.30	11.00	10.45	3.70	29.28	A	10.80	A	A
38	90.80	6.20	6.83	10.8	17.55	A	7.36	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
39	101.20	5.20	5.14	0.40	15.24	A	5.83	A	A
40	99.00	6.60	6.67	2.60	18.50	A	7.21	A	A
41	73.54	1.95	2.65	28.0	8.80	N	3.82	A	N
42	84.00	6.10	7.26	17.6	17.32	N	7.77	A	W
43	93.60	2.60	2.78	8.00	9.86	A	3.91	A	A
44	106.00	40.00	37.74	4.40	103.45	A	37.84	N	W
46	113.80	6.49	5.70	12.2	18.23	A	6.33	A	A
47	98.80	7.30	7.39	2.80	20.17	A	7.89	A	A
49	2.04	0.34	16.46	99.5	7.28	N	16.69	N	N
50	103.99	5.77	5.55	2.39	16.55	A	6.20	A	A
51	91.40	9.52	10.42	10.2	25.60	A	10.77	A	A
52	86.50	3.50	4.05	15.1	11.56	N	4.90	A	W
54	132.00	8.30	6.29	30.4	22.60	N	6.87	A	N
55	89.00	5.00	5.62	12.6	14.78	A	6.26	A	A
56	84.00	6.00	7.14	17.6	17.08	N	7.66	A	W
57	116.10	2.50	2.15	14.5	9.68	N	3.50	A	W
58	99.60	0.50	0.50	2.00	7.34	A	2.80	A	A
59	118.00	10.00	8.47	16.4	26.79	A	8.91	A	A
60	110.70	2.10	1.90	9.10	9.03	N	3.35	A	W
62	96.00	3.00	3.13	5.60	10.59	A	4.17	A	A
63	82.90	7.53	9.08	18.7	20.73	A	9.49	A	A
64	94.59	3.69	3.90	7.01	11.95	A	4.78	A	A
65	95.20	6.20	6.51	6.40	17.55	A	7.07	A	A
67	95.80	3.10	3.24	5.80	10.78	A	4.25	A	A
69	93.70	5.10	5.44	7.90	15.01	A	6.10	A	A
70	53.90	5.82	10.80	47.70	16.66	N	11.14	A	N
71	90.93	6.37	7.01	10.6	17.95	A	7.53	A	A
72	88.00	3.00	3.41	13.60	10.59	N	4.38	A	W
73	90.60	4.70	5.19	11.00	14.11	A	5.87	A	A
74	95.00	5.00	5.26	6.60	14.78	A	5.94	A	A
76	93.40	4.18	4.48	8.20	12.98	A	5.26	A	A
77	63.50	3.80	5.98	38.10	12.18	N	6.59	A	N
78	93.00	12.00	12.90	8.60	31.79	A	13.19	A	A
79	93.30	4.00	4.29	8.30	12.60	A	5.10	A	A
80	130.40	4.00	3.07	28.80	12.60	N	4.12	A	N
81	90.89	7.72	8.49	10.71	21.19	A	8.93	A	A
82	57.70	4.20	7.28	43.90	13.02	N	7.78	A	N
83	101.70	5.70	5.60	0.10	16.38	A	6.25	A	A
84	97.68	5.44	5.57	3.92	15.79	A	6.21	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
85	106.00	4.00	3.77	4.40	12.60	A	4.67	A	A
86	15.53	0.29	1.87	86.07	7.26	N	3.33	A	N
87	84.00	5.00	5.95	17.60	14.78	N	6.56	A	W
88	89.31	3.28	3.67	12.29	11.13	N	4.59	A	W
89	86.53	1.73	2.00	15.07	8.49	N	3.40	A	W
90	99.87	7.67	7.68	1.73	21.07	A	8.16	A	A
91	95.77	4.41	4.60	5.83	13.48	A	5.37	A	A
92	99.90	4.70	4.70	1.70	14.11	A	5.45	A	A
93	96.10	3.10	3.23	5.50	10.78	A	4.24	A	A
94	98.00	4.90	5.00	3.60	14.56	A	5.71	A	A
95	83.50	4.00	4.79	18.10	12.60	N	5.53	A	W
96	100.30	2.70	2.69	1.30	10.04	A	3.85	A	A
97	87.80	3.50	3.99	13.80	11.56	N	4.85	A	W
98	86.00	6.00	6.98	15.60	17.08	A	7.50	A	A
99	140.00	10.62	7.59	38.40	28.34	N	8.07	A	N
100	117.40	4.00	3.41	15.80	12.60	N	4.38	A	W
101	89.70	6.38	7.11	11.89	17.96	A	7.62	A	A
102	87.10	2.26	2.59	14.50	9.28	N	3.79	A	W
103	89.40	4.47	5.00	12.20	13.61	A	5.71	A	A
104	90.10	2.90	3.22	11.50	10.40	N	4.24	A	W
105	92.30	4.90	5.31	9.30	14.56	A	5.98	A	A
106	103.30	11.60	11.23	1.70	30.79	A	11.56	A	A
107	100.24	9.24	9.22	1.36	24.91	A	9.62	A	A
108	98.27	1.58	1.61	3.33	8.29	A	3.19	A	A
109	101.00	2.30	2.28	0.60	9.35	A	3.58	A	A
110	121.00	8.50	7.02	19.40	23.09	A	7.55	A	A
112	100.20	1.00	1.00	1.40	7.67	A	2.93	A	A
114	100.80	11.00	10.91	0.80	29.28	A	11.26	A	A
115	99.10	3.50	3.53	2.50	11.56	A	4.48	A	A
116	78.00	13.50	17.31	23.60	35.57	A	17.53	N	N
117	95.00	2.90	3.05	6.60	10.40	A	4.11	A	A
118	84.07	5.56	6.61	17.53	16.06	N	7.16	A	W
119	93.00	4.00	4.30	8.60	12.60	A	5.11	A	A
120	73.20	8.10	11.07	28.40	22.11	N	11.40	A	N
121	91.70	5.50	6.00	9.90	15.92	A	6.60	A	A
123	96.00	5.10	5.31	5.60	15.01	A	5.98	A	A
124	89.62	2.82	3.15	11.98	10.25	N	4.18	A	W
125	102.60	0.60	0.58	1.00	7.39	A	2.82	A	A
126	95.80	2.30	2.40	5.80	9.35	A	3.65	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
127	98.13	1.37	1.40	3.47	8.04	A	3.09	A	A

Data evaluation of sample 03




Fig. I-05: S-shape chart of sample 03.




Fig. I-06: z-score chart of sample 03.

Data evaluation of sample 03

Target Value: 52.8 ± 1.4 Bq/kg

TABLE I-03: DATA EVALUATION OF SAMPLE 03.

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
1	49.30	2.50	5.07	3.50	7.39	A	5.72	A	A
2	47.40	4.20	8.86	5.40	11.42	A	9.25	A	A
3	57.80	4.50	7.79	5.00	12.16	A	8.22	A	A
4	44.70	1.90	4.25	8.10	6.09	N	5.01	A	W
5	53.00	1.60	3.02	0.20	5.49	A	4.02	A	A
6	11.80	1.50	12.71	41.00	5.29	N	12.99	A	N
7	52.10	1.92	3.69	0.70	6.13	A	4.54	A	A
8	50.30	1.00	1.99	2.50	4.44	A	3.31	A	A
9	48.00	1.30	2.71	4.80	4.93	A	3.79	A	A
12	41.90	4.70	11.22	10.90	12.65	A	11.53	A	A
13	32.00	3.00	9.38	20.80	8.54	N	9.74	A	N
14	45.50	1.80	3.96	7.30	5.88	N	4.76	A	W
15	50.43	2.60	5.16	2.37	7.62	A	5.80	A	A
16	50.66	1.49	2.94	2.14	5.27	A	3.96	A	A
17	42.20	5.30	12.56	10.60	14.14	A	12.84	A	A
18	61.60	3.90	6.33	8.80	10.69	A	6.86	A	A
19	47.97	3.23	6.73	4.83	9.08	A	7.24	A	A
20	95.00	5.00	5.26	42.20	13.40	N	5.89	A	N
21	51.75	3.15	6.09	1.05	8.89	A	6.64	A	A
22	42.64	3.11	7.29	10.16	8.80	N	7.76	A	W
23	85.00	8.00	9.41	32.20	20.95	N	9.78	A	N
24	46.81	2.63	5.62	5.99	7.69	A	6.21	A	A
25	52.20	1.10	2.11	0.60	4.59	A	3.39	A	A
26	47.19	6.38	13.52	5.61	16.85	A	13.78	A	A
27	51.00	2.60	5.10	1.80	7.62	A	5.75	A	A
28	60.00	8.00	13.33	7.20	20.95	A	13.59	A	A
29	43.60	1.16	2.66	9.20	4.69	N	3.76	A	W
30	58.70	1.73	2.95	5.90	5.74	N	3.96	A	W
31	48.27	0.88	1.82	4.53	4.27	N	3.22	A	W
33	59.00	3.00	5.08	6.20	8.54	A	5.73	A	A
34	50.90	2.20	4.32	1.90	6.73	A	5.07	A	A
35	48.40	1.60	3.31	4.40	5.49	A	4.24	A	A
36	21.00	2.00	9.52	31.80	6.30	N	9.89	A	N
37	65.80	2.70	4.10	13.00	7.85	N	4.89	A	N
38	48.30	3.30	6.83	4.50	9.25	A	7.33	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
39	54.50	3.00	5.50	1.70	8.54	A	6.11	A	A
40	51.00	3.40	6.67	1.80	9.49	A	7.17	A	A
41	42.20	1.25	2.96	10.60	4.84	N	3.98	A	N
42	46.70	3.30	7.07	6.10	9.25	A	7.55	A	A
43	50.70	1.50	2.96	2.10	5.29	A	3.97	A	A
44	9.00	38.00	422.22	43.80	98.11	A	422.2	N	N
46	56.21	3.14	5.59	3.41	8.87	A	6.19	A	A
47	51.40	2.90	5.64	1.40	8.31	A	6.23	A	A
49	9.27	1.18	12.76	43.53	4.73	N	13.03	A	N
50	56.86	3.99	7.02	4.06	10.91	A	7.50	A	A
51	50.10	5.20	10.38	2.70	13.89	A	10.71	A	A
52	46.96	1.93	4.11	5.84	6.15	A	4.89	A	A
54	70.90	4.70	6.63	18.10	12.65	N	7.14	A	N
55	46.00	2.00	4.35	6.80	6.30	N	5.09	A	W
56	45.00	3.00	6.67	7.80	8.54	A	7.17	A	A
57	54.10	2.50	4.62	1.30	7.39	A	5.33	A	A
58	51.70	0.40	0.77	1.10	3.76	A	2.76	A	A
59	64.00	5.00	7.81	11.20	13.40	A	8.25	A	A
60	62.10	1.30	2.09	9.30	4.93	N	3.38	A	W
62	47.00	2.00	4.26	5.80	6.30	A	5.01	A	A
63	45.55	4.97	10.91	7.25	13.32	A	11.23	A	A
64	46.97	1.69	3.60	5.83	5.66	N	4.47	A	W
65	48.20	3.10	6.43	4.60	8.78	A	6.96	A	A
67	52.60	1.80	3.42	0.20	5.88	A	4.33	A	A
69	47.60	2.00	4.20	5.20	6.30	A	4.97	A	A
70	15.11	1.56	10.34	37.69	5.41	N	10.68	A	N
71	44.45	3.11	7.00	8.35	8.80	A	7.48	A	A
72	47.00	2.00	4.26	5.80	6.30	A	5.01	A	A
73	50.50	2.70	5.35	2.30	7.85	A	5.97	A	A
74	49.00	3.00	6.12	3.80	8.54	A	6.67	A	A
76	49.30	2.23	4.52	3.50	6.79	A	5.24	A	A
77	31.60	4.00	12.66	21.20	10.93	N	12.93	A	N
78	50.20	6.70	13.35	2.60	17.66	A	13.61	A	A
79	49.10	2.00	4.07	3.70	6.30	A	4.86	A	A
80	70.00	2.35	3.36	17.20	7.06	N	4.28	A	N
81	48.05	4.08	8.49	4.75	11.13	A	8.90	A	A
82	37.80	3.30	8.73	15.00	9.25	N	9.12	A	N
83	54.30	3.00	5.52	1.50	8.54	A	6.13	A	A
84	52.17	3.29	6.31	0.63	9.22	A	6.84	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
85	51.00	2.00	3.92	1.80	6.30	A	4.73	A	A
86	11.94	0.24	2.01	40.86	3.66	N	3.33	A	N
87	44.10	2.20	4.99	8.70	6.73	N	5.65	A	W
88	46.91	1.77	3.77	5.89	5.82	N	4.61	A	W
89	49.14	0.98	2.00	3.66	4.41	A	3.32	A	A
90	32.32	4.81	14.88	20.48	12.92	N	15.12	N	N
91	51.20	2.38	4.65	1.60	7.12	A	5.35	A	A
92	51.70	2.70	5.22	1.10	7.85	A	5.86	A	A
93	50.30	2.00	3.98	2.50	6.30	A	4.78	A	A
94	46.70	3.50	7.49	6.10	9.73	A	7.95	A	A
95	42.00	2.50	5.95	10.80	7.39	N	6.52	A	N
96	51.90	1.50	2.89	0.90	5.29	A	3.92	A	A
97	45.70	2.40	5.25	7.10	7.17	A	5.88	A	A
98	55.00	4.00	7.27	2.20	10.93	A	7.74	A	A
99	72.31	5.75	7.95	19.51	15.27	N	8.38	A	N
100	61.20	1.90	3.10	8.40	6.09	N	4.08	A	W
101	72.09	5.67	7.87	19.29	15.08	N	8.30	A	N
102	45.00	1.28	2.84	7.80	4.89	N	3.89	A	W
103	54.30	3.80	7.00	1.50	10.45	A	7.48	A	A
104	54.50	1.80	3.30	1.70	5.88	A	4.24	A	A
105	48.80	2.60	5.33	4.00	7.62	A	5.95	A	A
106	58.20	6.50	11.17	5.40	17.15	A	11.48	A	A
107	53.93	4.48	8.31	1.13	12.11	A	8.72	A	A
108	49.95	0.87	1.74	2.85	4.25	A	3.17	A	A
109	51.80	1.30	2.51	1.00	4.93	A	3.65	A	A
110	56.20	3.90	6.94	3.40	10.69	A	7.43	A	A
112	51.60	0.70	1.36	1.20	4.04	A	2.98	A	A
114	56.50	5.70	10.09	3.70	15.14	A	10.43	A	A
115	53.20	1.80	3.38	0.40	5.88	A	4.30	A	A
116	42.30	7.30	17.26	10.50	19.18	A	17.46	N	W
117	50.90	1.60	3.14	1.90	5.49	A	4.11	A	A
118	43.53	2.82	6.48	9.27	8.12	N	7.00	A	W
119	58.00	2.00	3.45	5.20	6.30	A	4.35	A	A
120	39.00	4.30	11.03	13.80	11.67	N	11.34	A	N
121	44.50	2.60	5.84	8.30	7.62	N	6.42	A	W
123	50.70	2.90	5.72	2.10	8.31	A	6.30	A	A
124	49.16	0.92	1.87	3.64	4.32	A	3.25	A	A
125	52.80	0.20	0.38	0.00	3.65	A	2.68	A	A
126	50.60	1.30	2.57	2.20	4.93	A	3.69	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
127	52.65	1.09	2.08	0.15	4.58	A	3.37	A	A

Data evaluation of sample 04




Fig. I-07: S-shape chart of sample 04.




Fig. I-08: z-score chart of sample 04.

Data evaluation of sample 04

Target Value: 101.6 ± 2.8 Bq/kg

TABLE I-04: DATA EVALUATION OF SAMPLE 04

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
1	96.00	4.90	5.10	5.60	14.56	A	5.80	A	A
2	89.80	7.90	8.80	11.80	21.62	A	9.22	A	A
3	96.70	7.30	7.55	4.90	20.17	A	8.04	A	A
4	83.30	3.40	4.08	18.30	11.36	N	4.92	A	W
5	103.50	2.50	2.42	1.90	9.68	A	3.66	A	A
6	17.00	2.10	12.35	84.60	9.03	N	12.66	A	N
7	103.00	3.00	2.91	1.40	10.59	A	4.01	A	A
8	95.20	2.10	2.21	6.40	9.03	A	3.53	A	A
9	108.00	4.00	3.70	6.40	12.60	A	4.62	A	A
12	83.00	9.30	11.20	18.60	25.06	A	11.54	A	A
13	67.00	6.00	8.96	34.60	17.08	N	9.37	A	N
14	85.90	2.80	3.26	15.70	10.22	N	4.27	A	W
15	91.70	4.60	5.02	9.90	13.89	A	5.72	A	A
16	97.57	2.53	2.59	4.03	9.74	A	3.79	A	A
17	82.90	8.90	10.74	18.70	24.07	A	11.08	A	A
18	95.60	5.80	6.07	6.00	16.62	A	6.66	A	A
19	83.97	3.92	4.67	17.63	12.43	N	5.42	A	W
20	160.00	7.00	4.38	58.40	19.45	N	5.17	A	N
21	100.97	6.15	6.09	0.63	17.43	A	6.69	A	A
22	78.28	5.61	7.17	23.32	16.18	N	7.68	A	N
23	140.00	24.00	17.14	38.40	62.34	A	17.36	N	N
24	96.91	4.29	4.43	4.69	13.22	A	5.21	A	A
25	99.30	2.10	2.11	2.30	9.03	A	3.47	A	A
26	91.49	12.16	13.29	10.11	32.19	A	13.57	A	A
27	101.30	5.10	5.03	0.30	15.01	A	5.74	A	A
28	110.00	15.00	13.64	8.40	39.37	A	13.91	A	A
29	46.82	0.86	1.84	54.78	7.56	N	3.31	A	N
30	119.40	3.09	2.59	17.80	10.76	N	3.78	A	W
31	91.30	1.40	1.53	10.30	8.08	N	3.15	A	W
33	109.00	7.00	6.42	7.40	19.45	A	6.99	A	A
34	98.90	4.50	4.55	2.70	13.67	A	5.32	A	A
35	90.60	3.00	3.31	11.00	10.59	N	4.31	A	W
36	45.00	2.00	4.44	56.60	8.88	N	5.23	A	N
37	122.70	5.10	4.16	21.10	15.01	N	4.99	A	N
38	88.20	6.00	6.80	13.40	17.08	A	7.34	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
39	100.20	5.10	5.09	1.40	15.01	A	5.79	A	A
40	100.00	6.70	6.70	1.60	18.73	A	7.24	A	A
41	90.68	2.46	2.71	10.92	9.62	N	3.87	A	W
42	84.90	5.90	6.95	16.70	16.85	A	7.48	A	A
43	95.20	2.60	2.73	6.40	9.86	A	3.88	A	A
44	25.00	38.00	152.00	76.60	98.31	A	152.0	N	N
46	116.70	6.49	5.56	15.10	18.24	A	6.21	A	A
47	98.60	7.70	7.81	3.00	21.14	A	8.28	A	A
49	8.82	1.08	12.22	92.78	7.74	N	12.53	A	N
50	107.98	8.07	7.47	6.38	22.04	A	7.97	A	A
51	97.10	10.09	10.39	4.50	27.02	A	10.75	A	A
52	70.70	2.88	4.07	30.90	10.36	N	4.92	A	N
54	129.00	8.50	6.59	27.40	23.09	N	7.14	A	N
55	91.00	5.00	5.49	10.60	14.78	A	6.15	A	A
56	92.00	7.00	7.61	9.60	19.45	A	8.09	A	A
57	102.80	3.40	3.31	1.20	11.36	A	4.31	A	A
58	99.60	0.50	0.50	2.00	7.34	A	2.80	A	A
59	117.00	10.00	8.55	15.40	26.79	A	8.98	A	A
60	113.10	2.20	1.95	11.50	9.19	N	3.37	A	W
62	79.00	2.00	2.53	22.60	8.88	N	3.74	A	N
63	85.22	4.71	5.53	16.38	14.14	N	6.18	A	W
64	95.81	3.53	3.68	5.79	11.62	A	4.60	A	A
65	94.40	6.10	6.46	7.20	17.32	A	7.03	A	A
67	97.20	3.10	3.19	4.40	10.78	A	4.22	A	A
69	86.80	2.00	2.30	14.80	8.88	N	3.59	A	W
70	6.69	0.70	10.46	94.91	7.45	N	10.82	A	N
71	85.94	6.02	7.00	15.66	17.13	A	7.53	A	A
72	93.00	4.00	4.30	8.60	12.60	A	5.11	A	A
73	90.60	4.50	4.97	11.00	13.67	A	5.68	A	A
74	98.00	5.00	5.10	3.60	14.78	A	5.80	A	A
76	93.20	4.17	4.47	8.40	12.96	A	5.25	A	A
77	82.50	3.70	4.48	19.10	11.97	N	5.26	A	W
78	94.00	13.00	13.83	7.60	34.31	A	14.10	A	A
79	96.50	4.00	4.15	5.10	12.60	A	4.98	A	A
80	127.20	3.70	2.91	25.60	11.97	N	4.01	A	N
81	91.29	7.75	8.49	10.31	21.26	A	8.93	A	A
82	62.90	3.40	5.41	38.70	11.36	N	6.07	A	N
83	96.10	5.30	5.52	5.50	15.46	A	6.17	A	A
84	92.87	5.23	5.63	8.73	15.31	A	6.27	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
85	101.00	4.00	3.96	0.60	12.60	A	4.82	A	A
86	25.54	0.34	1.33	76.06	7.28	N	3.06	A	N
87	86.70	4.30	4.96	14.90	13.24	N	5.67	A	W
88	89.91	3.31	3.68	11.69	11.19	N	4.60	A	W
89	93.85	1.88	2.00	7.75	8.70	A	3.41	A	A
90	60.37	6.19	10.25	41.23	17.53	N	10.62	A	N
91	89.44	3.13	3.50	12.16	10.84	N	4.45	A	W
92	102.90	4.60	4.47	1.30	13.89	A	5.25	A	A
93	103.10	3.30	3.20	1.50	11.17	A	4.22	A	A
94	94.00	4.80	5.11	7.60	14.34	A	5.80	A	A
95	73.60	3.40	4.62	28.00	11.36	N	5.38	A	N
96	103.80	2.80	2.70	2.20	10.22	A	3.86	A	A
97	85.10	4.20	4.94	16.50	13.02	N	5.65	A	W
98	110.00	8.00	7.27	8.40	21.87	A	7.78	A	A
99	145.21	11.15	7.68	43.61	29.66	N	8.16	A	N
100	122.20	3.60	2.95	20.60	11.77	N	4.03	A	N
101	118.78	8.68	7.31	17.18	23.53	A	7.81	A	A
102	83.30	2.06	2.47	18.30	8.97	N	3.70	A	W
103	92.00	4.60	5.00	9.60	13.89	A	5.71	A	A
104	86.20	2.60	3.02	15.40	9.86	N	4.09	A	W
105	93.40	5.00	5.35	8.20	14.78	A	6.02	A	A
106	101.30	11.30	11.15	0.30	30.04	A	11.49	A	A
107	99.38	7.76	7.81	2.22	21.28	A	8.28	A	A
108	99.46	1.42	1.43	2.14	8.10	A	3.10	A	A
109	104.00	2.40	2.31	2.40	9.51	A	3.59	A	A
110	107.00	7.50	7.01	5.40	20.65	A	7.53	A	A
112	98.70	1.00	1.01	2.90	7.67	A	2.94	A	A
114	99.70	9.90	9.93	1.90	26.54	A	10.31	A	A
115	95.20	2.90	3.05	6.40	10.40	A	4.11	A	A
116	77.60	13.50	17.40	24.00	35.57	A	17.61	N	N
117	96.60	2.90	3.00	5.00	10.40	A	4.08	A	A
118	79.52	5.22	6.56	22.08	15.28	N	7.12	A	N
119	97.00	5.00	5.15	4.60	14.78	A	5.85	A	A
120	74.70	8.20	10.98	26.90	22.36	N	11.32	A	N
121	90.40	5.40	5.97	11.20	15.69	A	6.58	A	A
123	92.90	5.10	5.49	8.70	15.01	A	6.14	A	A
124	91.90	1.92	2.09	9.70	8.76	N	3.46	A	W
125	91.60	0.50	0.55	10.00	7.34	N	2.81	A	W
126	96.87	2.50	2.58	4.73	9.68	A	3.78	A	A

Lab code	Rep. Value	Reported Unc.	Unc. [%]	A1	A2	Trueness	P	Precision	Final Score
127	102.50	1.67	1.63	0.90	8.41	A	3.20	A	A

Data evaluation of sample 05, Blank




Fig. I-09: Graphical presentation of the 'Blank' sample results.

Data evaluation of sample 05, Blank

Target value: <0.1 Bq/kg

TABLE I-05: DATA EVALUATION OF THE 'BLANK' - SAMPLE 05

Lab code	Rep. Value	Reported Unc.	Final Score	Lab code	Rep. Value	Reported Unc.	Final Score
1	0.00	0.00	A	65	0.01	0.01	A
2	0.00	0.07	A	67	<0.073		A
3	0.52	0.07	N	69	<0.5		A
4	0.01	0.00	A	70	0.00	0.00	A
5	0.07	0.02	A	71	0.07	0.01	A
6	<0.55		A	72	0.10		A
7	0.00	0.01	A	73	0.19	0.07	N
8	0.01	0.00	A	74	0.13	0.04	A
9	0.28	0.08	N	76	0.00	0.04	A
12	0.03	0.03	A	77	0.07	1.60	N
13	0.73	0.07	N	78	<0.09		A
14	0.01	0.01	A	79	<0.1	0.03	A
15	<0.90		A	80	0.78	0.09	N
16	0.07	0.04	A	81	<0.12		A
17	0.41	0.06	N	82	0.05	0.02	A
18			-	83	0.00	0.00	A
19	0.00	0.00	A	84	0.44	0.18	N
20	4.00	1.00	N	85	3.70	2.00	N
21			-	86	0.00	0.02	A
22	0.04	0.02	A	87	0.04	0.02	A
23	27.00	5.00	N	88	0.00	0.00	A
24	0.06	0.03	A	89	0.02	0.00	A
25	<0.02		A	90	0.00	1.89	A
26	0.01	0.00	A	91	0.01	0.03	A
27	<0.16		A	92	0.02	0.00	A
28	0.00	20.00	-	93	0.60	0.30	N
29	0.02	0.01	A	94	0.00	1.10	A
30	0.20	0.02	N	95	0.42	0.04	N
31	<0.07		A	96	0.00	0.03	A
33	0.07	0.04	A	97	0.01	0.01	A
34	0.20	0.10	A	98	1.30	0.30	N
35	0.01	0.00	A	99	0.38	0.11	N
36	0.00	2.00	-	100	0.00	0.00	-
37	0.00	0.00	A	101	2.45	0.38	N
38	<0.04		A	102	0.00	0.00	A

Lab code	Rep. Value	Reported Unc.	Final Score	Lab code	Rep. Value	Reported Unc.	Final Score
39	0.04	0.03	A	103	<0.17		A
40	0.00			104	0.03	0.01	A
41	0.08	0.03	A	105	0.03	0.01	A
42	0.11	0.01	A	106	0.40	0.31	A
44	0.00	38.00	N	107	0.00	0.01	A
46	0.00	0.00	A	108	0.00	0.01	A
47	0.00	0.00	A	109	<0.1		A
49	0.14	0.05	A	110	3.28	0.28	N
50	0.15	0.22	A	112	0.06		A
51	<0.08		A	114	28.70	2.90	N
52	0.10	0.03	A	115	0.00	0.10	A
54	<0.2		A	116	0.10		A
55	0.03		A	117	<0.025		A
56	0.00	0.00	A	118	6.76	0.49	N
57	<0.15		A	119	<0.0045		A
58	0.00	1.00	N	120	0.05	0.01	A
59	3.20	0.50	N	121	<0.09		A
60	0.06	0.12	A	123	0.05	0.07	A
62	0.64	0.10	N	124	0.02	0.00	A
63	0.03	0.03	A	125	0.02	0.00	A
64	<0.01		A	126	0.00	0.00	A

APPENDIX II. DATA EVALUATION TABLES SORTED BY LABORATORY CODE

All results listed in this Appendix are expressed in Bq/kg units at a reference date set to 2007-April-01. The abbreviations and calculation formulas used in the individual evaluation tables are explained in Section 3 of this report.

The individual laboratory evaluation reports are presented in ascending order of the laboratory code.

Laboratory No. 3

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z-Score	U-Test	Ratio	A1	A2	True	P	Prec	Final Score
Sample 01	52.80	1.40	52.80	4.10	7.77	0.00	0.0	0.0	1.0	0.00	11.18	A	8.21	A	A
Sample 02	101.60	2.80	99.00	7.70	7.78	-2.56	-0.26	-0.32	0.97	2.60	21.14	A	8.25	A	A
Sample 03	52.80	1.40	57.80	4.50	7.79	9.47	0.95	1.06	1.09	5.00	12.16	A	8.22	A	A
Sample 04	101.60	2.80	96.70	7.30	7.55	-4.82	-0.48	-0.63	0.95	4.90	20.17	A	8.04	A	A
Sample 05	0.10	0.01	0.52	0.07											N

Laboratory No. 4

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z-Score	U-Test	Ratio	A1	A2	True	P	Prec	Final Score
Sample 01	52.80	1.40	48.80	2.00	4.10	-7.58	-0.76	-1.64	0.92	4.00	6.30	A	4.88	A	A
Sample 02	101.60	2.80	81.60	3.30	4.04	-19.69	-1.97	-4.62	0.8	20.00	11.17	N	4.89	A	W
Sample 03	52.80	1.40	44.70	1.90	4.25	-15.34	-1.53	-3.43	0.85	8.10	6.09	N	5.01	A	W
Sample 04	101.60	2.80	83.30	3.40	4.08	-18.01	-1.8	-4.15	0.82	18.30	11.36	N	4.92	A	W
Sample 05	0.10	0.01	0.01	0.00											A

Laboratory No. 7

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec	Final Score
Sample 01	52.80	1.40	52.60	1.90	3.61	-0.38	-0.04	-0.08	1.0	0.20	6.09	A	4.48	A	A
Sample 02	101.60	2.80	100.00	3.35	3.35	-1.57	-0.16	-0.37	0.98	1.60	11.26	A	4.34	A	A
Sample 03	52.80	1.40	52.10	1.92	3.69	-1.33	-0.13	-0.29	0.99	0.70	6.13	A	4.54	A	A
Sample 04	101.60	2.80	103.00	3.00	2.91	1.38	0.14	0.34	1.01	1.40	10.59	A	4.01	A	A
Sample 05	0.10	0.01	0.00	0.01											A

Laboratory No. 8

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec	Final Score
Sample 01	52.80	1.40	48.30	1.00	2.07	-8.52	-0.85	-2.62	0.91	4.50	4.44	N	3.36	A	W
Sample 02	101.60	2.80	95.40	2.00	2.10	-6.10	-0.61	-1.8	0.94	6.20	8.88	A	3.46	A	A
Sample 03	52.80	1.40	50.30	1.00	1.99	-4.73	-0.47	-1.45	0.95	2.50	4.44	A	3.31	A	A
Sample 04	101.60	2.80	95.20	2.10	2.21	-6.30	-0.63	-1.83	0.94	6.40	9.03	A	3.53	A	A
Sample 05	0.10	0.01	0.01	0.00											A

Laboratory No. 13

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	36.00	3.00	8.33	-31.82	-3.18	-5.07	0.68	16.80	8.54	N	8.74	A	N
Sample 02	101.60	2.80	72.00	6.00	8.33	-29.13	-2.91	-4.47	0.71	29.60	17.08	N	8.78	A	N
Sample 03	52.80	1.40	32.00	3.00	9.38	-39.39	-3.94	-6.28	0.61	20.80	8.54	N	9.74	A	N
Sample 04	101.60	2.80	67.00	6.00	8.96	-34.06	-3.41	-5.23	0.66	34.60	17.08	N	9.37	A	N
Sample 05	0.10	0.01	0.73	0.07											N

Laboratory No. 14

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	51.00	1.80	3.53	-3.41	-0.34	-0.79	0.97	1.80	5.88	A	4.41	A	A
Sample 02	101.60	2.80	83.90	2.80	3.34	-17.42	-1.74	-4.47	0.83	17.70	10.22	N	4.33	A	W
Sample 03	52.80	1.40	45.50	1.80	3.96	-13.83	-1.38	-3.2	0.86	7.30	5.88	N	4.76	A	W
Sample 04	101.60	2.80	85.90	2.80	3.26	-15.45	-1.55	-3.96	0.85	15.70	10.22	N	4.27	A	W
Sample 05	0.10	0.01	0.01	0.01											A

Laboratory No. 15

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	45.60	2.35	5.15	-13.64	-1.36	-2.63	0.86	7.20	7.06	N	5.80	A	W
Sample 02	101.60	2.80	92.93	4.68	5.04	-8.53	-0.85	-1.59	0.91	8.67	14.07	A	5.74	A	A
Sample 03	52.80	1.40	50.43	2.60	5.16	-4.49	-0.45	-0.8	0.96	2.37	7.62	A	5.80	A	A
Sample 04	101.60	2.80	91.70	4.60	5.02	-9.74	-0.97	-1.84	0.9	9.90	13.89	A	5.72	A	A
Sample 05	0.10	0.01	<0.90												A

Laboratory No. 16

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	58.92	1.42	2.41	11.59	1.16	3.07	1.12	6.12	5.14	N	3.58	A	W
Sample 02	101.60	2.80	102.01	2.56	2.51	0.41	0.04	0.11	1.0	0.41	9.79	A	3.73	A	A
Sample 03	52.80	1.40	50.66	1.49	2.94	-4.06	-0.41	-1.05	0.96	2.14	5.27	A	3.96	A	A
Sample 04	101.60	2.80	97.57	2.53	2.59	-3.97	-0.4	-1.07	0.96	4.03	9.74	A	3.79	A	A
Sample 05	0.10	0.01	0.07	0.04											A

Laboratory No. 17

Sample code	Target Value	Lab. Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U- Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	43.00	1.40	5.40	12.56	-18.56	-1.86	-1.76	0.81	9.80	14.39	12.84	A	12.84	A	A
Sample 02	101.60	75.80	2.80	7.90	10.42	-25.39	-2.54	-3.08	0.75	25.80	21.62	10.78	N	10.78	A	N
Sample 03	52.80	42.20	1.40	5.30	12.56	-20.08	-2.01	-1.93	0.8	10.60	14.14	12.84	A	12.84	A	A
Sample 04	101.60	82.90	2.80	8.90	10.74	-18.41	-1.84	-2.0	0.82	18.70	24.07	11.08	A	11.08	A	A
Sample 05	0.10	0.41	0.01	0.06												N

Laboratory No. 18

Sample code	Target Value	Lab. Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	59.80	1.40	4.70	7.86	13.26	1.33	1.43	1.13	7.00	12.65	8.29	A	8.29	A	A
Sample 02	101.60	109.00	2.80	7.00	6.42	7.28	0.73	0.98	1.07	7.40	19.45	6.99	A	6.99	A	A
Sample 03	52.80	61.60	1.40	3.90	6.33	16.67	1.67	2.12	1.17	8.80	10.69	6.86	A	6.86	A	A
Sample 04	101.60	95.60	2.80	5.80	6.07	-5.91	-0.59	-0.93	0.94	6.00	16.62	6.66	A	6.66	A	A
Sample 05	0.10	0.01	0.01													-

Laboratory No. 19

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	54.02	3.76	6.96	2.31	0.23	0.3	1.02	1.22	10.35	A	7.45	A	A
Sample 02	101.60	2.80	107.28	7.35	6.85	5.59	0.56	0.72	1.06	5.68	20.29	A	7.38	A	A
Sample 03	52.80	1.40	47.97	3.23	6.73	-9.15	-0.91	-1.37	0.91	4.83	9.08	A	7.24	A	A
Sample 04	101.60	2.80	83.97	3.92	4.67	-17.35	-1.74	-3.66	0.83	17.63	12.43	N	5.42	A	W
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 20

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	99.00	5.00	5.05	87.50	8.75	8.9	1.88	46.20	13.40	N	5.70	A	N
Sample 02	101.60	2.80	144.00	7.00	4.86	41.73	4.17	5.62	1.42	42.40	19.45	N	5.59	A	N
Sample 03	52.80	1.40	95.00	5.00	5.26	79.92	7.99	8.13	1.8	42.20	13.40	N	5.89	A	N
Sample 04	101.60	2.80	160.00	7.00	4.38	57.48	5.75	7.75	1.57	58.40	19.45	N	5.17	A	N
Sample 05	0.10	0.01	4.00	1.00											N

Laboratory No. 21

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	49.41	3.01	6.09	-6.42	-0.64	-1.02	0.94	3.39	8.56	A	6.64	A	A
Sample 02	101.60	2.80	101.15	6.16	6.09	-0.44	-0.04	-0.07	1.0	0.45	17.46	A	6.68	A	A
Sample 03	52.80	1.40	51.75	3.15	6.09	-1.99	-0.2	-0.3	0.98	1.05	8.89	A	6.64	A	A
Sample 04	101.60	2.80	100.97	6.15	6.09	-0.62	-0.06	-0.09	0.99	0.63	17.43	A	6.69	A	A
Sample 05	0.10	0.01													-

Laboratory No. 22

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	40.98	2.98	7.27	-22.39	-2.24	-3.59	0.78	11.82	8.49	N	7.74	A	N
Sample 02	101.60	2.80	80.55	5.78	7.18	-20.72	-2.07	-3.28	0.79	21.05	16.57	N	7.69	A	N
Sample 03	52.80	1.40	42.64	3.11	7.29	-19.24	-1.92	-2.98	0.81	10.16	8.80	N	7.76	A	W
Sample 04	101.60	2.80	78.28	5.61	7.17	-22.95	-2.3	-3.72	0.77	23.32	16.18	N	7.68	A	N
Sample 05	0.10	0.01	0.04	0.02											A

Laboratory No. 23

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	102.00	10.00	9.80	93.18	9.32	4.87	1.93	49.20	26.05	N	10.16	A	N
Sample 02	101.60	2.80	148.00	12.00	8.11	45.67	4.57	3.77	1.46	46.40	31.79	N	8.56	A	N
Sample 03	52.80	1.40	85.00	8.00	9.41	60.98	6.1	3.96	1.61	32.20	20.95	N	9.78	A	N
Sample 04	101.60	2.80	140.00	24.00	17.14	37.80	3.78	1.59	1.38	38.40	62.34	A	17.36	N	N
Sample 05	0.10	0.01	27.00	5.00											N

Laboratory No. 24

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	53.76	3.09	5.75	1.82	0.18	0.28	1.02	0.96	8.75	A	6.33	A	A
Sample 02	101.60	2.80	85.75	4.34	5.06	-15.60	-1.56	-3.07	0.84	15.85	13.33	N	5.76	A	W
Sample 03	52.80	1.40	46.81	2.63	5.62	-11.34	-1.13	-2.01	0.89	5.99	7.69	A	6.21	A	A
Sample 04	101.60	2.80	96.91	4.29	4.43	-4.62	-0.46	-0.92	0.95	4.69	13.22	A	5.21	A	A
Sample 05	0.10	0.01	0.06	0.03											A

Laboratory No. 25

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	52.00	1.10	2.12	-1.52	-0.15	-0.45	0.98	0.80	4.59	A	3.39	A	A
Sample 02	101.60	2.80	98.80	2.10	2.13	-2.76	-0.28	-0.8	0.97	2.80	9.03	A	3.48	A	A
Sample 03	52.80	1.40	52.20	1.10	2.11	-1.14	-0.11	-0.34	0.99	0.60	4.59	A	3.39	A	A
Sample 04	101.60	2.80	99.30	2.10	2.11	-2.26	-0.23	-0.66	0.98	2.30	9.03	A	3.47	A	A
Sample 05	0.10	0.01	<0.02												A

Laboratory No. 26

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	48.02	6.66	13.87	-9.05	-0.91	-0.7	0.91	4.78	17.56	A	14.12	A	A
Sample 02	101.60	2.80	97.98	14.44	14.74	-3.56	-0.36	-0.25	0.96	3.62	37.95	A	14.99	A	A
Sample 03	52.80	1.40	47.19	6.38	13.52	-10.63	-1.06	-0.86	0.89	5.61	16.85	A	13.78	A	A
Sample 04	101.60	2.80	91.49	12.16	13.29	-9.95	-1.0	-0.81	0.9	10.11	32.19	A	13.57	A	A
Sample 05	0.10	0.01	0.01	0.00											A

Laboratory No. 27

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	53.10	2.70	5.08	0.57	0.06	0.1	1.01	0.30	7.85	A	5.73	A	A
Sample 02	101.60	2.80	101.60	5.10	5.02	0.00	0.0	0.0	1.0	0.00	15.01	A	5.73	A	A
Sample 03	52.80	1.40	51.00	2.60	5.10	-3.41	-0.34	-0.61	0.97	1.80	7.62	A	5.75	A	A
Sample 04	101.60	2.80	101.30	5.10	5.03	-0.30	-0.03	-0.05	1.0	0.30	15.01	A	5.74	A	A
Sample 05	0.10	0.01	<0.16												A

Laboratory No. 28

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	65.00	10.00	15.38	23.11	2.31	1.21	1.23	12.20	26.05	A	15.61	N	N
Sample 02	101.60	2.80	110.00	15.00	13.64	8.27	0.83	0.55	1.08	8.40	39.37	A	13.91	A	A
Sample 03	52.80	1.40	60.00	8.00	13.33	13.64	1.36	0.89	1.14	7.20	20.95	A	13.59	A	A
Sample 04	101.60	2.80	110.00	15.00	13.64	8.27	0.83	0.55	1.08	8.40	39.37	A	13.91	A	A
Sample 05	0.10	0.01	0.00	20.00											-

Laboratory No. 29

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	41.64	1.03	2.47	-21.14	-2.11	-6.42	0.79	11.16	4.48	N	3.63	A	N
Sample 02	101.60	2.80	55.19	1.11	2.01	-45.68	-4.57	-15.41	0.54	46.41	7.77	N	3.41	A	N
Sample 03	52.80	1.40	43.60	1.16	2.66	-17.42	-1.74	-5.06	0.83	9.20	4.69	N	3.76	A	W
Sample 04	101.60	2.80	46.82	0.86	1.84	-53.92	-5.39	-18.7	0.46	54.78	7.56	N	3.31	A	N
Sample 05	0.10	0.01	0.02	0.01											A

Laboratory No. 30

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	58.15	1.65	2.84	10.13	1.01	2.47	1.1	5.35	5.58	A	3.88	A	A
Sample 02	101.60	2.80	116.50	2.66	2.28	14.67	1.47	3.86	1.15	14.90	9.96	N	3.58	A	W
Sample 03	52.80	1.40	58.70	1.73	2.95	11.17	1.12	2.65	1.11	5.90	5.74	N	3.96	A	W
Sample 04	101.60	2.80	119.40	3.09	2.59	17.52	1.75	4.27	1.18	17.80	10.76	N	3.78	A	W
Sample 05	0.10	0.01	0.20	0.02											N

Laboratory No. 31

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	48.12	0.88	1.83	-8.86	-0.89	-2.83	0.91	4.68	4.27	N	3.22	A	W
Sample 02	101.60	2.80	89.92	1.32	1.47	-11.50	-1.15	-3.77	0.89	11.68	7.99	N	3.12	A	W
Sample 03	52.80	1.40	48.27	0.88	1.82	-8.58	-0.86	-2.74	0.91	4.53	4.27	N	3.22	A	W
Sample 04	101.60	2.80	91.30	1.40	1.53	-10.14	-1.01	-3.29	0.9	10.30	8.08	N	3.15	A	W
Sample 05	0.10	0.01	<0.07												A

Laboratory No. 33

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	58.00	3.00	5.17	9.85	0.98	1.57	1.1	5.20	8.54	A	5.81	A	A
Sample 02	101.60	2.80	113.00	7.00	6.19	11.22	1.12	1.51	1.11	11.40	19.45	A	6.78	A	A
Sample 03	52.80	1.40	59.00	3.00	5.08	11.74	1.17	1.87	1.12	6.20	8.54	A	5.73	A	A
Sample 04	101.60	2.80	109.00	7.00	6.42	7.28	0.73	0.98	1.07	7.40	19.45	A	6.99	A	A
Sample 05	0.10	0.01	0.07	0.04											A

Laboratory No. 34

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U- Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.70	2.30	4.54	-3.98	-0.4	-0.78	0.96	2.10	6.95	A	5.25	A	A
Sample 02	101.60	2.80	100.20	3.60	3.59	-1.38	-0.14	-0.31	0.99	1.40	11.77	A	4.53	A	A
Sample 03	52.80	1.40	50.90	2.20	4.32	-3.60	-0.36	-0.73	0.96	1.90	6.73	A	5.07	A	A
Sample 04	101.60	2.80	98.90	4.50	4.55	-2.66	-0.27	-0.51	0.97	2.70	13.67	A	5.32	A	A
Sample 05	0.10	0.01	0.20	0.10											A

Laboratory No. 35

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	47.90	1.70	3.55	-9.28	-0.93	-2.22	0.91	4.90	5.68	A	4.43	A	A
Sample 02	101.60	2.80	93.60	3.20	3.42	-7.87	-0.79	-1.88	0.92	8.00	10.97	A	4.39	A	A
Sample 03	52.80	1.40	48.40	1.60	3.31	-8.33	-0.83	-2.07	0.92	4.40	5.49	A	4.24	A	A
Sample 04	101.60	2.80	90.60	3.00	3.31	-10.83	-1.08	-2.68	0.89	11.00	10.59	N	4.31	A	W
Sample 05	0.10	0.01	0.01	0.00											A

Laboratory No. 36

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	20.00	2.00	10.00	-62.12	-6.21	-13.44	0.38	32.80	6.30	N	10.35	A	N
Sample 02	101.60	2.80	46.00	2.00	4.35	-54.72	-5.47	-16.16	0.45	55.60	8.88	N	5.15	A	N
Sample 03	52.80	1.40	21.00	2.00	9.52	-60.23	-6.02	-13.03	0.4	31.80	6.30	N	9.89	A	N
Sample 04	101.60	2.80	45.00	2.00	4.44	-55.71	-5.57	-16.45	0.44	56.60	8.88	N	5.23	A	N
Sample 05	0.10	0.01	0.00	2.00											-

Laboratory No. 37

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	54.10	2.90	5.36	2.46	0.25	0.4	1.02	1.30	8.31	A	5.98	A	A
Sample 02	101.60	2.80	105.30	11.00	10.45	3.64	0.36	0.33	1.04	3.70	29.28	A	10.80	A	A
Sample 03	52.80	1.40	65.80	2.70	4.10	24.62	2.46	4.27	1.25	13.00	7.85	N	4.89	A	N
Sample 04	101.60	2.80	122.70	5.10	4.16	20.77	2.08	3.63	1.21	21.10	15.01	N	4.99	A	N
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 38

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	45.90	3.20	6.97	-13.07	-1.31	-1.98	0.87	6.90	9.01	A	7.46	A	A
Sample 02	101.60	2.80	90.80	6.20	6.83	-10.63	-1.06	-1.59	0.89	10.80	17.55	A	7.36	A	A
Sample 03	52.80	1.40	48.30	3.30	6.83	-8.52	-0.85	-1.26	0.91	4.50	9.25	A	7.33	A	A
Sample 04	101.60	2.80	88.20	6.00	6.80	-13.19	-1.32	-2.02	0.87	13.40	17.08	A	7.34	A	A
Sample 05	0.10	0.01	<0.045												A

Laboratory No. 39

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	55.00	3.00	5.45	4.17	0.42	0.66	1.04	2.20	8.54	A	6.06	A	A
Sample 02	101.60	2.80	101.20	5.20	5.14	-0.39	-0.04	-0.07	1.0	0.40	15.24	A	5.83	A	A
Sample 03	52.80	1.40	54.50	3.00	5.50	3.22	0.32	0.51	1.03	1.70	8.54	A	6.11	A	A
Sample 04	101.60	2.80	100.20	5.10	5.09	-1.38	-0.14	-0.24	0.99	1.40	15.01	A	5.79	A	A
Sample 05	0.10	0.01	0.04	0.03											A

Laboratory No. 40

Sample code	Target Value	Lab. Value	Unc	Lab. Unc.	Lab. Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	50.00	1.40	3.40	6.80	-5.30	-0.53	-0.76	0.95	2.80	9.49	A	7.30	A	A
Sample 02	101.60	99.00	2.80	6.60	6.67	-2.56	-0.26	-0.36	0.97	2.60	18.50	A	7.21	A	A
Sample 03	52.80	51.00	1.40	3.40	6.67	-3.41	-0.34	-0.49	0.97	1.80	9.49	A	7.17	A	A
Sample 04	101.60	100.00	2.80	6.70	6.70	-1.57	-0.16	-0.22	0.98	1.60	18.73	A	7.24	A	A
Sample 05	0.10	0.00	0.01											-	-

Laboratory No. 41

Sample code	Target Value	Lab. Value	Unc	Lab. Unc.	Lab. Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	48.36	1.40	1.33	2.75	-8.41	-0.84	-2.3	0.92	4.44	4.98	A	3.82	A	A
Sample 02	101.60	73.54	2.80	1.95	2.65	-27.62	-2.76	-8.22	0.72	28.06	8.80	N	3.82	A	N
Sample 03	52.80	42.20	1.40	1.25	2.96	-20.08	-2.01	-5.65	0.8	10.60	4.84	N	3.98	A	N
Sample 04	101.60	90.68	2.80	2.46	2.71	-10.75	-1.07	-2.93	0.89	10.92	9.62	N	3.87	A	W
Sample 05	0.10	0.08	0.01	0.03											A

Laboratory No. 42

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	45.10	3.20	7.10	-14.58	-1.46	-2.2	0.85	7.70	9.01	A	7.57	A	A
Sample 02	101.60	2.80	84.00	6.10	7.26	-17.32	-1.73	-2.62	0.83	17.60	17.32	N	7.77	A	W
Sample 03	52.80	1.40	46.70	3.30	7.07	-11.55	-1.16	-1.7	0.88	6.10	9.25	A	7.55	A	A
Sample 04	101.60	2.80	84.90	5.90	6.95	-16.44	-1.64	-2.56	0.84	16.70	16.85	A	7.48	A	A
Sample 05	0.10	0.01	0.11	0.01											A

Laboratory No. 43

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	49.90	1.60	3.21	-5.49	-0.55	-1.36	0.95	2.90	5.49	A	4.16	A	A
Sample 02	101.60	2.80	93.60	2.60	2.78	-7.87	-0.79	-2.09	0.92	8.00	9.86	A	3.91	A	A
Sample 03	52.80	1.40	50.70	1.50	2.96	-3.98	-0.4	-1.02	0.96	2.10	5.29	A	3.97	A	A
Sample 04	101.60	2.80	95.20	2.60	2.73	-6.30	-0.63	-1.67	0.94	6.40	9.86	A	3.88	A	A
Sample 05	0.10	0.01													-

Laboratory No. 44

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	36.00	38.00	105.56	-31.82	-3.18	-0.44	0.68	16.80	98.11	A	105.59	N	N
Sample 02	101.60	2.80	106.00	40.00	37.74	4.33	0.43	0.11	1.04	4.40	103.45	A	37.84	N	W
Sample 03	52.80	1.40	9.00	38.00	422.22	-82.95	-8.3	-1.15	0.17	43.80	98.11	A	422.2	N	N
Sample 04	101.60	2.80	25.00	38.00	152.00	-75.39	-7.54	-2.01	0.25	76.60	98.31	A	152.0	N	N
Sample 05	0.10	0.01	0.00	38.00											N

Laboratory No. 46

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	62.35	3.57	5.73	18.08	1.81	2.49	1.18	9.55	9.90	A	6.31	A	A
Sample 02	101.60	2.80	113.80	6.49	5.70	12.00	1.2	1.73	1.12	12.20	18.23	A	6.33	A	A
Sample 03	52.80	1.40	56.21	3.14	5.59	6.45	0.65	0.99	1.06	3.41	8.87	A	6.19	A	A
Sample 04	101.60	2.80	116.70	6.49	5.56	14.86	1.49	2.14	1.15	15.10	18.24	A	6.21	A	A
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 47

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	51.10	2.70	5.28	-3.22	-0.32	-0.56	0.97	1.70	7.85	A	5.91	A	A
Sample 02	101.60	2.80	98.80	7.30	7.39	-2.76	-0.28	-0.36	0.97	2.80	20.17	A	7.89	A	A
Sample 03	52.80	1.40	51.40	2.90	5.64	-2.65	-0.27	-0.43	0.97	1.40	8.31	A	6.23	A	A
Sample 04	101.60	2.80	98.60	7.70	7.81	-2.95	-0.3	-0.37	0.97	3.00	21.14	A	8.28	A	A
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 49

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	6.43	0.86	13.32	-87.81	-8.78	-28.25	0.12	46.37	4.24	N	13.58	A	N
Sample 02	101.60	2.80	2.04	0.34	16.46	-98.00	-9.8	-35.31	0.02	99.56	7.28	N	16.69	N	N
Sample 03	52.80	1.40	9.27	1.18	12.76	-82.44	-8.24	-23.75	0.18	43.53	4.73	N	13.03	A	N
Sample 04	101.60	2.80	8.82	1.08	12.22	-91.32	-9.13	-30.92	0.09	92.78	7.74	N	12.53	A	N
Sample 05	0.10	0.01	0.14	0.05											A

Laboratory No. 50

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	56.57	4.08	7.21	7.14	0.71	0.87	1.07	3.77	11.13	A	7.68	A	A
Sample 02	101.60	2.80	103.99	5.77	5.55	2.35	0.24	0.37	1.02	2.39	16.55	A	6.20	A	A
Sample 03	52.80	1.40	56.86	3.99	7.02	7.69	0.77	0.96	1.08	4.06	10.91	A	7.50	A	A
Sample 04	101.60	2.80	107.98	8.07	7.47	6.28	0.63	0.75	1.06	6.38	22.04	A	7.97	A	A
Sample 05	0.10	0.01	0.15	0.22											A

Laboratory No. 51

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	48.30	5.02	10.39	-8.52	-0.85	-0.86	0.91	4.50	13.45	A	10.73	A	A
Sample 02	101.60	2.80	91.40	9.52	10.42	-10.04	-1.0	-1.03	0.9	10.20	25.60	A	10.77	A	A
Sample 03	52.80	1.40	50.10	5.20	10.38	-5.11	-0.51	-0.5	0.95	2.70	13.89	A	10.71	A	A
Sample 04	101.60	2.80	97.10	10.09	10.39	-4.43	-0.44	-0.43	0.96	4.50	27.02	A	10.75	A	A
Sample 05	0.10	0.01	<0.08												A

Laboratory No. 52

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	34.60	1.42	4.10	-34.47	-3.45	-9.13	0.66	18.20	5.14	N	4.89	A	N
Sample 02	101.60	2.80	86.50	3.50	4.05	-14.86	-1.49	-3.37	0.85	15.10	11.56	N	4.90	A	W
Sample 03	52.80	1.40	46.96	1.93	4.11	-11.06	-1.11	-2.45	0.89	5.84	6.15	A	4.89	A	A
Sample 04	101.60	2.80	70.70	2.88	4.07	-30.41	-3.04	-7.69	0.7	30.90	10.36	N	4.92	A	N
Sample 05	0.10	0.01	0.10	0.03											A

Laboratory No. 54

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	71.20	4.90	6.88	34.85	3.48	3.61	1.35	18.40	13.15	N	7.38	A	N
Sample 02	101.60	2.80	132.00	8.30	6.29	29.92	2.99	3.47	1.3	30.40	22.60	N	6.87	A	N
Sample 03	52.80	1.40	70.90	4.70	6.63	34.28	3.43	3.69	1.34	18.10	12.65	N	7.14	A	N
Sample 04	101.60	2.80	129.00	8.50	6.59	26.97	2.7	3.06	1.27	27.40	23.09	N	7.14	A	N
Sample 05	0.10	0.01	<0.2												A

Laboratory No. 55

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	46.00	3.00	6.52	-12.88	-1.29	-2.05	0.87	6.80	8.54	A	7.04	A	A
Sample 02	101.60	2.80	89.00	5.00	5.62	-12.40	-1.24	-2.2	0.88	12.60	14.78	A	6.26	A	A
Sample 03	52.80	1.40	46.00	2.00	4.35	-12.88	-1.29	-2.79	0.87	6.80	6.30	N	5.09	A	W
Sample 04	101.60	2.80	91.00	5.00	5.49	-10.43	-1.04	-1.85	0.9	10.60	14.78	A	6.15	A	A
Sample 05	0.10	0.01	0.03												A

Laboratory No. 56

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	53.00	4.00	7.55	0.38	0.04	0.05	1.0	0.20	10.93	A	8.00	A	A
Sample 02	101.60	2.80	84.00	6.00	7.14	-17.32	-1.73	-2.66	0.83	17.60	17.08	N	7.66	A	W
Sample 03	52.80	1.40	45.00	3.00	6.67	-14.77	-1.48	-2.36	0.85	7.80	8.54	A	7.17	A	A
Sample 04	101.60	2.80	92.00	7.00	7.61	-9.45	-0.94	-1.27	0.91	9.60	19.45	A	8.09	A	A
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 57

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z-Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	65.60	3.00	4.57	24.24	2.42	3.87	1.24	12.80	8.54	N	5.29	A	N
Sample 02	101.60	2.80	116.10	2.50	2.15	14.27	1.43	3.86	1.14	14.50	9.68	N	3.50	A	W
Sample 03	52.80	1.40	54.10	2.50	4.62	2.46	0.25	0.45	1.02	1.30	7.39	A	5.33	A	A
Sample 04	101.60	2.80	102.80	3.40	3.31	1.18	0.12	0.27	1.01	1.20	11.30	A	4.31	A	A
Sample 05	0.10	0.01	<0.15								6				A

Laboratory No. 58

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z-Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	51.70	0.40	0.77	-2.08	-0.21	-0.76	0.98	1.10	3.76	A	2.76	A	A
Sample 02	101.60	2.80	99.60	0.50	0.50	-1.97	-0.2	-0.7	0.98	2.00	7.34	A	2.80	A	A
Sample 03	52.80	1.40	51.70	0.40	0.77	-2.08	-0.21	-0.76	0.98	1.10	3.76	A	2.76	A	A
Sample 04	101.60	2.80	99.60	0.50	0.50	-1.97	-0.2	-0.7	0.98	2.00	7.34	A	2.80	A	A
Sample 05	0.10	0.01	0.00	1.00											N

Laboratory No. 59

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	71.00	7.00	9.86	34.47	3.45	2.55	1.34	18.20	18.42	A	10.21	A	A
Sample 02	101.60	2.80	118.00	10.00	8.47	16.14	1.61	1.58	1.16	16.40	26.79	A	8.91	A	A
Sample 03	52.80	1.40	64.00	5.00	7.81	21.21	2.12	2.16	1.21	11.20	13.40	A	8.25	A	A
Sample 04	101.60	2.80	117.00	10.00	8.55	15.16	1.52	1.48	1.15	15.40	26.79	A	8.98	A	A
Sample 05	0.10	0.01	3.20	0.50											N

Laboratory No. 60

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	58.40	1.20	2.05	10.61	1.06	3.04	1.11	5.60	4.76	N	3.35	A	W
Sample 02	101.60	2.80	110.70	2.10	1.90	8.96	0.9	2.6	1.09	9.10	9.03	N	3.35	A	W
Sample 03	52.80	1.40	62.10	1.30	2.09	17.61	1.76	4.87	1.18	9.30	4.93	N	3.38	A	W
Sample 04	101.60	2.80	113.10	2.20	1.95	11.32	1.13	3.23	1.11	11.50	9.19	N	3.37	A	W
Sample 05	0.10	0.01	0.06	0.12											A

Laboratory No. 62

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	49.00	2.00	4.08	-7.20	-0.72	-1.56	0.93	3.80	6.30	A	4.87	A	A
Sample 02	101.60	2.80	96.00	3.00	3.13	-5.51	-0.55	-1.36	0.94	5.60	10.59	A	4.17	A	A
Sample 03	52.80	1.40	47.00	2.00	4.26	-10.98	-1.1	-2.38	0.89	5.80	6.30	A	5.01	A	A
Sample 04	101.60	2.80	79.00	2.00	2.53	-22.24	-2.22	-6.57	0.78	22.60	8.88	N	3.74	A	N
Sample 05	0.10	0.01	0.64	0.10											N

Laboratory No. 63

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	39.50	4.42	11.19	-25.19	-2.52	-2.87	0.75	13.30	11.96	N	11.50	A	N
Sample 02	101.60	2.80	82.90	7.53	9.08	-18.41	-1.84	-2.33	0.82	18.70	20.73	A	9.49	A	A
Sample 03	52.80	1.40	45.55	4.97	10.91	-13.73	-1.37	-1.4	0.86	7.25	13.32	A	11.23	A	A
Sample 04	101.60	2.80	85.22	4.71	5.53	-16.12	-1.61	-2.99	0.84	16.38	14.14	N	6.18	A	W
Sample 05	0.10	0.01	0.03	0.03											A

Laboratory No. 67

Sample code	Target Value	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	50.00	1.70	3.40	-5.30	-0.53	-1.27	0.95	2.80	5.68	A	4.31	A	A
Sample 02	101.60	95.80	3.10	3.24	-5.71	-0.57	-1.39	0.94	5.80	10.78	A	4.25	A	A
Sample 03	52.80	52.60	1.80	3.42	-0.38	-0.04	-0.09	1.0	0.20	5.88	A	4.33	A	A
Sample 04	101.60	97.20	3.10	3.19	-4.33	-0.43	-1.05	0.96	4.40	10.78	A	4.22	A	A
Sample 05	0.10	<0.073												A

Laboratory No. 69

Sample code	Target Value	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	46.00	1.40	3.04	-12.88	-1.29	-3.43	0.87	6.80	5.11	N	4.04	A	W
Sample 02	101.60	93.70	5.10	5.44	-7.78	-0.78	-1.36	0.92	7.90	15.0	A	6.10	A	A
Sample 03	52.80	47.60	2.00	4.20	-9.85	-0.98	-2.13	0.9	5.20	6.30	A	4.97	A	A
Sample 04	101.60	86.80	2.00	2.30	-14.57	-1.46	-4.3	0.85	14.8	8.88	N	3.59	A	W
Sample 05	0.10	<0.5												A

Laboratory No. 70

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	59.00	6.60	11.19	11.74	1.17	0.92	1.12	6.20	17.41	A	11.50	A	A
Sample 02	101.60	2.80	53.90	5.82	10.80	-46.95	-4.69	-7.39	0.53	47.70	16.66	N	11.14	A	N
Sample 03	52.80	1.40	15.11	1.56	10.34	-71.38	-7.14	-17.96	0.29	37.69	5.41	N	10.68	A	N
Sample 04	101.60	2.80	6.69	0.70	10.46	-93.42	-9.34	-32.88	0.07	94.91	7.45	N	10.82	A	N
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 71

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	47.37	3.32	7.01	-10.28	-1.03	-1.51	0.9	5.43	9.30	A	7.49	A	A
Sample 02	101.60	2.80	90.93	6.37	7.01	-10.50	-1.05	-1.53	0.89	10.67	17.95	A	7.53	A	A
Sample 03	52.80	1.40	44.45	3.11	7.00	-15.81	-1.58	-2.45	0.84	8.35	8.80	A	7.48	A	A
Sample 04	101.60	2.80	85.94	6.02	7.00	-15.41	-1.54	-2.36	0.85	15.66	17.13	A	7.53	A	A
Sample 05	0.10	0.01	0.07	0.01											A

Laboratory No. 72

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	47.50	1.90	4.00	-10.04	-1.0	-2.25	0.9	5.30	6.09	A	4.80	A	A
Sample 02	101.60	2.80	88.00	3.00	3.41	-13.39	-1.34	-3.31	0.87	13.60	10.59	N	4.38	A	W
Sample 03	52.80	1.40	47.00	2.00	4.26	-10.98	-1.1	-2.38	0.89	5.80	6.30	A	5.01	A	A
Sample 04	101.60	2.80	93.00	4.00	4.30	-8.46	-0.85	-1.76	0.92	8.60	12.60	A	5.11	A	A
Sample 05	0.10	0.01	0.10												A

Laboratory No. 73

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.60	2.70	5.34	-4.17	-0.42	-0.72	0.96	2.20	7.85	A	5.96	A	A
Sample 02	101.60	2.80	90.60	4.70	5.19	-10.83	-1.08	-2.01	0.89	11.00	14.11	A	5.87	A	A
Sample 03	52.80	1.40	50.50	2.70	5.35	-4.36	-0.44	-0.76	0.96	2.30	7.85	A	5.97	A	A
Sample 04	101.60	2.80	90.60	4.50	4.97	-10.83	-1.08	-2.08	0.89	11.00	13.67	A	5.68	A	A
Sample 05	0.10	0.01	0.19	0.07											N

Laboratory No. 74

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	48.00	3.00	6.25	-9.09	-0.91	-1.45	0.91	4.80	8.54	A	6.79	A	A
Sample 02	101.60	2.80	95.00	5.00	5.26	-6.50	-0.65	-1.15	0.94	6.60	14.78	A	5.94	A	A
Sample 03	52.80	1.40	49.00	3.00	6.12	-7.20	-0.72	-1.15	0.93	3.80	8.54	A	6.67	A	A
Sample 04	101.60	2.80	98.00	5.00	5.10	-3.54	-0.35	-0.63	0.96	3.60	14.78	A	5.80	A	A
Sample 05	0.10	0.01	0.13	0.04											A

Laboratory No. 76

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.23	2.26	4.50	-4.87	-0.49	-0.97	0.95	2.57	6.86	A	5.22	A	A
Sample 02	101.60	2.80	93.40	4.18	4.48	-8.07	-0.81	-1.63	0.92	8.20	12.98	A	5.26	A	A
Sample 03	52.80	1.40	49.30	2.23	4.52	-6.63	-0.66	-1.33	0.93	3.50	6.79	A	5.24	A	A
Sample 04	101.60	2.80	93.20	4.17	4.47	-8.27	-0.83	-1.67	0.92	8.40	12.96	A	5.25	A	A
Sample 05	0.10	0.01	0.00	0.04											A

Laboratory No. 77

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	35.00	3.10	8.86	-33.71	-3.37	-5.23	0.66	17.80	8.78	N	9.25	A	N
Sample 02	101.60	2.80	63.50	3.80	5.98	-37.50	-3.75	-8.07	0.63	38.10	12.18	N	6.59	A	N
Sample 03	52.80	1.40	31.60	4.00	12.66	-40.15	-4.02	-5.0	0.6	21.20	10.93	N	12.93	A	N
Sample 04	101.60	2.80	82.50	3.70	4.48	-18.80	-1.88	-4.12	0.81	19.10	11.97	N	5.26	A	W
Sample 05	0.10	0.01	0.07	1.60											N

Laboratory No. 78

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	49.30	6.60	13.39	-6.63	-0.66	-0.52	0.93	3.50	17.41	A	13.65	A	A
Sample 02	101.60	2.80	93.00	12.00	12.90	-8.46	-0.85	-0.7	0.92	8.60	31.79	A	13.19	A	A
Sample 03	52.80	1.40	50.20	6.70	13.35	-4.92	-0.49	-0.38	0.95	2.60	17.66	A	13.61	A	A
Sample 04	101.60	2.80	94.00	13.00	13.83	-7.48	-0.75	-0.57	0.93	7.60	34.31	A	14.10	A	A
Sample 05	0.10	0.01	<0.09												A

Laboratory No. 79

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	46.70	2.00	4.28	-11.55	-1.16	-2.5	0.88	6.10	6.30	A	5.04	A	A
Sample 02	101.60	2.80	93.30	4.00	4.29	-8.17	-0.82	-1.7	0.92	8.30	12.60	A	5.10	A	A
Sample 03	52.80	1.40	49.10	2.00	4.07	-7.01	-0.7	-1.52	0.93	3.70	6.30	A	4.86	A	A
Sample 04	101.60	2.80	96.50	4.00	4.15	-5.02	-0.5	-1.04	0.95	5.10	12.60	A	4.98	A	A
Sample 05	0.10	0.01	<0.1	0.03											A

Laboratory No. 80

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	73.88	2.17	2.94	39.92	3.99	8.16	1.4	21.08	6.66	N	3.96	A	N
Sample 02	101.60	2.80	130.40	4.00	3.07	28.35	2.83	5.9	1.28	28.80	12.60	N	4.12	A	N
Sample 03	52.80	1.40	70.00	2.35	3.36	32.58	3.26	6.29	1.33	17.20	7.06	N	4.28	A	N
Sample 04	101.60	2.80	127.20	3.70	2.91	25.20	2.52	5.52	1.25	25.60	11.97	N	4.01	A	N
Sample 05	0.10	0.01	0.78	0.09											N

Laboratory No. 81

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	48.81	4.14	8.48	-7.56	-0.76	-0.91	0.92	3.99	11.28	A	8.89	A	A
Sample 02	101.60	2.80	90.89	7.72	8.49	-10.54	-1.05	-1.3	0.89	10.71	21.19	A	8.93	A	A
Sample 03	52.80	1.40	48.05	4.08	8.49	-9.00	-0.9	-1.1	0.91	4.75	11.13	A	8.90	A	A
Sample 04	101.60	2.80	91.29	7.75	8.49	-10.15	-1.01	-1.25	0.9	10.31	21.26	A	8.93	A	A
Sample 05	0.10	0.01	<0.12												A

Laboratory No. 82

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	40.00	3.00	7.50	-24.24	-2.42	-3.87	0.76	12.80	8.54	N	7.95	A	N
Sample 02	101.60	2.80	57.70	4.20	7.28	-43.21	-4.32	-8.7	0.57	43.90	13.02	N	7.78	A	N
Sample 03	52.80	1.40	37.80	3.30	8.73	-28.41	-2.84	-4.18	0.72	15.00	9.25	N	9.12	A	N
Sample 04	101.60	2.80	62.90	3.40	5.41	-38.09	-3.81	-8.79	0.62	38.70	11.36	N	6.07	A	N
Sample 05	0.10	0.01	0.05	0.02											A

Laboratory No. 83

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.80	2.90	5.71	-3.79	-0.38	-0.62	0.96	2.00	8.31	A	6.29	A	A
Sample 02	101.60	2.80	101.70	5.70	5.60	0.10	0.01	0.02	1.0	0.10	16.38	A	6.25	A	A
Sample 03	52.80	1.40	54.30	3.00	5.52	2.84	0.28	0.45	1.03	1.50	8.54	A	6.13	A	A
Sample 04	101.60	2.80	96.10	5.30	5.52	-5.41	-0.54	-0.92	0.95	5.50	15.46	A	6.17	A	A
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 84

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.69	3.13	6.17	-4.00	-0.4	-0.62	0.96	2.11	8.85	A	6.72	A	A
Sample 02	101.60	2.80	97.68	5.44	5.57	-3.86	-0.39	-0.64	0.96	3.92	15.79	A	6.21	A	A
Sample 03	52.80	1.40	52.17	3.29	6.31	-1.19	-0.12	-0.18	0.99	0.63	9.22	A	6.84	A	A
Sample 04	101.60	2.80	92.87	5.23	5.63	-8.59	-0.86	-1.47	0.91	8.73	15.31	A	6.27	A	A
Sample 05	0.10	0.01	0.44	0.18											N

Laboratory No. 85

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	54.00	2.00	3.70	2.27	0.23	0.49	1.02	1.20	6.30	A	4.55	A	A
Sample 02	101.60	2.80	106.00	4.00	3.77	4.33	0.43	0.9	1.04	4.40	12.60	A	4.67	A	A
Sample 03	52.80	1.40	51.00	2.00	3.92	-3.41	-0.34	-0.74	0.97	1.80	6.30	A	4.73	A	A
Sample 04	101.60	2.80	101.00	4.00	3.96	-0.59	-0.06	-0.12	0.99	0.60	12.60	A	4.82	A	A
Sample 05	0.10	0.01	3.70	2.00											N

Laboratory No. 86

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	7.12	0.17	2.39	-86.52	-8.65	-32.39	0.13	45.68	3.64	N	3.57	A	N
Sample 02	101.60	2.80	15.53	0.29	1.87	-84.71	-8.47	-30.58	0.15	86.07	7.26	N	3.33	A	N
Sample 03	52.80	1.40	11.94	0.24	2.01	-77.39	-7.74	-28.77	0.23	40.86	3.66	N	3.33	A	N
Sample 04	101.60	2.80	25.54	0.34	1.33	-74.86	-7.49	-26.97	0.25	76.06	7.28	N	3.06	A	N
Sample 05	0.10	0.01	0.00	0.02											A

Laboratory No. 87

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	44.80	2.20	4.91	-15.15	-1.52	-3.07	0.85	8.00	6.73	N	5.58	A	W
Sample 02	101.60	2.80	84.00	5.00	5.95	-17.32	-1.73	-3.07	0.83	17.60	14.78	N	6.56	A	W
Sample 03	52.80	1.40	44.10	2.20	4.99	-16.48	-1.65	-3.34	0.84	8.70	6.73	N	5.65	A	W
Sample 04	101.60	2.80	86.70	4.30	4.96	-14.67	-1.47	-2.9	0.85	14.90	13.24	N	5.67	A	W
Sample 05	0.10	0.01	0.04	0.02											A

Laboratory No. 88

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	43.03	1.71	3.97	-18.50	-1.85	-4.42	0.81	9.77	5.70	N	4.78	A	W
Sample 02	101.60	2.80	89.31	3.28	3.67	-12.10	-1.21	-2.85	0.88	12.29	11.13	N	4.59	A	W
Sample 03	52.80	1.40	46.91	1.77	3.77	-11.16	-1.12	-2.61	0.89	5.89	5.82	N	4.61	A	W
Sample 04	101.60	2.80	89.91	3.31	3.68	-11.51	-1.15	-2.7	0.88	11.69	11.19	N	4.60	A	W
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 89

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.09	2.00	4.00	-5.14	-0.51	-1.11	0.95	2.71	6.30	A	4.80	A	A
Sample 02	101.60	2.80	86.53	1.73	2.00	-14.83	-1.48	-4.58	0.85	15.07	8.49	N	3.40	A	W
Sample 03	52.80	1.40	49.14	0.98	2.00	-6.93	-0.69	-2.14	0.93	3.66	4.41	A	3.32	A	A
Sample 04	101.60	2.80	93.85	1.88	2.00	-7.63	-0.76	-2.3	0.92	7.75	8.70	A	3.41	A	A
Sample 05	0.10	0.01	0.02	0.00											A

Laboratory No. 90

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	45.31	5.51	12.16	-14.19	-1.42	-1.32	0.86	7.49	14.67	A	12.45	A	A
Sample 02	101.60	2.80	99.87	7.67	7.68	-1.70	-0.17	-0.21	0.98	1.73	21.07	A	8.16	A	A
Sample 03	52.80	1.40	32.32	4.81	14.88	-38.79	-3.88	-4.09	0.61	20.48	12.92	N	15.12	N	N
Sample 04	101.60	2.80	60.37	6.19	10.25	-40.58	-4.06	-6.07	0.59	41.23	17.53	N	10.62	A	N
Sample 05	0.10	0.01	0.00	1.89											A

Laboratory No. 91

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Lab. Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.18	2.36	4.70	-4.96	-0.5	-0.96	0.95	2.62	7.07	A	5.39	A	A
Sample 02	101.60	2.80	95.77	4.41	4.60	-5.74	-0.57	-1.12	0.94	5.83	13.48	A	5.37	A	A
Sample 03	52.80	1.40	51.20	2.38	4.65	-3.03	-0.3	-0.58	0.97	1.60	7.12	A	5.35	A	A
Sample 04	101.60	2.80	89.44	3.13	3.50	-11.97	-1.2	-2.9	0.88	12.16	10.84	N	4.45	A	W
Sample 05	0.10	0.01	0.01	0.03											A

Laboratory No. 92

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Lab. Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	49.30	3.80	7.71	-6.63	-0.66	-0.86	0.93	3.50	10.45	A	8.15	A	A
Sample 02	101.60	2.80	99.90	4.70	4.70	-1.67	-0.17	-0.31	0.98	1.70	14.11	A	5.45	A	A
Sample 03	52.80	1.40	51.70	2.70	5.22	-2.08	-0.21	-0.36	0.98	1.10	7.85	A	5.86	A	A
Sample 04	101.60	2.80	102.90	4.60	4.47	1.28	0.13	0.24	1.01	1.30	13.89	A	5.25	A	A
Sample 05	0.10	0.01	0.02	0.00											A

Laboratory No. 93

Sample code	Target Value	Lab. Value	Unc	Lab. Unc.	Lab. Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	56.20	1.40	2.10	3.74	6.44	0.64	1.35	1.06	3.40	6.51	A	4.58	A	A
Sample 02	101.60	96.10	2.80	3.10	3.23	-5.41	-0.54	-1.32	0.95	5.50	10.78	A	4.24	A	A
Sample 03	52.80	50.30	1.40	2.00	3.98	-4.73	-0.47	-1.02	0.95	2.50	6.30	A	4.78	A	A
Sample 04	101.60	103.10	2.80	3.30	3.20	1.48	0.15	0.35	1.01	1.50	11.17	A	4.22	A	A
Sample 05	0.10	0.60	0.01	0.30											N

Laboratory No. 94

Sample code	Target Value	Lab. Value	Unc	Lab. Unc.	Lab. Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	51.30	1.40	3.60	7.02	-2.84	-0.28	-0.39	0.97	1.50	9.97	A	7.50	A	A
Sample 02	101.60	98.00	2.80	4.90	5.00	-3.54	-0.35	-0.64	0.96	3.60	14.56	A	5.71	A	A
Sample 03	52.80	46.70	1.40	3.50	7.49	-11.55	-1.16	-1.62	0.88	6.10	9.73	A	7.95	A	A
Sample 04	101.60	94.00	2.80	4.80	5.11	-7.48	-0.75	-1.37	0.93	7.60	14.34	A	5.80	A	A
Sample 05	0.10	0.00	0.01	1.10											A

Laboratory No. 95

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	42.90	2.10	4.90	-18.75	-1.87	-3.92	0.81	9.90	6.51	N	5.57	A	W
Sample 02	101.60	2.80	83.50	4.00	4.79	-17.81	-1.78	-3.71	0.82	18.10	12.60	N	5.53	A	W
Sample 03	52.80	1.40	42.00	2.50	5.95	-20.45	-2.05	-3.77	0.8	10.80	7.39	N	6.52	A	N
Sample 04	101.60	2.80	73.60	3.40	4.62	-27.56	-2.76	-6.36	0.72	28.00	11.36	N	5.38	A	N
Sample 05	0.10	0.01	0.42	0.04											N

Laboratory No. 96

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	52.10	1.50	2.88	-1.33	-0.13	-0.34	0.99	0.70	5.29	A	3.91	A	A
Sample 02	101.60	2.80	100.30	2.70	2.69	-1.28	-0.13	-0.33	0.99	1.30	10.04	A	3.85	A	A
Sample 03	52.80	1.40	51.90	1.50	2.89	-1.70	-0.17	-0.44	0.98	0.90	5.29	A	3.92	A	A
Sample 04	101.60	2.80	103.80	2.80	2.70	2.17	0.22	0.56	1.02	2.20	10.22	A	3.86	A	A
Sample 05	0.10	0.01	0.00	0.03											A

Laboratory No. 97

Sample code	Target Value	Lab. Value	Unc	Lab. Unc.	Lab. Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	45.60	1.40	2.40	5.26	-13.64	-1.36	-2.59	0.86	7.20	7.17	N	5.89	A	W
Sample 02	101.60	87.80	2.80	3.50	3.99	-13.58	-1.36	-3.08	0.86	13.80	11.56	N	4.85	A	W
Sample 03	52.80	45.70	1.40	2.40	5.25	-13.45	-1.34	-2.56	0.87	7.10	7.17	A	5.88	A	A
Sample 04	101.60	85.10	2.80	4.20	4.94	-16.24	-1.62	-3.27	0.84	16.50	13.02	N	5.65	A	W
Sample 05	0.10	0.01	0.01	0.01	0.01										A

Laboratory No. 98

Sample code	Target Value	Lab. Value	Unc	Lab. Unc.	Lab. Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	32.00	1.40	3.00	9.38	-39.39	-3.94	-6.28	0.61	20.80	8.54	N	9.74	A	N
Sample 02	101.60	86.00	2.80	6.00	6.98	-15.35	-1.54	-2.36	0.85	15.60	17.08	A	7.50	A	A
Sample 03	52.80	55.00	1.40	4.00	7.27	4.17	0.42	0.52	1.04	2.20	10.93	A	7.74	A	A
Sample 04	101.60	110.00	2.80	8.00	7.27	8.27	0.83	0.99	1.08	8.40	21.87	A	7.78	A	A
Sample 05	0.10	1.30	0.01	0.30											N

Laboratory No. 99

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	74.82	5.80	7.75	41.70	4.17	3.69	1.42	22.02	15.39	N	8.19	A	N
Sample 02	101.60	2.80	140.00	10.62	7.59	37.80	3.78	3.5	1.38	38.40	28.34	N	8.07	A	N
Sample 03	52.80	1.40	72.31	5.75	7.95	36.95	3.7	3.3	1.37	19.51	15.27	N	8.38	A	N
Sample 04	101.60	2.80	145.21	11.15	7.68	42.92	4.29	3.79	1.43	43.61	29.66	N	8.16	A	N
Sample 05	0.10	0.01	0.38	0.11											N

Laboratory No. 100

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	61.60	1.80	2.92	16.67	1.67	3.86	1.17	8.80	5.88	N	3.95	A	W
Sample 02	101.60	2.80	117.40	4.00	3.41	15.55	1.56	3.24	1.16	15.80	12.60	N	4.38	A	W
Sample 03	52.80	1.40	61.20	1.90	3.10	15.91	1.59	3.56	1.16	8.40	6.09	N	4.08	A	W
Sample 04	101.60	2.80	122.20	3.60	2.95	20.28	2.03	4.52	1.2	20.60	11.77	N	4.03	A	N
Sample 05	0.10	0.01	0.00	0.00											-

Laboratory No. 101

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	115.42	9.54	8.26	118.60	11.86	6.5	2.19	62.62	24.87	N	8.68	A	N
Sample 02	101.60	2.80	89.70	6.38	7.11	-11.71	-1.17	-1.71	0.88	11.89	17.96	A	7.62	A	A
Sample 03	52.80	1.40	72.09	5.67	7.87	36.53	3.65	3.3	1.37	19.29	15.08	N	8.30	A	N
Sample 04	101.60	2.80	118.78	8.68	7.31	16.91	1.69	1.88	1.17	17.18	23.53	A	7.81	A	A
Sample 05	0.10	0.01	2.45	0.38											N

Laboratory No. 102

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	45.70	1.35	2.95	-13.45	-1.34	-3.65	0.87	7.10	5.02	N	3.97	A	W
Sample 02	101.60	2.80	87.10	2.26	2.59	-14.27	-1.43	-4.03	0.86	14.50	9.28	N	3.79	A	W
Sample 03	52.80	1.40	45.00	1.28	2.84	-14.77	-1.48	-4.11	0.85	7.80	4.89	N	3.89	A	W
Sample 04	101.60	2.80	83.30	2.06	2.47	-18.01	-1.8	-5.26	0.82	18.30	8.97	N	3.70	A	W
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 103

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	48.60	2.92	6.01	-7.95	-0.8	-1.3	0.92	4.20	8.35	A	6.57	A	A
Sample 02	101.60	2.80	89.40	4.47	5.00	-12.01	-1.2	-2.31	0.88	12.20	13.61	A	5.71	A	A
Sample 03	52.80	1.40	54.30	3.80	7.00	2.84	0.28	0.37	1.03	1.50	10.45	A	7.48	A	A
Sample 04	101.60	2.80	92.00	4.60	5.00	-9.45	-0.94	-1.78	0.91	9.60	13.89	A	5.71	A	A
Sample 05	0.10	0.01	<0.17												A

Laboratory No. 104

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	48.30	1.70	3.52	-8.52	-0.85	-2.04	0.91	4.50	5.68	A	4.41	A	A
Sample 02	101.60	2.80	90.10	2.90	3.22	-11.32	-1.13	-2.85	0.89	11.50	10.40	N	4.24	A	W
Sample 03	52.80	1.40	54.50	1.80	3.30	3.22	0.32	0.75	1.03	1.70	5.88	A	4.24	A	A
Sample 04	101.60	2.80	86.20	2.60	3.02	-15.16	-1.52	-4.03	0.85	15.40	9.86	N	4.09	A	W
Sample 05	0.10	0.01	0.03	0.01											A

Laboratory No. 105

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	48.30	2.60	5.38	-8.52	-0.85	-1.52	0.91	4.50	7.62	A	6.00	A	A
Sample 02	101.60	2.80	92.30	4.90	5.31	-9.15	-0.92	-1.65	0.91	9.30	14.56	A	5.98	A	A
Sample 03	52.80	1.40	48.80	2.60	5.33	-7.58	-0.76	-1.35	0.92	4.00	7.62	A	5.95	A	A
Sample 04	101.60	2.80	93.40	5.00	5.35	-8.07	-0.81	-1.43	0.92	8.20	14.78	A	6.02	A	A
Sample 05	0.10	0.01	0.03	0.01											A

Laboratory No. 106

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	57.70	6.40	11.09	9.28	0.93	0.75	1.09	4.90	16.90	A	11.40	A	A
Sample 02	101.60	2.80	103.30	11.60	11.23	1.67	0.17	0.14	1.02	1.70	30.79	A	11.56	A	A
Sample 03	52.80	1.40	58.20	6.50	11.17	10.23	1.02	0.81	1.1	5.40	17.15	A	11.48	A	A
Sample 04	101.60	2.80	101.30	11.30	11.15	-0.30	-0.03	-0.03	1.0	0.30	30.04	A	11.49	A	A
Sample 05	0.10	0.01	0.40	0.31											A

Laboratory No. 107

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	52.40	2.39	4.56	-0.76	-0.08	-0.14	0.99	0.40	7.15	A	5.28	A	A
Sample 02	101.60	2.80	100.24	9.24	9.22	-1.34	-0.13	-0.14	0.99	1.36	24.91	A	9.62	A	A
Sample 03	52.80	1.40	53.93	4.48	8.31	2.14	0.21	0.24	1.02	1.13	12.11	A	8.72	A	A
Sample 04	101.60	2.80	99.38	7.76	7.81	-2.19	-0.22	-0.27	0.98	2.22	21.28	A	8.28	A	A
Sample 05	0.10	0.01	0.00	0.01											A

Laboratory No. 108

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	51.23	0.94	1.83	-2.97	-0.3	-0.93	0.97	1.57	4.35	A	3.22	A	A
Sample 02	101.60	2.80	98.27	1.58	1.61	-3.28	-0.33	-1.04	0.97	3.33	8.29	A	3.19	A	A
Sample 03	52.80	1.40	49.95	0.87	1.74	-5.40	-0.54	-1.73	0.95	2.85	4.25	A	3.17	A	A
Sample 04	101.60	2.80	99.46	1.42	1.43	-2.11	-0.21	-0.68	0.98	2.14	8.10	A	3.10	A	A
Sample 05	0.10	0.01	0.00	0.01											A

Laboratory No. 109

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U- Test	Ratio	A1	A2	True P	Prec .	Final Score	
Sample 01	52.80	1.40	52.20	1.30	2.49	-1.14	-0.11	-0.31	0.99	0.60	4.93	A	3.64	A	A
Sample 02	101.60	2.80	101.00	2.30	2.28	-0.59	-0.06	-0.17	0.99	0.60	9.35	A	3.58	A	A
Sample 03	52.80	1.40	51.80	1.30	2.51	-1.89	-0.19	-0.52	0.98	1.00	4.93	A	3.65	A	A
Sample 04	101.60	2.80	104.00	2.40	2.31	2.36	0.24	0.65	1.02	2.40	9.51	A	3.59	A	A
Sample 05	0.10	0.01	<0.1												A

Laboratory No. 110

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test Ratio	A1	A2	True P	Prec.	Final Score		
Sample 01	52.80	1.40	63.50	4.40	6.93	20.27	2.03	2.32	1.2	10.70	11.91	A	7.42	A	A
Sample 02	101.60	2.80	121.00	8.50	7.02	19.09	1.91	2.17	1.19	19.40	23.09	A	7.55	A	A
Sample 03	52.80	1.40	56.20	3.90	6.94	6.44	0.64	0.82	1.06	3.40	10.69	A	7.43	A	A
Sample 04	101.60	2.80	107.00	7.50	7.01	5.31	0.53	0.67	1.05	5.40	20.65	A	7.53	A	A
Sample 05	0.10	0.01	3.28	0.28											N

Laboratory No. 112

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	53.50	0.70	1.31	1.33	0.13	0.45	1.01	0.70	4.04	A	2.96	A	A
Sample 02	101.60	2.80	100.20	1.00	1.00	-1.38	-0.14	-0.47	0.99	1.40	7.67	A	2.93	A	A
Sample 03	52.80	1.40	51.60	0.70	1.36	-2.27	-0.23	-0.77	0.98	1.20	4.04	A	2.98	A	A
Sample 04	101.60	2.80	98.70	1.00	1.01	-2.85	-0.29	-0.98	0.97	2.90	7.67	A	2.94	A	A
Sample 05	0.10	0.01	0.06												A

Laboratory No. 114

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	60.00	6.00	10.00	13.64	1.36	1.17	1.14	7.20	15.90	A	10.35	A	A
Sample 02	101.60	2.80	100.80	11.00	10.91	-0.79	-0.08	-0.07	0.99	0.80	29.28	A	11.26	A	A
Sample 03	52.80	1.40	56.50	5.70	10.09	7.01	0.7	0.63	1.07	3.70	15.14	A	10.43	A	A
Sample 04	101.60	2.80	99.70	9.90	9.93	-1.87	-0.19	-0.18	0.98	1.90	26.54	A	10.31	A	A
Sample 05	0.10	0.01	28.70	2.90											N

Laboratory No. 115

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	51.50	1.90	3.69	-2.46	-0.25	-0.55	0.98	1.30	6.09	A	4.54	A	A
Sample 02	101.60	2.80	99.10	3.50	3.53	-2.46	-0.25	-0.56	0.98	2.50	11.56	A	4.48	A	A
Sample 03	52.80	1.40	53.20	1.80	3.38	0.76	0.08	0.18	1.01	0.40	5.88	A	4.30	A	A
Sample 04	101.60	2.80	95.20	2.90	3.05	-6.30	-0.63	-1.59	0.94	6.40	10.40	A	4.11	A	A
Sample 05	0.10	0.01	0.00	0.10											A

Laboratory No. 116

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	42.70	7.10	16.63	-19.13	-1.91	-1.4	0.81	10.10	18.67	A	16.84	N	W
Sample 02	101.60	2.80	78.00	13.50	17.31	-23.23	-2.32	-1.71	0.77	23.60	35.57	A	17.53	N	N
Sample 03	52.80	1.40	42.30	7.30	17.26	-19.89	-1.99	-1.41	0.8	10.50	19.18	A	17.46	N	W
Sample 04	101.60	2.80	77.60	13.50	17.40	-23.62	-2.36	-1.74	0.76	24.00	35.57	A	17.61	N	N
Sample 05	0.10	0.01	0.10												A

Laboratory No. 117

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	49.10	1.70	3.46	-7.01	-0.7	-1.68	0.93	3.70	5.68	A	4.36	A	A
Sample 02	101.60	2.80	95.00	2.90	3.05	-6.50	-0.65	-1.64	0.94	6.60	10.40	A	4.11	A	A
Sample 03	52.80	1.40	50.90	1.60	3.14	-3.60	-0.36	-0.89	0.96	1.90	5.49	A	4.11	A	A
Sample 04	101.60	2.80	96.60	2.90	3.00	-4.92	-0.49	-1.24	0.95	5.00	10.40	A	4.08	A	A
Sample 05	0.10	0.01	<0.025												A

Laboratory No. 118

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	47.62	2.94	6.17	-9.81	-0.98	-1.59	0.9	5.18	8.40	A	6.72	A	A
Sample 02	101.60	2.80	84.07	5.56	6.61	-17.25	-1.73	-2.82	0.83	17.53	16.06	N	7.16	A	W
Sample 03	52.80	1.40	43.53	2.82	6.48	-17.56	-1.76	-2.94	0.82	9.27	8.12	N	7.00	A	W
Sample 04	101.60	2.80	79.52	5.22	6.56	-21.73	-2.17	-3.73	0.78	22.08	15.28	N	7.12	A	N
Sample 05	0.10	0.01	6.76	0.49											N

Laboratory No. 119

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	49.00	2.00	4.08	-7.20	-0.72	-1.56	0.93	3.80	6.30	A	4.87	A	A
Sample 02	101.60	2.80	93.00	4.00	4.30	-8.46	-0.85	-1.76	0.92	8.60	12.60	A	5.11	A	A
Sample 03	52.80	1.40	58.00	2.00	3.45	9.85	0.98	2.13	1.1	5.20	6.30	A	4.35	A	A
Sample 04	101.60	2.80	97.00	5.00	5.15	-4.53	-0.45	-0.8	0.95	4.60	14.78	A	5.85	A	A
Sample 05	0.10	0.01	<0.004												A

5

Laboratory No. 120

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	38.70	4.30	11.11	-26.70	-2.67	-3.12	0.73	14.10	11.67	N	11.42	A	N
Sample 02	101.60	2.80	73.20	8.10	11.07	-27.95	-2.8	-3.31	0.72	28.40	22.11	N	11.40	A	N
Sample 03	52.80	1.40	39.00	4.30	11.03	-26.14	-2.61	-3.05	0.74	13.80	11.67	N	11.34	A	N
Sample 04	101.60	2.80	74.70	8.20	10.98	-26.48	-2.65	-3.1	0.74	26.90	22.36	N	11.32	A	N
Sample 05	0.10	0.01	0.05	0.01											A

Laboratory No. 121

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	52.50	3.30	6.29	-0.57	-0.06	-0.08	0.99	0.30	9.25	A	6.82	A	A
Sample 02	101.60	2.80	91.70	5.50	6.00	-9.74	-0.97	-1.6	0.9	9.90	15.92	A	6.60	A	A
Sample 03	52.80	1.40	44.50	2.60	5.84	-15.72	-1.57	-2.81	0.84	8.30	7.62	N	6.42	A	W
Sample 04	101.60	2.80	90.40	5.40	5.97	-11.02	-1.1	-1.84	0.89	11.20	15.69	A	6.58	A	A
Sample 05	0.10	0.01	<0.09												A

Laboratory No. 123

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.00	2.60	5.20	-5.30	-0.53	-0.95	0.95	2.80	7.62	A	5.84	A	A
Sample 02	101.60	2.80	96.00	5.10	5.31	-5.51	-0.55	-0.96	0.94	5.60	15.01	A	5.98	A	A
Sample 03	52.80	1.40	50.70	2.90	5.72	-3.98	-0.4	-0.65	0.96	2.10	8.31	A	6.30	A	A
Sample 04	101.60	2.80	92.90	5.10	5.49	-8.56	-0.86	-1.5	0.91	8.70	15.01	A	6.14	A	A
Sample 05	0.10	0.01	0.05	0.07											A

Laboratory No. 124

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	49.36	1.05	2.13	-6.52	-0.65	-1.97	0.93	3.44	4.52	A	3.40	A	A
Sample 02	101.60	2.80	89.62	2.82	3.15	-11.79	-1.18	-3.01	0.88	11.98	10.25	N	4.18	A	W
Sample 03	52.80	1.40	49.16	0.92	1.87	-6.89	-0.69	-2.17	0.93	3.64	4.32	A	3.25	A	A
Sample 04	101.60	2.80	91.90	1.92	2.09	-9.55	-0.95	-2.86	0.9	9.70	8.76	N	3.46	A	W
Sample 05	0.10	0.01	0.02	0.00											A

Laboratory No. 125

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	46.70	0.20	0.43	-11.55	-1.16	-4.31	0.88	6.10	3.65	N	2.69	A	W
Sample 02	101.60	2.80	102.60	0.60	0.58	0.98	0.1	0.35	1.01	1.00	7.39	A	2.82	A	A
Sample 03	52.80	1.40	52.80	0.20	0.38	0.00	0.0	0.0	1.0	0.00	3.65	A	2.68	A	A
Sample 04	101.60	2.80	91.60	0.50	0.55	-9.84	-0.98	-3.52	0.9	10.00	7.34	N	2.81	A	W
Sample 05	0.10	0.01	0.02	0.00											A

Laboratory No. 126

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	50.60	1.30	2.57	-4.17	-0.42	-1.15	0.96	2.20	4.93	A	3.69	A	A
Sample 02	101.60	2.80	95.80	2.30	2.40	-5.71	-0.57	-1.6	0.94	5.80	9.35	A	3.65	A	A
Sample 03	52.80	1.40	50.60	1.30	2.57	-4.17	-0.42	-1.15	0.96	2.20	4.93	A	3.69	A	A
Sample 04	101.60	2.80	96.87	2.50	2.58	-4.66	-0.47	-1.26	0.95	4.73	9.68	A	3.78	A	A
Sample 05	0.10	0.01	0.00	0.00											A

Laboratory No. 127

Sample code	Target Value	Unc	Lab. Value	Lab. Unc.	Unc. [%]	Rel. Bias	Z- Score	U-Test	Ratio	A1	A2	True	P	Prec.	Final Score
Sample 01	52.80	1.40	54.48	0.79	1.44	3.18	0.32	1.05	1.03	1.68	4.14	A	3.02	A	A
Sample 02	101.60	2.80	98.13	1.37	1.40	-3.42	-0.34	-1.11	0.97	3.47	8.04	A	3.09	A	A
Sample 03	52.80	1.40	52.65	1.09	2.08	-0.29	-0.03	-0.09	1.0	0.15	4.58	A	3.37	A	A
Sample 04	101.60	2.80	102.50	1.67	1.63	0.89	0.09	0.28	1.01	0.90	8.41	A	3.20	A	A
Sample 05	0.10	0.01													-

APPENDIX III. LIST OF PARTICIPATING LABORATORIES

ARGENTINA	JORGE MARIANO DIODATI AUTORIDAD REGULATORIA NUCLEAR APOYO CIENTIFICO TECNICO MEDICIONES AMBIENTALES AVENIDA DEL LIBERTADOR 8250 RA-1429, BUENOS AIRES ARGENTINA
AUSTRALIA	SANDRA SDRAULIG RADIOANALYTICAL SERVICES AUSTRALIAN RADIATION PROTECTION AND NUCLEAR SAFETY AGENCY (ARPANSA) LOWER PLENTY ROAD 619 YALLAMBIE, VICTORIA 3085 AUSTRALIA
	ATUN ZAWADZKI INSTITUTE FOR ENVIRONMENTAL RESEARCH AUSTRALIAN NUCLEAR SCIENCE AND TECHNOLOGY ORGANIZATION BUILDING 34, NEW ILLAWARRA RD MENAI, 2234 NEW SOUTH WALES AUSTRALIA
AUSTRIA	KATZLBERGER CHRISTIAN ABT. F. STRAHLENSCHUTZ UND RADIOCHEMIE AGRICULTURE RESEARCH VIENNA, AUSTRIAN AGENCY FOR HEALTH AND FOOD SAFETY SPARGELFELDSTRASSE 191 A-1226, VIENNA AUSTRIA
BELGIUM	CECILE DELPORTE SCIENTIFIC INSTITUTE OF PUBLIC HEALTH (FEDERAL PUBLIC SERVICE HEALTH, FOOD CHAIN SECURITY AND ENVIRONMENT) SECTION RADIOACTIVITY JULIETTE WYTSMAN STREET, 14 B-1050 BRUSSELS BELGIUM
	UWE WAETJEN ANALYTICAL CHEMISTRY EUROPEAN COMMISSION (E.C.) RETIESEWEG 111 B-2440, GEEL BELGIUM
BRAZIL	JOSELENE DE OLIVEIRA CENTRO DE METROLOGIA DAS RADIACOES LABORATORIO DE RADIOMETRIA AMBIENTAL INSTITUTO DE PESQUISAS ENERGETICAS E NUCLEARES IPEN/CNEN/SP AV. PROF. LINEU PRESTES 2242 CEP 05508-000, SAO PAULO, SP BRAZIL

MS. LIGIA JULIAO
SEMIN - SERVICO DE MONITORACAO
INDIVIDUAL INTERNA
INSTITUTO DE RADIOPROTECAO E DOSIMETRIA -
C.N.E.N.
AVENIDA SALVADOR ALLENDE S/N
RECREIO DOS BANDEIRANTES
22780-160, RIO DE JANEIRO
BRAZIL

BULGARIA

M.KATRANKIEVA
EXECUTIVE ENVIRONMENTAL RADIOLOGICAL
LABORATORY
EXECUTIVE ENVIRONMENTAL AGENCY
REGIONAL LABORATORY - VRATZA
81 EXARH JOSIF STR. 3000, VRATZA
BULGARIA

MS. ROSITZA KUKewa
EXECUTIVE ENVIRONMENT AGENCY,
MINISTRY OF ENVIRONMENT AND WATER
136 TZAR BORIS III, BLVD.
BG-1618, SOFIA
BULGARIA

BELARUS

MAKAREVICH VALERY, MATUSEVICH ZHANA
RESEARCH DEPARTMENT OF RADIATIVE METROLOGY
(RDRM)
BELARUSSIAN STATE INSTITUTE OF METROLOGY
STAROVILENSKI TRAKT 93
220053, MINSK
BELARUS

CANADA

MS KATHY NIELSEN
DIRECTOR, SLOWPOKE-2 FACILITY
DEPARTMENT OF CHEMISTRY AND
CHEMICAL ENGINEERING
ROYAL MILITARY COLLEGE
ROOM 2092 - 11 GENERAL CRERAR CRESENT
STN FORCES 17000
KINGSTON, ONTARIO K7K 7B4
CANADA

DAVID A. LOBB
DEPARTMENT OF SOIL SCIENCE
UNIVERSITY OF MANITOBA
13 FREEDMAN CRES.
WINNIPEG, R3T 2N2
CANADA

DR. GARY H. KRAMER
NATIONAL CALIBRATION REFERENCE CENTRE
FOR BIOASSAY AND IN VIVO MONITORING
RADIATION PROTECTION BUREAU
HEALTH CANADA
775 BROOKFIELD ROAD
OTTAWA, K1A 1C1, ONTARIO
CANADA

MR. DALE SANDERS
ATTN. GAIL SHUTIAK
SRC ANALYTICAL LABORATORIES
SASKATCHEWAN RESEARCH COUNCIL
15 INNOVATION BOULEVARD, SASKATOON
SASKATCHEWA, S7N 4N1
CANADA

MR. PETER NELSON
CANADIAN NUCLEAR SAFETY COMMISSION
BUILDING NO. 7 200 TUNNEYS PASTURE DRIVEWAY
OTTAWA, K1A 0L2
ONTARIO
CANADA

CHINA

MR. JINZHOU DU
STATE KEY LABORATORY OF ESTUARINE
AND COASTAL RESEARCH
EAST CHINA NORMAL UNIVERSITY
3663# ZHANGSHAN ROAD-NORTH
SHANGHAI, 200062
CHINA

MR. ZHIYUAN CHANG
DEPARTMENT OF RADIOCHEMISTRY
CHINA INSTITUTE OF ATOMIC ENERGY
P.O.BOX 275 (8)
BEIJING, 102413
CHINA

CZECH REPUBLIC

DIANA IVANOVVA
THE T.G. MASARYK WATER RESEARCH
INSTITUTE OF PRAGUE
PODBABSKA 30
CZ-160 62, PRAGUE
CZECH REPUBLIC

DENMARK

SVEN NIELSEN
RADIATION RESEARCH DEPARTMENT
BUILDING 204
RISOE NATIONAL LABORATORY
FREDERIKSBORGVEJ, 399
DK-4000, ROSKILDE
DENMARK

FINLAND

PIA VESTERBACKA
RADIONUCLIDE ANALYTICS / NAL
RADIATION AND NUCLEAR SAFETY AUTHORITY
(STUK)
LAIPPATIE 4
P.O. BOX 14, FIN-00880, HELSINKI
FINLAND

FRANCE

MR. PATRICE BOISSON
INSTITUT DE RADIOPROTECTION ET DE SURETE
NUCLEAIRE (IRSN)
31, RUE DE L ECLUSE
F-78116, LE VESINET CEDEX
FRANCE

MS. MELANIE OSMOND
SERVICE DE TRAITEMENT DES
ECHANTILLONS ET DE METROLOGIE
POUR L ENVIRONNEMENT
INSTITUT DE RADIOPROTECTION ET DE SURETE
NUCLEAIRE (IRSN)
31, RUE DE L ECLUSE
F-78116, LE VESINET CEDEX
FRANCE

GERMANY

DR. SIEGURD MOEBIUS
CENTER FOR ADVANCED TECHNOLOGICAL AND
ENVIRONMENTAL TRAINING
RESEARCH CENTER KARLSRUHE
HERMANN-VON-HELMHOLTZ-PLATZ
D-76344 EGGENSTEIN-LEOPOLDSHAFEN
GERMANY

MS. MARTINA HARTMANN
BUNDESAMT FUER STRAHLENSCHUTZ
KOEPEINCKER ALLEE 120-130
D-10318, BERLIN
GERMANY

DR. THOMAS ERNST
NLWKN NIEDERSACHSEN
AN DER SCHARLAKE 39
31135 HILDESHEIM
GERMANY

DR. UDO GERSTMANN
GSF NATIONAL RESEARCH CENTER FOR
ENVIRONMENT AND HEALTH INSTITUTE OF
RADIATION PROTECTION RADLAB
INGOLSTAEDTER LANDSTRASSE 1 85764
D-85758, NEUHERBERG
GERMANY

MS. LENA JOHANSSON
ZENTRUM FUER STRAHLENSCHUTZ
UNIVERSITY OF HANNOVER
HERRENHAEUSERSTRASSE 2
D-30419, HANNOVER
GERMANY

DR. INGRID SCHAEFER
VEREIN FUER KERNVERFAHRENSTECHNIK
UND ANALYTIK ROSSENDORF
BAUTZNER LANDSTRASSE 128
D-01314, DRESDEN,
GERMANY

GREECE

DR. STAMOULIS KONSTANTINOS
NUCLEAR PHYSICS LABORATORY
PHYSICS DEPARTMENT
UNIVERSITY OF IOANNINA
45110, IOANNINA,
GREECE

HUNGARY

NORA VAJDA
RADIOCHEMISTRY
INSTITUTE OF NUCLEAR TECHNIQUES
TECHNICAL UNIVERSITY OF BUDAPEST
MUEGYETEM RAKPART 9
H-1111, BUDAPEST
HUNGARY

MR. GYULA SZABO
NATIONAL RESEARCH INSTITUTE OF
RADIOBIOLOGY AND RADIOHYGIENE
ANNA U. 5
H-1775, BUDAPEST, 22
HUNGARY

INDIA

DR D. D. RAO
ENVIRONMENTAL SURVEY LABORATORY
HEALTH PHYSICS DIVISION
TAPS COLONY, TAPP
TARAPUR ATOMIC POWER STATION
VIA BOISAR
PALGHAR (TALUG), THANE DISTRICT
TARAPUR, 401 504
MAHARASHTRA
INDIA

DR. V. KANNAN
GOVERNMENT OF INDIA,
DEPARTMENT OF ATOMIC ENERGY
BHABHA ATOMIC RESEARCH CENTRE
HEALTH PHYSICS DIVISION
ENVIRONMENTAL SURVEY LABORATORY
ATOMIC ENERGY TOWNSHIP
SADRAS COLONY WEST
KALPAKKAM KANCHEEPURAM DISTRICT
TAMILNADU STATE PIN:603102
INDIA

INDONESIA

ALI ARMAN
NATIONAL NUCLEAR ENERGY AGENCY
EARTH SCIENCE AND ENVIRONMENT
CENTRE FOR THE APPLICATION OF ISOTOPES AND
RADIATION / BATAN
JALAN CINERE PASAR JUMAT
KOTAK POS 7002
JAKARTA, 12070, INDONESIA

DADONG ISKANDAR
CENTRE FOR TECHNOLOGY OF
RADIATION SAFETY AND METROLOGY
PTKMR-BATAN
NATIONAL NUCLEAR ENERGY AGENCY
JALAN CUNERE PASAR JUMAT
JAKARTA, 12070, INDONESIA

IRELAND

MR. DAVID POLLARD
RADIOLOGICAL PROTECTION
INSTITUTE OF IRELAND
3 CLONSKEAGH SQ., CLONSKEAGH RD
DUBLIN, 14
IRELAND

ITALY

DESIDERI DONATELLA
INSTITUTE OF GENERAL AND
INORGANIC CHEMISTRY
UNIVERSITY OF URBINO
PIAZZA RINASCIMENTO, 6 I-61029, URBINO,
ITALY

RITA OCONE
RIS LAB - SETTORE RADIOATTIVITA ARTIFICIALE
APAT (ENVIRONMENTAL PROTECTION AGENCY), VIA
DI CASTEL ROMANO, 100
00128, ROME
ITALY

PAOLO BATTISTI
ENEA CR CASACCIA BAS ION-IRP
INSTITUTE OF RADIATION PROTECTION
VIA ANGUILLARESE KM 1+300
00123 SANTA MARIA DI GALERIA
ROME
ITALY

MONTALTO MARIO
ISTITUTO DI RADIOPROTEZIONE
LABORATORIO SALUGGIA
ENEA RESEARCH CENTRE
STRADA PER CRESCENTINO
I-13040, SALUGGIA (VERCELLI)
ITALY

RONGONI ALBA AND SAETTA DANIELA
DIPARTIMENTO DI SCIENZE CHIRURGICHE,
RADIOLOGICHE, ODONTOSTOMATOLOGICHE
MEDICO LEGALI
POLICLINICO MONTELUCE
UNIVERSITA DEGLI STUDI DI PERUGIA
VIA BRUNAMONTI, 06100, PERUGIA
ITALY

MAGNONI MAURO
ARPA PIEMONTE
CENTRO REGIONALE RADIAZIONI
VIA JERVIS, 30, 10015, IVREA (TORINO)
ITALY

DR. MASSIMO CAPPAL
SETTORE RADIAZIONI IONIZZANTI E
NON IONIZZANTI
C.R.R. SARDEGNA
P.M.P.-AREA FISICA
AZIENDA US.L. 8 CAGLIARI
VIALE F. CIUSA, 6/8
09131, CAGLIARI, ITALY

ROSELLA RUSCONI
ARPA LOMBARDIA - SEZIONE RADIOPROTEZIONE
VIA JUVARA 22
I-20129, MILANO
ITALY

MS. LAURA PORZIO
CENTRO REGIONALE PER LE RADIAZIONI
IONIZZANTI E NON IONIZZANTI
STRUTTURA SEMPLICE 21.02 RAD. IONIZZANTI
MONITORAGGIO E CONTROLLO SITI NUCLEARI
ARPA PIEMONTE
VIA TRINO, 89, 13100, VERCELLI
ITALY

JAPAN

MR HIROTO FUKUSHIMA
JAPAN CHEMICAL ANALYSIS CENTER
295-3, SANNO-CHO, INAGE-KU
CHIBA, 263-0002
JAPAN

MR. KUNIO SHIRAISHI
RESEARCH CENTER FOR
RADIATION EMERGENCY MEDICINE,
DEPARTMENT OF RADIATION DOSIMETRY
NATIONAL INSTITUTE OF RADIOLOGICAL SCIENCES
4-9-1, ANAGAWA
CHIBA, 263
JAPAN

JORDAN

DR. MAMOUN MAKALEH
JORDAN ATOMIC ENERGY COMMISSION
SHAFI BADRAN
MARJ FARAS - P.BOX 70
AMMAN, 11934
JORDAN

KOREA, REPUBLIC OF

GI HOON HONG
MARINE BIOGEOCHEMISTRY AND ISOTOPES
KOREAN OCEAN RESEARCH AND DEVELOPMENT
INSTITUTE (KORDI)
ANSAN P.O. BOX 29
SEOUL, 425-600
KOREA, REPUBLIC OF

LEE, CHANG WOO
ENVIRONMENTAL RADIATION ASSESMT.
KOREA ATOMIC ENERGY RESEARCH INSTITUTE
150 DUCKJIN-DONG, YUSUNG-KU
TAEJON, 305-353
KOREA, REPUBLIC OF

LITHUANIA

ARUNAS GUDELIS
NUCLEAR AND ENVIRONMENTAL
RADIOACTIVITY RESEARCH LABORATORY
INSTITUTE OF PHYSICS
SAVANORIU AV. 231
LT-02300, VILNIUS
LITHUANIA, REPUBLIC OF

MALAYSIA
 DR ZAHARUDIN AHMAD
 RADIOCHEMISTRY AND ENVIRONMENT GROUP
 MALAYSIAN NUCLEAR AGENCY(NUCLEAR
 MALAYSIA)
 SELANGOR DARUL EHSAN
 BANGI, 43000, KAJANG, SELANGOR D.E.,
 MALAYSIA

MONACO
 MAI KHANH PHAM
 INTERNATIONAL ATOMIC ENERGY AGENCY
 MARINE ENVIRONMENT LABORATORY
 4, QUAI ANTOINE 1 ER
 98000, MONACO, CEDEX
 MONACO

MOROCCO
 BENMANSOUR
 DEPARTEMENT DES APPLICATIONS AUX SCIENCES
 DE LA TERRE ET DE L ENVIRONNEMENT
 CENTRE NATIONAL DE L ENERGIE DES SCIENCES ET
 DES TECHNIQUES NUCLEAIRES (CNESTEN)
 B.P. 1382, 65 RUE TANSIFT- AGDAL
 RABAT 10001
 MOROCCO

NETHERLANDS
 DR. PJM KWAKMAN
 LABORATORY FOR RADIATION RESEARCH
 RIVM
 ANTONIE VAN LEEUWENHOEKLAAN 9
 NL-3720 BA, BILTHOVEN
 THE NETHERLANDS

MR. FRITS MOET
 BUILDING 015 RM 005
 NRG PETTEN
 WESTERDUINWEG 3
 1755 ZG, PETTEN
 THE NETHERLANDS

W.J.C. OKX
 REACTOR INSTITUTE DELFT
 DELFT UNIVERSITY OF TECHNOLOGY
 MEKELWEG 15
 NL-2629 JB, DELFT
 THE NETHERLANDS

NEW ZEALAND
 NIKOLAUS HERMANSPAHN
 NATIONAL RADIATION LABORATORY
 108 VICTORIA ST.
 CHRISTCHURCH
 NEW ZEALAND

POLAND
 ZOFIA PIETRZAK-FLIS, PH.D.,D.SC.
 DEPARTMENT OF RADIATION HYGIENE
 CENTRAL LABORATORY FOR RADIOLOGICAL
 PROTECTION
 UL. KONWALIOWA 7
 PL-03-194, WARSAW - ZERAN
 POLAND

PORTUGAL

FERNANDO P. CARVALHO
DEPTO. DE PROTECCAO RADIOLOGICA E SEGURANCA
NUCLEAR
INSTITUTO TECNOLOGICO E NUCLEAR
ESTRADA NACIONAL, 10
P-2686-953, SACAVERM,
PORTUGAL

ROMANIA

DR.OLGA IACOB
RADIATION PROTECTION DEPARTMENT
INSTITUTE OF PUBLIC HEALTH AND
MEDICAL RESEARCH
VICTOR BABES STREET 14
700465, IASI
ROMANIA

FLORENTINA CONSTANTIN, DECEBAL POPA
PUBLIC HEALTH AUTHORITY FROM BUCHAREST
LABORATORY OF IONISING RADIATION HYGIENE
BUCHAREST
INTRAREA RECONSTRUCTIEI NR.6,
SECTOR 3,
COD 031726,
BUCHAREST
ROMANIA

EMANUELA CINCU
ACTIVA-N LABORATORY
DEPARTMENT OF APPLIED PHYSICS,
IFIN-HH
STR. ATOMISTILOR NR. 407
R-077125, BUCHAREST,
MAGURELE
ROMANIA

RUSSIAN FEDERATION

ALEXANDER ERMAKOV
RADIOCHEMICAL LABORATORY
SIA "RADON"
7TH ROSTOVSKY LN., 2/14
119121,
MOSCOW
RUSSIAN FEDERATION

SAUDI ARABIA

ASHRAF KHATER
COLLEGE OF SCIENCES
PHYSICS DEPARTMENT
KING SAUD UNIVERSITY
P.O. BOX 2455, RIYADH, 11451
SAUDI ARABIA

KHALID A. ALEISSA
DEPARTMENT OF HEALTH AND ENVIRONMENTAL
STUDIES
INSTITUTE OF ATOMIC ENERGY RESEARCH
KING ABDULAZIZ CITY FOR SCIENCE AND
TECHNOLOGY, KING ABDULLAH ROAD
P.O. BOX 6086, RIYADH, 11442
SAUDI ARABIA

SLOVENIA

BORUT SMODIS
ENVIRONMENTAL SCIENCES DIVISION
INSTITUTE JOZEF STEFAN
JAMOVA 39
1000, LJUBLJANA
SLOVENIA

SOUTH AFRICA

MS. IMMANDA LOUW
RADIOANALYSIS
SOUTH AFRICAN NUCLEAR ENERGY CORPORATION
(NECSA)
BUILDING 1600
PO BOX 582, CHURCH STREET WEST
PRETORIA, 0001,
SOUTH AFRICA

SPAIN

DRA. CATALINA GASCO LEONARTE
ED. 3A
RADIOECOLOGIA DEL MEDIO ACUATICO
C I E M A T
ED. 70
AV. COMPLUTENSE, 22
E-28040, MADRID
SPAIN

JOSEFINA ORTIZ
LAB. DE RADIOACTIVIDAD AMBIENTAL
UNIVERSIDAD POLITECNICA DE VALENCIA
CAMINO DE VERA 14
APARTADO 22012
E-46021, VALENCIA
SPAIN

JUAN P. BOLIVAR
DEPT. OF FISICA APLICADA, FACULTAD DE CIENCIAS
UNIVERSIDAD DE HUELVA
CAMPUS DE EL CARMEN
E-21071, HUELVA
SPAIN

ELISA NAVARRO ANGLES
LABORATORIO DE RADIOACTIVIDAD AMBIENTAL
UNIVERSIDAD DE VALENCIA
EDIFICIO DE INVESTIGACION
C/ DR. MOLINER 50
E-46100, VALENCIA
BURJASSOT
SPAIN

ISABEL VALLES MURCIANO
RADIOCHEMISTRY LABORATORY
INSTITUTO DE TECNICAS ENERGETICAS
UNIVERSIDAD POLITECNICA
AVENIDA DIAGONAL 647
E-08028, BARCELONA
SPAIN

JOSE ANTONIO SUAREZ NAVARRO
UNIDAD DE ANALISIS Y CARACTERIZACI3N DE
RESIDUOS
C I E M A T
ED. 70
AV. COMPLUTENSE, 22
E-28040, MADRID
SPAIN

DR. GUILLERMO MANJON
DEPARTAMENTO DE FISICA APLICADA II
UNIVERSIDAD DE SEVILLA
AVENIDA REINA MERCEDES 2
APARTADO 1065
E-41012, SEVILLA
SPAIN

LUIS SANTIAGO QUINDOS PONCELA
CATEDRA DE FISICA MEDICA
FACULTAD DE MEDICINA
UNIVERSIDAD DE CANTABRIA
C/ CARDENAL HERRERA ORIA S/N
E-39011, SANTANDER
SPAIN

LUIS MIGUEL NIETO CALZADA
LABORATORIO DE INVESTIGACION
EN BAJA RADIOACTIVIDAD
DEPARTAMENTO DE FISICA TEORICA, ATOMICA Y
NUCLEAR
FACULTAD DE CIENCIAS
UNIVERSIDAD DE VALLADOLID
EDIFICIO I+D PASEO BELEN 3 C.P.
47071, VALLADOLID,
SPAIN

PROF. F. LEGARDA
DPT. INGENIERIA NUCLEAR Y
MECANICA DE FLUIDOS
ESCUELA TECNICA SUPERIOR
DE INGENIERIA
ALDA. URQUIJO, S/N
48013, BILBAO
SPAIN

SWEDEN

ELIS HOLM
DEPARTMENT OF MEDICAL RADIATION PHYSICS
LUND UNIVERSITY HOSPITAL
S-221 85, LUND
SWEDEN

MATS ISAKSSON
DEPARTMENT OF RADIATION PHYSICS
GOETEBORG UNIVERSITY
SAHLGREN UNIVERSITY HOSPITAL
GULASTRAEKET 2B
SE-413 45, GOETEBORG,
SWEDEN

MR. NILS ADDO
STUDSVIK NUCLEAR AB
S-611 82, NYKOEPIG
SWEDEN

DR. HAKAN B.L. PETTERSON
DEPARTMENT OF RADIATION PHYSICS
FACULTY OF HEALTH SCIENCES
LINKOEPING UNIVERSITY
HALSOUNIVERSITETET
S-518 85,
LINKOEPING,
SWEDEN

SWITZERLAND

DR. M.H.A. JUNGCK
ABTEILUNG STRAHLENSCHUTZ
BUNDESAMT FUER GESUNDHEITSWESSEN
SCHWARZENBURGSTR.165
3097, LIEBEFELD
SWITZERLAND

FROIDEVAUX PASCAL
GSE
INSTITUT DE RADIOPHYSIQUE APPLIQUEE
RUE DU GRAND-PRE 1
CH-1007, LAUSANNE
SWITZERLAND

MR. STEFAN ROELLIN
FEDERAL OFFICE FOR CIVIL PROTECTION
SPIEZ LABORATORY
CH-3700 SPIEZ
SWITZERLAND

SYRIAN ARAB REPUBLIC

PROF. MOHAMMAD SAID AL-MASRI
DEPARTMENT OF RADIATION PROTECTION
ATOMIC ENERGY COMMISSION OF SYRIA
DAMASCUS
SYRIAN ARAB REPUBLIC

DR.RAFAAT AL-MEREY
QUALITY ASSURANCE OFFICE
ATOMIC ENERGY COMMISSION OF SYRIA
17 NISAN STREET, P.O.BOX:6091
DAMASCUS
SYRIAN ARAB REPUBLIC

THAILAND

MS. KANITHA SRISUKSAWAD
THAILAND INSTITUTE OF NUCLEAR TECHNOLOGY
(TINT)
16 VIBHAVADI RANGSIT,
BANGKOK, 10900
THAILAND

TURKEY

MS. HILAL HAZNEDAROGLU
TURKISH ATOMIC ENERGY AUTHORITY
CEKMECE NUCLEAR RESEARCH AND TRAINING
CENTER
ALTINSEHIR YOLU 5.KM
34303, HALKALI, ISTANBUL, TURKEY

TAIWAN

MR. CHING CHUNG HUANG
TAIWAN RADIATION MONITORING CENTER,
ATOMIC ENERGY COUNCIL
823 CHERNG-CHING ROAD
KAOHSIUNG, 833
TAIWAN, ROC

UNITED KINGDOM

MICHAEL JOHN
CHEMICAL AND METTALURGICAL SERVICE
DEPARTMENT A396
SPRINGFIELDS FUELS LTD
WESTINGHOUSE
SALWICK, PRESTON, PR4 0XJ
UNITED KINGDOM

TERRY GINGELL
DSTL, ENVIRONMENTAL SCIENCES DEPARTMENT
C/O INSTITUTE OF NAVAL MEDICINE
CRESCENT ROAD
GOSPORT, PO12 2DL
UNITED KINGDOM

PHILLIP WARWICK
GEOSCIENCES ADVISORY UNIT
NATIONAL OCEANOGRAPHY CENTRE
EUROPEAN WAY
SOUTHAMPTON, SO14 3ZH
UNITED KINGDOM

MR TONY DELL
RADIOCHEMISTRY GROUP
VETERINARY LABORATORIES AGENCY
WOODHAM LANE, NEW HAW
ADDLESTONE, SURREY KT15 3NB
UNITED KINGDOM

DR ARVIC HARMS
NATIONAL PHYSICAL LABORATORY
QUALITY OF LIFE
G6-A5, HAMPTON ROAD
TEDDINGTON, TW11 0LW
MIDDLESEX
UNITED KINGDOM

ROGER BENZING
RADIOCHEMISTRY
HARWELL SCIENTIFICS
HARWELL SCIENCE AND INNOVATION CAMPUS, 551
SOUTH, BECQUEREL AVENUE
DIDCOT, OX11 OTB
OXFORDSHIRE,
UNITED KINGDOM

KEITH BRADSHAW
NIRAS
AMEC NNC LIMITED
THE RENAISSANCE CENTRE
601 FARADAY STREET
WARRINGTON, WA3 6GN
UNITED KINGDOM

MR. RICHARD C. WILSON
WESTLAKES SCIENTIFIC CONSULTING LTD
THE PRINCESS ROYAL BUILDING
WESTLAKES SCIENCE & TECHNOLOGY PARK
MOOR ROW
WHITEHAVEN, CUMBRIA CA243LN
UNITED KINGDOM

UKRAINE

MICHAEL BUZINNY
HEAD OF RADIATION HYGIENE MONITORING
LABORATORY
THE MARZEYEV INSTITUTE OF HYGIENE AND
MEDICAL ECOLOGY
ACADEMY OF MEDICAL SCIENCES OF UKRAINE
50 POPUDRENKO STR.
02094, KIEV
UKRAINE

TSYGANKOV MYKOLA
DOSIMETRY AND RADIATION
HYGIENE DEPT.
RESEARCH CENTRE FOR RADIATION MEDICINE
MELNIKOVA STREET 53
254050, KIEV
UKRAINE

UNITED STATES OF AMERICA

LIAQUAT HUSAIN
DEPARTMENT OF ENVIRONMENTAL DISEASE
PREVENTION
N.Y STATE DEPT OF HEALTH
WADSWORTH CENTER, EMPIRE STATE PLAZA
ALBANY, 12201-0509
UNITED STATES OF AMERICA

KENNETH G.W. INN
DEPARTMENT OF IONIZING RADIATION
NATIONAL INSTITUTE OF STANDARDS AND
TECHNOLOGY
100 BUREAU DRIVE,
GAITHERSBURG, MD 20899-8462
UNITED STATES OF AMERICA

ARTHUR C. LUCAS
LUCAS NEWMAN SCIENCE
AND TECHNOLOGIES, INC.
1110 INNOVATION WAY
STILLWATER, 74074
UNITED STATES OF AMERICA

BERND KAHN
ENVIRONMENTAL RADIATION BRANCH
GTRI
430 TENTH ST. BLDG. N-003
ATLANTA, GA 30332-0871
UNITED STATES OF AMERICA

BRADLEY K. ESSER
ENVIRONMENTAL MONITORING RADIOANALYTICAL
LABORATORY, CHEMISTRY & MATERIAL
SCIENCE DIRECTORATE
UNIVERSITY OF CALIFORNIA

LAWRENCE LIVERMORE NATIONAL LABORATORY
7000 EAST AVENUE
LIVERMORE, CA 94550
UNITED STATES OF AMERICA

MS. DENNI GRUNSTRA
GEL LABORATORIES, LLC
2040 SAVAGE ROAD
P.O.BOX 30712
CHARLESTON, SC 29407
UNITED STATES OF AMERICA

JOSEPH ESCALANTE
BUREAU OF RADIATION CONTROL
FLORIDA DEPARTMENT OF HEALTH
2044 ALL CHILDRENS WAY
ORLANDO, FL 32818
UNITED STATES OF AMERICA

MELISSA MANNION
EBERLINE SERVICES
2030 WRIGHT AVE.
RICHMOND, CA 94804-0040
UNITED STATES OF AMERICA

MYUNG CHUL JO
EH&S, MS 328
UNIVERSITY OF NEVADA
MAILSTOP 328 ARF ROOM 218, RENO, NV 89557
UNITED STATES OF AMERICA

S. KUSUM PERERA, PH.D.
SANITATION & RADIATION LABORATORY BRANCH,
CALIFORNIA DEPARTMENT OF HEALTH SERVICES
850, MARINA BAY PARKWAY, G 164
RICHMOND, CA 94804-6403
UNITED STATES OF AMERICA

THEODORE E. RAHON
COPHYSICS CORPORATION
1242 ROUTE 208, MONROE, NY 10950
UNITED STATES OF AMERICA

CAROLYN WONG
HAZARDS CONTROL DEPARTMENT
RADIATION SAFETY SECTION
BIOASSAY LABORATORY
LAWRENCE LIVERMORE NATIONAL LABORATORY
7000 EAST AVENUE
LIVERMORE, CA 94550
UNITED STATES OF AMERICA

MR. CONG WEI
RADIONUCLIDES SECTION
ANALYTICAL BRANCH
US FOOD AND DRUG ADMINISTRATION
WINCHESTER ENGINEERING AND
ANALYTICAL CENTER
109 HOLTON STREET
WINCHESTER, MA 01890-1152
UNITED STATES OF AMERICA

REFERENCES

- [1] HARRISON, J; LEGGETT, R; LLOYD, D; Plonium-210 as a poison, *Journal of Radiological Protection*, 27 (2007) 17-40.
- [2] ISO/IEC 17025:2005, *General Requirements for the Competence of Testing and Calibration Laboratories*, ISO, Geneva, Switzerland (2005).
- [3] SHAKHASHIRO, A; FAJGELJ, A; SANSONE, U; Comparison of Different Approaches to Evaluate Proficiency Test Data. Presented and accepted in the publications of the International Workshop on Combining and Reporting Analytical Results. The Role of (metrological) Traceability and (measurement) Uncertainty for Comparing Analytical Results, Rome 6-8 March, (2006).
- [4] BROOKES, CJ; BETTELEY, IG; LOXTON, SM; *Fundamentals of Mathematics and Statistics*, Wiley (1979).

CONTRIBUTORS TO DRAFTING AND REVIEW

C.K. Kim,	International Atomic Energy Agency
G. Kis-Benedek,	International Atomic Energy Agency
P. Martin	International Atomic Energy Agency
U. Sansone,	International Atomic Energy Agency
A. Shakhashiro,	International Atomic Energy Agency
A. Trinkl.	International Atomic Energy Agency



Where to order IAEA publications

In the following countries IAEA publications may be purchased from the sources listed below, or from major local booksellers. Payment may be made in local currency or with UNESCO coupons.

Australia

DA Information Services, 648 Whitehorse Road, Mitcham Victoria 3132
Telephone: +61 3 9210 7777 • Fax: +61 3 9210 7788
Email: service@dadirect.com.au • Web site: <http://www.dadirect.com.au>

Belgium

Jean de Lannoy, avenue du Roi 202, B-1190 Brussels
Telephone: +32 2 538 43 08 • Fax: +32 2 538 08 41
Email: jean.de.lannoy@infoboard.be • Web site: <http://www.jean-de-lannoy.be>

Canada

Bernan Associates, 4611-F Assembly Drive, Lanham, MD 20706-4391, USA
Telephone: 1-800-865-3457 • Fax: 1-800-865-3450
Email: order@bernan.com • Web site: <http://www.bernan.com>

Renouf Publishing Company Ltd., 1-5369 Canotek Rd., Ottawa, Ontario, K1J 9J3
Telephone: +613 745 2665 • Fax: +613 745 7660
Email: order.dept@renoufbooks.com • Web site: <http://www.renoufbooks.com>

China

IAEA Publications in Chinese: China Nuclear Energy Industry Corporation, Translation Section, P.O. Box 2103, Beijing

Czech Republic

Suweco CZ, S.R.O. Klecakova 347, 180 21 Praha 9
Telephone: +420 26603 5364 • Fax: +420 28482 1646
Email: nakup@suweco.cz • Web site: <http://www.suweco.cz>

Finland

Akateeminen Kirjakauppa, PL 128 (Keskuskatu 1), FIN-00101 Helsinki
Telephone: +358 9 121 41 • Fax: +358 9 121 4450
Email: akatilaus@akateeminen.com • Web site: <http://www.akateeminen.com>

France

Form-Edit, 5, rue Janssen, P.O. Box 25, F-75921 Paris Cedex 19
Telephone: +33 1 42 01 49 49 • Fax: +33 1 42 01 90 90 • Email: formedit@formedit.fr

Lavoisier SAS, 145 rue de Provigny, 94236 Cachan Cedex
Telephone: + 33 1 47 40 67 02 • Fax +33 1 47 40 67 02
Email: romuald.verrier@lavoisier.fr • Web site: <http://www.lavoisier.fr>

Germany

UNO-Verlag, Vertriebs- und Verlags GmbH, August-Bebel-Allee 6, D-53175 Bonn
Telephone: +49 02 28 949 02-0 • Fax: +49 02 28 949 02-22
Email: info@uno-verlag.de • Web site: <http://www.uno-verlag.de>

Hungary

Librotrade Ltd., Book Import, P.O. Box 126, H-1656 Budapest
Telephone: +36 1 257 7777 • Fax: +36 1 257 7472 • Email: books@librotrade.hu

India

Allied Publishers Group, 1st Floor, Dubash House, 15, J. N. Heredia Marg, Ballard Estate, Mumbai 400 001,
Telephone: +91 22 22617926/27 • Fax: +91 22 22617928
Email: alliedpl@vsnl.com • Web site: <http://www.alliedpublishers.com>

Bookwell, 2/72, Nirankari Colony, Delhi 110009
Telephone: +91 11 23268786, +91 11 23257264 • Fax: +91 11 23281315
Email: bookwell@vsnl.net

Italy

Libreria Scientifica Dott. Lucio di Biasio "AEIOU", Via Coronelli 6, I-20146 Milan
Telephone: +39 02 48 95 45 52 or 48 95 45 62 • Fax: +39 02 48 95 45 48

Japan

Maruzen Company, Ltd., 13-6 Nihonbashi, 3 chome, Chuo-ku, Tokyo 103-0027
Telephone: +81 3 3275 8582 • Fax: +81 3 3275 9072
Email: journal@maruzen.co.jp • Web site: <http://www.maruzen.co.jp>

Korea, Republic of

KINS Inc., Information Business Dept. Samho Bldg. 2nd Floor, 275-1 Yang Jae-dong SeoCho-G, Seoul 137-130
Telephone: +02 589 1740 • Fax: +02 589 1746
Email: sj8142@kins.co.kr • Web site: <http://www.kins.co.kr>

Netherlands

De Lindeboom Internationale Publicaties B.V., M.A. de Ruyterstraat 20A, NL-7482 BZ Haaksbergen
Telephone: +31 (0) 53 5740004 • Fax: +31 (0) 53 5729296
Email: books@delindeboom.com • Web site: <http://www.delindeboom.com>

Martinus Nijhoff International, Koraaalrood 50, P.O. Box 1853, 2700 CZ Zoetermeer
Telephone: +31 793 684 400 • Fax: +31 793 615 698 • Email: info@nijhoff.nl • Web site: <http://www.nijhoff.nl>

Swets and Zeitlinger b.v., P.O. Box 830, 2160 SZ Lisse
Telephone: +31 252 435 111 • Fax: +31 252 415 888 • Email: info@swets.nl • Web site: <http://www.swets.nl>

New Zealand

DA Information Services, 648 Whitehorse Road, MITCHAM 3132, Australia
Telephone: +61 3 9210 7777 • Fax: +61 3 9210 7788
Email: service@dadirect.com.au • Web site: <http://www.dadirect.com.au>

Slovenia

Cankarjeva Založba d.d., Kopitarjeva 2, SI-1512 Ljubljana
Telephone: +386 1 432 31 44 • Fax: +386 1 230 14 35
Email: import.books@cankarjeva-z.si • Web site: <http://www.cankarjeva-z.si/uvoz>

Spain

Díaz de Santos, S.A., c/ Juan Bravo, 3A, E-28006 Madrid
Telephone: +34 91 781 94 80 • Fax: +34 91 575 55 63 • Email: compras@diazdesantos.es
carmela@diazdesantos.es • barcelona@diazdesantos.es • julio@diazdesantos.es
Web site: <http://www.diazdesantos.es>

United Kingdom

The Stationery Office Ltd, International Sales Agency, PO Box 29, Norwich, NR3 1 GN
Telephone (orders): +44 870 600 5552 • (enquiries): +44 207 873 8372 • Fax: +44 207 873 8203
Email (orders): book.orders@tso.co.uk • (enquiries): book.enquiries@tso.co.uk • Web site: <http://www.tso.co.uk>

On-line orders:

DELTA Int. Book Wholesalers Ltd., 39 Alexandra Road, Addlestone, Surrey, KT15 2PQ
Email: info@profbooks.com • Web site: <http://www.profbooks.com>

Books on the Environment:

Earthprint Ltd., P.O. Box 119, Stevenage SG1 4TP
Telephone: +44 1438748111 • Fax: +44 1438748844
Email: orders@earthprint.com • Web site: <http://www.earthprint.com>

United Nations (UN)

Dept. 1004, Room DC2-0853, First Avenue at 46th Street, New York, N.Y. 10017, USA
Telephone: +800 253-9646 or +212 963-8302 • Fax: +212 963-3489
Email: publications@un.org • Web site: <http://www.un.org>

United States of America

Bernan Associates, 4611-F Assembly Drive, Lanham, MD 20706-4391
Telephone: 1-800-865-3457 • Fax: 1-800-865-3450
Email: order@bernan.com • Web site: <http://www.bernan.com>

Renouf Publishing Company Ltd., 812 Proctor Ave., Ogdensburg, NY, 13669
Telephone: +888 551 7470 (toll-free) • Fax: +888 568 8546 (toll-free)
Email: order.dept@renoufbooks.com • Web site: <http://www.renoufbooks.com>

Orders and requests for information may also be addressed directly to:

Sales and Promotion Unit, International Atomic Energy Agency

Wagramer Strasse 5, P.O. Box 100, A-1400 Vienna, Austria
Telephone: +43 1 2600 22529 (or 22530) • Fax: +43 1 2600 29302
Email: sales.publications@iaea.org • Web site: <http://www.iaea.org/books>

INTERNATIONAL ATOMIC ENERGY AGENCY
VIENNA
ISSN 2074-7659