

IAEA Analytical Quality in Nuclear Applications Series No. 56

Worldwide Interlaboratory Comparison on the Determination of Trace Elements in Oyster Biota Sample IAEA-470



IAEA

International Atomic Energy Agency

WORLDWIDE INTERLABORATORY
COMPARISON ON THE
DETERMINATION OF TRACE ELEMENTS
IN OYSTER BIOTA SAMPLE IAEA-470

The following States are Members of the International Atomic Energy Agency:

AFGHANISTAN	GERMANY	PALAU
ALBANIA	GHANA	PANAMA
ALGERIA	GREECE	PAPUA NEW GUINEA
ANGOLA	GRENADA	PARAGUAY
ANTIGUA AND BARBUDA	GUATEMALA	PERU
ARGENTINA	GUYANA	PHILIPPINES
ARMENIA	HAITI	POLAND
AUSTRALIA	HOLY SEE	PORTUGAL
AUSTRIA	HONDURAS	QATAR
AZERBAIJAN	HUNGARY	REPUBLIC OF MOLDOVA
BAHAMAS	ICELAND	ROMANIA
BAHRAIN	INDIA	RUSSIAN FEDERATION
BANGLADESH	INDONESIA	RWANDA
BARBADOS	IRAN, ISLAMIC REPUBLIC OF	SAINT VINCENT AND
BELARUS	IRAQ	THE GRENADINES
BELGIUM	IRELAND	SAN MARINO
BELIZE	ISRAEL	SAUDI ARABIA
BENIN	ITALY	SENEGAL
BOLIVIA, PLURINATIONAL STATE OF	JAMAICA	SERBIA
BOSNIA AND HERZEGOVINA	JAPAN	SEYCHELLES
BOTSWANA	JORDAN	SIERRA LEONE
BRAZIL	KAZAKHSTAN	SINGAPORE
BRUNEI DARUSSALAM	KENYA	SLOVAKIA
BULGARIA	KOREA, REPUBLIC OF	SLOVENIA
BURKINA FASO	KUWAIT	SOUTH AFRICA
BURUNDI	KYRGYZSTAN	SPAIN
CAMBODIA	LAO PEOPLE'S DEMOCRATIC REPUBLIC	SRI LANKA
CAMEROON	LATVIA	SUDAN
CANADA	LEBANON	SWEDEN
CENTRAL AFRICAN REPUBLIC	LESOTHO	SWITZERLAND
CHAD	LIBERIA	SYRIAN ARAB REPUBLIC
CHILE	LIBYA	TAJIKISTAN
CHINA	LIECHTENSTEIN	THAILAND
COLOMBIA	LITHUANIA	THE FORMER YUGOSLAV REPUBLIC OF MACEDONIA
CONGO	LUXEMBOURG	TOGO
COSTA RICA	MADAGASCAR	TRINIDAD AND TOBAGO
CÔTE D'IVOIRE	MALAWI	TUNISIA
CROATIA	MALAYSIA	TURKEY
CUBA	MALI	TURKMENISTAN
CYPRUS	MALTA	UGANDA
CZECH REPUBLIC	MARSHALL ISLANDS	UKRAINE
DEMOCRATIC REPUBLIC OF THE CONGO	MAURITANIA	UNITED ARAB EMIRATES
DENMARK	MAURITIUS	UNITED KINGDOM OF GREAT BRITAIN AND NORTHERN IRELAND
DJIBOUTI	MEXICO	UNITED REPUBLIC OF TANZANIA
DOMINICA	MONACO	UNITED STATES OF AMERICA
DOMINICAN REPUBLIC	MONGOLIA	URUGUAY
ECUADOR	MONTENEGRO	UZBEKISTAN
EGYPT	MOROCCO	VANUATU
EL SALVADOR	MOZAMBIQUE	VENEZUELA, BOLIVARIAN REPUBLIC OF
ERITREA	MYANMAR	VIET NAM
ESTONIA	NAMIBIA	YEMEN
ESWATINI	NEPAL	ZAMBIA
ETHIOPIA	NETHERLANDS	ZIMBABWE
FIJI	NEW ZEALAND	
FINLAND	NICARAGUA	
FRANCE	NIGER	
GABON	NIGERIA	
GEORGIA	NORWAY	
	OMAN	
	PAKISTAN	

The Agency's Statute was approved on 23 October 1956 by the Conference on the Statute of the IAEA held at United Nations Headquarters, New York; it entered into force on 29 July 1957. The Headquarters of the Agency are situated in Vienna. Its principal objective is "to accelerate and enlarge the contribution of atomic energy to peace, health and prosperity throughout the world".

WORLDWIDE INTERLABORATORY
COMPARISON ON THE
DETERMINATION OF TRACE ELEMENTS
IN OYSTER BIOTA SAMPLE IAEA-470

COPYRIGHT NOTICE

All IAEA scientific and technical publications are protected by the terms of the Universal Copyright Convention as adopted in 1952 (Berne) and as revised in 1972 (Paris). The copyright has since been extended by the World Intellectual Property Organization (Geneva) to include electronic and virtual intellectual property. Permission to use whole or parts of texts contained in IAEA publications in printed or electronic form must be obtained and is usually subject to royalty agreements. Proposals for non-commercial reproductions and translations are welcomed and considered on a case-by-case basis. Enquiries should be addressed to the IAEA Publishing Section at:

Marketing and Sales Unit, Publishing Section
International Atomic Energy Agency
Vienna International Centre
PO Box 100
1400 Vienna, Austria
fax: +43 1 26007 22529
tel.: +43 1 2600 22417
email: sales.publications@iaea.org
www.iaea.org/books

For further information on this publication, please contact:

IAEA Environment Laboratories, Monaco
Marine Environmental Studies Laboratory
International Atomic Energy Agency
4a Quai Antoine 1er, 98000
Principality of Monaco

WORLDWIDE INTERLABORATORY COMPARISON ON THE DETERMINATION
OF TRACE ELEMENTS IN OYSTER BIOTA SAMPLE IAEA-470

IAEA, VIENNA, 2018
IAEA/AQ/56
ISSN 2074-7659

© IAEA, 2018

Printed by the IAEA in Austria
December 2018

FOREWORD

The primary goal of the IAEA Environment Laboratories is to help Member States understand, monitor and protect the marine environment. Thus, the major impact on marine ecosystems exerted by large coastal cities is an issue of primary concern for the IAEA and its Member States.

Marine pollution assessments depend on accurate knowledge of contaminant concentrations in various environmental compartments. Since the early 1970s, the IAEA Environment Laboratories have been assisting national laboratories and regional laboratory networks in the analysis of radionuclides, trace elements and organic contaminants in marine samples through global interlaboratory comparisons and regional proficiency tests, the production of marine certified reference materials and the development of analytical methods for analysing trace elements and organic pollutants in marine samples. The Marine Environmental Studies Laboratory in Monaco, part of the IAEA Environment Laboratories, actively assists Member States in organizing interlaboratory comparisons designed to monitor and demonstrate the performance and analytical capabilities of the participating laboratories, and to identify gaps and problem areas where further developments are needed.

The present publication summarizes the results of the IAEA-470 interlaboratory comparison on the determination of trace elements in oyster sample.

The IAEA wishes to thank the Korea Institute of Science and Technology for providing the raw sample material. The IAEA is also grateful to the the Government of Monaco for its support. The IAEA officers responsible for this publication were S. Azemard and E. Vasileva of the IAEA Environment Laboratories.

EDITORIAL NOTE

This publication has been prepared from the original material as submitted by the contributors and has not been edited by the editorial staff of the IAEA. The views expressed remain the responsibility of the contributors and do not necessarily reflect those of the IAEA or the governments of its Member States.

Neither the IAEA nor its Member States assume any responsibility for consequences which may arise from the use of this publication. This publication does not address questions of responsibility, legal or otherwise, for acts or omissions on the part of any person.

The use of particular designations of countries or territories does not imply any judgement by the publisher, the IAEA, as to the legal status of such countries or territories, of their authorities and institutions or of the delimitation of their boundaries.

The mention of names of specific companies or products (whether or not indicated as registered) does not imply any intention to infringe proprietary rights, nor should it be construed as an endorsement or recommendation on the part of the IAEA.

The IAEA has no responsibility for the persistence or accuracy of URLs for external or third party Internet web sites referred to in this publication and does not guarantee that any content on such web sites is, or will remain, accurate or appropriate.

CONTENTS

1. INTRODUCTION	1
2. SCOPE OF THE INTERLABORATORY COMPARISON	1
3. DESCRIPTION OF THE MATERIAL	2
4. EVALUATION OF ANALYTICAL PERFORMANCE	4
5. RESULTS AND DISCUSSION	4
5.1. Overview of the results	4
5.2. Laboratory performances	5
5.3. Analytical methods	13
5.4. Sample treatment, impact of water content, use of crm and recovery corrections....	14
6. RECOMMENDATIONS	15
7. CONCLUSIONS.....	15
APPENDIX I.....	17
APPENDIX II	41
REFERENCES.....	61
LIST OF PARTICIPANTS	63
CONTRIBUTORS TO THE DRAFTING AND REVIEW.....	69

1 INTRODUCTION

The Marine Environmental Studies Laboratory (MESL) of the IAEA's Environment Laboratories (IAEA-NAEL) has the programmatic responsibility to provide assistance to Member States' laboratories in maintaining and improving the reliability of analytical measurement results, both in trace elements and organic pollutants. This is accomplished through the provision of certified reference materials of marine origin, validated analytical procedures, training in the implementation of internal quality control, and through the evaluation of measurement performance by the organization of worldwide and regional interlaboratory comparisons. IAEA's subprogram 'Reference Products for Science and Trade' represents an important benchmark in upgrading the quality of laboratory performances and assessing the validity of the analytical methods used for marine monitoring studies in the IAEA Member States.

For more than thirty years, the MESL has conducted worldwide laboratory performance studies, also known as interlaboratory comparison [1, 2]. The results have been used to evaluate laboratory performance with respect to a wide range of organic [3, 4] and inorganic pollutants, including methylmercury [5, 6].

The goal of interlaboratory comparison is to demonstrate the measurement capabilities of laboratories participating in interlaboratory comparisons (ILCs) and targeted proficiency tests (PTs). The results from ILCs or PTs are of crucial interest for laboratories as they provide clear information of their measurement capabilities. NAEL Interlaboratory Comparison (ILC) and Proficiency Test (PT) schemes involve comparison of participant's results with an assigned value, which usually is delivered as a consensus value from the overall population of reported by participating laboratories test results.

These exercises are designed to monitor and demonstrate the performance and analytical capabilities of the participating laboratories, and to identify gaps and problem areas where further development is needed.

2 SCOPE OF THE INTERLABORATORY COMPARISON

The present interlaboratory comparison was designed to evaluate the measurement performance of participating laboratories for the analysis of trace elements in marine biota sample (oyster).

Invitation letters have been sent to 39 laboratories, previously participated in an IAEA ILC. Positive responses were received from 34 laboratories in 20 Member States and samples were duly dispatched to them.

Each participating laboratory received one bottle of the marine biota material, accompanied by an information sheet and a reporting form. Using the procedures, routinely applied in their laboratories, participants were requested to determine as many elements as possible from the following list: copper, cadmium, lead, lithium, zinc, mercury, methyl mercury nickel, vanadium, arsenic, silver, aluminium, iron, manganese, chromium, cobalt, tin and strontium. The organisers were also interested in receiving results for any other elements that the participating laboratories were willing to provide.

In total 30 laboratories (reporting 34 datasets) from 19 countries participated in this ILC and reported results for 35 elements.

The data reported by the laboratories, together with the technical and statistical evaluations of the results for each element, are included in this report. All results were treated confidentially and each laboratory was identified with a code number for anonymity.

3 DESCRIPTION OF THE MATERIAL

The interlaboratory comparison sample, distributed in this exercise was a Certified Reference Material (CRM) IAEA-470. All details about homogeneity, stability and characterization can be found in the certification report [7].

Reported measurement results were first evaluated with Kernel density plots [8]. Certified mass fraction values for selected trace elements and methylmercury, used in the evaluation process of reported results are shown in Table 1.

The assigned values for 5 additional elements, not included in Table 1, were calculated as described in the ISO 13528 standard, Annex C.21 [9]. Calculated robust mean and robust standard deviations are presented in Tables 2 and 3.

The uncertainties associated with the assigned values were calculated according to the ISO 35 standard [10]. The relative combined uncertainty of the assigned value of the CRM consists of uncertainty related to the characterization (u_{char}), between bottle heterogeneity (u_{hom}) and long-term stability (u_{stab}). Above mentioned contributions were combined to estimate the expanded uncertainty, as described in the Eq. (1).

$$U = k \times \sqrt{u_{char}^2 + u_{ho}^2 + u_{stab}^2} \quad (1)$$

Where k : coverage factor 2, representing a level of confidence of about 95%.

u_{hom} – uncertainty on homogeneity of oyster sample is set at 3.5% and the estimation is based on the obtained results from homogeneity study.

u_{stab} - uncertainty on stability during the storage period was estimated at 3.5%.

More details on stability and homogeneity study for the IAEA-470 sample are described in the certification report [7].

u_{char} – uncertainty on characterisation of oyster sample are calculated following the Eq. (2)

$$u_{char} = 1.25 \times \frac{s^*}{\sqrt{n}} \quad (2)$$

Where s^* is the robust standard deviation, calculated as described in the ISO 13528 [9] and n is the number of datasets used for calculation of the robust mean.

For Al and Ni obtained expanded uncertainties were higher than 25%. The reasons for the high uncertainties were the broad distribution of reported data for Al and insufficient homogeneity for Ni [7], respectively. As a consequence, those values were given as informative values and were not used for the evaluation of measurement performances of laboratories, participating in the IAEA-470 interlaboratory comparison.

TABLE 1. CERTIFIED VALUES FOR THE IAEA-470 OYSTER CRM

Element	Unit	Certified value ¹	Expanded uncertainty ($k=2$) ²
Ag	mg kg ⁻¹	1.29	0.10
As	mg kg ⁻¹	11.9	0.9
Ca	mg kg ⁻¹	2430	282
Cd	mg kg ⁻¹	3.14	0.24
CH ₃ Hg	μg kg ⁻¹ as Hg	5.22	0.87
Co	mg kg ⁻¹	0.201	0.025
Cr	mg kg ⁻¹	0.965	0.113
Cu	mg kg ⁻¹	146	13
Fe	mg kg ⁻¹	131	12
Hg	μg kg ⁻¹	21.1	2.1
Mg	mg kg ⁻¹	3080	390
Mn	mg kg ⁻¹	66.7	5.3
Na	mg kg ⁻¹	19.7 10 ³	2.3 10 ³
Pb	mg kg ⁻¹	0.361	0.053
Rb	mg kg ⁻¹	5.14	0.59
Se	mg kg ⁻¹	3.06	0.33
Sr	mg kg ⁻¹	20.6	1.6
V	mg kg ⁻¹	0.899	0.130
Zn	mg kg ⁻¹	727	48

TABLE 2. ASSIGNED VALUE FOR THE IAEA-470 OYSTER SAMPLE

Element	Unit	Assigned value ¹	Expanded uncertainty ($k=2$) ²
K	mg kg ⁻¹	12720	2150
Li	mg kg ⁻¹	0.532	0.094

TABLE 3. INFORMATIVE VALUE IN IAEA-470 OYSTER SAMPLE

Element	Unit	Informative value ¹	Expanded uncertainty ($k=2$) ²
Al	mg kg ⁻¹	72.9	19.9
Ni	mg kg ⁻¹	0.849	0.291

4 EVALUATION OF ANALYTICAL PERFORMANCE

The individual laboratory performance was expressed in terms of z -score and Zeta-score, in accordance with the ISO 13528 [9].

The determination of target standard deviation was done on the basis of the outcome from the previous ILCs organised by the MESL for the same population of laboratories and similar sample matrix. The standard deviation for the proficiency assessment, σ_p , was fixed to 12.5 % from the assigned values. The appropriateness of this level of variability of results was confirmed by calculation of the robust standard deviation of the participants' results and the uncertainty of the assigned values for the respective measurants.

z -score, calculated following the Eq. (3), effectively expresses the difference between the mean of the laboratory and the assigned value in units of target standard deviation (σ_p).

Zeta-score, calculated according to the Eq. (4), indicates the agreement of the result of the individual participant with the ILC assigned value within the respective uncertainties. The denominator in the Eq. (4) is calculated from the combined uncertainty of the assigned value and the measurement uncertainty reported by the participating laboratory.

$$z = \frac{x_{lab} - x_{ref}}{\sigma_p} \quad (3)$$

$$\text{Zeta} = \frac{x_{lab} - x_{ref}}{\sqrt{u_{x_{lab}}^2 + u_{x_{ref}}^2}} \quad (4)$$

Where:

x_{lab} : Reported results by participant (express as the mean of multiple determination)

x_{ref} : Certified value or assigned value

σ_p : Target standard deviation

$u_{x_{lab}}$: Standard uncertainty reported by participant

u_{ref} : Standard uncertainty of certified (or assigned) value

The interpretation of a laboratory's performance was evaluated according to the following generally accepted limits:

$ z \text{ or Zeta } \leq 2$	Satisfactory
$2 < z \text{ or Zeta } < 3$	Questionable
$ z \text{ or Zeta } \geq 3$	Unsatisfactory

5 RESULTS AND DISCUSSION

5.1 OVERVIEW OF THE RESULTS

Thirty laboratories provided results for the analysis of the IAEA-470 oyster sample by the final deadline. Thirty-three sets of measurement results were submitted (some laboratories reported

data generated by multiple techniques) comprising 407 results for 35 different elements. z -scores and Zeta-scores were evaluated only for 21 elements.

From the 21 elements evaluated, 11 elements (As, Cd, Co, Cr, Cu, Fe, Hg, Mn, Pb, V and Zn) have been reported by more than 50% of participants.

For elements with more than 5 reported datasets, graphical presentations of z -score and Zeta-scores (except for Al and Ni) are presented in Appendix I, together with a Kernel density plot (if more than 8 data have been reported) and short summary on the statistical evaluation of results for the reported elements [8].

All reported data are presented in the Appendix II. For editorial purposes some results have been rounded to appropriate number of significant figures.

5.2 LABORATORY PERFORMANCES

5.2.1 z -scores:

Table 4 shows the z -scores of laboratories for all reported elements. The overall performance (z -scores) of participating laboratories is graphically presented on Figures 1 and 3.

The z -score compares the participant's deviation from the reference value with the target standard deviation (σ_p for proficiency assessment). σ_p is defined by the ILC organizer as the maximum acceptable standard deviation (25% in this case) of the assigned value for the investigated trace elements.

In total from the calculated 352 z -scores, 93% were $|z| \leq 2$, 96% $|z| < 3$, and only 3% were considered as unsatisfactory with $|z| > 3$.

Among 33 datasets, 17 (52%) were with $|z| \leq 2$ for all reported results and 25 (76%) with $|z| < 3$. All participants reported at least 50% of their measurement results with $|z| \leq 2$.

As shown on Figure 3 the part of accepted measurement results was 65% or higher for each of the evaluated elements. Methyl mercury, total mercury and lead appeared to be the elements reported with the highest number of unsatisfactory z -scores. This shows that the accurate determination of those elements in biological material is still challenging analytical task, particularly at low mass fraction levels.

This is underlined by the fact that a large part (72%) of $|z| > 3$ is positive, demonstrating positive bias. These biased results may arise from the contamination during sample preparation (e.g., digestion) step or from the instrumental steps. The laboratories concerned should carefully check sample preparation procedures (e.g., quality chemical reagents used) and try to improve their working laboratory environment. Laboratories should also develop an effective scheme for cleaning of the lab ware and regularly control this process.

Erroneous calibration standards could be another source of bias. It is important to note that losses in the working standard solutions at low concentration level can be the reason for the overestimation of the concentrations of investigated elements. Only standard solutions (CRMs) with stated SI traceability should be used for calibration purposes and in addition they should not be stored for an extended period of time.

Laboratories with questionable and unsatisfactory results should also carefully check all analytical procedures, laboratory equipment and instrumentation.

5.2.2 Zeta-scores:

Table 5 shows the overall performance (Zeta-scores) of laboratories for different trace elements. The overall performance (Zeta-scores) of participating laboratories are graphically presented on Figures 2 and 4.

The Zeta-score shows the level of agreement of the laboratory result with the assigned value within the respective uncertainties. The denominator in the Eq. (2) is the combined uncertainty of the assigned value and the uncertainty of measurement result, reported by the individual laboratory.

Five laboratories (15%) didn't report uncertainties and Zeta-scores for their results were not calculated.

As it can be easily seen in Figures 1 and 2, the comparison of laboratories performances evaluated with z -score and Zeta-score clearly indicate that the number of unsatisfactory Zeta-scores is considerably higher than the number of unsatisfactory z -scores (3% for z -scores and 8% for Zeta-scores). Only 10 laboratories (30%) could report 100% of their results with $|z|$ and $|Zeta-scores| \leq 2$. As the Zeta-score is the parameter, reflecting all parts of the measurement procedure, laboratories with unsatisfactory Zeta-scores should invest additional efforts in the proper evaluation of measurement uncertainty of reported measurement results. Overall, obtained results indicates that they are still remaining problems with the realistic estimation of measurement uncertainty.

If should be mentioned here that an unsatisfactory Zeta-score can be also caused by wrong estimation of the mass fraction of the respective trace element.







FIG. 1. The z -scores of results reported by the participants per laboratory.



|Zeta| > 3,
 2 < |Zeta| < 3,
 |Zeta| ≤ 2
 FIG. 2. The Zeta-scores of results reported by the participants per laboratory.



|z| ≥ 3,
 2 < |z| < 3,
 |z| ≤ 2
 FIG. 3. The z-scores of results reported by the participants per element.



Zeta ≥ 3,
 2 < Zeta < 3,
 Zeta ≤ 2

FIG. 4. The Zeta-scores of results reported by the participants per element.

TABLE 4. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) BY ELEMENTS

Lab code	Ag	As	Ca	Cd	CH ₃ Hg	Co	Cr	Cu	Fe	Hg	K	Li	Mg	Mn	Na	Pb	Rb	Se	Sr	V	Zn
1	0.11	0.31	0.63	0.08	0.08	-0.04	0.29	-0.11	0.37	-1.18			0.49	0.94		-0.47	1.75			0.19	0.51
2	0.11	0.31	0.63	0.45	0.45	0.48	2.12	0.35	-0.41			0.04	0.33	0.00	-0.83	0.33			0.19	0.69	0.32
3		-0.61		0.10	0.10		1.50	0.88	0.43					0.61		14.47			-1.38	-2.57	0.83
4		-0.94		-0.10	-0.10	-1.55		-0.88						-0.53							-0.48
5		0.34	0.33	0.36	0.14	-0.08	-1.72	0.00	-0.37	-2.39			-0.13	-0.24		-0.13	0.26			-0.16	0.30
6		-0.07		-0.56	-0.56			-0.16	0.00	2.24				-0.05		-0.62					-0.50
7	-1.77	0.47	0.14	0.31	0.31	0.28	-0.07	-0.41	0.07	-1.18	0.38		0.08	0.77	0.63	-0.71	-0.16		1.01	0.32	0.06
8	0.43	-0.09	-0.49		-3.09	-0.69	0.10		0.12	-0.45	-0.89			0.35	0.35		-0.02	0.41			0.32
8B										-0.38											
9	-0.87	-1.14	-0.62	-1.25	-1.25	-0.16	-1.01	-0.82	-0.12	-0.04		-1.75	-0.29	-0.65		-0.13	-1.05		-0.12	-0.06	-1.32
10		-0.07		-0.54	-0.54		-1.20	0.00	0.24					-0.20		12.61					-1.42
12					0.11					-0.91											
13	-0.33	-0.32		0.31	0.31	-0.09	0.00	-0.07	-0.46					0.14		0.20	-0.11	0.14	-0.22	-0.47	0.51
14		-0.47	0.35	-1.73	-1.73	6.57	-1.04		-0.49	0.99					-0.35	0.44	-0.53	-0.42			0.17
15		0.58	0.39	0.81	1.96	0.18	-0.51	0.48	0.49	0.47	1.39		0.63	0.79	1.02	0.20	0.42	0.36	-0.04	0.58	0.56
16	-0.25	-0.74	-0.40	0.38	0.38	-0.92	-0.97	0.38	0.24	-5.57			-0.87	0.16		-0.31	-0.37		-0.82	-0.69	0.15
17				-1.35	-1.35			0.48	1.35	0.17				-0.32							
18		0.79		0.31	0.31	-0.06	-0.35	0.01	-0.23	2.30		1.05		1.70					2.54	0.18	0.36
19	0.12			0.23	0.23			0.58		-0.04						0.20					0.75
20					1.26					-2.39											
22A	-0.68	0.07				-0.76	0.21		-0.19						-0.04		-0.84				-0.30
22B				-1.71	-1.71	1.75		-0.33								-1.13			-1.01		-1.17
23		0.47		-0.10	-0.10			0.62								3.08					
25	1.12	0.07	-0.95						0.31		-0.96		-0.69	-0.44	0.23		0.25	0.37		-0.88	0.14
26	1.55	1.09	0.47	1.45	2.90	0.73	-0.02	0.82	1.36	1.47			1.58	1.34		0.43	1.65	1.61	1.61	0.53	0.95
27				-1.50	6.13	6.73	0.12	-2.73	-1.77	5.27				-1.72		-3.12				-0.35	-1.14

TABLE 4. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) BY ELEMENTS (cont.)

Lab code	Ag	As	Ca	Cd	CH ₃ Hg	Co	Cr	Cu	Fe	Hg	K	Li	Mg	Mn	Na	Pb	Rb	Se	Sr	V	Zn
28	-0.56	-0.85	-0.12	-0.32		-0.84	0.35	-0.51	2.30	-0.35	-0.54	0.92	-0.19	-0.78	-0.08	-1.43		-1.09	-0.67	-0.17	-0.39
30		1.75	0.16	0.00	3.80	-1.23		0.16	-0.31	-0.53	0.61		-0.05	0.07	0.97	0.42		0.34	-0.12		-0.10
31				-0.18				-0.71								0.75					0.02
32				-0.28		2.40		-0.21	0.64					0.34					0.43	2.34	-0.65
33	-0.28	-0.01		0.97			-1.57	-0.17		-1.74				-0.19						-1.19	0.33
34		0.47			-0.18	-1.47	0.54	0.22	0.00	0.53	0.37			0.79				-0.18	0.19	-1.01	0.13
35	0.12	0.27		0.69	-0.64	0.36		0.38		0.08				0.02		0.02					0.66
36					0.21					1.02											

TABLE 5. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (ZETA-SCORE) BY ELEMENTS

Lab code	Ag	As	Ca	Cd	CH3Hg	Co	Cr	Cu	Fe	Hg	K	Li	Mg	Mn	Na	Pb	Rb	Se	Sr	V	Zn
1	0.03	0.71	0.90	0.17	-0.05	0.32	-0.23	0.58	-1.14	0.06	0.77	1.70	0.59	0.01	-1.50	0.54	2.74	0.51	0.20	1.15	
2	0.03	0.71	0.90	0.17	0.84	3.45	0.82	-0.87	-1.14	0.06	0.59	0.01	-1.50	0.54	2.74	0.51	0.20	1.15	0.20	1.15	
3																					
4																					
5																					
6																					
7																					
8																					
8B																					
9																					
10																					
12																					
13																					
14																					
15																					
16																					
17																					
18																					
19																					
20																					
22A																					
22B																					
23																					
25																					
26																					

5.3 ANALYTICAL METHODS

As can be concluded from Figure 5, a wide range of analytical methodologies was used to provide results for the determination of trace elements mass fractions in oyster IAEA-470 sample. Generally, they can be divided to three groups: non-destructive techniques (NAA); plasma spectrometric methods (ICP-MS and ICP-OES) and atomic absorption spectrometry methods, representing 12%, 65% and 15% respectively. Abbreviation used in the figure 5 and appendix III are shown in Table 6.

TABLE 6. INSTRUMENTAL TECHNIQUES USED IN THE IAEA-470 INTERLABORATORY COMPARISON AND THEIR ABBREVIATIONS

Method code	Instrumental technique
Method code	Instrumental technique
CV-AAS	Cold Vapour Atomic Absorption Spectrometry
CV-AFS	Cold Vapour Atomic Fluorescence Spectrometry
ET-AAS	Electro Thermal Atomic Absorption Spectrometry
F-AAS	Flame Atomic Absorption Spectrometry
GC-AFS	Gas Chromatography Atomic Fluorescence Spectrometry
GC-ECD	Gas Chromatography Electron Capture Detector
HPLC-AFS	High Performance Liquid Chromatography Atomic Fluorescence Spectrometry
ICP-MS	Inductively Coupled Plasma Mass Spectrometry
ICP-OES	Inductively Coupled Plasma Optical Emission Spectrometry
ID-ICP-MS	Isotope Dilution Inductively Coupled Plasma Mass Spectrometry
NAA	Neutron Activation Analysis
Solid AAS	Solid Atomic Absorption Spectrometry
SV	Stripping voltammetry




FIG. 5. Graphical distribution of instrumental techniques.

5.4 SAMPLE TREATMENT, IMPACT OF WATER CONTENT, USE OF CRM AND RECOVERY CORRECTIONS

In the IAEA-470 interlaboratory comparison, a wide range of sample preparation methods and instrumental set-ups were applied.

All laboratories stated using a CRM for quality control. Only one participating laboratory didn't report the results for the analysed CRM as requested in the reporting form. Five laboratories reported quality control results for selected elements only. An important principle for the selection of reference material by laboratories is the principle for matrix and concentration range matching. CRMs used in the IAEA-470 ILC were well selected, as all laboratories used biota of marine origin (i.e. TORT-2 lobster hepatopancreas from NRCC, IAEA 452 scallop from IAEA, NIST 1566 oyster from NIST etc).

Only 6 laboratories (18%) implemented correction for recovery for at least part of their reported results. 14 participants reported results for the obtained recovery rates. Most of recoveries reported were in the range of $100 \pm 15\%$.

High proportion of the laboratories that didn't correct for recovery, still reported satisfactory results. They have correctly estimated that the recovery achieved by them was not significantly different from 100%.

IAEA-470 sample was subject to freeze drying as part of its preparation procedure. At the time of bottling, the moisture content of the oyster sample was around 6.5%. Depending on local storage conditions and humidity levels the ILC sample might absorb moisture from the laboratory environment. Consequently, the users were advised to make a separate determination of the moisture content in the oyster sample. As the moisture is operationally dependent parameter, a recommended procedure on moisture content determination in the test sample was preliminary provided to the participants [10]. The moisture content reported by the

laboratories that applied correction factor was in the range of 0.02 to 13%. Only 1 participant didn't correct for moisture content.

6 RECOMMENDATIONS

Participants got recommendation to review their data element by element, appraising whether the z -scores and Zeta-scores are less than or equal to 2. The use of z -scores and Zeta-scores evaluation parameters will help to identify systematic errors in the measurement results (*e.g.* from calibration, reagent contamination or incomplete digestion) and should ultimately improve data quality.

Some laboratories still need to improve quality assurance / quality control procedures. Interlaboratory comparisons represent only one aspect of data quality assurance and can only provide occasional indicators of data reliability. Another valuable approach is through the regular analysis of certified reference materials, and by plotting the resulting data on a quality control chart. This provides continuous feedback to the analyst and is an essential tool for monitoring data quality and assuring acceptable results in future exercises.

A full catalogue of available IAEA reference materials is published regularly and can be consulted on the IAEA website: <http://www.iaea.org/programmes/aqcs>

7 CONCLUSIONS

Although the overall performance of the laboratories is satisfactory, it should be pointed that for elements present at low to ultra-low concentration levels (*i.e.* Pb, Hg, MeHg) in oyster is problematic. It is a subject of concern as these elements are toxic and the accurate determination of their concentration in seafood is of crucial importance.

An extra effort is still needed in the evaluation of measurement uncertainties, associated with the results, since the number of unsatisfactory Zeta-scores was systematically higher than the number of unsatisfactory z -scores for reported trace elements. The uncertainty associated with measurement results is of paramount importance in the frame of different regulations and international agreements, so it is fundamental for one laboratory to be able to report a sound uncertainty statement.


In general, laboratories should keep in mind that uncertainties evaluated only on the basis of measurement precision are frequently underestimated. In many cases, they just reflect variations coming from the measurement step and usually do not include uncertainty contributors, coming from other sources, namely uncertainty on recovery, procedural blank, moisture content etc.

The sample used in IAEA-470 interlaboratory comparison was further certified by expert laboratories for the mass fractions of 19 trace elements. Participants in the ILC could retain their ILC samples and request from the IAEA reference sheets with the certified mass fractions value of the IAEA-470 oyster sample.

APPENDIX I
PERFORMANCE EVALUATION BY ELEMENT IN IAEA-470

Reported data for Ag in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	77%	8%	15%


X_{cert} :	1.29 mg kg ⁻¹
$U_{cert} (k=2)$:	0.10 mg kg ⁻¹
$2\sigma_p$:	0.32 mg kg ⁻¹
Number of results:	13
Number of method:	3

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score



Reported data for Al in IAEA-470

Kernel density Plot.




Summary of results:

X_{info} :	72.9 mg kg ⁻¹
$U_{\text{info}} (k=2)$:	19.9 mg kg ⁻¹
Number of results:	11
Number of method:	4


Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{\text{lab}} \pm U_{\text{lab}}$; - - - $X_{\text{Cert}} \pm 2\sigma_p$; - - - $X_{\text{Cert}} \pm U_{\text{Cert}}(k=2)$



Reported data for As in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	86%	10%	5%


\bar{X}_{cert} :	11.9 mg kg ⁻¹
$U_{cert} (k=2)$:	0.9 mg kg ⁻¹
$2\sigma_p$:	2.98 mg kg ⁻¹
Number of results:	24
Number of method:	4

Reported results and expanded uncertainties.

— \bar{X}_{Cert} ; $\bullet \bar{X}_{lab} \pm U_{lab}$; - - - $\bar{X}_{Cert} \pm 2\sigma_p$; - - - $\bar{X}_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Ca in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	100%	0%	0%


X_{cert} :	2430 mg kg ⁻¹
$U_{cert} (k=2)$:	282 mg kg ⁻¹
$2\sigma_p$:	607 mg kg ⁻¹
Number of results:	12
Number of method:	3

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Cd in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	82%	5%	14%


X_{cert} :	3.14 mg kg ⁻¹
$U_{cert} (k=2)$:	0.24 mg kg ⁻¹
$2\sigma_p$:	0.79 mg kg ⁻¹
Number of results:	26
Number of method:	5

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{lab}$; $--- X_{Cert} \pm 2\sigma_p$; $--- X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for CH₃Hg in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	64%	9%	27%
Zeta-score	75%	0%	25%


X_{cert} :	5.22 $\mu\text{g kg}^{-1}$ as Hg
$U_{cert} (k=2)$:	0.87 $\mu\text{g kg}^{-1}$ as Hg
$2\sigma_p$:	1.31 $\mu\text{g kg}^{-1}$ as Hg
Number of results:	11
Number of method:	7

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{lab}$; $\cdots X_{Cert} \pm 2\sigma_p$; $\cdots X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Co in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	85%	5%	10%
Zeta-score	79%	11%	11%


X_{cert} :	0.20 mg kg ⁻¹
$U_{cert} (k=2)$:	0.03 mg kg ⁻¹
$2\sigma_p$:	0.05 mg kg ⁻¹
Number of results:	21
Number of method:	3

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Cr in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	95%	5%	0%
Zeta-score	88%	6%	6%


X_{cert} :	0.97 mg kg ⁻¹
$U_{cert} (k=2)$:	0.11 mg kg ⁻¹
$2\sigma_p$:	0.24 mg kg ⁻¹
Number of results:	19
Number of method:	4

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Cu in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	96%	4%	0%
Zeta-score	91%	9%	0%


X_{cert} :	146 mg kg ⁻¹
$U_{cert} (k=2)$:	13 mg kg ⁻¹
$2\sigma_p$:	36.5 mg kg ⁻¹
Number of results:	26
Number of method:	5

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Fe in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	96%	4%	0%
Zeta-score	89%	0%	11%


X_{cert} :	131 mg kg ⁻¹
$U_{cert} (k=2)$:	12 mg kg ⁻¹
$2\sigma_p$:	33 mg kg ⁻¹
Number of results:	23
Number of method:	5

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Hg in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	73%	18%	9%
Zeta-score	61%	22%	17%


X_{cert} :	21.1 $\mu\text{g kg}^{-1}$
$U_{cert} (k=2)$:	2.1 $\mu\text{g kg}^{-1}$
$2\sigma_p$:	5.3 $\mu\text{g kg}^{-1}$
Number of results:	23
Number of method:	5

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$



Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for K in IAEA-470


Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	100%	0%	0%


X_{ass} :	12720 mg kg ⁻¹
$U_{\text{ass}} (k=2)$:	2150 mg kg ⁻¹
$2\sigma_p$:	3180 mg kg ⁻¹
Number of results:	6
Number of method:	3

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{\text{lab}} \pm U_{\text{lab}}$; - - - $X_{\text{Cert}} \pm 2\sigma_p$; - - - $X_{\text{Cert}} \pm U_{\text{Cert}} (k=2)$



Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Li in IAEA-470


Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	100%	0%	0%


X_{ass} :	0.532 mg kg ⁻¹
$U_{\text{ass}} (k=2)$:	0.094 mg kg ⁻¹
$2\sigma_p$:	0.13 mg kg ⁻¹
Number of results:	6
Number of method:	2

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{\text{lab}} \pm U_{\text{lab}}$; - - - $X_{\text{Cert}} \pm 2\sigma_p$; - - - $X_{\text{Cert}} \pm U_{\text{Cert}} (k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Mg in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	90%	0%	10%


X_{cert} :	3080 mg kg ⁻¹
$U_{cert} (k=2)$:	390 mg kg ⁻¹
$2\sigma_p$:	770 mg kg ⁻¹
Number of results:	11
Number of method:	3

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Mn in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	84%	11%	5%


X_{cert} :	66.7 mg kg ⁻¹
$U_{cert} (k=2)$:	5.3 mg kg ⁻¹
$2\sigma_p$:	16.7 mg kg ⁻¹
Number of results:	23
Number of method:	

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Na in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	100%	0%	0%


X_{cert} :	19700 mg kg ⁻¹
$U_{cert} (k=2)$:	2300 mg kg ⁻¹
$2\sigma_p$:	4925 mg kg ⁻¹
Number of results:	9
Number of method:	3

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Ni in IAEA-470

Kernel density Plot.




Summary of results:

X_{info} :	0.849 mg kg ⁻¹
$U_{\text{info}} (k=2)$:	0.291 mg kg ⁻¹
Number of results:	18
Number of method:	2


Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{\text{lab}}$; - - - $X_{\text{Cert}} \pm 2\sigma_p$; - - - $X_{\text{Cert}} \pm U_{\text{Cert}}(k=2)$



Reported data for Pb in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	81%	0%	19%
Zeta-score	83%	6%	11%


X_{cert} :	0.361 mg kg ⁻¹
$U_{cert} (k=2)$:	0.053 mg kg ⁻¹
$2\sigma_p$:	0.09 mg kg ⁻¹
Number of results:	21
Number of method:	3

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$



Performance evaluation: ■ z-score ■ Zeta-score



Reported data for Rb in IAEA-470


Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	100%	0%	0%


X_{cert} :	5.14 mg kg ⁻¹
$U_{\text{cert}} (k=2)$:	0.59 mg kg ⁻¹
$2\sigma_p$:	1.29 mg kg ⁻¹
Number of results:	5
Number of method:	2

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm X_{\text{lab}} \pm U_{\text{lab}}$; - - - $X_{\text{Cert}} \pm 2\sigma_p$; - - - $X_{\text{Cert}} \pm U_{\text{Cert}} (k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Se in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	79%	14%	7%


X_{cert} :	3.06 mg kg ⁻¹
$U_{cert} (k=2)$:	0.33 mg kg ⁻¹
$2\sigma_p$:	0.77 mg kg ⁻¹
Number of results:	15
Number of method:	4

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Sr in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	93%	7%	0%
Zeta-score	77%	15%	8%


X_{cert} :	20.6mg kg ⁻¹
$U_{cert} (k=2)$:	1.6 mg kg ⁻¹
$2\sigma_p$:	5.15 mg kg ⁻¹
Number of results:	15
Number of method:	3

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for V in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	88%	12%	0%
Zeta-score	77%	15%	8%


X_{cert} :	0.90 mg kg ⁻¹
$U_{cert} (k=2)$:	0.13 mg kg ⁻¹
$2\sigma_p$:	0.22 mg kg ⁻¹
Number of results:	17
Number of method:	4

Reported results and expanded uncertainties.

— X_{Cert} ; • $X_{lab} \pm U_{lab}$; - - - $X_{Cert} \pm 2\sigma_p$; - - - $X_{Cert} \pm U_{Cert}(k=2)$




Performance evaluation: ■ z-score ■ Zeta-score.



Reported data for Zn in IAEA-470

Kernel density Plot.




Summary of results:

	Satisfactory	Questionable	Unsatisfactory
z-score	100%	0%	0%
Zeta-score	92%	0%	8%


X_{cert} :	727 mg kg ⁻¹
$U_{cert} (k=2)$:	48 mg kg ⁻¹
$2\sigma_p$:	182 mg kg ⁻¹
Number of results:	27
Number of method:	6

Reported results and expanded uncertainties.

— X_{Cert} ; $\bullet \pm U_{lab}$; $--- X_{Cert} \pm 2\sigma_p$; $--- X_{Cert} \pm U_{Cert}(k=2)$



Performance evaluation: ■ z-score ■ Zeta-score.



APPENDIX II
REPORTED RESULTS BY PARTICIPANTS

TABLE 7. REPORTED RESULTS BY PARTICIPANTS

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
2	Ag	mg kg-1	ICP-MS	1.31	0.66	1.32	2.00	0.1	0.0	NIST 1566b
7	Ag	mg kg-1	ICP-MS	1.01	0.05			-1.8	-3.9	
8	Ag	mg kg-1	NAA	1.36	0.07	0.14	2.00	0.4	0.8	NIST 1547
9	Ag	mg kg-1	ICP-MS	1.15	0.23	0.46	2.00	-0.9	-0.6	
13	Ag	mg kg-1	ICP-MS	1.24	0.19	0.37	2.00	-0.3	-0.3	DORM-2
16	Ag	mg kg-1	ICP-MS	1.25	0.07	0.14	2.00	-0.2	-0.5	DOLT-4
19	Ag	mg kg-1	ICP-MS	1.31	0.10	0.20	2.00	0.1	0.2	NIST 1566b
25	Ag	mg kg-1	NAA	1.47	0.04	0.07	2.00	1.1	2.9	NIST 1566b
26	Ag	mg kg-1	ICP-MS	1.54	0.02	0.03	2.00	1.6	4.7	NIST 1566b
28	Ag	mg kg-1	ICP-MS	1.20	0.02	0.04	2.00	-0.6	-1.7	ESA-2
33	Ag	mg kg-1	ICP-OES	1.25	0.13	0.26	2.00	-0.3	-0.3	DOLT-4
35	Ag	mg kg-1	ICP-MS	1.31	0.08	0.16	2.00	0.1	0.2	NIST 2976
22A	Ag	mg kg-1	NAA	1.18	0.06	0.11	2.00	-0.7	-1.5	DOLT-4
1	Al	mg kg-1	ICP-MS	58.6	4.1	8.1	2.0	-1.6	-1.3	NIST 1515
5	Al	mg kg-1	ICP-MS	49.0	1.6	3.2	2.0	-2.6	-2.4	NIST 1547
9	Al	mg kg-1	ICP-OES	72.3	9.4	18.8	2.0	-0.1	0.0	
10	Al	mg kg-1	ET-AAS	74.2	9.3	18.6	2.0	0.1	0.1	NIST 2976
13	Al	mg kg-1	ICP-MS	87.8	8.8	17.6	2.0	1.6	1.1	DORM-2
15	Al	mg kg-1	ICP-MS	83.7	7.5	15.1	2.0	1.2	0.9	TORT-3
16	Al	mg kg-1	ICP-MS	83.9	0.6	1.1	2.0	1.2	1.1	NIST 1548
25	Al	mg kg-1	NAA	95.0	2.5	5.0	2.0	2.4	2.2	NIST 1566b
28	Al	mg kg-1	ICP-MS	228.0	2.0	4.0	2.0	17	15	ESA-2
30	Al	mg kg-1	ICP-MS	47.5				-2.8		TORT-2

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
33	Al	mg kg-1	ICP-OES	40.5				-3.5		TORT-2
1	As	mg kg-1	ICP-MS	12.6	0.9	1.7	2.0	0.5	0.7	TORT-2
2	As	mg kg-1	ICP-MS	12.4	0.5	1.0	2.0	0.3	0.7	NIST 1566b
3	As	mg kg-1	ICP-OES	11.0				-0.6		IAEA-452
4	As	mg kg-1	ICP-MS	10.5	0.2	0.3	2.0	-0.9	-2.9	TORT-2
5	As	mg kg-1	ICP-MS	12.4	0.1	0.3	2.0	0.3	1.1	DORM-3
6	As	mg kg-1	ET-AAS	11.8	0.3	0.7	2.0	-0.1	-0.2	DORM-3
7	As	mg kg-1	ICP-MS	12.6	0.5	5.0	2.0	0.5	1.0	TORT-2
8	As	mg kg-1	NAA	11.8	0.5	1.0	2.0	-0.1	-0.2	NIST 1547
9	As	mg kg-1	ICP-MS	10.2	2.6	5.1	2.0	-1.1	-0.6	IAEA-SL-1
10	As	mg kg-1	ET-AAS	11.8	1.5	3.0	2.0	-0.1	-0.1	NIST 2976
13	As	mg kg-1	ICP-MS	11.4	1.1	2.3	2.0	-0.3	-0.4	DORM-2
14	As	mg kg-1	NAA	11.2	0.4	0.8	2.0	-0.5	-1.2	NIST 2976
15	As	mg kg-1	ICP-MS	12.8	4.3	8.7	2.0	0.6	0.2	TORT-3
16	As	mg kg-1	ICP-MS	10.8	0.1	0.2	2.0	-0.7	-2.4	NIST 1548
18	As	mg kg-1	ICP-MS	13.1	0.6	1.2	2.0	0.8	1.6	IAEA-407
23	As	mg kg-1	ICP-MS	12.6	0.5	0.9	2.0	0.5	1.1	ERM-CE278K
25	As	mg kg-1	NAA	12.0	0.3	0.6	2.0	0.1	0.2	NIST 1566b
26	As	mg kg-1	ICP-MS	13.5	0.2	0.4	2.0	1.1	3.3	NIST 1566b
28	As	mg kg-1	ICP-MS	10.6	0.6	1.2	2.0	-0.9	-1.7	ESA-2
30	As	mg kg-1	ICP-MS	14.5				1.7		TORT-2
33	As	mg kg-1	ICP-OES	11.9	0.5	1.0	2.0	0.0	0.0	TORT-2

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
34	As	mg kg-1	ET-AAS	12.6	1.0	2.0	2.0	0.5	0.6	TORT-2
35	As	mg kg-1	ICP-MS	12.3		1.5	2.0	0.3		NIST 2976
22A	As	mg kg-1	NAA	12.0	0.7	1.4	2.0	0.1	0.1	DOLT-4
15	B	mg kg-1	Other	20.5	2.6	5.1	2.0			TORT-3
30	B	mg kg-1	ICP-OES	20.9						TORT-2
7	Ba	mg kg-1	ICP-MS	0.554	0.010					
15	Ba	mg kg-1	ICP-MS	0.494	0.037	0.074	2			TORT-3
28	Ba	mg kg-1	ICP-MS	0.910	0.040	0.080	2			ESA-2
8	Br	mg kg-1	NAA	195.3	7.0	14.0	2.0			NIST 1547
14	Br	mg kg-1	NAA	203.0	7.0	14.0	2.0			NIST 2976
25	Br	mg kg-1	NAA	194.0	4.5	9.0	2.0			NIST 1566b
22A	Br	mg kg-1	NAA	170.0	8.0	16.0	2.0			IAEA-407
1	Ca	mg kg-1	ICP-MS	2620	156	312	2	0.6	0.9	NIST 1515
5	Ca	mg kg-1	ICP-MS	2530	8	15	2	0.3	0.7	NIST 1547
7	Ca	mg kg-1	ICP-MS	2473	115	495	2	0.1	0.2	TORT-2
8	Ca	mg kg-1	NAA	2282	104	208	2	-0.5	-0.8	NIST 1547
9	Ca	mg kg-1	ICP-OES	2243	337	673	2	-0.6	-0.5	
14	Ca	mg kg-1	NAA	2537	120	240	2	0.4	0.6	NIST 2976
15	Ca	mg kg-1	ICP-MS	2548	192	384	2	0.4	0.5	TORT-3
16	Ca	mg kg-1	ICP-MS	2310	134	267	2	-0.4	-0.6	NIST 1548
25	Ca	mg kg-1	NAA	2140	214	430	2	-1.0	-1.1	NIST 1566b
26	Ca	mg kg-1	ICP-MS	2571	32	64	2	0.5	1.0	NIST 1566b

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
28	Ca	mg kg-1	ICP-MS	2393	30	60	2	-0.1	-0.3	ESA-2
30	Ca	mg kg-1	ICP-OES	2480				0.2		TORT-2
1	Cd	mg kg-1	ICP-MS	3.17	0.13	0.25	2.00	0.1	0.2	TORT-2
2	Cd	mg kg-1	ICP-MS	3.32	0.11	0.21	2.00	0.4	1.1	NIST 1566b
3	Cd	mg kg-1	ET-AAS	3.18				0.1		IAEA-452
4	Cd	mg kg-1	ICP-MS	3.10	0.11	0.22	2.00	-0.1	-0.2	TORT-2
5	Cd	mg kg-1	ICP-MS	3.28	0.04	0.08	2.00	0.4	1.1	DORM-3
6	Cd	mg kg-1	F-AAS	2.92	0.07	0.14	2.00	-0.6	-1.6	NIST 1566b
7	Cd	mg kg-1	ICP-MS	3.26	0.18	1.30	2.00	0.3	0.6	TORT-2
9	Cd	mg kg-1	ICP-MS	2.65	0.67	1.33	2.00	-1.2	-0.7	IAEA-SL-1
10	Cd	mg kg-1	ET-AAS	2.93	0.37	0.73	2.00	-0.5	-0.5	NIST 2976
13	Cd	mg kg-1	ICP-MS	3.26	0.33	0.65	2.00	0.3	0.4	IAEA-452
14	Cd	mg kg-1	ET-AAS	2.46	0.10	0.19	2.00	-1.7	-4.4	NIST 2976
15	Cd	mg kg-1	ICP-MS	3.46	0.29	0.58	2.00	0.8	1.0	TORT-3
16	Cd	mg kg-1	ICP-MS	3.29	0.19	0.38	2.00	0.4	0.7	TORT-2
17	Cd	mg kg-1	ET-AAS	2.61			2.00	-1.4		DOLT-2
18	Cd	mg kg-1	ICP-MS	3.26	0.16	0.31	1.96	0.3	0.6	IAEA-140
19	Cd	mg kg-1	ICP-MS	3.23	0.24	0.48	2.00	0.2	0.3	NIST 1566b
23	Cd	mg kg-1	ICP-MS	3.10	0.10	0.20	2.00	-0.1	-0.3	ERM-CE278K
26	Cd	mg kg-1	ICP-MS	3.71	0.05	0.09	2.00	1.4	4.4	NIST 1566b
27	Cd	mg kg-1	F-AAS	2.55				-1.5		TORT-2
28	Cd	mg kg-1	ICP-MS	3.01	0.07	0.14	2.00	-0.3	-0.9	ESA-2

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
30	Cd	mg kg-1	ICP-MS	3.14				0.0		TORT-2
31	Cd	mg kg-1	SV	3.07	0.03	0.06	2.00	-0.2	-0.6	NIST 1566b
32	Cd	mg kg-1	ICP-MS	3.03	0.12	0.24	2.00	-0.3	-0.6	NIST 1566a
33	Cd	mg kg-1	ICP-OES	3.52	0.11	0.21	2.00	1.0	2.4	TORT-2
35	Cd	mg kg-1	ICP-MS	3.41	0.20	0.40	2.00	0.7	1.2	NIST 2976
22B	Cd	mg kg-1	ICP-MS	2.47	0.06	0.12	2.00	-1.7	-5.0	NIST 1566B
15	Ce	mg kg-1	Other	0.025	0.003	0.006	2			TORT-3
5	CH ₃ Hg	µg kg-1 as Hg	CV-AFS	5.31	0.13	0.26	2.00	0.1	0.2	DOLT-4
8	CH ₃ Hg	µg kg-1 as Hg	CV-AFS	3.21	0.14	0.28	2.00	-3.1	-4.4	T-38
12	CH ₃ Hg	µg kg-1 as Hg	CV-AFS	5.29	0.53	1.06	2.00	0.1	0.1	DORM-2
15	CH ₃ Hg	µg kg-1 as Hg	GC-AFS	6.50	0.64	1.29	2.00	2.0	1.6	TORT-3
20	CH ₃ Hg	µg kg-1 as Hg	GC-ECD	6.04				1.3		IAEA-452
26	CH ₃ Hg	µg kg-1 as Hg	HPLC-AFS	7.11	0.24	0.48	2.00	2.9	3.8	NIST 1566b
27	CH ₃ Hg	µg kg-1 as Hg	CV-AAS	9.22				6.1		TORT-2
30	CH ₃ Hg	µg kg-1 as Hg	Solid-AAS	7.70				3.8		TORT-2
34	CH ₃ Hg	µg kg-1 as Hg	Solid-AAS	5.10	0.40	0.80	2.00	-0.2	-0.2	IAEA-452
35	CH ₃ Hg	µg kg-1 as Hg	ID-ICP-MS	4.80	0.10	0.20	2.00	-0.6	-0.9	IAEA-461
36	CH ₃ Hg	µg kg-1 as Hg	GC-AFS	5.36	0.40	0.80	2.00	0.2	0.2	IAEA-452
1	Co	mg kg-1	ICP-MS	0.200	0.015	0.030	2	0.0	-0.1	
2	Co	mg kg-1	ICP-MS	0.213	0.007	0.014	2	0.5	0.8	NIST 1566b

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
4	Co	mg kg-1	ICP-MS	0.162	0.008	0.015	2	-1.6	-2.7	TORT-2
5	Co	mg kg-1	ICP-MS	0.199	0.005	0.011	2	-0.1	-0.1	NIST 1547
7	Co	mg kg-1	ICP-MS	0.208	0.007			0.3	0.5	TORT-2
8	Co	mg kg-1	NAA	0.184	0.008	0.016	2	-0.7	-1.2	NIST 1547
9	Co	mg kg-1	ICP-MS	0.197	0.040	0.079	2	-0.2	-0.1	IAEA-SL-1
13	Co	mg kg-1	ICP-MS	0.199	0.020	0.040	2	-0.1	-0.1	DORM-2
14	Co	mg kg-1	NAA	0.366	0.010	0.020	2	6.6	10.3	NIST 2976
15	Co	mg kg-1	ICP-MS	0.206	0.018	0.035	2	0.2	0.2	TORT-3
16	Co	mg kg-1	ICP-MS	0.178	0.011	0.021	2	-0.9	-1.4	TORT-2
18	Co	mg kg-1	ICP-MS	0.200	0.014	0.027	2	-0.1	-0.1	IAEA-407
26	Co	mg kg-1	ICP-MS	0.219	0.002	0.005	2	0.7	1.4	NIST 1566b
27	Co	mg kg-1	ET-AAS	0.370				6.7		TORT-2
28	Co	mg kg-1	ICP-MS	0.180	0.010	0.020	2	-0.8	-1.3	ESA-2
30	Co	mg kg-1	ICP-MS	0.170				-1.2		TORT-2
32	Co	mg kg-1	ICP-MS	0.261	0.013	0.026	2	2.4	3.3	NIST 1566a
34	Co	mg kg-1	ET-AAS	0.164	0.016	0.033	2	-1.5	-1.8	TORT-2
35	Co	mg kg-1	ICP-MS	0.210	0.012	0.025	2	0.4	0.5	NIST 2976
22A	Co	mg kg-1	NAA	0.182	0.008	0.016	2	-0.8	-1.3	DOLT-4
22B	Co	mg kg-1	ICP-MS	0.245	0.009	0.018	2	1.8	2.9	NIST 1566B
1	Cr	mg kg-1	ICP-MS	1.00	0.10	0.19	2.00	0.3	0.3	TORT-2
2	Cr	mg kg-1	ICP-MS	1.22	0.05	0.10	2.00	2.1	3.5	NIST 1566b
3	Cr	mg kg-1	ICP-OES	1.15				1.5		IAEA-140

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
5	Cr	mg kg-1	ICP-MS	0.758	0.060	0.120	2	-1.7	-2.5	DORM-3
7	Cr	mg kg-1	ICP-MS	0.956	0.047			-0.1	-0.1	TORT-2
8	Cr	mg kg-1	NAA	0.977	0.060	0.120	2	0.1	0.1	NIST 1547
9	Cr	mg kg-1	ICP-MS	0.843	0.085	0.169	2	-1.0	-1.2	IAEA-SL-1
10	Cr	mg kg-1	ET-AAS	0.820	0.103	0.205	2	-1.2	-1.2	NIST 2976
13	Cr	mg kg-1	ICP-MS	0.965	0.145	0.290	2	0.0	0.0	DORM-2
14	Cr	mg kg-1	NAA	0.840	0.060	0.120	2	-1.0	-1.5	NIST 2976
15	Cr	mg kg-1	ICP-MS	0.903	0.146	0.291	2	-0.5	-0.4	TORT-3
16	Cr	mg kg-1	ICP-MS	0.848	0.049	0.098	2	-1.0	-1.6	TORT-2
18	Cr	mg kg-1	ICP-MS	0.923	0.051	0.099	2	-0.4	-0.6	IAEA-407
26	Cr	mg kg-1	ICP-MS	0.963	0.011	0.022	2	0.0	0.0	
27	Cr	mg kg-1	ET-AAS	0.980				0.1		TORT-2
28	Cr	mg kg-1	ICP-MS	1.01	0.02	0.04	2.00	0.3	0.7	ESA-2
33	Cr	mg kg-1	ICP-OES	0.776	0.135	0.270	2	-1.6	-1.3	TORT-2
34	Cr	mg kg-1	ET-AAS	1.03	0.10	0.20	2.00	0.5	0.6	TORT-2
22A	Cr	mg kg-1	NAA	0.990	0.120	0.240	2	0.2	0.2	DORM-2
8	Cs	mg kg-1	NAA	0.034	0.001	0.002	2			NIST 1547
14	Cs	mg kg-1	NAA	0.027	0.005	0.010	2			NIST 2976
1	Cu	mg kg-1	ICP-MS	144.0	5.5	11.0	2.0	-0.1	-0.2	TORT-2
2	Cu	mg kg-1	ICP-MS	152.4	4.3	8.6	2.0	0.4	0.8	NIST 1566b
3	Cu	mg kg-1	ICP-OES	162.0				0.9		
4	Cu	mg kg-1	ICP-MS	130.0	4.5	9.0	2.0	-0.9	-2.0	TORT-2

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
5	Cu	mg kg-1	ICP-MS	146.0	3.4	6.9	2.0	0.0	0.0	DORM-3
6	Cu	mg kg-1	F-AAS	143.0	2.0	3.9	2.0	-0.2	-0.4	DORM-3
7	Cu	mg kg-1	ICP-MS	138.6	4.4	69.3	2.0	-0.4	-0.9	TORT-2
9	Cu	mg kg-1	ICP-MS	131.0	13.0	26.0	2.0	-0.8	-1.0	IAEA-SL-1
10	Cu	mg kg-1	ET-AAS	146.0	18.3	36.5	2.0	0.0	0.0	NIST 2976
13	Cu	mg kg-1	ICP-MS	144.8	10.9	21.7	2.0	-0.1	-0.1	IAEA-452
15	Cu	mg kg-1	ICP-MS	154.7	11.9	23.8	2.0	0.5	0.6	TORT-3
16	Cu	mg kg-1	ICP-MS	153.0	8.5	17.0	2.0	0.4	0.7	TORT-2
17	Cu	mg kg-1	ET-AAS	154.8			2.0	0.5		DOLT-2
18	Cu	mg kg-1	F-AAS	146.2	10.3	20.3	2.0	0.0	0.0	IAEA-407
19	Cu	mg kg-1	ICP-MS	156.5	7.8	15.7	2.0	0.6	1.0	NIST 1566b
23	Cu	mg kg-1	ICP-MS	157.3	6.2	12.4	2.0	0.6	1.3	ERM-CE278K
26	Cu	mg kg-1	ICP-MS	161.0	2.0	4.0	2.0	0.8	2.2	NIST 1566b
27	Cu	mg kg-1	F-AAS	96.2				-2.7		TORT-2
28	Cu	mg kg-1	ICP-MS	136.7	3.0	6.0	2.0	-0.5	-1.3	ESA-2
30	Cu	mg kg-1	ICP-OES	149.0				0.2		TORT-2
31	Cu	mg kg-1	SV	133.0	2.0	4.0	2.0	-0.7	-1.9	NIST 1566b
32	Cu	mg kg-1	ICP-MS	142.1	5.7	11.4	2.0	-0.2	-0.5	NIST 1566a
33	Cu	mg kg-1	ICP-OES	142.9	4.6	9.2	2.0	-0.2	-0.4	TORT-2
34	Cu	mg kg-1	F-AAS	150.0	8.0	15.0	2.0	0.2	0.4	TORT-2
35	Cu	mg kg-1	ICP-MS	153.0	9.0	18.0	2.0	0.4	0.6	NIST 2976
22B	Cu	mg kg-1	ICP-MS	140.0	3.0	6.0	2.0	-0.3	-0.8	NIST 1566B

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
1	Fe	mg kg-1	ICP-MS	137.0	8.5	17.0	2.0	0.4	0.6	TORT-2
2	Fe	mg kg-1	ICP-MS	124.3	4.8	9.6	2.0	-0.4	-0.9	NIST 1566b
3	Fe	mg kg-1	ICP-OES	138.0				0.4		IAEA-452
5	Fe	mg kg-1	ICP-MS	125.0	3.0	5.9	2.0	-0.4	-0.9	DORM-3
6	Fe	mg kg-1	F-AAS	131.0	1.6	3.2	2.0	0.0	0.0	DORM-3
7	Fe	mg kg-1	ICP-MS	132.1	1.2			0.1	0.2	TORT-2
8	Fe	mg kg-1	NAA	133.0	5.0	10.0	2.0	0.1	0.3	NIST 1547
9	Fe	mg kg-1	ICP-MS	129.0	18.0	36.0	2.0	-0.1	-0.1	IAEA-SL-1
10	Fe	mg kg-1	ET-AAS	135.0	16.9	33.8	2.0	0.2	0.2	NIST 2976
13	Fe	mg kg-1	ICP-MS	123.5	9.3	18.5	2.0	-0.5	-0.7	DORM-2
14	Fe	mg kg-1	NAA	123.0	4.0	8.0	2.0	-0.5	-1.3	NIST 2976
15	Fe	mg kg-1	ICP-OES	139.1	15.8	31.6	2.0	0.5	0.5	TORT-3
16	Fe	mg kg-1	ICP-MS	135.0	7.5	15.0	2.0	0.2	0.4	GBW10014
17	Fe	mg kg-1	ET-AAS	153.2				1.4		DOLT-2
18	Fe	mg kg-1	F-AAS	127.3	6.9	13.5	2.0	-0.2	-0.4	IAEA-407
25	Fe	mg kg-1	NAA	136.0	4.9	10.0	2.0	0.3	0.6	NIST 1566b
26	Fe	mg kg-1	ICP-MS	153.3	1.4	2.7	2.0	1.4	3.6	NIST 1566b
27	Fe	mg kg-1	F-AAS	102.0				-1.8		TORT-2
28	Fe	mg kg-1	ICP-MS	168.7	2.0	4.0	2.0	2.3	6.0	ESA-2
30	Fe	mg kg-1	ICP-OES	126.0				-0.3		TORT-2
32	Fe	mg kg-1	ICP-MS	141.4	8.5	17.0	2.0	0.6	1.0	NIST 1566a
34	Fe	mg kg-1	F-AAS	131.0	12.0	23.0	2.0	0.0	0.0	TORT-2

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
22A	Fe	mg kg-1	NAA	127.9	4.8	9.6	2.0	-0.2	-0.4	DOLT-4
1	Hg	µg kg-1	CV-AFS	18.0	2.5	5.0	2.0	-1.2	-1.1	TORT-2
5	Hg	µg kg-1	CV-AFS	14.8	0.2	0.4	2.0	-2.4	-5.9	DORM-3
6	Hg	µg kg-1	CV-AFS	27.0	2.0	4.0	2.0	2.2	2.6	DORM-3
7	Hg	µg kg-1	ICP-MS	18.0	1.0	7.0	2.0	-1.2	-2.1	TORT-2
8	Hg	µg kg-1	Solid-AAS	19.9	0.5	1.0	2.0	-0.5	-1.0	NIST 1566b
9	Hg	µg kg-1	CV-AAS	21.0	8.0	16.0	2.0	0.0	0.0	IAEA-SL-1
12	Hg	µg kg-1	CV-AFS	18.7	1.7	3.4	2.0	-0.9	-1.2	DORM-3
14	Hg	µg kg-1	CV-AAS	23.7	2.1	4.3	2.0	1.0	1.1	NIST 1566b
15	Hg	µg kg-1	CV-AFS	22.3	3.4	6.7	2.0	0.5	0.3	TORT-3
16	Hg	µg kg-1	ICP-MS	6.41	0.35	0.70	2.00	-5.6	-13.3	TORT-2
17	Hg	µg kg-1	Solid-AAS	21.5			2.0	0.2		IAEA-452
18	Hg	µg kg-1	CV-AAS	27.2	1.6	3.1	2.0	2.3	3.2	IAEA-140
19	Hg	µg kg-1	CV-AAS	21.0	2.0	3.0	2.0	0.0	0.0	IAEA-142
20	Hg	µg kg-1	Solid-AAS	14.8				-2.4		NIST 2976
26	Hg	µg kg-1 as Hg	ICP-MS	25.0	0.9	1.7	2.0	1.5	2.8	
27	Hg	µg kg-1 as Hg	CV-AAS	35.0				5.3		TORT-2
28	Hg	µg kg-1	Solid-AAS	20.2	0.2	0.4	2.0	-0.4	-0.9	ESA-2
30	Hg	µg kg-1	Solid-AAS	19.7				-0.5		TORT-2
33	Hg	µg kg-1	Solid-AAS	16.5	1.2	2.4	2.0	-1.7	-2.9	TORT-2
34	Hg	µg kg-1	Solid-AAS	22.5	0.9	1.8	2.0	0.5	1.0	NIST 2976
35	Hg	µg kg-1	ID-ICP-MS	21.3		1.1	2.0	0.1		

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
36	Hg	$\mu\text{g kg}^{-1}$	CV-AFS	23.8	1.6	3.2	2.0	1.0	1.4	IAEA-452
8B	Hg	$\mu\text{g kg}^{-1}$	CV-AAS	20.1	1.1	2.2	2.0	-0.4	-0.7	NIST 1566b
7	K	mg kg^{-1}	ICP-MS	13332	600	4000	2	0.4	0.5	TORT-2
8	K	mg kg^{-1}	NAA	11305	714	1428	2	-0.9	-1.1	NIST 1547
15	K	mg kg^{-1}	ICP-OES	14942	1302	2603	2	1.4	1.3	TORT-3
25	K	mg kg^{-1}	NAA	11200	258	511	2	-1.0	-1.4	NIST 1566b
28	K	mg kg^{-1}	ICP-MS	11867	400	800	2	-0.5	-0.7	ESA-2
30	K	mg kg^{-1}	ICP-OES	13700				0.6		TORT-2
2	Li	mg kg^{-1}	ICP-MS	0.535	0.015	0.031	2	0.0	0.1	NIST 1566b
9	Li	mg kg^{-1}	ICP-MS	0.416	0.083	0.166	2	-1.7	-1.2	
16	Li	mg kg^{-1}	ICP-MS	0.474	0.028	0.055	2	-0.9	-1.1	GBW10014
18	Li	mg kg^{-1}	ICP-MS	0.602	0.029	0.056	2	1.0	1.3	IAEA-407
28	Li	mg kg^{-1}	ICP-MS	0.593	0.020	0.040	2	0.9	1.2	ESA-2
34	Li	mg kg^{-1}	ET-AAS	0.557	0.045	0.090	2	0.4	0.4	TORT-2
1	Mg	mg kg^{-1}	ICP-MS	3270	153	305	2	0.5	0.8	NIST 1515
2	Mg	mg kg^{-1}	ICP-MS	3208	91	182	2	0.3	0.6	NIST 1566b
5	Mg	mg kg^{-1}	ICP-MS	3030	73	146	2	-0.1	-0.2	NIST 1547
7	Mg	mg kg^{-1}	ICP-MS	3112	124	622	2	0.1	0.1	TORT-2
9	Mg	mg kg^{-1}	ICP-OES	2968	445	890	2	-0.3	-0.2	
15	Mg	mg kg^{-1}	ICP-MS	3322	260	521	2	0.6	0.7	TORT-3
16	Mg	mg kg^{-1}	ICP-MS	2960	170	340	2	-0.3	-0.5	NIST 1548
25	Mg	mg kg^{-1}	NAA	2815	101	203	2	-0.7	-1.2	NIST 1566b

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
26	Mg	mg kg-1	ICP-MS	3687	46	93	2	1.6	3.0	NIST 1566b
28	Mg	mg kg-1	ICP-MS	3007	60	120	2	-0.2	-0.4	ESA-2
30	Mg	mg kg-1	ICP-OES	3060				-0.1		TORT-2
1	Mn	mg kg-1	ICP-MS	74.5	3.8	7.5	2.0	0.9	1.7	TORT-2
2	Mn	mg kg-1	ICP-MS	66.7	1.7	3.4	2.0	0.0	0.0	NIST 1566b
3	Mn	mg kg-1	ICP-OES	71.8				0.6		IAEA-140
4	Mn	mg kg-1	ICP-MS	62.3	2.2	4.4	2.0	-0.5	-1.3	TORT-2
5	Mn	mg kg-1	ICP-MS	64.7	0.4	0.7	2.0	-0.2	-0.7	NIST 1547
6	Mn	mg kg-1	F-AAS	66.3	1.7	3.3	2.0	0.0	-0.1	NIST 1566b
7	Mn	mg kg-1	ICP-MS	73.1	2.6			0.8	1.7	TORT-2
9	Mn	mg kg-1	ICP-OES	61.3	9.2	18.4	2.0	-0.6	-0.6	IAEA-SL-1
10	Mn	mg kg-1	ET-AAS	65.0	8.1	16.3	2.0	-0.2	-0.2	NIST 2976
13	Mn	mg kg-1	ICP-MS	67.9	3.4	6.8	2.0	0.1	0.3	IAEA-452
15	Mn	mg kg-1	ICP-MS	73.3	5.0	10.1	2.0	0.8	1.2	TORT-3
16	Mn	mg kg-1	ICP-MS	68.0	3.9	7.8	2.0	0.2	0.3	GBW10014
17	Mn	mg kg-1	ET-AAS	64.0				-0.3		DOLT-2
18	Mn	mg kg-1	ICP-MS	80.8	4.2	8.2	2.0	1.7	2.9	IAEA-140
25	Mn	mg kg-1	NAA	63.0	6.3	13.0	2.0	-0.4	-0.5	NIST 1566b
26	Mn	mg kg-1	ICP-MS	77.9	0.8	1.5	2.0	1.3	4.0	NIST 1566b
27	Mn	mg kg-1	F-AAS	52.4				-1.7		TORT-2
28	Mn	mg kg-1	ICP-MS	60.2	1.0	2.0	2.0	-0.8	-2.3	ESA-2
30	Mn	mg kg-1	ICP-OES	67.3				0.1		TORT-2

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
32	Mn	mg kg-1	ICP-MS	69.5	2.8	5.6	2.0	0.3	0.7	NIST 1566a
33	Mn	mg kg-1	ICP-OES	65.1	3.5	7.0	2.0	-0.2	-0.4	TORT-2
34	Mn	mg kg-1	F-AAS	73.3	5.9	11.7	2.0	0.8	1.0	TORT-2
35	Mn	mg kg-1	ICP-MS	66.9	4.0	8.0	2.0	0.0	0.0	NIST 2976
15	Mo	mg kg-1	ICP-MS	0.316	0.051	0.101	2			TORT-3
30	Mo	mg kg-1	ICP-MS	0.490						TORT-2
2	Na	mg kg-1	ICP-MS	17660	720	1430	2	-0.8	-1.5	NIST 1566b
7	Na	mg kg-1	ICP-MS	21248	718	8499	2	0.6	1.1	TORT-2
8	Na	mg kg-1	NAA	20553	829	1658	2	0.3	0.6	NIST 1547
14	Na	mg kg-1	NAA	18828	339	678	2	-0.4	-0.7	NIST 2976
15	Na	mg kg-1	ICP-OES	22213	1658	3315	2	1.0	1.2	TORT-3
25	Na	mg kg-1	NAA	20276	487	810	2	0.2	0.5	NIST 1566b
28	Na	mg kg-1	ICP-MS	19500	200	400	2	-0.1	-0.2	ESA-2
30	Na	mg kg-1	ICP-OES	22100				1.0		TORT-2
22A	Na	mg kg-1	NAA	19600	900	1800	2	0.0	-0.1	IAEA-407
1	Ni	mg kg-1	ICP-MS	0.770	0.075	0.150	2	-0.7	-0.5	TORT-2
2	Ni	mg kg-1	ICP-MS	0.876	0.026	0.052	2	0.3	0.2	NIST 1566b
3	Ni	mg kg-1	ET-AAS	0.802				-0.4		IAEA-140
5	Ni	mg kg-1	ICP-MS	0.660	0.010	0.020	2	-1.8	-1.3	DORM-3
7	Ni	mg kg-1	ICP-MS	0.820	0.005			-0.3	-0.2	TORT-2
9	Ni	mg kg-1	ICP-MS	0.758	0.133	0.265	2	-0.9	-0.5	IAEA-SL-1
10	Ni	mg kg-1	ET-AAS	0.950	0.119	0.238	2	1.0	0.5	NIST 2976

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
13	Ni	mg kg-1	ICP-MS	0.820	0.082	0.164	2	-0.3	-0.2	DORM-2
15	Ni	mg kg-1	ICP-MS	0.776	0.174	0.348	2	-0.7	-0.3	TORT-3
16	Ni	mg kg-1	ICP-MS	0.713	0.041	0.082	2	-1.3	-0.9	NIST 1548
18	Ni	mg kg-1	ICP-MS	0.963	0.122	0.240	2	1.1	0.6	IAEA-407
19	Ni	mg kg-1	ICP-MS	0.920	0.090	0.180	2	0.7	0.4	NIST 1566b
26	Ni	mg kg-1	ICP-MS	0.817	0.011	0.022	2	-0.3	-0.2	NIST 1566b
27	Ni	mg kg-1	ET-AAS	0.800				-0.5		TORT-2
28	Ni	mg kg-1	ICP-MS	0.913	0.060	0.120	2	0.6	0.4	ESA-2
30	Ni	mg kg-1	ICP-MS	0.690				-1.5		TORT-2
32	Ni	mg kg-1	ICP-MS	0.840	0.042	0.084	2	-0.1	-0.1	NIST 1566a
34	Ni	mg kg-1	ET-AAS	0.891	0.090	0.180	2	0.4	0.2	TORT-2
1	Pb	mg kg-1	ICP-MS	0.340	0.030	0.060	2	-0.5	-0.5	TORT-2
2	Pb	mg kg-1	ICP-MS	0.376	0.008	0.016	2	0.3	0.5	NIST 1566b
3	Pb	mg kg-1	ET-AAS	1.01				14		IAEA-140
5	Pb	mg kg-1	ICP-MS	0.355	0.005	0.010	2	-0.1	-0.2	NIST 1547
6	Pb	mg kg-1	ET-AAS	0.333	0.011	0.022	2	-0.6	-1.0	DORM-3
7	Pb	mg kg-1	ICP-MS	0.329	0.019	0.165	2	-0.7	-1.0	TORT-2
9	Pb	mg kg-1	ICP-MS	0.355	0.045	0.089	2	-0.1	-0.1	IAEA-SL-1
10	Pb	mg kg-1	ET-AAS	0.930	0.116	0.233	2	13	4.8	NIST 2976
13	Pb	mg kg-1	ICP-MS	0.370	0.046	0.093	2	0.2	0.2	DORM-2
14	Pb	mg kg-1	ET-AAS	0.381	0.038	0.076	2	0.4	0.4	NIST 2976
15	Pb	mg kg-1	ICP-MS	0.370	0.133	0.265	2	0.2	0.1	TORT-3

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
16	Pb	mg kg-1	ICP-MS	0.347	0.020	0.040	2	-0.3	-0.4	TORT-2
19	Pb	mg kg-1	ICP-MS	0.370	0.030	0.060	2	0.2	0.2	NIST 1566b
23	Pb	mg kg-1	ICP-MS	0.500	0.030	0.060	2	3.1	3.5	ERM-CE278K
26	Pb	mg kg-1	ICP-MS	0.380	0.003	0.007	2	0.4	0.7	NIST 1566b
27	Pb	mg kg-1	ET-AAS	0.220				-3.1		TORT-2
28	Pb	mg kg-1	ICP-MS	0.297	0.006	0.012	2	-1.4	-2.4	ESA-2
30	Pb	mg kg-1	ICP-MS	0.380				0.4		TORT-2
31	Pb	mg kg-1	SV	0.395	0.050	0.100	2	0.8	0.6	NIST 1566b
35	Pb	mg kg-1	ICP-MS	0.362	0.020	0.043	2	0.0	0.0	NIST 2976
22B	Pb	mg kg-1	ICP-MS	0.310	0.013	0.026	2	-1.1	-1.7	NIST 1566B
8	Rb	mg kg-1	NAA	5.13	0.19	0.38	2.00	0.0	0.0	NIST 1547
13	Rb	mg kg-1	ICP-MS	5.07	0.38	0.76	2.00	-0.1	-0.1	IAEA-452
14	Rb	mg kg-1	NAA	4.80	0.20	0.40	2.00	-0.5	-1.0	NIST 2976
15	Rb	mg kg-1	Other	5.41	0.51	1.02	2.00	0.4	0.5	TORT-3
25	Rb	mg kg-1	NAA	5.30	0.14	0.30	2.00	0.2	0.5	NIST 1566b
15	Sb	mg kg-1	Other	0.013	0.002	0.005	2			TORT-3
8	Sc	mg kg-1	NAA	0.026	0.001	0.002	2			NIST 1547
14	Sc	mg kg-1	NAA	0.026	0.0002	0.0004	2			NIST 2976
1	Se	mg kg-1	ICP-MS	3.73	0.18	0.36	2.00	1.8	2.7	TORT-2
5	Se	mg kg-1	ICP-MS	3.16	0.13	0.26	2.00	0.3	0.5	DORM-3
7	Se	mg kg-1	ICP-MS	3.00	0.02	1.50	2.00	-0.2	-0.4	TORT-2
8	Se	mg kg-1	NAA	3.22	0.13	0.26	2.00	0.4	0.7	NIST 1547

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
9	Se	mg kg-1	ICP-MS	2.66	0.67	1.33	2.00	-1.0	-0.6	IAEA-SL-1
13	Se	mg kg-1	ICP-MS	3.11	0.31	0.62	2.00	0.1	0.1	DORM-2
14	Se	mg kg-1	NAA	2.90	0.20	0.40	2.00	-0.4	-0.6	NIST 2976
15	Se	mg kg-1	ICP-MS	3.20	0.34	0.68	2.00	0.4	0.4	TORT-3
16	Se	mg kg-1	ICP-MS	2.92	0.17	0.34	2.00	-0.4	-0.6	TORT-2
25	Se	mg kg-1	NAA	3.20	0.08	0.16	2.00	0.4	0.8	NIST 1566b
26	Se	mg kg-1	ICP-MS	3.69	0.08	0.16	2.00	1.6	3.5	NIST 1566b
28	Se	mg kg-1	AFS	2.64	0.02	0.04	2.00	-1.1	-2.5	ESA-2
30	Se	mg kg-1	ICP-MS	3.19				0.3		TORT-2
34	Se	mg kg-1	ET-AAS	2.99	0.30	0.60	2.00	-0.2	-0.2	TORT-2
22A	Se	mg kg-1	NAA	2.74	0.15	0.30	2.00	-0.8	-1.4	DOLT-4
7	Sn	mg kg-1	ICP-MS	0.034	0.0005					
9	Sn	mg kg-1	ICP-MS	0.316	0.063	0.126	2			
16	Sn	mg kg-1	ICP-MS	0.000036	0.000002	0.000004	2			NIST 1548
2	Sr	mg kg-1	ICP-MS	21.1	0.5	1.1	2.0	0.2	0.5	NIST 1566b
3	Sr	mg kg-1	ICP-OES	17.1				-1.4		IAEA-140
7	Sr	mg kg-1	ICP-MS	23.2	1.1			1.0	2.0	TORT-2
9	Sr	mg kg-1	ICP-MS	20.3	3.1	6.1	2.0	-0.1	-0.1	IAEA-SL-1
13	Sr	mg kg-1	ICP-MS	20.0	1.5	3.0	2.0	-0.2	-0.3	IAEA-452
15	Sr	mg kg-1	ICP-MS	20.5	1.6	3.1	2.0	0.0	-0.1	TORT-3
16	Sr	mg kg-1	ICP-MS	18.5	1.1	2.1	2.0	-0.8	-1.6	GBW10014
18	Sr	mg kg-1	ICP-MS	27.1	2.2	4.2	2.0	2.5	2.8	IAEA-407

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
26	Sr	mg kg-1	ICP-MS	24.8	0.3	0.7	2.0	1.6	4.8	NIST 1566b
28	Sr	mg kg-1	ICP-MS	18.9	0.5	1.0	2.0	-0.7	-1.8	ESA-2
30	Sr	mg kg-1	ICP-MS	20.3				-0.1		TORT-2
32	Sr	mg kg-1	ICP-MS	21.7	0.9	1.7	2.0	0.4	0.9	NIST 1566a
34	Sr	mg kg-1	ET-AAS	21.1	1.7	3.4	2.0	0.2	0.3	TORT-2
35	Sr	mg kg-1	ICP-MS	22.3	1.4	2.7	2.0	0.7	1.1	NIST 2976
22B	Sr	mg kg-1	ICP-MS	18.0	0.4	0.8	2.0	-1.0	-2.9	NIST 1566B
30	Ti	mg kg-1	ICP-MS	10.3						TORT-2
15	Tl	mg kg-1	Other	0.006	0.001	0.002	2			TORT-3
8	U	mg kg-1	NAA	0.146	0.005	0.010	2			NIST 1547
15	U	mg kg-1	ICP-MS	0.164	0.036	0.073	2			TORT-3
32	U	mg kg-1	ICP-MS	0.171	0.007	0.014	2			NIST 1566a
1	V	mg kg-1	ICP-MS	0.920	0.085	0.170	2	0.2	0.2	
2	V	mg kg-1	ICP-MS	0.977	0.037	0.073	2	0.7	1.0	NIST 1566b
3	V	mg kg-1	ET-AAS	0.610				-2.6		IAEA-140
5	V	mg kg-1	ICP-MS	0.881	0.065	0.130	2	-0.2	-0.2	NIST 1547
7	V	mg kg-1	ICP-MS	0.935	0.036	0.374		0.3	0.5	TORT-2
9	V	mg kg-1	ICP-MS	0.892	0.112	0.223	2	-0.1	-0.1	IAEA-SL-1
13	V	mg kg-1	ICP-MS	0.846	0.085	0.169	2	-0.5	-0.5	IAEA-452
15	V	mg kg-1	ICP-MS	0.964	0.201	0.402	2	0.6	0.3	TORT-3
16	V	mg kg-1	ICP-MS	0.822	0.048	0.095	2	-0.7	-1.0	TORT-2
18	V	mg kg-1	ICP-MS	0.919	0.062	0.121	2	0.2	0.2	IAEA-407

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
25	V	mg kg-1	NAA	0.800	0.020	0.040	2	-0.9	-1.5	NIST 1566b
26	V	mg kg-1	ICP-MS	0.958	0.010	0.019	2	0.5	0.9	NIST 1566b
27	V	mg kg-1	ET-AAS	0.860				-0.3		TORT-2
28	V	mg kg-1	ICP-MS	0.880	0.020	0.040	2	-0.2	-0.3	ESA-2
32	V	mg kg-1	ICP-MS	1.16	0.05	0.09	2.00	2.3	3.3	NIST 1566a
33	V	mg kg-1	ICP-OES	0.766	0.094	0.189	2	-1.2	-1.2	TORT-2
34	V	mg kg-1	ET-AAS	0.785	0.078	0.157	2	-1.0	-1.1	TORT-2
1	Zn	mg kg-1	ICP-MS	773.0	32.0	64.0	2.0	0.5	1.2	TORT-2
2	Zn	mg kg-1	ICP-MS	756.0	24.2	48.4	2.0	0.3	0.9	NIST 1566b
3	Zn	mg kg-1	ICP-OES	802.0				0.8		IAEA-452
4	Zn	mg kg-1	ICP-MS	683.0	14.8	29.6	2.0	-0.5	-1.6	TORT-2
5	Zn	mg kg-1	ICP-MS	754.0	14.4	28.8	2.0	0.3	1.0	DORM-3
6	Zn	mg kg-1	F-AAS	682.0	20.3	40.6	2.0	-0.5	-1.4	DORM-3
7	Zn	mg kg-1	ICP-MS	732.1	29.6	366.1	2.0	0.1	0.1	TORT-2
8	Zn	mg kg-1	NAA	755.7	26.0	52.0	2.0	0.3	0.8	NIST 1547
9	Zn	mg kg-1	ICP-OES	607.0	61.0	121.0	2.0	-1.3	-1.8	IAEA-SL-1
10	Zn	mg kg-1	ET-AAS	598.0	74.8	149.5	2.0	-1.4	-1.6	NIST 2976
13	Zn	mg kg-1	ICP-MS	773.5	58.0	116.0	2.0	0.5	0.7	IAEA-452
14	Zn	mg kg-1	NAA	742.0	31.0	62.0	2.0	0.2	0.4	NIST 2976
15	Zn	mg kg-1	ICP-MS	778.1	54.5	109.0	2.0	0.6	0.9	TORT-3
16	Zn	mg kg-1	ICP-MS	741.0	43.0	86.0	2.0	0.2	0.3	TORT-2
18	Zn	mg kg-1	F-AAS	759.5	43.7	85.6	2.0	0.4	0.7	IAEA-407

TABLE 7. REPORTED RESULTS BY PARTICIPANTS (cont.)

Lab Code	Element	Unit	Instrumental Method	Lab Mean	Lab u	Lab U	k	z-score	Zeta-score	QC
19	Zn	mg kg-1	ICP-MS	795.0	40.0	80.0	2.0	0.7	1.5	NIST 1566b
25	Zn	mg kg-1	NAA	740.0	18.5	37.0	2.0	0.1	0.4	NIST 1566b
26	Zn	mg kg-1	ICP-MS	813.6	8.4	16.9	2.0	1.0	3.4	NIST 1566b
27	Zn	mg kg-1	F-AAS	623.0				-1.1		TORT-2
28	Zn	mg kg-1	ICP-MS	691.7	7.0	14.0	2.0	-0.4	-1.4	ESA-2
30	Zn	mg kg-1	ICP-OES	718.0				-0.1		TORT-2
31	Zn	mg kg-1	SV	729.0	8.6	17.2	2.0	0.0	0.1	NIST 1566b
32	Zn	mg kg-1	ICP-MS	668.1	26.7	53.4	2.0	-0.6	-1.6	NIST 1566a
33	Zn	mg kg-1	ICP-OES	756.7	27.5	54.9	2.0	0.3	0.8	TORT-2
34	Zn	mg kg-1	F-AAS	739.0	37.0	74.0	2.0	0.1	0.3	TORT-2
22A	Zn	mg kg-1	NAA	700.0	18.0	36.0	2.0	-0.3	-0.9	DOLT-4
22B	Zn	mg kg-1	ICP-MS	621.0	13.0	26.0	2.0	-1.2	-3.9	NIST 1566B

REFERENCES

- [1] CARVALHO, F.P., The inter-agency programme on marine pollution, IAEA Bulletin, **40**, 3 (1998) 7–10.
- [2] COQUERY, M., et al., The IAEA worldwide intercomparison exercises (1990-1997): determination of trace elements in marine sediments and biological samples, *Sci Total Environ.*, **30** (1999) 237–238, 501–8.
- [3] VILLENEUVE, J-P., DE MORA, S., CATTINI, C., Determination of organochlorinated compounds and petroleum hydrocarbons in fish-homogenate sample IAEA-406: Results from a worldwide interlaboratory study, *TrAC Trends in Analytical Chemistry*, **23**, 7 (2004) 501–510.
- [4] VILLENEUVE, J-P., DE MORA, S., CATTINI, C., Global intercomparison for the determination of chlorinated pesticides, PCBs and petroleum hydrocarbons in sediment sample IAEA-417, *Environ Technol.*, Nov; **23**, 11 (2002) 1203–17.
- [5] WYSE, E.J., et al., Characterization of trace elements and methylmercury in an estuarine sediment reference material, IAEA-405, *Journal of Environmental Monitoring*, **6** (2004) 48–57.
- [6] VASILEVA E., AZEMARD, S., BUSTAMANTE, J. OH, J., BETTI, M., Certification of IAEA-452 Biota Sample for Trace Elements and Methyl Mercury, *Accreditation and Quality Assurance*, **16**, (2011), 439-447.
- [7] INTERNATIONAL ATOMIC ENERGY AGENCY, Certification of Trace Element Mass Fractions in IAEA-470 Oyster Sample, IAEA Analytical Quality in Nuclear Applications Series No. **XX**, IAEA/AQ/XX, IAEA, Vienna (2016) In press
- [8] ROYAL SOCIETY OF CHEMISTRY, Statistical Subcommittee of the Analytical Methods Committee (AMC), AMC Technical Brief: Representing data distributions with Kernel density estimates” 2006, www.rsc.org/amc.
- [9] INTERNATIONAL ORGANISATION FOR STANDARDISATION, Guide 13528 2005, Statistical Methods for Use in Proficiency Testing by Interlaboratory Comparisons, ISO, Geneva, Switzerland.
- [10] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION, ISO Guide 35:2006, Reference Materials – General and Statistical Principles for Certification, ISO, Geneva (2006).
- [10] VASSILEVA E., C. QUÉTEL C, Influence of the Correction for Moisture Content on the Quality of the Certification (Uncertainty estimation) of Cd, Cu and Pb Mass Fractions in Rice, *Food Chemistry*, Volume **106**, (2008), 1485-1490.

LIST OF PARTICIPANTS

AUSTRIA

Katona, R. IAEA Laboratories Seibersdorf
Reaktorstrasse 1
A-2444 Seibersdorf

BRAZIL

Araújo, P. Centro de Tecnologia Mineral
Laboratorio de Especiacao de Mercurio Ambiental
Av Pedro Calmon 900 - Cicade Universitaria
Rio de Janeiro RJ - CEP 21941-508

Castilhos, Z. Centro de Tecnologia Mineral
Laboratorio de Especiacao de Mercurio Ambiental
Av Pedro Calmon 900 - Cicade Universitaria
Rio de Janeiro RJ - CEP 21941-508

Domingos, L. Centro de Tecnologia Mineral
Laboratorio de Especiacao de Mercurio Ambiental
Av Pedro Calmon 900 - Cicade Universitaria
Rio de Janeiro RJ - CEP 21941-508

Favaro, D. Comissao Nacional de Energia Nuclear
- IPEN-CNEN/SP
LAN/CRPq Travessa R n0 400
Cidade Universitária
05508-170 São Paulo - SP

Vasconcellos, M. Comissao Nacional de Energia Nuclear –
IPEN-CNEN/SP
LAN/CRPq Travessa R n0 400
Cidade Universitária
05508-170 São Paulo - SP

CANADA

Cram, J. ALS Environmental (Vancouver)
8081 Lougheed Highway Suite 100
Burnaby BC
V5A 1W9

Flett, R. Flett Research Ltd.
440 Desalaberry Ave
Winnipeg
Manitoba R2L 0Y7

Franco, H. ALS Environmental (Vancouver)
8081 Lougheed Highway Suite 100
Burnaby BC
V5A 1W9

Gilbert, D Flett Research Ltd.
440 Desalaberry Ave
Winnipeg
Manitoba R2L 0Y7

Harold, K. ALS Environmental (Winnipeg) –
Quality Systems Coordinator
12-1329 Niakwa Road East E
Winnipeg
MB R2J 3T4

Xiang, W. Flett Research Ltd.
440 Desalaberry Ave
Winnipeg
Manitoba R2L 0Y7

CHILE

Andonie, O. Comisión Chilena de Energía Nuclear
Nueva Bilbao 12501, Las Condes
7600713 Las Condes

CHINA

Lin, Z. National Marine Environmental Monitoring Center
NMEMC
No 42, Linghe Street, Shahekou Distric
P.O. Box 303, DALIAN 116023

Xiaowei, L. Key Lab. Of Food Safety Risk Assessment
China National Center for Food Safety Risk
Assessment (CFSA)
Room 310, Building 2, N°37 Guangqu Road,
Chaoyang District
100022 Beijing

Yao, Z., National Marine Environmental Monitoring Center
NMEMC
No 42, Linghe Street, Shahekou Distric
P.O. Box 303, Dalian 116023

CROATIA

Zurga, P. Institute of Public Health of Primorsko
goranska county
Kresimirova 52A
51000 Rijeka

FINLAND

Sara-Aho, T. SYKE, Finnish Environment Institute
Environmental Measurement and Testing Laboratory
Hakuninmaantie 6
FI-00430 Helsinki

France

Auger, D. IFREMER - Centre de Nantes
Rue de l'Île d' Yeu
B.P. 21105
F-44311 Nantes Cedex 03

Gourves, P.Y. Station marine
Equipe E.A
Place du Docteur Peyneau
33120 Arcachon

Guerin, T. ANSES - LSA - CIME unit
Laboratoire de sécurité des aliments de
Maisons-Alfort
23, Avenue du Général de Gaulle
94706 Maisons-Alfort Cedex

Noel, L. ANSES - LSA - CIME unit
Laboratoire de sécurité des aliments de
Maisons-Alfort
23, Avenue du Général de Gaulle
94706 Maisons-Alfort Cedex

GERMANY

Manthey-Karl, M. Max Rubner-Institut
Inst. Für Sicherheit und Qualität bei Milch und Fisch
Palmaille 9
22767 Hamburg

ISRAEL

Herut, B. Israel Oceanographic & Limnological Research,
Tel Shikmona,
P.O. Box 8030
31080 Haifa

Kress, N. Israel Oceanographic & Limnological Research,
Tel Shikmona,
P.O. Box 8030
31080 Haifa

Shefer, E. Israel Oceanographic & Limnological Research,
Tel Shikmona,
P.O. Box 8030
31080 Haifa

ITALY

Belli, M. ISPRA
Via di Castel Romano, 100
00128 Roma

NORWAY

Kristoffersen, M.E. NIFES
P.B. 2029
nordes N- 5817
Bergen

PERU

Bedregal, P. Instituto Peruano de Energía Nuclear
División de Técnicas Analíticas Nucleares (TANU)
Av. Canadá 1470
San Borja Lima 41

POLAND

Chajduk, E. Inst. of Nuclear Chemistry and Technology
Department of Analytical Chemistry
Dorodna 16
03-195 Warsaw

PORTUGAL

Dr Eduarda Pereira Central Labotory of Analysis (LCA)
University of Aveiro
3810-193 Aveiro

RUSSIAN FEDERATION

Gladilovich, D.B. LUMEX-MARKETING LLC
P.O. Box 1234
190000 St Petersburg

Glushchenkova, G.F. CJSC "Water Research and Control Center"
Komsomola st. 9-K
195009 St Petersburg

Kochetkov, A. Federal State Budgetary Institution " Research and
Production Association "Typhoon"
Pobedy St 4
Obninsk Kaluga region
249038

Lyapunov, S. Geological Institute of Russian Academy of Sciences
Pyzhevsky Line 7
119017 Moscow

Matveeva, I.S. State Oceanographic Inst.
Kropotkinsky Lane 6
119034 Moscow

Okina, O. Geological Institute of Russian Academy of Sciences
Pyzhevsky Line 7
119017 Moscow

Selivanova, T.Y. CJSC "Water Research and Control Center"
Komsomola st. 9-K
195009 St Petersburg

SLOVENIA

Horvat, M. Josef Stefan Institute
Dept. of Environmental Sciences
Jamova 39
1000 Ljubljana

Jacimovi, R. Josef Stefan Institute
Dept. of Environmental Sciences
Jamova 39
1000 Ljubljana

UNITED KINGDOM

Fisher, A.

School of Geography, Earth and Environmental
Science
University of Plymouth
Drake Circus
Plymouth, Devon PL4 8AA

UNITED STATES OF AMERICA

Mr Frank Mc Farland

Brooks Rand Labs
3958 6th Ave. N.W.
Seattle
Washington 98107

CONTRIBUTORS TO THE DRAFTING AND REVIEW

The following persons, all from the IAEA, contributed to the preparation and editorial correction of this report:

S. Azemard	Laboratory Technician
E. Vasileva-Veleva	Research Scientist
L. Barilaro-Hamonic	Team Assistant



IAEA

International Atomic Energy Agency

No. 25

ORDERING LOCALLY

In the following countries, IAEA priced publications may be purchased from the sources listed below or from major local booksellers.

Orders for unpriced publications should be made directly to the IAEA. The contact details are given at the end of this list.

CANADA

Renouf Publishing Co. Ltd

22-1010 Polytek Street, Ottawa, ON K1J 9J1, CANADA

Telephone: +1 613 745 2665 • Fax: +1 643 745 7660

Email: order@renoufbooks.com • Web site: www.renoufbooks.com

Bernan / Rowman & Littlefield

15200 NBN Way, Blue Ridge Summit, PA 17214, USA

Tel: +1 800 462 6420 • Fax: +1 800 338 4550

Email: orders@rowman.com Web site: www.rowman.com/bernan

CZECH REPUBLIC

Suweco CZ, s.r.o.

Sestupná 153/11, 162 00 Prague 6, CZECH REPUBLIC

Telephone: +420 242 459 205 • Fax: +420 284 821 646

Email: nakup@suweco.cz • Web site: www.suweco.cz

FRANCE

Form-Edit

5 rue Janssen, PO Box 25, 75921 Paris CEDEX, FRANCE

Telephone: +33 1 42 01 49 49 • Fax: +33 1 42 01 90 90

Email: formedit@formedit.fr • Web site: www.form-edit.com

GERMANY

Goethe Buchhandlung Teubig GmbH

Schweitzer Fachinformationen

Willstätterstrasse 15, 40549 Düsseldorf, GERMANY

Telephone: +49 (0) 211 49 874 015 • Fax: +49 (0) 211 49 874 28

Email: kundenbetreuung.goethe@schweitzer-online.de • Web site: www.goethebuch.de

INDIA

Allied Publishers

1st Floor, Dubash House, 15, J.N. Heredi Marg, Ballard Estate, Mumbai 400001, INDIA

Telephone: +91 22 4212 6930/31/69 • Fax: +91 22 2261 7928

Email: alliedpl@vsnl.com • Web site: www.alliedpublishers.com

Bookwell

3/79 Nirankari, Delhi 110009, INDIA

Telephone: +91 11 2760 1283/4536

Email: bkwell@nde.vsnl.net.in • Web site: www.bookwellindia.com

ITALY

Libreria Scientifica "AEIOU"

Via Vincenzo Maria Coronelli 6, 20146 Milan, ITALY
Telephone: +39 02 48 95 45 52 • Fax: +39 02 48 95 45 48
Email: info@libreriaaeiou.eu • Web site: www.libreriaaeiou.eu

JAPAN

Maruzen-Yushodo Co., Ltd

10-10 Yotsuyasakamachi, Shinjuku-ku, Tokyo 160-0002, JAPAN
Telephone: +81 3 4335 9312 • Fax: +81 3 4335 9364
Email: bookimport@maruzen.co.jp • Web site: www.maruzen.co.jp

RUSSIAN FEDERATION

Scientific and Engineering Centre for Nuclear and Radiation Safety

107140, Moscow, Malaya Krasnoselskaya st. 2/8, bld. 5, RUSSIAN FEDERATION
Telephone: +7 499 264 00 03 • Fax: +7 499 264 28 59
Email: secnrs@secnrs.ru • Web site: www.secnrs.ru

UNITED STATES OF AMERICA

Bernan / Rowman & Littlefield

15200 NBN Way, Blue Ridge Summit, PA 17214, USA
Tel: +1 800 462 6420 • Fax: +1 800 338 4550
Email: orders@rowman.com • Web site: www.rowman.com/bernan

Renouf Publishing Co. Ltd

812 Proctor Avenue, Ogdensburg, NY 13669-2205, USA
Telephone: +1 888 551 7470 • Fax: +1 888 551 7471
Email: orders@renoufbooks.com • Web site: www.renoufbooks.com

Orders for both priced and unpriced publications may be addressed directly to:

Marketing and Sales Unit
International Atomic Energy Agency
Vienna International Centre, PO Box 100, 1400 Vienna, Austria
Telephone: +43 1 2600 22529 or 22530 • Fax: +43 1 2600 29302 or +43 1 26007 22529
Email: sales.publications@iaea.org • Web site: www.iaea.org/books

