

IAEA Analytical Quality in Nuclear Applications Series No. 66

Worldwide Interlaboratory Comparison on the Determination of Organochlorine Compounds, Polybrominated Diphenyl Ethers and Polycyclic Aromatic Hydrocarbons in a Fish Sample

ILC-IAEA-MESL-2019-01-OC



IAEA

International Atomic Energy Agency

WORLDWIDE INTERLABORATORY
COMPARISON ON THE DETERMINATION
OF ORGANOCHLORINE COMPOUNDS,
POLYBROMINATED DIPHENYL ETHERS
AND POLYCYCLIC AROMATIC
HYDROCARBONS IN A FISH SAMPLE

The following States are Members of the International Atomic Energy Agency:

| | | |
|------------------------|---------------------------|--------------------------|
| AFGHANISTAN | GEORGIA | OMAN |
| ALBANIA | GERMANY | PAKISTAN |
| ALGERIA | GHANA | PALAU |
| ANGOLA | GREECE | PANAMA |
| ANTIGUA AND BARBUDA | GRENADA | PAPUA NEW GUINEA |
| ARGENTINA | GUATEMALA | PARAGUAY |
| ARMENIA | GUYANA | PERU |
| AUSTRALIA | HAITI | PHILIPPINES |
| AUSTRIA | HOLY SEE | POLAND |
| AZERBAIJAN | HONDURAS | PORTUGAL |
| BAHAMAS | HUNGARY | QATAR |
| BAHRAIN | ICELAND | REPUBLIC OF MOLDOVA |
| BANGLADESH | INDIA | ROMANIA |
| BARBADOS | INDONESIA | RUSSIAN FEDERATION |
| BELARUS | IRAN, ISLAMIC REPUBLIC OF | RWANDA |
| BELGIUM | IRAQ | SAINT LUCIA |
| BELIZE | IRELAND | SAINT VINCENT AND |
| BENIN | ISRAEL | THE GRENADINES |
| BOLIVIA, PLURINATIONAL | ITALY | SAMOA |
| STATE OF | JAMAICA | SAN MARINO |
| BOSNIA AND HERZEGOVINA | JAPAN | SAUDI ARABIA |
| BOTSWANA | JORDAN | SENEGAL |
| BRAZIL | KAZAKHSTAN | SERBIA |
| BRUNEI DARUSSALAM | KENYA | SEYCHELLES |
| BULGARIA | KOREA, REPUBLIC OF | SIERRA LEONE |
| BURKINA FASO | KUWAIT | SINGAPORE |
| BURUNDI | KYRGYZSTAN | SLOVAKIA |
| CAMBODIA | LAO PEOPLE'S DEMOCRATIC | SLOVENIA |
| CAMEROON | REPUBLIC | SOUTH AFRICA |
| CANADA | LATVIA | SPAIN |
| CENTRAL AFRICAN | LEBANON | SRI LANKA |
| REPUBLIC | LESOTHO | SUDAN |
| CHAD | LIBERIA | SWEDEN |
| CHILE | LIBYA | SWITZERLAND |
| CHINA | LIECHTENSTEIN | SYRIAN ARAB REPUBLIC |
| COLOMBIA | LITHUANIA | TAJIKISTAN |
| COMOROS | LUXEMBOURG | THAILAND |
| CONGO | MADAGASCAR | TOGO |
| COSTA RICA | MALAWI | TRINIDAD AND TOBAGO |
| CÔTE D'IVOIRE | MALAYSIA | TUNISIA |
| CROATIA | MALI | TURKEY |
| CUBA | MALTA | TURKMENISTAN |
| CYPRUS | MARSHALL ISLANDS | UGANDA |
| CZECH REPUBLIC | MAURITANIA | UKRAINE |
| DEMOCRATIC REPUBLIC | MAURITIUS | UNITED ARAB EMIRATES |
| OF THE CONGO | MEXICO | UNITED KINGDOM OF |
| DENMARK | MONACO | GREAT BRITAIN AND |
| DJIBOUTI | MONGOLIA | NORTHERN IRELAND |
| DOMINICA | MONTENEGRO | UNITED REPUBLIC |
| DOMINICAN REPUBLIC | MOROCCO | OF TANZANIA |
| ECUADOR | MOZAMBIQUE | UNITED STATES OF AMERICA |
| EGYPT | MYANMAR | URUGUAY |
| EL SALVADOR | NAMIBIA | UZBEKISTAN |
| ERITREA | NEPAL | VANUATU |
| ESTONIA | NETHERLANDS | VENEZUELA, BOLIVARIAN |
| ESWATINI | NEW ZEALAND | REPUBLIC OF |
| ETHIOPIA | NICARAGUA | VIET NAM |
| FIJI | NIGER | YEMEN |
| FINLAND | NIGERIA | ZAMBIA |
| FRANCE | NORTH MACEDONIA | ZIMBABWE |
| GABON | NORWAY | |

The Agency's Statute was approved on 23 October 1956 by the Conference on the Statute of the IAEA held at United Nations Headquarters, New York; it entered into force on 29 July 1957. The Headquarters of the Agency are situated in Vienna. Its principal objective is "to accelerate and enlarge the contribution of atomic energy to peace, health and prosperity throughout the world".

WORLDWIDE INTERLABORATORY
COMPARISON ON THE DETERMINATION
OF ORGANOCHLORINE COMPOUNDS,
POLYBROMINATED DIPHENYL ETHERS
AND POLYCYCLIC AROMATIC
HYDROCARBONS IN A FISH SAMPLE

ILC-IAEA-MESL-2019-01-OC

COPYRIGHT NOTICE

All IAEA scientific and technical publications are protected by the terms of the Universal Copyright Convention as adopted in 1952 (Berne) and as revised in 1972 (Paris). The copyright has since been extended by the World Intellectual Property Organization (Geneva) to include electronic and virtual intellectual property. Permission to use whole or parts of texts contained in IAEA publications in printed or electronic form must be obtained and is usually subject to royalty agreements. Proposals for non-commercial reproductions and translations are welcomed and considered on a case-by-case basis. Enquiries should be addressed to the IAEA Publishing Section at:

Marketing and Sales Unit, Publishing Section
International Atomic Energy Agency
Vienna International Centre
PO Box 100
1400 Vienna, Austria
fax: +43 1 26007 22529
tel.: +43 1 2600 22417
email: sales.publications@iaea.org
www.iaea.org/publications

For further information on this publication, please contact:

IAEA Marine Environmental Laboratories, Monaco
Marine Environmental Studies Laboratory
International Atomic Energy Agency
4a Quai Antoine 1er, 98000
Principality of Monaco

WORLDWIDE INTERLABORATORY COMPARISON ON THE DETERMINATION
OF ORGANOCHLORINE COMPOUNDS, POLYBROMINATED DIPHENYL ETHERS
AND POLYCYCLIC AROMATIC HYDROCARBONS IN A FISH SAMPLE
IAEA, VIENNA, 2022
IAEA/AQ/66
ISSN 2074-7659

© IAEA, 2022
Printed by the IAEA in Austria
March 2022

FOREWORD

The marine environment, as a global sink of persistent chemical mixtures, needs to be regularly monitored to assess the status of pollution and its impact on different marine compartments. Policy makers need access to reliable, accurate and comparable analytical data to enable sound management of the marine environment.

Among the goals of the IAEA Marine Environmental Laboratories is to assist Member States in understanding, monitoring and protecting the marine environment. In this scope, the Marine Environmental Studies Laboratory (MESL) in Monaco acts as an analytical support centre for the analysis of organic contaminants and trace elements in the marine environment. This is achieved by organizing worldwide laboratory performance studies, also known as interlaboratory comparisons (ILCs), by producing certified reference materials and by conducting training courses on the analysis of contaminants in marine samples.

This publication summarizes the results of the ILC-IAEA-MESL-2019-01-OC Fish Sample ILC on the determination of organic contaminants listed as persistent organic pollutants (POPs) in the Stockholm Convention as well as other POPs listed as priority substances, such as polycyclic aromatic hydrocarbons included in many environmental monitoring programmes. The goal of the ILCs is to demonstrate the measurement capabilities of the laboratories participating in the ILC exercises. The results, used to evaluate laboratory performance with respect to a wide range of pollutants, contribute to the reliability and comparability of worldwide analytical data. The ILC was coordinated by the IAEA through the Marine Environmental Studies Laboratory.

The IAEA is grateful to the Government of Monaco for the support provided to its Marine Environmental Laboratories and to all laboratories participating in this ILC exercise. The IAEA officers responsible for this publication were R. Cassi, D. Huertas and I. Tolosa of the Division of IAEA Marine Environmental Laboratories.

EDITORIAL NOTE

This publication has been prepared from the original material as submitted by the contributors and has not been edited by the editorial staff of the IAEA. The views expressed remain the responsibility of the contributors and do not necessarily reflect those of the IAEA or the governments of its Member States.

This publication has not been edited by the editorial staff of the IAEA. It does not address questions of responsibility, legal or otherwise, for acts or omissions on the part of any person.

The use of particular designations of countries or territories does not imply any judgement by the publisher, the IAEA, as to the legal status of such countries or territories, of their authorities and institutions or of the delimitation of their boundaries.

The mention of names of specific companies or products (whether or not indicated as registered) does not imply any intention to infringe proprietary rights, nor should it be construed as an endorsement or recommendation on the part of the IAEA.

The contributors are responsible for having obtained the necessary permission for the IAEA to reproduce, translate or use material from sources already protected by copyrights.

The IAEA has no responsibility for the persistence or accuracy of URLs for external or third party Internet web sites referred to in this publication and does not guarantee that any content on such web sites is, or will remain, accurate or appropriate.

CONTENTS

| | | |
|----|--|-----|
| 1. | INTRODUCTION..... | 1 |
| 2. | SCOPE OF THE INTERLABORATORY COMPARISON..... | 1 |
| 3. | DESCRIPTION OF THE MATERIAL..... | 2 |
| 4. | MOISTURE CONTENT..... | 2 |
| 5. | HOMOGENEITY TEST..... | 2 |
| 6. | SAMPLE DISPATCH AND DATA RETURN..... | 5 |
| 7. | RESULTS EVALUATION..... | 7 |
| | 7.1 MOISTURE AND LIPIDS CONTENT..... | 7 |
| | 7.2 ANALYTICAL METHODS..... | 8 |
| | 7.3 ASSIGNED VALUES AND ASSOCIATED UNCERTAINTIES..... | 11 |
| | 7.3.1 Evaluation criteria..... | 15 |
| 8. | LABORATORIES PERFORMANCES..... | 17 |
| | 8.1 OVERVIEW OF THE PCB RESULTS..... | 17 |
| | 8.2 OVERVIEW OF THE ORGANOCHLORINE PESTICIDES RESULTS..... | 18 |
| | 8.3 OVERVIEW OF THE PBDE RESULTS..... | 19 |
| | 8.4 OVERVIEW OF PAH RESULTS..... | 19 |
| 9. | RECOMMENDATIONS AND CONCLUSIONS..... | 49 |
| | APPENDIX I: RESULTS REPORTED BY PARTICIPANTS..... | 51 |
| | I.1. RESULTS REPORTED FOR PCB CONGENERS..... | 51 |
| | I.2. RESULTS REPORTED FOR ORGANOCHLORINE PESTICIDES..... | 71 |
| | I.3. RESULTS REPORTED FOR PBDES..... | 89 |
| | I.4. RESULTS REPORTED FOR PAHS..... | 95 |
| | APPENDIX II: REPRESENTATION OF LABORATORY PERFORMANCES..... | 109 |
| | II.1. LABORATORIES PERFORMANCES FOR PCB CONGENERS..... | 109 |
| | II.2. LABORATORIES PERFORMANCES FOR ORGANOCHLORINE PESTICIDES..... | 144 |
| | II.3. LABORATORIES PERFORMANCES FOR PBDEs..... | 157 |
| | II.4. LABORATORIES PERFORMANCES FOR PAHs..... | 165 |
| | REFERENCES..... | 175 |
| | LIST OF PARTICIPANTS..... | 177 |
| | CONTRIBUTORS TO DRAFTING AND REVIEW OF THE REPORT..... | 187 |
| | ANNEX: DOCUMENTS SENT TO PARTICIPANTS..... | 189 |

1. INTRODUCTION

One of the major programs of the International Atomic Energy Agency's Environment Laboratories (IAEA-NAEL) is to assist Member States in the use of nuclear and non-nuclear analytical techniques to understand, monitor and protect the environment. The Marine Environmental Studies Laboratory (MESL) of the IAEA-EL has conducted worldwide interlaboratory comparisons (ILCs) on trace organic contaminants for over forty years as part of its programmatic program to assist Member States in maintaining and improving the reliability of their analytical measurements in marine pollution.

The results from previous ILCs on persistent organic contaminants (POPs) [1, 2] have revealed some serious difficulties for many laboratories to obtain comparable data, in particular in difficult matrixes, such as biota samples. The accurate and precise determination of halogenated compounds and polycyclic aromatic hydrocarbons in marine samples is essential for assessing the pollution in coastal and ocean environments. To help improve the quality of analytical data, MESL has organized a new worldwide ILC in a marine fish to follow the progress of the performance of the participant laboratories in the analysis of persistent organic contaminants in seafood samples.

This report describes and evaluates results from the IAEA-MESL-2019-01-OC Interlaboratory comparison study, for the chemical analysis of organochlorine compounds (PCBs and organochlorine pesticides), polybrominated diphenyl ethers (PBDEs) and polycyclic aromatic hydrocarbons (PAHs) in tuna fish sample (IAEA-435).

2. SCOPE OF THE INTERLABORATORY COMPARISON

The scope of this worldwide ILC is to provide a valuable way for analyst participants to anonymously allow them to check the accuracy of their results and identify any analytical gap or problem that might require improvement or corrective action. Furthermore, it can be an opportunity to compare the data reported with their associated analytical procedures.

The target compounds of this ILC includes organic contaminants listed as Persistent Organic Pollutants (POPs) in the Stockholm Convention, as well as other POPs registered as priority substances (PSs), such as polycyclic aromatic hydrocarbons (PAHs) included in many environmental monitoring programmes.

Invitation letters for the present ILC were sent to 326 laboratories from 86 Member States, previously participated or expressed interest in participating in the IAEA ILCs. Positive responses were received from 79 laboratories from 43 Member States.

3. DESCRIPTION OF THE MATERIAL

A sample of 150 kg of Tuna (*Thunus thynnus*) was collected from the Mediterranean Sea in 2004. This sample was freeze-dried, ground, and sieved at 250 μm using a stainless-steel sieve. This powder was homogenized by mixing it in a stainless-steel rotating drum for two weeks. Then, aliquots of about 30 g were packed into glass bottles with aluminum screw caps, labeled IAEA-435 and sealed with Teflon tape. This material was previously used for a world-wide ILC in 2004-2005 and reported values were used for producing the IAEA 435 reference material which provides information values for petroleum hydrocarbons, PCBs and organochlorine pesticides [2].

Before dispatching the ILC samples to participating laboratories, the original labels were removed and replaced by labels reporting only the name of the exercise: (IAEA-MESL-2019-01-OC).

4. MOISTURE CONTENT

The moisture content of the lyophilized material, determined by drying an aliquot to a constant mass at 105 $^{\circ}\text{C}$, was found to be $5.0 \pm 0.7\%$ at the time of the preparation of this sample in 2004 [2] and $7.0 \pm 0.1\%$ at the time of dispatching the bottles kept at 20 $^{\circ}\text{C}$. As the moisture content may change with the ambient humidity and temperature, it was recommended that participants determine the water content of the ILC sample from a separate sub-sample (not the one taken for analysis of contaminants), by drying to a constant weight for at least 8 hours at 105 $^{\circ}\text{C}$ at the time of analysis in the laboratory.

5. HOMOGENEITY TEST

The homogeneity (between and within bottles) of the material was assessed for organochlorine compounds and polybrominated diphenyl ethers by determining the concentration of these compounds in 10 bottles randomly selected from the bottling of the bulk dry powder. Each bottle unit was extracted and analyzed in duplicate, resulting in 2 independent data values by bottle unit. A one-way variance analysis of the results indicated that the material can be

considered homogeneous. All measurements were performed in MESL, IAEA Organic Chemistry Laboratory using previously validated methods. For organochlorine and PBDEs compounds, subsamples of 3 g were extracted with a microwave oven, purified with sulphuric acid, fractionated by Florisil SPE column, and determined by GC-ECD in dual column and GC-MS/MS for confirmation.

Grubbs-tests at 95% and 99% confidence levels were performed to identify potentially outlying individual results as well as outlying bottle means. Few unit means were detected as outliers at 95% confidence levels for PCB 170, PCB 201, BDE154 and BDE153, but they were retained as they followed a normal distribution.

Quantification of between-unit homogeneity was estimated according to ISO Guide 35 [3] by analysis of variance (ANOVA) which can separate the between-unit variation (s_{bb}) from the within-unit variation (s_{wb}). The latter is equivalent to the method repeatability if the individual aliquots are representative for the whole unit. ANOVA allows the calculation of within-unit standard deviation s_{wb} and also between-units standard deviation s_{bb} :

$$s_{wb} = u_{wb} = \sqrt{MS_{wb}} \quad (1)$$

$$s_{bb} = u_{bb} = \sqrt{\frac{MS_{bb} - MS_{wb}}{n}} \quad (2)$$

s_{bb} and s_{wb} are estimates of the true standard deviations and are therefore subject to random fluctuations. In some cases, the mean square between groups (MS_{bb}) can be smaller than the mean squares within groups (MS_{wb}), resulting in negative arguments under the square root used for the estimation of the between-unit variation. In this case, u_{bb}^* , the maximum between-unit variability that could be masked by method repeatability, was calculated as described by Linsinger et al. [4]:

$$u_{bb}^* = \frac{s_{wb}}{\sqrt{n}} \sqrt[4]{\frac{2}{v_{MS_{wb}}}} \quad (3)$$

Where: n is the number of replicate sub-samples per bottle; and $v_{MS_{wb}}$ is the degrees of freedom of MS_{wb} .

As presented in Table 1, the between-unit variations (s_{bb} and u_{bb}^*) for most of the selected compounds were between 1.2 and 4.1%, small enough to ensure the homogeneity of the material at 3 g sample size. The within-unit variation (s_{wb}) derived from the ANOVA calculation range between and 3.3 and 8.7%, which was similar to the typical method

repeatability (s_{meas} : 3.9-8.9%) derived from relative standard deviation of six replicates from the same bottle in the same batch.

The uncertainty u_{hom} associated with inhomogeneity of the material was estimated according to the ISO Guide 35 [3] by Eq. 4:

$$u_{hom} = \sqrt{u_{wb}^2 + u_{bb}^2} \quad (4)$$

where the u_{wb} was derived from ANOVA and the u_{bb} was taken as the maximum values of the between-unit variations (s_{bb} and u_{bb}^*). As it is presented in Table 1, the uncertainty contribution related to inhomogeneity was estimated to range from 3.9 to 9.0 %, and, thus, we set the uncertainty associated with inhomogeneity at 8 % for all analytes.

TABLE THE ESTIMATE OF INHOMOGENEITY CONTRIBUTIONS TO THE TOTAL UNCERTAINTY FOR THE SELECTED PCBs, ORGANOCHLORINE PESTICIDES AND PBDEs COMPOUNDS AND REPEATABILITY OF THE METHOD

| Compound | S_{wb} | S_{bb} | u_{bb}^* | Repeatability | u_{hom} |
|----------|----------|---------------|------------|---------------|-----------|
| | % | % | % | % | % |
| PCB 28 | 4.7 | ¹⁾ | 2.3 | 4.7 | 5.3 |
| PCB 31 | 6.1 | ¹⁾ | 3.0 | 5.7 | 6.8 |
| PCB 52 | 3.3 | 2.1 | 1.6 | 3.9 | 3.9 |
| PCB 99 | 4.9 | 2.5 | 2.4 | 5.4 | 5.5 |
| PCB 101 | 4.0 | 1.2 | 1.9 | 4.1 | 4.2 |
| PCB 118 | 5.9 | 4.3 | 2.9 | 7.2 | 7.3 |
| PCB 126 | 4.9 | 1.3 | 2.4 | 4.9 | 5.1 |
| PCB 138 | 6.9 | ¹⁾ | 3.4 | 6.6 | 7.7 |
| PCB 149 | 5.0 | ¹⁾ | 2.4 | 4.8 | 5.5 |
| PCB 151 | 6.9 | ¹⁾ | 3.4 | 6.5 | 7.7 |
| PCB 153 | 6.6 | ¹⁾ | 3.2 | 6.3 | 7.3 |
| PCB 169 | 5.7 | 3.1 | 2.8 | 6.3 | 6.5 |
| PCB 170 | 5.0 | 3.7 | 2.4 | 6.0 | 6.2 |
| PCB 174 | 3.8 | 1.7 | 1.9 | 4.1 | 4.2 |
| PCB 180 | 5.0 | 2.8 | 2.4 | 5.5 | 5.7 |
| PCB 183 | 4.6 | ¹⁾ | 2.3 | 4.4 | 5.2 |
| PCB 187 | 4.5 | ¹⁾ | 2.2 | 4.1 | 5.0 |
| PCB 194 | 5.7 | 1.9 | 2.8 | 6.2 | 6.0 |
| PCB 195 | 6.9 | ¹⁾ | 3.3 | 6.9 | 7.6 |
| PCB 201 | 5.3 | 3.8 | 2.6 | 6.3 | 6.5 |
| PCB 206 | 4.2 | 2.0 | 2.0 | 6.4 | 4.6 |
| PCB 209 | 6.1 | ¹⁾ | 3.0 | 5.3 | 6.8 |
| pp'DDT | 6.6 | 5.5 | 3.2 | 6.0 | 8.6 |
| BDE 154 | 7.8 | 3.2 | 3.7 | 8.4 | 8.4 |
| BDE 153 | 8.7 | 2.2 | 4.1 | 8.9 | 9.0 |

¹⁾Not defined due to negative argument under the square root

6. SAMPLE DISPATCH AND DATA RETURN

The ILC sample labelled “ILC-IAEA-MESL-2019-01-OC, Organic Contaminants in Fish Sample’ was distributed to 79 laboratories worldwide between November and December 2019.

Due to the unfortunate pandemic situation, only a total of 51 laboratories from 28 countries reported results.

Each participating laboratory received one bottle of the test sample, accompanied by one information sheet and instruction for the use of the IAEA-NAEL on-line results reporting system. Participants were requested to determine PCBs, organochlorine pesticides, PBDEs and PAHs as many compounds as possible from the list reported in Annex, using the analytical procedures established in their laboratories.

Participants were requested to make at least three independent determinations and to report the results together with a short description of the method used. Participants were also required to provide proof of traceability of obtained results to the International System of Units (SI), via standard calibration solutions and CRM applied, as a part of their analytical procedures.

It was also requested that concentrations must be calculated on a dry mass basis and given as ng g^{-1} , leaving as many significant figures as justified by the precision of the method used.

Participants were advised that if a compound was not detected by the method used, the corresponding limit of detection should have been given rather than the statement “not detected”.

For each class of compounds, participants were required also to report information about:

- Water content (in %);
- Total lipids (mg/g);
- Extractable lipids (mg/g);
- Extraction technique;
- Quantity extracted;
- Solvent used;
- Samples clean up;
- Sample fractionation;
- Detection method;
- Injector;
- GC or HPLC column;
- Results quantification;
- Statement on traceability of the obtained measurement results (standards and reference materials used);
- QA/QC;
- Surrogates spiked before extraction;

- Internal standards spiked before injection;
- Use of CRMs;
- Use of validated method;
- Accreditation.

Participants were requested to report their data and information using an on-line reporting. The initial deadline for reporting data was set on the 3rd of March 2020. However, due to the pandemic spread of Covid-19 disease and subsequent multiple lockdowns, it was agreed with participants to postpone the deadline to the 15th of December 2020.

7. RESULTS EVALUATION

Results reported by participants are presented in Appendix I (Tables 19 to 107).

For the purpose of this report, each laboratory is represented by a code number which remains unchanged throughout the data tables. The numbers do not correspond to the sequence of laboratories in the list of participants given at the end of the report, so that the anonymity of each laboratory is preserved.

Along with the computed arithmetic mean of the replicate determinations reported by participants, Tables 19 to 107 report also fundamental analytical and quality parameters like the instrument type, the use of surrogates, internal standards and Reference Materials and the existence of a quality assurance/quality control system in the laboratory.

7.1 MOISTURE AND LIPIDS CONTENT

Water content (in %) as well as Extractable lipids (mg/g) and/or Total lipids (mg/g) were parameters that participants had to report for each group of analytes. Unfortunately, many participants provided these type of information only partially. Moreover, some participants misinterpreted the indications about reporting units and some of the data provided were incoherent. As an indication of water content and extractable lipids it was decided to use data issued from the analysis performed in MESL, IAEA Organic Chemistry Laboratory. Water content was found to be equal to $7.0 \pm 0.1\%$ while extractable lipids were found to be equal to 136 ± 10 mg/g.

7.2 ANALYTICAL METHODS

Figure 1 shows a graphic representation of the techniques used by participants to extract the test sample for each group of compounds: Organochlorine Compounds, PBDEs and PAHs. Among the different techniques Soxhlet, Accelerated Solvent Extraction (ASE) and Microwave assisted extraction (MW) are by far the most used in every group of compounds.

Figure 2 exhibits a graphic representation of the type of instrument used by participants to analyze the test sample for each group of compounds: Organochlorine Compounds, PBDEs and PAHs. As expected, a clear predominance of gas chromatographic-mass spectrometry techniques were used for all different groups of compounds.

Table 2 reports abbreviations used in Figure 1 and 2 and along this report.

TABLE 2. ABBREVIATIONS OF TECHNIQUES AND INSTRUMENTS USED BY PARTICIPANTS FOR THE EXTRACTION AND ANALYSIS OF THE TEST SAMPLE

| Abbreviation | |
|---------------------|---|
| ASE | Accelerated Solvent Extraction |
| QUECHERS | Quick, Easy, Cheap, Effective, Rugged and Safe Solid Phase Extraction |
| MW | Microwave Assisted Extraction |
| Shaking (s/l) | Mechanical Shaking Solid/Liquid Extraction |
| GC/ECD | Gas Chromatography – Electron Capture Detector |
| GC/ECD & GC/MS | Gas Chromatography – ECD and peak confirmation with GC/MS |
| GC/ECD - Dual | Gas Chromatography – ECD with dual columns confirmation |
| GC/MS | Gas chromatography–mass spectrometry |
| GC/MS EI | Gas chromatography–mass spectrometry, Electron Ionization |
| GC/MS NCI | Gas chromatography–mass spectrometry, Negative Chemical Ionization |
| GC/MSMS | Gas Chromatography - Tandem Mass Spectrometry |
| GC/HRMS | Gas Chromatography - High Resolution Mass Spectrometry |
| HPLC-Fluo | High Performance Liquid Chromatography - Fluorescence Detector |
| HPLC-UV | High Performance Liquid Chromatography – Ultraviolet Detector |




FIG. 1. Graphic representation of the techniques used by participants to extract the test sample




FIG. 2. Graphic representation of the type of instrument used by participants to analyze the test sample

7.3 ASSIGNED VALUES AND ASSOCIATED UNCERTAINTIES

The assigned values for organic contaminants in the ILC-IAEA-MESL-2019-01-OC sample were established according to the requirements of the ISO 17043 standard [5] using the statistical robust methods described in ISO 13528 [6] for use in proficiency testing by interlaboratory comparisons. Kernel density plot was also used as an appropriate tool to visualize and identify modes in the distribution of the results [7].

The first estimation of the robust average values was performed considering all reported values. Then, reported results that diverged more than $\pm 50\%$ from the robust average were considered outliers and discarded, and new robust average was calculated with the remaining results for getting the final assigned values and associated uncertainties. No assigned values were calculated for compounds where most participating laboratories using mass spectrometry provided values lower than their detection limit.

The uncertainties associated with the new assigned values of the ILC-IAEA-MESL-2019-01-OC sample were conducted according to ISO Guide 35 [3], combining the standard uncertainties associated with the characterization (u_{char}), homogeneity (u_{hom}) and long-term stability (u_{stab}). Because the uncertainty component derived from the long-term stability was insignificant and assumed to be zero, the final expanded uncertainty was a combination of the other two contributions using the law of propagation of uncertainty as shown:

$$U_a = k \times \sqrt{u_{char}^2 + u_{ho}^2} \quad (5)$$

where k is the coverage factor of 2, representing a confidence level of 95%

u_{hom} was set at 8%

u_{char} was calculated as described in ISO 13528 [6] using:

$$u_{char} = 1.25 \times \frac{s^*}{\sqrt{p}} \quad (6)$$

where: s^* is the robust standard deviation and p is the number of participating laboratories.

The new assigned values and their associated uncertainties for the target PCBs, organochlorinated pesticides, PBDE congeners and PAHs are shown in Tables 3, 4, 5 and 6 respectively. For data with number of numerical observations ≤ 3 , the robust mean is given as an indicative value and no evaluation was performed. For compounds where most participating laboratories using mass spectrometry provided values lower than their detection limit, an indicative value as “<” than the assigned value derived from reference laboratories was attributed.

TABLE 3. ASSIGNED VALUES AND ASSOCIATED UNCERTAINTIES FOR PCBs IN THE ILC-IAEA-MESL-2019-01-OC

| Compound | No. Participants | No. Results accepted | Robust mean \bar{x}^* $\mu\text{g kg}^{-1}$ | u_{char} (k=1) (%) | $U_{\text{a total}}$ (k=2) (%) |
|----------|------------------|----------------------|--|-----------------------------------|--------------------------------------|
| PCB 8 | 4 | 3 | <0.30* | | |
| PCB 18 | 9 | 5 | 0.71 | 17.5 | 39 |
| PCB 28 | 33 | 17 | 1.31 | 11.6 | 28 |
| PCB 31 | 14 | 10 | 0.62 | 11.2 | 27 |
| PCB 44 | 9 | 8 | 1.75 | 12.8 | 30 |
| PCB 49 | 4 | 4 | 2.15 | 19.4 | 42 |
| PCB 52 | 34 | 24 | 3.90 | 7.7 | 22 |
| PCB 66 | 4 | 3 | 7.12 | 26.5 | 55 |
| PCB 70 | 3 | 2 | 5.12 | 16.0 | 36 |
| PCB 74 | 3 | 3 | 3.61 | 18.4 | 40 |
| PCB 87 | 4 | 4 | 6.54 | 9.1 | 24 |
| PCB 95 | 10 | 5 | 5.23 | 9.4 | 25 |
| PCB 97 | 2 | 2 | 5.34 | 10.3 | 26 |
| PCB 99 | 11 | 8 | 17.4 | 17.5 | 38 |
| PCB 101 | 38 | 28 | 23.3 | 6.9 | 21 |
| PCB 105 | 16 | 13 | 8.00 | 6.4 | 21 |
| PCB 110 | 11 | 9 | 12.4 | 13.7 | 32 |
| PCB 118 | 29 | 20 | 28.4 | 5.4 | 19 |
| PCB 128 | 10 | 5 | 8.43 | 8.0 | 23 |
| PCB 138 | 36 | 24 | 68.1 | 5.4 | 19 |
| PCB 146 | 7 | 4 | 18.2 | 3.7 | 18 |
| PCB 149 | 18 | 11 | 27.3 | 8.1 | 23 |
| PCB 151 | 10 | 6 | 9.29 | 11.1 | 27 |
| PCB 153 | 35 | 24 | 90.0 | 5.4 | 19 |
| PCB 156 | 15 | 13 | 3.85 | 4.4 | 18 |
| PCB 170 | 14 | 9 | 11.9 | 6.7 | 21 |
| PCB 174 | 2 | 2 | 7.67 | 5.6 | 20 |
| PCB 177 | 10 | 7 | 8.16 | 12.1 | 29 |
| PCB 180 | 36 | 24 | 34.8 | 5.6 | 20 |
| PCB 183 | 11 | 7 | 9.51 | 16.3 | 36 |
| PCB 187 | 12 | 6 | 29.5 | 10.9 | 27 |
| PCB 194 | 12 | 10 | 4.30 | 10.5 | 26 |
| PCB 195 | 4 | 4 | 1.62 | 17.8 | 39 |
| PCB 201 | 3 | 2 | 3.58 | 17.5 | 38 |
| PCB 206 | 5 | 3 | 3.31 | 4.4 | 18 |
| PCB 209 | 9 | 5 | 1.94 | 5.1 | 19 |

*For compounds “<”, no assigned value was calculated as most participating laboratories using mass spectrometry provided values lower than their detection limit.

TABLE 4. ASSIGNED VALUES AND ASSOCIATED UNCERTAINTIES FOR THE ORGANOCHLORINE PESTICIDE COMPOUNDS IN THE ILC-IAEA-MESL-2019-01-OC

| Compound | No. Participants | No. Results accepted | Robust mean \bar{x}^* $\mu\text{g kg}^{-1}$ | u_{char} (k=1) (%) | $U_{\text{a total}}$ (k=2) (%) |
|-------------------------|------------------|----------------------|--|-----------------------------------|--------------------------------------|
| HCB | 27 | 17 | 2.89 | 7.1 | 21 |
| α -HCH | 22 | 9 | <0.15* | | |
| β -HCH | 22 | 8 | <0.25* | | |
| γ -HCH- Lindane | 27 | 11 | <0.15* | | |
| δ -HCH | 16 | 6 | <0.15* | | |
| <i>pp'</i> -DDD | 29 | 14 | 13.1 | 5.2 | 19 |
| <i>pp'</i> -DDE | 34 | 25 | 121 | 7.0 | 21 |
| <i>pp'</i> -DDT | 29 | 20 | 18.2 | 7.0 | 21 |
| <i>op'</i> -DDE | 17 | 12 | 2.23 | 6.0 | 20 |
| <i>op'</i> -DDD | 16 | 11 | 2.95 | 9.2 | 24 |
| <i>op'</i> -DDT | 18 | 14 | 7.48 | 7.7 | 22 |
| Heptachlor | 24 | 12 | <0.16* | | |
| Aldrin | 24 | 10 | <0.10* | | |
| Dieldrin | 24 | 11 | 5.28 | 13.0 | 31 |
| Endrin | 19 | 7 | <0.15* | | |
| <i>cis</i> -Chlordane | 15 | 9 | 6.53 | 11.7 | 28 |
| <i>trans</i> -Chlordane | 16 | 4 | 0.38 | 6.8 | 21 |
| <i>cis</i> -Nonachlor | 4 | 4 | 11.3 | 0.5 | 16 |
| <i>trans</i> -Nonachlor | 5 | 5 | 20.5 | 9.0 | 24 |
| α -Endosulfan | 20 | 6 | <0.30* | | |
| β -Endosulfan | 16 | 4 | <0.10* | | |
| Endosulfan sulfate | 10 | 4 | <0.30* | | |
| Heptachlor epoxide | 22 | 7 | 0.81 | 13.4 | 31 |

*For compounds "< ", no assigned value was calculated as most participating laboratories using mass spectrometry provided values lower than their detection limit.

TABLE 5. ASSIGNED VALUES AND ASSOCIATED UNCERTAINTIES FOR PBDEs IN THE ILC-IAEA-MESL-2019-01-OC

| Compound | No. Participants | No. Results accepted | Robust mean x^* $\mu\text{g kg}^{-1}$ | u_{char} (k=1) (%) | U_a total (k=2) (%) |
|----------|------------------|----------------------|--|-----------------------------------|-----------------------------|
| BDE 17 | 4 | 4 | <0.10 | | |
| BDE 28 | 18 | 14 | 1.21 | 8.3 | 23 |
| BDE 47 | 18 | 16 | 22.4 | 4.6 | 18 |
| BDE 49 | 5 | 4 | 2.07 | 7.9 | 22 |
| BDE 66 | 7 | 7 | 2.55 | 11.3 | 28 |
| BDE 85 | 6 | 6 | <0.10* | | |
| BDE 99 | 18 | 12 | 1.40 | 6.4 | 21 |
| BDE 100 | 18 | 14 | 7.30 | 9.0 | 24 |
| BDE 153 | 17 | 12 | 0.61 | 7.8 | 22 |
| BDE 154 | 18 | 9 | 4.68 | 8.8 | 24 |
| BDE 183 | 8 | 6 | <0.10* | | |
| BDE 209 | 6 | 4 | <1.00* | | |

*For compounds “<”, no assigned value was calculated as most participating laboratories using mass spectrometry provided values lower than their detection limit.

TABLE 6. ASSIGNED VALUES AND ASSOCIATED UNCERTAINTIES FOR PAHs IN THE ILC-IAEA-MESL-2019-01-OC

| Compound | No. Participants | No. Results accepted | Robust mean x^* $\mu\text{g kg}^{-1}$ | u_{char} (k=1) (%) | U_a total (k=2) (%) |
|---------------------------------|------------------|----------------------|--|-----------------------------------|-----------------------------|
| Naphthalene | 20 | 8 | 18.6 | 13.5 | 31 |
| 1-Methylnaphthalene | 7 | 7 | 14.5 | 9.6 | 25 |
| 2-Methylnaphthalene | 8 | 6 | 18.4 | 15.5 | 35 |
| Acenaphthylene | 18 | 5 | 0.94 | 7.5 | 22 |
| Acenaphthene | 21 | 7 | 1.62 | 17.6 | 39 |
| Fluorene | 22 | 9 | 4.04 | 10.9 | 27 |
| Phenanthrene | 23 | 12 | 11.8 | 11.7 | 28 |
| Anthracene | 21 | 9 | <0.7* | | |
| 1-Methylphenanthrene | 3 | 2 | 2.73 | 22.1 | 47 |
| 2-Methylphenanthrene | 2 | 2 | 4.97 | 26.5 | 55 |
| Benz[<i>a</i>]anthracene | 23 | 12 | <0.7* | | |
| Chrysene (+Triphenylene) | 22 | 10 | <0.6* | | |
| Fluoranthene | 22 | 10 | <2* | | |
| Pyrene | 21 | 9 | <2* | | |
| Benzo[<i>a</i>]pyrene | 23 | 12 | <0.6* | | |
| Benzo[<i>e</i>]pyrene | 9 | 6 | <0.6* | | |
| Benzo[<i>b+j</i>]fluoranthene | 22 | 11 | <0.7* | | |
| Benzo[<i>g,h,i</i>]perylene | 22 | 11 | <0.6* | | |

*For compounds “<”, no assigned value was calculated as most participating laboratories using mass spectrometry provided values lower than their detection limit.

7.3.1 Evaluation criteria

The performance of each participant was evaluated with the z -score which expresses the difference between the mean of the laboratory and the assigned value in the same unit. The z -score represents a simple method of giving each participant a normalized performance score for the measurement bias of the respective measurement result. The standard deviation for the proficiency assessment (also called target standard deviation), σ_p , was set to be fit for purpose, according to the Harmonized Protocol [8] and was fixed to 12.5 % of the assigned values. The determination of the target standard deviation was done on the basis of the outcome of previous ILCs organized by MESL, for the same group of laboratories. Considering that the uncertainty (μ_a , $k=1$) in the assigned value x_a is not negligible ($\mu_a \geq 0.3 \sigma_p$) [6], we combined the target standard deviation for proficiency assessment (σ_p), usually set at 12.5% with the target uncertainty of the assigned value (μ_a) for the calculation of the “Total error” according to the following formula:

$$Total\ error_a = \sqrt{u_a^2 + \sigma_p^2} \quad (7)$$

For the assessment of the laboratory performances based on ISO/IEC 17043:2010 [5] and ISO 13528 [6]:

a modified z_i -score is calculated taking in account the total error:

$$Z_i = \frac{x_i - x^*}{total\ error_a} \quad (8)$$

A Zeta-score, taking in account the measurement uncertainties reported by the respective participant are calculated following the Eq. (9),

$$Zeta = \frac{x_i - x^*}{\sqrt{u_a^2 + u_{x_{lab}}^2}} \quad (9)$$

where:

z_i is the z -score of laboratory i for the respective analyte.

x_i is the reported result of laboratory i for that analyte, expressed as the mean of multiple determinations.

x^* is the assigned value for the respective analyte.

$total\ error$ is the target standard deviation for proficiency assessment (σ_p) combined with the uncertainty of the assigned value (μ_a).

u_{lab} is the standard uncertainty ($k=1$) reported by participant

The acceptability of a laboratory's performance was evaluated according to the following internationally accepted limits [8]:

$$\begin{aligned} |z \text{ or Zeta}| &\leq 2.0 \text{ satisfactory} \\ 2.0 < |z \text{ or Zeta}| &< 3.0 \text{ questionable} \\ |z \text{ or Zeta}| &\geq 3.0 \text{ unsatisfactory} \end{aligned}$$

For data with the number of numerical observations ≤ 3 , the robust mean was given as an indicative value and no evaluation was performed. Reported values as "<LD" from participants were evaluated as "Consistent (C)" or "Inconsistent (I)" to provide some extra information about their LD and their application to the exercise:

For those compounds with an assigned value, if a laboratory reported a "<LD" result, it has been evaluated by calculating a range based on the following criteria: Assigned Value $- 2 \times$ Total Error and Assigned Value $+ 6 \times$ Total Error. If the absolute value of their result was inside this range, it has been evaluated as "Consistent (C)", if not then as "Inconsistent (I)".

When the assigned value is "<LD", the reported data of the laboratory was assigned "Consistent (C)" if the reported value was $\leq 6 \times$ LD. Otherwise it was considered "Inconsistent (I)".

8. LABORATORIES PERFORMANCES

In total, 51 laboratories provided results for the analysis of at least one group of compounds by the final deadline. Among the data sets, 39 were submitted for PCBs, 37 for organochlorine pesticides, 18 for PBDEs and 27 for PAHs. The z-scores were calculated for 30 PCB congeners, 13 organochlorine pesticides, 8 PBDEs and 7 PAH compounds.

All reported data are compiled in Appendix I. For editorial purposes some results have been rounded to appropriate number of significant figures.

Graphical presentations of results obtained by participating laboratories; z scores and summary of the statistical evaluation as well as Kernel density plot [7] (when more than 8 results) are presented in Appendix II.

8.1 OVERVIEW OF THE PCB RESULTS

In total, 39 datasets were submitted for PCB congeners with number of analytes reported by participants ranging from 1 to 36. Not surprisingly the seven most abundant PCB congeners (PCB 28, 52, 101, 118, 153, 138, and 180) were the most reported.

Assigned values were calculated for 30 out of the 36 PCB congeners that participants could report. For PCBs 70, 74, 97, 174 and 201, only indicative values were provided as the number of reported values were ≤ 3 . For PCB 8 congener no assigned value was calculated since most participants reported values “<” than their detection limits.

Figure 3 illustrates the overall performances of the laboratories, as a z-scores, for each individual PCB congener. Among the 453 z-scores calculated for PCB congeners, 59% were found to be acceptable, 13% questionable and 28% were unsatisfactory.

Most participants reported uncertainties along with their measurements of the test sample and as shown in Figure 4, the corresponding Zeta-scores distribution followed closely the one for z-scores. Among the 318 Zeta-scores calculated for PCB congeners, 61% were found to be acceptable, 12% questionable and 27% were unsatisfactory.

Figures 5 and 6 shows the laboratories performances (z- and Zeta- scores, respectively) for the analyses of PCB congeners by laboratory code. Tables 7 and 8 report the overall calculated z-scores for PCB congeners. Tables 9 and 10 reports the overall calculated Zeta-scores for PCB congeners.

A large majority of participants extracted the test sample using either Soxhlet or Microwave Assisted Extraction while the most employed instrumental techniques were GC/MS EI and GC/MSMS.

The majority of participants used Certified Reference Materials, spiked surrogates before extraction and internal standards before injection of the test sample. Although most participants used validated methods, only a minority were accredited for the analysis of PCB congeners.

Overall, the comparability between the participating laboratories was relatively satisfactory for the major PCB congeners, whereas higher variability was observed for minor PCB congeners that may co-elute with other isomers on the commonly 5% phenyl-substituted methyl polysiloxane GC stationary phase used by the participants. In this aspect, the use of GC columns with minimum length of 50 m or at least two columns with stationary phases of different polarity is recommended.

8.2 OVERVIEW OF THE ORGANOCHLORINE PESTICIDES RESULTS

In total, 37 datasets were submitted for organochlorine pesticides with number of analytes reported by participants ranging from 2 to 23.

Assigned values were calculated for 13 out of the 23 organochlorine pesticides that participants could report. For ten analytes, no assigned values were calculated since most participants reported values “<” than their detection limits.

Figures 7 and 8 represent a graphical representation of laboratories performances (*z*- and Zeta-scores, respectively) by compound for organochlorine pesticides. Among the 219 *z*-scores calculated for organochlorine pesticides, 59% were found to be acceptable, 13% questionable and 28% were unsatisfactory. The Zeta-scores distribution for the participants that reported uncertainties along with their measurements of the test sample followed closely the one for *z*-scores. Among the 145 Zeta-scores calculated for organochlorine pesticides, 63% were found to be acceptable, 11% questionable and 26% were unsatisfactory.

Figures 9 and 10 represent a graphical representation of laboratories performances (*z*- and Zeta-scores, respectively) for the analyses of organochlorine pesticides by laboratory.

Since most of the time the analysis of PCB congeners and organochlorine pesticides are carried out simultaneously, similarly to what observed for PCBs, a large majority of participants extracted the test sample using either Soxhlet or Microwave Assisted Extraction while the most employed instrumental techniques were GC/MS EI and GC/MSMS.

The majority of participants used Certified Reference Materials, surrogate standards before extracting the sample and internal standards before the instrumental analysis. Although most participants used validated methods, only a minority were accredited for the analysis of organochlorine pesticides.

Tables 11 and 12 report the overall calculated *z*-scores for organochlorine pesticides. Table 13 reports the overall calculated Zeta-scores for organochlorine pesticides.

8.3 OVERVIEW OF THE PBDE RESULTS

In total, 18 datasets were submitted for PBDEs with number of analytes reported by participants ranging from 6 to 12. Assigned values were calculated for 8 out of the 12 BDEs that participants could report. For 4 BDE analytes no assigned values were calculated since most participants reported values “<” than their detection limits.

Figures 11 and 12 show the overall laboratory performances (*z*- and Zeta- scores, respectively) for each individual BDE congener. Among the 117 *z*-scores calculated for PBDEs, 62% were found to be acceptable, 15% questionable and 22% were unsatisfactory. Most of the participants reported uncertainties along with their measurements of the test sample and the corresponding Zeta-scores distribution followed closely the one for *z*-scores. Among the 86 Zeta-scores calculated for PBDEs, 64% were found to be acceptable, 10% questionable and 26% were unsatisfactory.

Figures 13 and 14 illustrates the laboratory performances (*z*- and Zeta- scores, respectively) for the analyses of PBDEs by laboratory code. Table 14 reports the overall calculated *z*-scores for PBDEs. Table 15 reports the overall calculated Zeta-scores for PBDEs

A large majority of participants extracted the test sample using either accelerated solvent extraction or Soxhlet while the most employed instrumental techniques were GC/MSMS and GC/HRMS. Also many of the participants used Certified Reference Materials, surrogates before the extraction of the sample, and internal standards before the instrumental analysis. Although most participants used validated methods, only one third were accredited for the analysis of PBDEs.

.

8.4 OVERVIEW OF PAH RESULTS

For PAHs, 27 datasets covering between 1 and 19 analytes were submitted by the participants. Assigned values were calculated for 7 out of the 25 PAH analytes than participants could report. Indicative values were provided for 1-methylphenantrene and 2-methylphenantrene as the number of reported values were ≤ 3 . For 16 PAH analytes, no assigned values were calculated since most participants reported values “<” than their detection limits.

Figures 15 and 16 shows the laboratory performances (*z*- and Zeta- scores, respectively) for each individual PAH compound. Among the 101 *z*-scores calculated for PAHs, 42% were found to be acceptable, 16% questionable and 43% were unsatisfactory. Most participants reported uncertainties along with their measurements of the test sample and the corresponding Zeta-scores distribution followed closely the one for *z*-scores. Among the 64 Zeta-scores calculated for PAHs, 55% were found to be acceptable, 11% questionable and 34% unsatisfactory.

Figures 17 and 18 represent a graphical representation of laboratories performances (*z*- and Zeta- scores, respectively) for the analyses of PAHs by laboratory. Table 16 and 17 reports the

overall calculated z-scores for PAHs. Table 18 reports the overall calculated Zeta-scores for PAHs. The values reported for each individual PAH compound were very variable, expected due to the relatively low concentrations of these chemicals in the whole tissue and the high lipid content of the fish material.

Most of the participants extracted the test sample using either Accelerated Solvent Extraction or Soxhlet while the most employed instrumental techniques were GC/MS EI and GC/MSMS. Less than half of participants used Certified Reference Materials, surrogate standards before extracting the sample and internal standards before the instrumental analysis. Similarly, less than half of the participants used validated methods, and only one third were accredited for the analysis of PAHs. Overall, the comparability between the participating laboratories was poor, probably due to the combination of high lipid content, low analyte concentrations and the variations in the analytical procedures used by the participants.




FIG. 3. Percentage distribution of the z-scores of results reported by the participants per PCB congener.




FIG. 4. Percentage distribution of the Zeta-scores of results reported by participants per PCB congener




FIG. 5. Percentage distribution of the z-scores of PCBs results reported by participants per laboratory.




FIG. 6. Percentage distribution of the Zeta-scores of PCBs results reported by participants per laboratory.

TABLE 7. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCES (z-SCORE) FOR PCBs (8 to 118).

| Lab Code | PCB 8 | PCB 18 | PCB 28 | PCB 31 | PCB 44 | PCB 49 | PCB 52 | PCB 66 | PCB 87 | PCB 95 | PCB 99 | PCB 101 | PCB 105 | PCB 110 | PCB 118 |
|----------|-------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|---------|---------|---------|---------|
| 1 | | | | | | | -4.21 | | | | | | | | |
| 2 | | | C | | | | I | | | | 1.33 | | | | -0.21 |
| 3 | | -1.19 | | | | | -1.45 | | | | | -0.85 | -2.71 | | -0.77 |
| 4 | | -1.64 | | | | | | | | | | -5.64 | | | -6.10 |
| 5 | | -1.52 | | | | | 0.46 | | | | | 0.13 | | | -0.43 |
| 7 | | 1.90 | | | | | 0.80 | | | | | 0.47 | | | -1.73 |
| 8 | | 16.5 | | | | | 97.7 | | | | | 6.48 | | | |
| 10 | | 1.81 | | | | | 0.28 | | | 4.08 | 0.79 | 2.09 | 0.70 | 1.18 | 1.31 |
| 14 | | | | | | | -1.96 | | | -2.01 | -2.10 | -3.05 | | -1.21 | |
| 15 | | -0.42 | | | | | 1.13 | | | | | 1.51 | -0.95 | | -0.44 |
| 17 | 0.80 | -3.40 | | -1.94 | -0.58 | | 0.85 | | | | | 2.15 | | | |
| 23 | I | 51.4 | | | | | 104 | | | | | 4.14 | | | 29.2 |
| 24 | | -2.18 | | | | | -0.97 | | | | | -0.12 | | | |
| 25 | | -2.55 | | -1.12 | | | -1.29 | | | -0.60 | -1.67 | -2.23 | -0.27 | -1.20 | -1.33 |
| 26 | | I | | | | | I | | | | | I | | | I |
| 27 | | 4.35 | | -1.35 | | | -1.38 | | | | | -1.13 | 0.14 | | 0.32 |
| 35 | | 2.28 | | | | | 0.70 | | | | | 1.44 | | | 3.20 |
| 36 | | -0.08 | 0.40 | 2.51 | -2.35 | | -2.72 | | | -0.82 | -2.03 | -2.29 | -2.46 | -1.78 | -5.05 |
| 40 | C | -2.14 | -1.25 | 0.58 | -1.05 | 0.08 | 0.08 | -0.20 | 0.16 | 2.38 | 0.53 | 1.51 | 1.76 | 3.17 | 1.77 |
| 45 | | -2.97 | | | | | 19.1 | | | | | -5.96 | | | -4.42 |
| 47 | | 15.7 | | | | | | | | | | -2.34 | | | -2.30 |
| 48 | 1.32 | -2.68 | -0.41 | -0.50 | | | 2.63 | | | | | -0.18 | | | 1.24 |
| 49 | | | | | | | 0.15 | | | | | -1.32 | | | -0.46 |
| 50 | | -2.59 | | | | | -1.61 | | | 9.07 | 0.50 | -0.87 | 0.24 | 0.89 | 0.29 |
| 53 | | | | | | | | | | | | -1.43 | | | |
| 56 | | 1.72 | | | | | -1.61 | | | | | 0.23 | | | 3.18 |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data.

TABLE 7. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCES (z-SCORE) FOR PCBs (8 to 118) (Cont.)

| Lab Code | PCB 8 | PCB 18 | PCB 28 | PCB 31 | PCB 44 | PCB 49 | PCB 52 | PCB 66 | PCB 87 | PCB 95 | PCB 99 | PCB 101 | PCB 105 | PCB 110 | PCB 118 |
|----------|-------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|---------|---------|---------|---------|
| 57 | | | 0.05 | | | | -0.70 | | | | | 2.06 | | | 1.62 |
| 62 | | | -3.96 | -0.35 | | | -1.80 | | | | | 1.69 | | | |
| 63 | | -2.56 | -2.35 | 1.64 | 2.09 | -2.01 | -3.51 | | -2.52 | -0.27 | -3.97 | -4.00 | -5.66 | -0.57 | -4.16 |
| 65 | | | -3.82 | | | | -4.89 | | | -4.02 | -3.52 | -4.66 | 0.42 | -3.84 | 0.36 |
| 66 | | | 13.3 | 32.2 | | | 2.63 | | | | -1.30 | -2.19 | | -0.47 | -1.74 |
| 67 | C | I | 17.5 | 59.8 | I | | 17.5 | 23.4 | | | | 12.9 | 49.0 | | 30.5 |
| 71 | | | | | | | | | | | | -3.33 | | | |
| 72 | | 4.23 | 1.11 | 8.09 | 1.72 | 1.36 | 2.33 | -0.93 | 0.87 | 6.28 | 1.31 | 2.58 | 4.65 | 2.21 | 1.55 |
| 73 | | | 9.84 | | | | 6.56 | | | | | -1.79 | | | |
| 74 | | | -2.33 | -0.11 | | | -0.63 | | | | | 0.40 | 0.01 | | 0.17 |
| 76 | | | | | | | | | | 17.5 | 14.1 | 9.24 | | | |
| 78 | | C | 3.08 | C | 0.14 | | 1.13 | | | | | 0.52 | 0.83 | | 0.66 |
| 79 | C | -0.21 | -1.81 | 1.28 | 0.70 | -0.07 | 3.13 | 1.54 | -0.11 | 1.79 | 1.26 | 0.46 | 1.26 | 0.95 | -0.60 |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data.

TABLE 8. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCES (z-SCORE) FOR PCBs (128 to 209)

| Lab Code | PCB 128 | PCB 138 | PCB 146 | PCB 149 | PCB 151 | PCB 153 | PCB 156 | PCB 170 | PCB 177 | PCB 180 | PCB 183 | PCB 187 | PCB 194 | PCB 195 | PCB 206 | PCB 209 |
|----------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|
| 2 | 0.67 | | | | 0.01 | | | | | -0.47 | | | | | | C |
| 3 | -0.29 | | | | -0.34 | | -2.04 | | | -2.00 | | | | | | |
| 4 | -5.83 | | | | -5.74 | | | | | -5.12 | | | | | | |
| 5 | 1.65 | | | | 0.04 | | | | | 0.31 | | | | | | |
| 7 | | | | | -0.54 | | | | | -0.81 | | | | | | |
| 8 | 6.21 | | | | 0.83 | | | | | 1.53 | | | | | | |
| 10 | -0.36 | | 0.29 | -0.08 | 0.28 | 1.17 | 0.22 | 0.36 | 1.62 | 1.13 | 0.09 | 0.28 | | | | |
| 14 | -1.20 | | | -3.60 | | -1.02 | | | | | -1.12 | | | | | |
| 15 | 1.42 | | | | 1.52 | | 0.75 | | | 0.39 | | | | | | |
| 17 | 0.42 | | | 1.54 | 2.14 | | 2.98 | | | 3.03 | | | 0.64 | | | 0.20 |
| 23 | 3.55 | | | | 8.39 | | | | | 3.95 | | | | | | 353 |
| 24 | -1.57 | | | | -2.55 | | | | | -1.46 | | | | | | |
| 25 | 0.29 | -0.86 | 24.6 | -0.85 | -1.01 | -0.23 | 0.09 | -0.06 | -1.05 | 0.30 | -1.31 | -0.99 | | | | |
| 26 | I | | | | I | | | | | I | | | | | | |
| 27 | -0.37 | | | -0.27 | -0.27 | | -0.52 | -0.72 | | -1.39 | | -0.15 | -0.86 | | | |
| 35 | -0.14 | | | | 0.28 | | | | | 3.03 | | | | | | |
| 36 | -4.57 | -1.53 | -0.34 | -2.08 | -1.73 | -1.55 | -0.68 | -2.09 | -1.81 | -2.29 | -1.39 | -1.19 | | | | |
| 40 | 0.32 | 3.56 | 0.29 | 1.66 | 0.77 | 1.31 | 0.71 | -1.04 | 1.12 | 0.64 | 2.17 | 2.98 | -1.20 | -1.30 | -0.06 | 0.55 |
| 45 | -5.96 | | | | -5.72 | | | | | -6.22 | | | | | | -6.24 |
| 47 | -2.91 | | | | -2.58 | | | | | -0.81 | | | | | | |
| 48 | | | | -3.65 | | | | | | 1.93 | | | 1.41 | | | |
| 49 | -0.32 | | | -0.99 | | | | | | -0.64 | | | -1.04 | | | |
| 50 | 0.80 | -0.79 | -0.24 | 0.60 | 1.22 | 0.98 | 0.21 | 0.31 | -0.16 | 1.23 | 1.32 | 1.59 | | | | |
| 53 | 1.32 | | | | -0.40 | | | | | -0.69 | | | | | | |
| 56 | 2.03 | | | | -0.31 | | | | | 1.20 | | | | | | |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data

TABLE 8. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCES (z-SCORE) FOR PCBs (128 to 209) (Cont)

| Lab Code | PCB 128 | PCB 138 | PCB 146 | PCB 149 | PCB 151 | PCB 153 | PCB 156 | PCB 170 | PCB 177 | PCB 180 | PCB 183 | PCB 187 | PCB 194 | PCB 195 | PCB 206 | PCB 209 |
|----------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|
| 57 | -4.40 | -4.56 | -4.61 | -4.05 | -2.19 | -3.36 | -0.03 | -3.04 | -2.57 | -3.35 | -2.51 | 0.28 | -4.81 | -2.04 | | |
| 62 | 1.41 | -4.00 | -2.71 | -4.78 | 0.06 | -4.85 | -3.74 | -3.64 | -3.41 | 9.99 | 0.96 | -3.84 | 0.96 | | | |
| 63 | -0.67 | -5.20 | -5.08 | -0.89 | 18.0 | 9.61 | 3.86 | 2.06 | 3.00 | 3.00 | 2.02 | 1.30 | 2.35 | 3.68 | | |
| 65 | -4.09 | -2.56 | -4.46 | -4.25 | -2.80 | -3.71 | 3.33 | 4.08 | 4.22 | 0.13 | 0.13 | 4.22 | 2.02 | 1.30 | 2.35 | 3.68 |
| 66 | -3.83 | 16.2 | -3.99 | I | -4.50 | 9.61 | 3.86 | 2.06 | 3.00 | 0.13 | 0.22 | 4.22 | 2.02 | 1.30 | 2.35 | 3.68 |
| 71 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 | -3.99 |
| 72 | 4.35 | 2.75 | 3.40 | 3.31 | 2.81 | 4.08 | 3.33 | 4.08 | 4.22 | 0.13 | 0.22 | 4.22 | 2.02 | 1.30 | 2.35 | 3.68 |
| 73 | 0.52 | 0.52 | 0.52 | 0.52 | 0.97 | 0.97 | 0.97 | 0.97 | 0.97 | 0.97 | 0.97 | 0.97 | 0.97 | 0.97 | 0.97 | 0.97 |
| 74 | -0.48 | -0.48 | -0.48 | -0.48 | 1.05 | 0.04 | 0.04 | 0.04 | 0.04 | 0.04 | 0.04 | 0.04 | 0.04 | 0.04 | 0.04 | 0.04 |
| 76 | 8.87 | 13.6 | 13.1 | 13.0 | 13.0 | 5.44 | 19.2 | 10.89 | 17.0 | 17.0 | 17.0 | 17.0 | 17.0 | 17.0 | 17.0 | 17.0 |
| 78 | 0.54 | 0.54 | 0.54 | 0.54 | 0.89 | 0.65 | 0.13 | 0.13 | 0.13 | 0.13 | 0.13 | 0.13 | 0.13 | 0.13 | 0.13 | 0.13 |
| 79 | 3.79 | 0.66 | -0.07 | 0.13 | -0.09 | 0.92 | 0.76 | 0.30 | 0.23 | 0.03 | 0.23 | 0.48 | 0.04 | -0.28 | -0.30 | 0.22 |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data

TABLE 9. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCES (ZETA-SCORE) FOR PCBs (18 to 118)

| Lab Code | PCB 18 | PCB 28 | PCB 31 | PCB 44 | PCB 49 | PCB 52 | PCB 66 | PCB 87 | PCB 95 | PCB 99 | PCB 101 | PCB 105 | PCB 110 | PCB 118 |
|----------|--------|--------|--------|--------|--------|--------|--------|--------|--------|--------|---------|---------|---------|---------|
| 1 | | | | | | -5.98 | | | | | | | | |
| 2 | | | | | | | | | | 1.36 | | | | -0.25 |
| 7 | | 2.18 | | | | 0.99 | | | | 0.58 | | | | -2.57 |
| 8 | | 2.31 | | | | 5.03 | | | | 3.95 | | | | |
| 14 | | | | | | -1.78 | | | -1.88 | -2.15 | | | -1.05 | |
| 15 | | -0.30 | | | | 0.60 | | | | 0.75 | | -0.65 | | -0.28 |
| 17 | 0.92 | -4.52 | -2.67 | -0.57 | | 0.79 | | | | 2.77 | | | | |
| 24 | | -2.71 | | | | -1.44 | | | | -0.18 | | | | |
| 25 | | -3.38 | -1.47 | | | -1.90 | | | -0.84 | -1.99 | -3.39 | -0.31 | -1.52 | -2.04 |
| 35 | | 3.01 | | | | 1.04 | | | | 2.21 | | | | 4.90 |
| 36 | -0.06 | 0.20 | 1.97 | -1.87 | | -3.78 | | | -0.89 | -2.41 | -3.53 | -3.34 | -2.14 | -8.26 |
| 40 | -2.34 | -1.63 | 0.45 | -1.01 | 0.07 | 0.12 | -0.19 | 0.13 | 1.50 | 0.44 | 2.19 | 2.60 | 1.94 | 2.69 |
| 47 | | 3.69 | | | | | | | | | | | | -2.27 |
| 49 | | | | | | 0.02 | | | | -0.89 | | | | -0.33 |
| 50 | | -2.34 | | | | -1.09 | | | 2.02 | 0.29 | -0.52 | 0.12 | 0.47 | 0.14 |
| 53 | | | | | | | | | | -1.07 | | | | |
| 56 | | 0.13 | | | | -0.09 | | | | 0.17 | | | | 1.70 |
| 62 | | -5.01 | -0.22 | | | -1.18 | | | | 1.14 | | | | |
| 63 | -1.30 | -0.77 | 0.23 | 0.29 | -0.88 | -1.37 | | -0.76 | -0.05 | -4.29 | -1.81 | -6.87 | -0.13 | -1.84 |
| 65 | | -4.68 | | | | -6.90 | | | -5.06 | -4.10 | -6.47 | 0.27 | -4.66 | 0.22 |
| 66 | | 7.12 | 8.32 | | | 2.65 | | | | -1.86 | -2.84 | | -0.53 | -2.62 |
| 71 | | | | | | | | | | -4.37 | | | | |
| 72 | 4.88 | 1.44 | 9.20 | 2.20 | 1.58 | 3.44 | -1.01 | 1.21 | 8.63 | 1.54 | 3.93 | 7.14 | 2.78 | 2.49 |
| 76 | | | | | | | | | 7.66 | 7.26 | | | | |
| 79 | -0.23 | -2.17 | 1.27 | 0.73 | -0.08 | 2.79 | 1.50 | -0.12 | 1.67 | 1.29 | 0.50 | 1.29 | 0.97 | -0.72 |

TABLE 10. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (ZETA-SCORE) FOR PCBs (128 to 209)

| Lab Code | PCB 128 | PCB 138 | PCB 146 | PCB 149 | PCB 151 | PCB 153 | PCB 156 | PCB 170 | PCB 177 | PCB 180 | PCB 183 | PCB 187 | PCB 194 | PCB 195 | PCB 206 | PCB 209 |
|----------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|---------|
| 2 | 0.72 | | | | 0.01 | | | | | -0.55 | | | | | | |
| 7 | | | | | -0.73 | | | | | -1.10 | | | | | | |
| 8 | | 3.37 | | | 0.72 | | | | | 1.50 | | | | | | |
| 14 | -0.96 | | | -4.30 | | | | | | | -0.99 | | | | | |
| 15 | | 0.70 | | | 0.74 | | 0.40 | | | 0.22 | | | | | | |
| 17 | | 0.45 | | 1.74 | | 2.37 | | 2.76 | | 3.01 | | | 0.57 | | | 0.18 |
| 24 | | -2.57 | | | | -4.18 | | | | -2.37 | | | | | | |
| 25 | 0.42 | -1.13 | 12.8 | -1.24 | -1.26 | -0.33 | 0.09 | -0.06 | -1.30 | 0.41 | -1.58 | -1.32 | | | | |
| 35 | | -0.23 | | | | 0.46 | | | | 4.71 | | | | | | |
| 36 | -6.34 | -2.35 | -0.59 | -3.07 | -2.28 | -2.48 | -0.79 | -3.13 | -2.37 | -3.09 | -1.54 | -1.53 | | | | |
| 40 | 0.24 | 4.98 | 0.45 | 1.12 | 0.58 | 1.99 | 1.15 | -0.96 | 0.85 | 0.98 | 1.49 | 1.83 | -1.13 | -1.30 | -0.05 | 0.84 |
| 47 | | -3.17 | | | | -2.66 | | | | -0.64 | | | | | | |
| 49 | | -0.40 | | -0.72 | | | | | | -0.57 | | | -0.15 | | | |
| 50 | 0.38 | -0.44 | -0.12 | 0.29 | 0.58 | 0.43 | 0.10 | 0.15 | -0.09 | 0.52 | 0.68 | 0.71 | | | | |
| 53 | | 0.65 | | | | -0.25 | | | | -0.45 | | | | | | |
| 56 | | 3.14 | | | | -0.50 | | | | 1.51 | | | | | | |
| 62 | | 1.68 | | | | -7.35 | | | | -9.88 | | | | | | |
| 63 | -0.13 | -1.66 | | -1.39 | -0.97 | -1.71 | -0.51 | -1.18 | -0.01 | -0.92 | -1.21 | -1.51 | -0.81 | 0.06 | -2.74 | -0.47 |
| 65 | -5.21 | -7.87 | -7.71 | -5.98 | -5.44 | -6.82 | 0.04 | -6.93 | -4.53 | | -3.96 | -4.89 | | | | |
| 66 | | -4.07 | | -1.22 | | -4.48 | -10.4 | -5.27 | | -5.57 | | | -4.79 | | | |
| 71 | | -5.65 | | | | -6.72 | | | | -5.23 | | | | | | |
| 72 | 5.61 | 4.38 | | 4.92 | 4.41 | 4.48 | 6.37 | 4.75 | 4.76 | 3.18 | 3.53 | 5.55 | 2.65 | 1.50 | 2.48 | 5.23 |
| 76 | | 5.83 | 7.19 | 7.07 | 6.87 | | | | | 7.88 | 1.51 | 7.88 | | | | |
| 79 | 3.21 | 0.71 | | -0.08 | 0.14 | -0.10 | 0.98 | 0.80 | 0.32 | 0.04 | 0.24 | 0.51 | 0.05 | -0.30 | -0.35 | 0.25 |




FIG. 7. Percentage distribution of the z-scores of results reported by participants per organochlorine pesticides compound.




FIG. 8. Percentage distribution of the Zeta-scores of results reported by participants per organochlorine pesticides compound.



FIG. 9. Percentage distribution of the z-scores of results reported for organochlorine pesticides by participants per laboratory.



FIG. 10. Percentage distribution of the Zeta-scores of results reported for organochlorine pesticides by participants per laboratory.

Table 11. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) FOR ORGANOCHLORINE PESTICIDES

| Lab Code | HCB | α -HCH | β -HCH | γ -HCH (Lindane) | pp' -DDD | pp' -DDE | pp' -DDT | δ -HCH | op -DDE | op -DDD | op -DDT | Heptachlor |
|----------|-------|---------------|--------------|-------------------------|------------|------------|------------|---------------|-----------|-----------|-----------|------------|
| 1 | | | | | -5.92 | -5.92 | -5.28 | | | | | |
| 2 | C | I | I | I | I | -0.68 | I | I | C | C | I | I |
| 3 | -1.91 | | C | C | -2.96 | -0.08 | -2.04 | | | | | |
| 4 | -5.26 | | | C | | | | C | | | | C |
| 5 | | C | C | C | 0.11 | -1.09 | 0.39 | | | | | |
| 7 | -0.25 | I | C | C | -2.56 | -0.43 | | | 0.82 | 16.7 | 10.3 | I |
| 10 | 1.74 | | | | | | -0.02 | | | | -0.41 | C |
| 14 | I | I | I | I | 12.4 | 5.39 | | I | I | | | I |
| 15 | 0.97 | I | I | I | | 1.80 | | I | -0.61 | -0.76 | | I |
| 17 | -5.30 | C | C | C | 0.25 | 2.42 | 0.52 | C | 0.22 | 2.67 | 0.04 | |
| 23 | | I | I | I | 14.4 | -5.83 | 7.53 | C | | | | I |
| 24 | -2.10 | | | | | -1.98 | | | | | | |
| 25 | -0.63 | C | C | C | 2.21 | 2.06 | 1.87 | C | -0.12 | 0.74 | -0.34 | C |
| 26 | | C | C | C | I | I | I | | | | | C |
| 27 | -1.61 | C | | C | -0.73 | -1.01 | -2.43 | | | | -2.77 | |
| 35 | 10.2 | C | C | I | 5.90 | -0.75 | 2.17 | C | 1.04 | 0.93 | 1.81 | C |
| 37 | | | | | | -1.14 | 0.70 | | | | | |
| 40 | 0.22 | C | C | C | -0.48 | 1.17 | -0.24 | C | -0.76 | -1.59 | -0.74 | C |
| 42 | -1.70 | C | C | C | -0.80 | 0.84 | 2.04 | | -2.60 | -2.75 | 0.66 | C |
| 45 | | C | C | C | -4.95 | -5.85 | -5.63 | C | | | | I |
| 50 | 0.22 | | | | 0.01 | 0.80 | -0.43 | | -0.37 | -1.97 | -0.93 | |
| 53 | 0.60 | | | | | -0.53 | | | | | | |
| 54 | 1.25 | I | I | I | 4.20 | 1.91 | 1.16 | I | 0.25 | 1.35 | 2.19 | I |
| 55 | 48.6 | I | I | I | | -1.48 | I | I | I | | C | I |
| 56 | 1.97 | C | C | C | 5.67 | 2.41 | 7.33 | C | -5.97 | 0.70 | 7.03 | C |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data.

Table 11. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) FOR ORGANOCHLORINE PESTICIDES (Cont.)

| Lab Code | HCB | α -HCH | β -HCH | γ -HCH (Lindane) | pp' -DDD | pp' -DDE | pp' -DDT | δ -HCH | op' -DDE | op' -DDD | op' -DDT | Heptachlor |
|----------|-------|---------------|--------------|-------------------------|------------|------------|------------|---------------|------------|------------|------------|------------|
| 57 | | | | | -0.12 | -2.17 | | | | | | C |
| 62 | -5.68 | | | | -3.81 | -5.51 | -2.92 | | | | | C |
| 63 | | | | | -5.36 | -5.33 | -5.15 | | | | | I |
| 66 | -1.54 | C | C | | 0.54 | -1.88 | 0.60 | C | 0.53 | 4.71 | 2.45 | C |
| 67 | 214 | I | I | | 147 | 157 | 141 | | | | | I |
| 71 | | | | | -3.13 | -2.02 | -1.50 | | | | | |
| 72 | 0.86 | C | C | | -0.56 | 1.83 | -0.28 | | 1.99 | 0.09 | 1.24 | C |
| 75 | 0.57 | | | | 2.92 | -0.89 | 0.32 | | | 6.57 | -0.91 | |
| 76 | 9.34 | | | | 16.9 | 18.5 | -1.08 | | | | | |
| 78 | 0.51 | I | I | | 1.86 | 1.08 | 1.15 | I | -1.11 | -0.36 | 0.45 | I |
| 79 | -3.70 | C | C | | -0.74 | -0.23 | -1.27 | C | 9.89 | -1.29 | -2.60 | C |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data.

TABLE 12 OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) FOR ORGANOCHLORINE PESTICIDES

| Lab Code | Aldrin | Dieldrin | Endrin | cis-Chlordane | trans-Chlordane | cis-Nonachlor | trans-Nonachlor | α-Endosulfan | β-Endosulfan | Endosulfan sulfate | Heptachlor epoxide |
|----------|--------|----------|--------|---------------|-----------------|---------------|-----------------|--------------|--------------|--------------------|--------------------|
| 1 | | 2.88 | | | | | | | | | |
| 2 | I | I | I | I | I | | | I | I | | I |
| 3 | C | -0.48 | C | | | | | | | | |
| 4 | | | C | | | | | | | | |
| 7 | I | -1.31 | I | | | | | I | | | 3.94 |
| 10 | | | | | | | | | | | -2.01 |
| 14 | I | | | C | I | | | I | I | I | I |
| 15 | I | C | I | -0.15 | I | | | | | | -4.13 |
| 17 | I | 0.50 | I | -3.01 | 71.3 | | | I | I | I | 47.9 |
| 23 | C | 1.59 | C | | | | | C | C | C | I |
| 25 | C | I | C | | | | | C | C | C | I |
| 26 | C | | | | | | | | | | I |
| 27 | C | | | -1.04 | I | | | C | C | | I |
| 35 | C | -4.89 | | | | | | | | | |
| 37 | | | | | | | | I | | | |
| 40 | C | 1.58 | C | 0.75 | -0.86 | 0.01 | 0.45 | C | I | | 0.20 |
| 42 | | | | -2.12 | 81.8 | | -1.67 | | | | |
| 45 | I | -0.72 | I | -4.75 | -0.27 | | | I | I | I | 22.1 |
| 47 | | 42.3 | I | | | | | | | | 282 |
| 50 | | 0.01 | | | | | | | | | |
| 53 | | | | | | | | | | | I |
| 54 | I | C | I | 3.27 | I | -0.05 | -0.10 | I | I | I | I |
| 55 | I | C | I | 39.3 | 269 | | | I | I | I | I |
| 56 | C | -2.96 | I | 1.26 | -2.40 | | | C | I | | -0.55 |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data

Table 12. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) FOR ORGANOCHLORINE PESTICIDES (Cont)

| Lab Code | Aldrin | Dieldrin | Endrin | cis-Chlordane | trans-Chlordane | cis-Nonachlor | trans-Nonachlor | α-Endosulfan | β-Endosulfan | Endosulfan sulfate | Heptachlor epoxide |
|----------|--------|----------|--------|---------------|-----------------|---------------|-----------------|--------------|--------------|--------------------|--------------------|
| 57 | I | | | | | | | I | I | C | -0.15 |
| 62 | I | 5.79 | I | | | | | C | | | |
| 63 | I | -3.04 | I | | -0.65 | | | C | I | | -1.37 |
| 66 | C | -2.15 | C | -0.89 | 18.3 | | | C | I | C | 4.89 |
| 67 | I | 141 | | | | | | | | | |
| 72 | C | 0.10 | | 1.50 | 0.22 | 1.50 | 1.03 | C | | | 1.00 |
| 76 | I | 187 | | | | | | | | | |
| 78 | I | I | I | 1.51 | 117 | | | I | I | I | I |
| 79 | C | -1.92 | C | -0.90 | 0.70 | -0.02 | -0.15 | C | C | C | -1.24 |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data

TABLE 13. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (ZETA-SCORE) FOR ORGANOCHLORINE PESTICIDES

| Lab Code | HCB | <i>pp'</i> -DDD | <i>pp'</i> -DDE | <i>pp'</i> -DDT | <i>op'</i> -DDE | <i>op'</i> -DDD | <i>op'</i> -DDT | Dieldrin | <i>cis</i> -Chlordane | <i>trans</i> -Chlordane | <i>cis</i> -Nonachlor | <i>trans</i> -Nonachlor | Heptachlor epoxide |
|----------|-------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|----------|-----------------------|-------------------------|-----------------------|-------------------------|--------------------|
| 1 | | -9.73 | -9.13 | -8.04 | | | | 2.31 | | | | | |
| 2 | | | -0.78 | | | | | | | | | | |
| 7 | -0.28 | -3.19 | -0.47 | | 1.00 | 4.38 | 7.48 | -1.31 | | | | | 2.35 |
| 14 | | 2.97 | 2.25 | | | | | | | | | | |
| 15 | 0.51 | | 0.87 | | -0.40 | -0.53 | | | -0.07 | | | | |
| 17 | -7.97 | 0.23 | 2.28 | 0.63 | 0.31 | 2.66 | 0.05 | | | | | | |
| 25 | -0.71 | 3.06 | 3.04 | 2.30 | -0.14 | 0.79 | -0.45 | 1.44 | | | | | |
| 35 | 9.28 | 8.82 | -1.13 | 3.10 | 1.50 | 1.28 | 2.56 | -6.29 | | | | | |
| 40 | 0.33 | -0.76 | 1.71 | -0.36 | -1.17 | -1.52 | -1.09 | 1.98 | 0.98 | -1.30 | 0.03 | 0.62 | 0.25 |
| 47 | | | | | | | | 5.23 | | | | | 5.78 |
| 50 | 0.11 | 0.01 | 0.37 | -0.24 | -0.20 | -1.48 | -0.56 | 0.01 | | | | | |
| 53 | 0.33 | | -0.34 | | | | | | | | | | |
| 54 | 0.62 | 1.54 | 0.91 | 0.61 | 0.14 | 0.72 | 1.00 | | 1.41 | | | | |
| 56 | 0.07 | 2.69 | 3.12 | 6.03 | -0.15 | 0.09 | 5.85 | -0.35 | 0.10 | -0.01 | -0.03 | -0.06 | -0.01 |
| 62 | -8.71 | -5.77 | -8.50 | -4.08 | | | | 4.75 | | | | | |
| 63 | | -4.56 | -5.30 | -4.43 | | | | -1.40 | | | | | |
| 66 | -1.82 | 0.59 | -2.64 | 0.71 | 0.61 | 4.71 | 2.14 | -2.72 | -1.14 | 6.14 | | | -0.37 |
| 71 | | -4.04 | -2.27 | -1.59 | | | | | | | | | 4.04 |
| 72 | 1.29 | -0.90 | 2.67 | -0.41 | 3.10 | 0.13 | 1.84 | 0.12 | 1.95 | 0.33 | 2.48 | 1.46 | 1.26 |
| 76 | | 6.76 | 14.01 | -1.16 | | | | 9.76 | | | | | |
| 79 | -5.37 | -0.75 | -0.27 | -1.58 | 5.68 | -1.56 | -3.50 | -2.29 | -1.03 | 0.77 | -0.03 | -0.17 | -1.44 |




FIG. 11. The z-scores of results reported by participants per PBDEs compound.




FIG. 12. Percentage distribution of the Zeta-scores of results reported by participants per PBDEs compound.



FIG. 13. Percentage distribution of the z-scores of PBDEs results reported by participants per laboratory.




FIG. 14. Percentage distribution of the Zeta-scores of PBDEs results reported by participants per laboratory.

TABLE 14. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) FOR PBDEs

| Lab Code | BDE 17 | BDE 28 | BDE 47 | BDE 49 | BDE 66 | BDE 85 | BDE 99 | BDE 100 | BDE 153 | BDE 154 | BDE 183 | BDE 209 |
|----------|--------|--------|--------|--------|--------|--------|--------|---------|---------|---------|---------|---------|
| 10 | | 1.37 | 0.88 | | | | -0.54 | 0.78 | 0.78 | 0.69 | | |
| 14 | | 1.19 | 0.29 | | | | -0.13 | 0.33 | 1.51 | 1.43 | C | |
| 17 | C | 2.15 | 0.11 | -0.76 | -1.22 | C | 0.15 | 1.60 | 0.36 | 0.99 | C | C |
| 27 | | 0.37 | 0.15 | | 0.38 | | 0.30 | 0.45 | 3.48 | 2.88 | | |
| 35 | | 0.74 | 0.48 | | | | 0.73 | 2.16 | 1.39 | 2.82 | | |
| 38 | C | 0.76 | 0.71 | 1.13 | 0.55 | C | 2.18 | 0.15 | 0.14 | 0.71 | C | C |
| 40 | C | 0.61 | -0.31 | -0.21 | -0.70 | C | -0.92 | -0.12 | 0.49 | -0.63 | C | C |
| 46 | | -5.86 | -6.41 | | | | -6.16 | -5.74 | -5.92 | -5.76 | | |
| 49 | | -0.62 | -2.47 | | | | 14.1 | -4.54 | | -4.33 | | |
| 50 | | -5.87 | -1.15 | | | | -6.17 | -2.93 | -5.95 | -3.29 | | |
| 56 | | -2.87 | 1.27 | | | | -3.67 | -5.30 | 0.19 | -5.14 | C | I |
| 65 | | -4.04 | -4.32 | | | | -4.31 | 7946 | -2.45 | -3.66 | | |
| 66 | | -2.39 | -2.24 | | | | -1.70 | -1.36 | -1.91 | 2.41 | C | |
| 72 | | 5.41 | 1.61 | 4.28 | 2.12 | C | 2.11 | 1.88 | 3.63 | 2.67 | C | |
| 74 | | 0.25 | -0.26 | 0.14 | -1.12 | C | -0.28 | 0.01 | 0.64 | 0.24 | C | C |
| 75 | | -1.41 | -3.10 | | | | -1.95 | -2.96 | -2.25 | -1.78 | | |
| 78 | | -0.53 | -0.43 | | | | C | -1.26 | I | -1.18 | | |
| 79 | C | -2.06 | 0.16 | | 0.17 | C | 0.64 | -0.41 | -1.19 | -0.67 | | C |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data.

TABLE 15. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (ZETA-SCORE) FOR PBDEs

| Lab Code | BDE 28 | BDE 47 | BDE 49 | BDE 66 | BDE 99 | BDE 100 | BDE 153 | BDE 154 |
|----------|--------|--------|--------|--------|--------|---------|---------|---------|
| 14 | 0.70 | 0.18 | | | -0.09 | 0.22 | 0.85 | 0.83 |
| 17 | 1.80 | 0.07 | -0.77 | -1.01 | 0.12 | 0.68 | 0.26 | 0.60 |
| 35 | 1.08 | 0.80 | | | 1.12 | 3.05 | 1.80 | 3.87 |
| 38 | 0.77 | 0.91 | 1.15 | 0.49 | 0.20 | 0.19 | 0.03 | 0.90 |
| 40 | 0.83 | -0.48 | -0.18 | -0.63 | -1.40 | -0.17 | 0.70 | -0.89 |
| 46 | -5.68 | -10.8 | | | -5.96 | -8.17 | -3.89 | -8.05 |
| 49 | -0.01 | -0.57 | | | 0.21 | -0.38 | | -0.23 |
| 50 | -8.67 | -0.67 | | | -9.73 | -2.60 | -8.92 | -3.15 |
| 56 | -0.18 | 0.67 | | | -0.33 | -1.33 | 0.00 | -0.90 |
| 65 | -5.12 | -5.73 | | | -5.69 | 9138 | -2.40 | -4.38 |
| 66 | -2.76 | -3.39 | | | -1.99 | -1.74 | -2.00 | 2.70 |
| 72 | 7.59 | 2.56 | 5.97 | 2.77 | 2.87 | 2.57 | 5.16 | 3.58 |
| 79 | -2.80 | 0.19 | | 0.18 | 0.69 | -0.47 | -1.44 | -0.78 |




FIG. 15. Percentage distribution of the z-scores of PAH results reported by participants per compound.




FIG. 16. Percentage distribution of the Zeta-scores of PAH results reported by participants per compound.




FIG. 17. Percentage distribution of the z-scores of PAH results reported by participants per laboratory.




FIG. 18. Percentage distribution of the Zeta-scores of PAH results reported by participants per laboratory.

TABLE 16. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) FOR PAHs

| Lab Code | Naphthalene | 1-Methylnaphthalene | 2-Methylnaphthalene | Acenaphthylene | Acenaphthene | Fluorene | Phenanthrene | Anthracene |
|----------|-------------|---------------------|---------------------|----------------|--------------|----------|--------------|------------|
| 4 | | | | | | | | I |
| 5 | 1.27 | | | 6.68 | 0.47 | 0.44 | 1.58 | C |
| 12 | -2.60 | -2.13 | -1.35 | 9.64 | 10.7 | -0.90 | -1.46 | C |
| 14 | | | | I | I | | | |
| 15 | 4.09 | 1.73 | 5.73 | I | I | 5.50 | 3.10 | I |
| 19 | | | | | | | -1.09 | C |
| 22 | -4.96 | | | -0.70 | -2.17 | -5.25 | -5.05 | |
| 25 | 12.9 | 2.82 | -2.09 | -0.47 | -1.48 | 0.72 | -0.55 | C |
| 26 | -4.85 | | | I | I | I | I | C |
| 35 | -0.64 | | | -4.39 | -2.16 | -2.78 | -0.17 | C |
| 36 | -4.36 | | | | -1.04 | -0.82 | 4.24 | |
| 37 | 29.2 | | | | | | | |
| 42 | -1.32 | | | | -2.46 | -0.72 | 1.91 | C |
| 45 | -4.98 | | | -5.99 | | | -5.29 | C |
| 47 | | | | | | 40.1 | 305 | I |
| 49 | | -0.31 | 1.01 | | 5.23 | -2.96 | -1.44 | |
| 52 | 1.32 | | | 8.07 | 2.39 | 2.50 | 2.87 | C |
| 55 | 86.2 | 0.47 | 1.24 | I | I | 17.4 | C | I |
| 56 | | | | -5.11 | -4.02 | -5.19 | -5.26 | C |
| 65 | -0.71 | | | 28 | 0.84 | 3.72 | 5.30 | C |
| 66 | 0.78 | | | 90.7 | 4.85 | 1.75 | 12.9 | C |
| 67 | 837 | | | C | C | I | 174 | C |
| 72 | -2.03 | -0.81 | -0.31 | -0.23 | 0.22 | -2.25 | 0.19 | C |
| 76 | 10.4 | | | 1668 | 711 | 15.2 | -3.04 | I |
| 78 | 1.33 | | C | I | 84.5 | I | -1.21 | C |
| 79 | -2.66 | -0.17 | 0.87 | -2.80 | -1.39 | -2.87 | -1.65 | C |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data.

TABLE 17. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (z-SCORE) FOR PAHs

| Lab Code | Benz[a]anthracene | Chrysene | Fluoranthene | Pyrene | Benzo[a]pyrene | Benzo[e]pyrene | Benzo[b+g]fluoranthene | Benzo[g,h,i]perylene |
|----------|-------------------|----------|--------------|--------|----------------|----------------|------------------------|----------------------|
| 4 | I | | I | | I | | I | I |
| 5 | C | C | C | C | C | | C | C |
| 12 | C | C | C | I | C | | C | C |
| 14 | | | | | | | | |
| 15 | I | I | I | C | I | | I | I |
| 19 | C | C | C | C | C | C | C | C |
| 22 | C | I | C | C | C | | C | C |
| 25 | C | C | C | C | C | C | C | C |
| 26 | C | C | C | C | C | C | C | C |
| 35 | C | C | C | C | C | C | C | C |
| 36 | I | I | I | C | | | | |
| 42 | C | C | C | C | C | | | C |
| 45 | C | C | C | C | C | C | C | C |
| 47 | I | I | | | I | | I | I |
| 52 | | | C | C | | | | I |
| 55 | I | I | C | C | I | | I | I |
| 56 | C | C | C | C | C | | C | C |
| 65 | C | C | C | C | C | | C | C |
| 66 | C | C | C | C | C | | C | C |
| 67 | I | C | I | I | I | I | C | C |
| 72 | C | C | C | C | C | C | C | C |
| 74 | C | C | C | | C | | C | C |
| 76 | I | C | C | I | I | | I | I |
| 78 | C | C | C | C | C | C | C | C |
| 79 | C | C | C | C | C | C | C | C |

Consistent (C) or Inconsistent (I) value after the assessment of <LD reported data.

TABLE 18. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (ZETA-SCORE) FOR PAHS

| Lab Code | Naphthalene | 1-Methylnaphthalene | 2-Methylnaphthalene | Acenaphthylene | Acenaphthene | Fluorene | Phenanthrene |
|----------|-------------|---------------------|---------------------|----------------|--------------|----------|--------------|
| 15 | 1.70 | 0.87 | 2.10 | | | 1.95 | 1.39 |
| 25 | 14.4 | 3.27 | -2.54 | -0.44 | -1.71 | 0.81 | -0.63 |
| 35 | -0.82 | | | -6.62 | -2.57 | -3.74 | -0.23 |
| 36 | -5.58 | | | | -0.66 | -0.94 | 0.94 |
| 47 | | | | | | 8.70 | 9.83 |
| 49 | | -0.13 | 0.58 | | 0.33 | -0.36 | -0.51 |
| 52 | 1.69 | | | 6.61 | 2.54 | 3.26 | 3.80 |
| 56 | | | | -0.18 | -0.12 | -2.85 | -0.71 |
| 65 | -0.58 | | | 3.62 | 0.62 | 1.73 | 2.16 |
| 66 | 0.88 | | | 15.6 | 3.68 | 1.79 | 11 |
| 72 | -2.59 | -1.10 | -0.38 | | 0.25 | -3.05 | 0.24 |
| 76 | 6.65 | | | 10.4 | 13.3 | 5.45 | |
| 79 | -3.11 | -0.15 | 0.75 | -3.35 | -1.48 | -3.46 | -1.78 |

9. RECOMMENDATIONS AND CONCLUSIONS

Participants are recommended to review their individual data with their corresponding z-scores and Zeta-scores. The performance values of z-scores and Zeta-scores should assist participant laboratories to identify any systematic error in their measurement results and then apply the consequent corrective actions to improve the data quality. Although in general more than half of the calculated z-scores for each group of analytes were satisfactory, the z-scores distribution of some of the laboratories showed an inconsistent pattern. In some cases, for the same group of compounds, excellent z-scores values were reported for some compounds while for others, z-scores were completely outlying. Such z-scores variation suggests the existence of possible underlying issues linked to both sample preparation and instrumental analysis. In this context, clean-up and fractionation are key sample pre-treatment steps that need to be developed, optimized and monitored especially in the analysis of complex biological matrices such as fish tissue. Systematic and random errors may also be due to contamination issues. Solvents and reagents used for sample preparation and analysis should be of the highest available purity and checked on regular bases. Special care should also be taken during the evaporation procedure of the solvent extracts to avoid dryness and losses of the more volatile contaminants. In this aspect, the use of internal standards/surrogates with physico-chemical properties similar to the target analytes, such as deuterated and ^{13}C labeled compounds for GC-MS, is fully recommended to compensate for these losses. Furthermore, to overcome the common problems of co-elution and interferences in chromatographic analyses, a multi-column technique for complete separation of all isomers of PCBs, PBDEs and PAHs, as well as multiple detection strategies is recommended.

Some laboratories reported results which differed by one order of magnitude from the assigned value. This may be due to a “reporting” mistake (for example: wrong unit conversion or wrong dataset reported) or due to more severe analytical issues which would require immediate root cause analysis and consequent corrective actions. These laboratories should verify their analytical procedures and their data reporting units.

According to the results provided, many of the participating laboratories, including some accredited according to the ISO/IEC 17025, need to improve their QA/QC procedures by using certified reference materials and replicate samples to produce accurate results. Reference materials must match the test sample matrix and must undergo the same exact procedure of the test sample to be as effective as possible to avoid inaccuracy and precision issues. It is also recommended that participants provide their data along with their estimates of uncertainty in accordance to the approach set in the basic Guide to the expression of uncertainty in measurement (GUM). Unfortunately, many laboratories did not report their uncertainties and some of the uncertainties reported by participants were not realistic. The knowledge on basic

principles of method validation, traceability and uncertainty of measurement results, are still limited and laboratories that lack proficiency in this area should take action.

Overall, the results indicate the need to improve the routine monitoring capabilities of persistent organic contaminants and for that interlaboratory comparisons for trace level of organic compounds in complex matrices is a valuable approach for monitoring the quality assurance of the data.

APPENDIX I: RESULTS REPORTED BY PARTICIPANTS

I.1. RESULTS REPORTED FOR PCB CONGENERS

TABLE 19. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 8

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 23 | GC/ECD | 108.28 | | | | | | Yes ³ | Yes |
| 40 ¹ | GC/HRMS | <0.3 | | | | Yes | No | Yes | Yes |
| 67 ¹ | GC/ECD | <0.15 | | | | | | | |
| 79 | GC/ECD ² | <0.33 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited

² GC/ECD and peak confirmation with GC/MS

³ Yes of a different matrix of the test sample

TABLE 20. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 18

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | 0.84 | 0.08 | 0.80 | 0.92 | Yes | Yes | Yes | No |
| 36 | GC/MS EI | 0.7 | 0.3 | -0.08 | -0.06 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 0.36 | 0.13 | -2.14 | -2.34 | Yes | No | Yes | Yes |
| 48 | GC/ECD | 0.93 | | 1.32 | | Yes | | Yes | Yes |
| 63 | GC/MS EI | 0.29 | 0.07 | -2.56 | -1.30 | No | No | Yes ³ | No |
| 67 ¹ | GC/ECD | <0.15 | | | | | | | |
| 72 ¹ | GC/MS EI | 1.4 | 0.07 | 4.23 | 4.88 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | <1 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 0.68 | 0.14 | -0.21 | -0.23 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited

² GC/ECD and peak confirmation with GC/MS

³ Yes of a different matrix of the test sample

TABLE 21. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 28

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <1.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 1.01 | | -1.19 | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 0.9 | | -1.64 | | Yes | No | No | Yes |
| 5 | GC/MSMS | 0.93 | | -1.52 | | | Yes | Yes | Yes |
| 7 | GC/ECD | 1.78 | 0.22 | 1.90 | 2.18 | No | No | Yes | Yes |
| 8 | GC/ECD | 5.38 | 3.5 | 16.5 | 2.31 | No | | | |
| 10 | GC/MSMS | 1.75 | | 1.81 | | Yes | Yes | No | |
| 15 ¹ | GC/MSMS | 1.20 | 0.6 | -0.42 | -0.30 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 0.47 | 0.04 | -3.40 | -4.52 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 14 | | 51.4 | | | | Yes ⁴ | Yes |
| 24 | GC/MS EI | 0.77 | 0.06 | -2.18 | -2.71 | No | Yes | Yes | Yes |
| 25 | GC/MSMS | 0.68 | 0.04 | -2.55 | -3.38 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 2.38 | | 4.35 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 1.87 | 0.07 | 2.28 | 3.01 | Yes | No | No | Yes |
| 36 | GC/MS EI | 1.41 | 0.91 | 0.40 | 0.20 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 1 | 0.07 | -1.25 | -1.63 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 0.57 | | -2.97 | | No | No | No | No |
| 47 ¹ | GC/ECD | 5.17 | 2.06 | 15.7 | 3.69 | No | No | No | Yes |
| 48 | GC/ECD | 0.65 | | -2.68 | | Yes | | Yes | Yes |
| 50 | GC/MSMS | 0.67 | 0.4 | -2.59 | -2.34 | Yes | Yes | No | |
| 56 | GC/HRMS | 1.73 | 6.73 | 1.72 | 0.13 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 1.32 | | 0.05 | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 0.33 | 0.15 | -3.96 | -5.01 | Yes | No | Yes | No |
| 63 | GC/MS EI | 0.73 | 0.06 | -2.35 | -0.77 | No | No | Yes ⁴ | No |
| 65 | GC/HRMS | 0.37 | 0.16 | -3.82 | -4.68 | Yes | Yes | Yes ⁴ | Yes |
| 66 | GC/MSMS | 4.57 | 0.83 | 13.3 | 7.12 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 5.61 | | 17.5 | | | | | |
| 72 ¹ | GC/MS EI | 1.58 | 0.09 | 1.11 | 1.44 | Yes | Yes | Yes | Yes |
| 73 | GC/ECD | 3.73 | | 9.84 | | No | No | No | Yes |
| 74 ¹ | GC/MSMS | 0.73 | | -2.33 | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 2.07 | | 3.08 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 0.86 | 0.17 | -1.81 | -2.17 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 22. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 31

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | 0.39 | 0.04 | -1.94 | -2.67 | Yes | Yes | Yes | No |
| 25 | GC/MSMS | 0.49 | 0.03 | -1.12 | -1.47 | Yes | | | Yes |
| 27 | GC/ECD | 0.47 | | -1.35 | | Yes | Yes | No | Yes |
| 36 | GC/MS EI | 0.91 | 0.23 | 2.51 | 1.97 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 0.69 | 0.24 | 0.58 | 0.45 | Yes | No | Yes | Yes |
| 48 | GC/ECD | 0.58 | | -0.4 | | Yes | | Yes | Yes |
| 62 | GC/MS EI | 0.58 | 0.41 | -0.35 | -0.22 | Yes | No | Yes | No |
| 63 | GC/MS EI | 0.81 | 0.08 | 1.64 | 0.23 | No | No | Yes ³ | No |
| 66 | GC/MSMS | 4.35 | 0.87 | 32.2 | 8.32 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 7.55 | | 59.8 | | | | | |
| 72 ¹ | GC/MS EI | 1.56 | 0.11 | 8.09 | 9.20 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 0.61 | | -0.11 | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | <1 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 0.77 | 0.15 | 1.28 | 1.27 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 23. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 44

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | 1.55 | 0.45 | -0.58 | -0.57 | Yes | Yes | Yes | No |
| 36 | GC/MS EI | 0.95 | 0.69 | -2.35 | -1.87 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 1.39 | 0.48 | -1.05 | -1.01 | Yes | No | Yes | Yes |
| 48 | GC/ECD | 1.58 | | -0.50 | | Yes | | Yes | Yes |
| 63 | GC/MS EI | 2.47 | 0.32 | 2.09 | 0.29 | No | No | Yes ³ | No |
| 67 ¹ | GC/ECD | <0.15 | | | | | | | |
| 72 ¹ | GC/MS EI | 2.34 | 0.09 | 1.72 | 2.20 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 1.80 | | 0.14 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 1.99 | 0.40 | 0.70 | 0.73 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 24. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 49

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 2.19 | 0.78 | 0.08 | 0.07 | Yes | No | Yes | Yes |
| 63 | GC/MS EI | 1.11 | 0.16 | -2.01 | -0.88 | No | No | Yes ³ | No |
| 72 ¹ | GC/MS EI | 2.86 | 0.11 | 1.36 | 1.58 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 2.11 | 0.42 | -0.07 | -0.08 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 25. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 52

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 1 | GC/MS EI | 1.15 | 0.30 | -4.21 | -5.98 | Yes | No | No | Yes |
| 2 ¹ | GC/ECD ² | <1.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 2.95 | | -1.45 | | No | Yes | Yes | Yes |
| 5 | GC/MSMS | 4.20 | | 0.46 | | | Yes | Yes | Yes |
| 7 | GC/ECD | 4.42 | 0.57 | 0.80 | 0.99 | No | No | Yes | Yes |
| 8 | GC/ECD | 67.6 | 25.3 | 97.7 | 5.03 | No | | | |
| 10 | GC/MSMS | 4.08 | | 0.28 | | Yes | Yes | No | |
| 14 | GC/MS EI | 2.62 | 1.14 | -1.96 | -1.78 | No | No | No | Yes |
| 15 ¹ | GC/MSMS | 4.64 | 2.32 | 1.13 | 0.60 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 4.45 | 1.10 | 0.8 | 0.79 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 71.5 | | 104 | | | | Yes ⁴ | Yes |
| 24 | GC/MS EI | 3.27 | 0.06 | -0.97 | -1.44 | No | Yes | Yes | Yes |
| 25 | GC/MSMS | 3.06 | 0.17 | -1.29 | -1.90 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 3.00 | | -1.38 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 4.36 | 0.12 | 0.70 | 1.04 | Yes | No | No | Yes |
| 36 | GC/MS EI | 2.12 | 0.36 | -2.72 | -3.78 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 3.95 | 0.25 | 0.08 | 0.12 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 16.3 | | 19.1 | | No | No | No | No |
| 48 | GC/ECD | 5.62 | | 2.63 | | Yes | | Yes | Yes |
| 49 ¹ | GC/MS EI | 3.99 | 10.2 | 0.15 | 0.02 | | No | Yes | Yes |
| 50 | GC/MSMS | 2.85 | 1.71 | -1.61 | -1.09 | Yes | Yes | No | |
| 56 | GC/HRMS | 2.85 | 24.1 | -1.61 | -0.09 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 3.44 | | -0.70 | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 2.72 | 2.24 | -1.80 | -1.18 | Yes | No | Yes | No |
| 63 | GC/MS EI | 1.61 | 0.23 | -3.51 | -1.37 | No | No | Yes ⁴ | No |
| 65 | GC/HRMS | 0.71 | 0.31 | -4.9 | -6.90 | Yes | Yes | Yes ⁴ | Yes |
| 66 | GC/MSMS | 5.62 | 0.95 | 2.6 | 2.65 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 15.3 | | 17.5 | | | | | |
| 72 ¹ | GC/MS EI | 5.42 | 0.17 | 2.33 | 3.44 | Yes | Yes | Yes | Yes |
| 73 | GC/ECD | 8.18 | | 6.56 | | No | No | No | Yes |
| 74 ¹ | GC/MSMS | 3.49 | | -0.63 | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 4.63 | | 1.13 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 5.94 | 1.19 | 3.13 | 2.79 | Yes | Yes | Yes | Yes |

¹Laboratory accredited²GC/ECD and peak confirmation with dual column³GC/ECD and peak confirmation with GC/MS⁴Yes of a different matrix of the test sample

TABLE 26. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 66

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 6.70 | 2.30 | -0.20 | -0.19 | Yes | No | Yes | Yes |
| 67 ¹ | GC/ECD | 57.7 | | 23.4 | | | | | |
| 72 ¹ | GC/MS EI | 5.12 | 0.17 | -0.93 | -1.01 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 10.5 | 2.09 | 1.54 | 1.50 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS

TABLE 27. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 70

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 5.70 | 2 | | | Yes | No | Yes | Yes |
| 63 | GC/MS EI | 1.75 | 0.17 | | | No | No | Yes ² | No |
| 72 ¹ | GC/MS EI | 4.54 | 0.11 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² Yes of a different matrix of the test sample

TABLE 28. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 74

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 4.35 | 1.53 | | | Yes | No | Yes | Yes |
| 63 | GC/MS EI | 2.76 | 0.35 | | | No | No | Yes ² | No |
| 72 ¹ | GC/MS EI | 3.72 | 0.09 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² Yes of a different matrix of the test sample

TABLE 29. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 87

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 6.72 | 2.40 | 0.16 | 0.13 | Yes | No | Yes | Yes |
| 63 | GC/MS EI | 3.67 | 0.53 | -2.52 | -0.76 | No | No | Yes ³ | No |
| 72 ¹ | GC/MS EI | 7.52 | 0.36 | 0.87 | 1.21 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 6.41 | 1.28 | -0.11 | -0.12 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 30. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 95

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 8.98 | | 4.08 | | Yes | Yes | No | |
| 14 | GC/MS EI | 3.38 | 1.48 | -2.01 | -1.88 | No | No | No | Yes |
| 25 | GC/MSMS | 4.68 | 0.21 | -0.60 | -0.84 | Yes | | | Yes |
| 36 | GC/MS EI | 4.47 | 1.10 | -0.82 | -0.89 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 7.41 | 2.60 | 2.38 | 1.50 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 13.6 | 8.14 | 9.1 | 2.02 | Yes | Yes | No | |
| 63 | GC/MS EI | 4.99 | 0.43 | -0.27 | -0.05 | No | No | Yes ² | No |
| 65 | GC/HRMS | 1.53 | 0.67 | -4.02 | -5.06 | Yes | Yes | Yes ² | Yes |
| 72 ¹ | GC/MS EI | 11 | 0.34 | 6.28 | 8.63 | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 21.33 | 4.00 | 17.5 | 7.66 | No | Yes | | Yes |

¹ Laboratory accredited² Yes of a different matrix of the test sample

TABLE 31. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 97

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 5.73 | 2.00 | | | Yes | No | Yes | Yes |
| 79 | GC/ECD ² | 4.95 | 0.99 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS

TABLE 32. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 99

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 20.5 | | 0.79 | | Yes | Yes | No | |
| 14 | GC/MS EI | 9.03 | 4.00 | -2.10 | -2.15 | No | No | No | Yes |
| 25 | GC/MSMS | 10.7 | 0.31 | -1.67 | -1.99 | Yes | | | Yes |
| 36 | GC/MS EI | 9.32 | 0.47 | -2.03 | -2.41 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 19.5 | 6.90 | 0.53 | 0.44 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 19.4 | 11.6 | 0.5 | 0.29 | Yes | Yes | No | |
| 63 | GC/MS EI | 1.57 | 0.10 | -3.97 | -4.29 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 3.38 | 1.49 | -3.52 | -4.10 | Yes | Yes | Yes ³ | Yes |
| 72 ¹ | GC/MS EI | 22.6 | 1.10 | 1.31 | 1.54 | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 73.7 | 14 | 14.1 | 7.26 | No | Yes | | Yes |
| 79 | GC/ECD ² | 24.5 | 5.32 | 1.79 | 1.67 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 33. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 101

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 2 ¹ | GC/ECD ² | 28.4 | 5.70 | 1.33 | 1.36 | No | Yes | No | Yes |
| 3 | GC/MSMS | 20.1 | | -0.85 | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 1.78 | | -5.64 | | Yes | No | No | Yes |
| 5 | GC/MSMS | 23.8 | | 0.13 | | | Yes | Yes | Yes |
| 7 | GC/ECD | 25.1 | 3.67 | 0.47 | 0.58 | No | No | Yes | Yes |
| 8 | GC/ECD | 48.1 | 11.5 | 6.5 | 3.95 | No | | | |
| 10 | GC/MSMS | 31.3 | | 2.09 | | Yes | Yes | No | |
| 14 | GC/MS EI | 11.7 | 5.20 | -3.05 | -3.25 | No | No | No | Yes |
| 15 ¹ | GC/MSMS | 29.1 | 14.6 | 1.51 | 0.75 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 31.6 | 3.32 | 2.2 | 2.77 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 39.1 | | 4.14 | | | | Yes ⁴ | Yes |
| 24 | GC/MS EI | 22.9 | 1.42 | -0.12 | -0.18 | No | Yes | Yes | Yes |
| 25 | GC/MSMS | 14.8 | 0.89 | -2.23 | -3.39 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 19 | | -1.13 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 28.8 | 0.51 | 1.44 | 2.21 | Yes | No | No | Yes |
| 36 | GC/MS EI | 14.6 | 0.41 | -2.29 | -3.53 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 29.1 | 1.80 | 1.51 | 2.19 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 0.54 | | -6.0 | | No | No | No | No |
| 47 ¹ | GC/ECD | 14.4 | | -2.34 | | No | No | No | Yes |
| 48 | GC/ECD | 22.6 | | -0.18 | | Yes | | Yes | Yes |
| 49 ¹ | GC/MS EI | 18.3 | 10.2 | -1.32 | -0.89 | | No | Yes | Yes |
| 50 | GC/MSMS | 20.0 | 12.0 | -0.87 | -0.52 | Yes | Yes | No | |
| 53 | GC/MSMS | 17.9 | 8.90 | -1.43 | -1.07 | | Yes | Yes | Yes |
| 56 | GC/HRMS | 24.2 | 9.05 | 0.23 | 0.17 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 31.2 | | 2.06 | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 29.8 | 12.7 | 1.7 | 1.14 | Yes | No | Yes | No |
| 63 | GC/MS EI | 8.06 | 1.24 | -4.0 | -1.81 | No | No | Yes ⁴ | No |
| 65 | GC/HRMS | 5.52 | 2.43 | -4.66 | -6.47 | Yes | Yes | Yes ⁴ | Yes |
| 66 | GC/MSMS | 18.4 | 2 | -1.30 | -1.86 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 72.8 | | 12.93 | | | | | |
| 71 | GC/ECD ³ | 10.6 | 2.88 | -3.33 | -4.37 | Yes | | No | Yes |
| 72 ¹ | GC/MS EI | 33.2 | 0.89 | 2.58 | 3.93 | Yes | Yes | Yes | Yes |
| 73 | GC/ECD | 16.5 | | -1.79 | | No | No | No | Yes |
| 74 ¹ | GC/MSMS | 24.9 | | 0.40 | | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 58.7 | 5 | 9.24 | | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | 25.3 | | 0.52 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 25.1 | 5.02 | 0.46 | 0.50 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 34. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 105

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 3 | GC/MSMS | 4.50 | | -2.71 | | No | Yes | Yes | Yes |
| 10 | GC/MSMS | 8.90 | | 0.70 | | Yes | Yes | No | |
| 15 ¹ | GC/MSMS | 6.77 | 3.38 | -0.95 | -0.65 | Yes | Yes | Yes | Yes |
| 25 | GC/MSMS | 7.65 | 1.51 | -0.27 | -0.31 | Yes | | | Yes |
| 27 | GC/ECD | 8.18 | | 0.14 | | Yes | Yes | No | Yes |
| 36 | GC/MS EI | 4.82 | 0.96 | -2.5 | -3.34 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 10.3 | 0.60 | 1.76 | 2.60 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 8.30 | 4.98 | 0.24 | 0.12 | Yes | Yes | No | |
| 63 | GC/MS EI | 0.68 | 0.15 | -5.66 | -6.87 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 8.54 | 3.76 | 0.4 | 0.27 | Yes | Yes | Yes ³ | Yes |
| 66 | GC/MSMS | 5.16 | 1.14 | -2.19 | -2.84 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 71.3 | | 48.96 | | | | | |
| 72 ¹ | GC/MS EI | 14 | 0.39 | 4.65 | 7.14 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 8.01 | | 0.01 | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 9.07 | | 0.83 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 9.63 | 1.93 | 1.26 | 1.29 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 35. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 110

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 15.4 | | 1.18 | | Yes | Yes | No | |
| 14 | GC/MS EI | 9.40 | 4.20 | -1.21 | -1.05 | No | No | No | Yes |
| 25 | GC/MSMS | 9.41 | 0.55 | -1.20 | -1.52 | Yes | | | Yes |
| 36 | GC/MS EI | 7.95 | 1.37 | -1.78 | -2.14 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 20.4 | 7.10 | 3.17 | 1.94 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 14.7 | 8.82 | 0.9 | 0.47 | Yes | Yes | No | |
| 63 | GC/MS EI | 11.0 | 2.96 | -0.57 | -0.13 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 2.79 | 1.23 | -3.84 | -4.66 | Yes | Yes | Yes ³ | Yes |
| 66 | GC/MSMS | 11.3 | 2.00 | -0.47 | -0.53 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 18 | 0.57 | 2.2 | 2.78 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 14.8 | 2.97 | 0.95 | 0.97 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 36. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 118

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 2 ¹ | GC/ECD ² | 27.5 | 5.50 | -0.21 | -0.25 | No | Yes | No | Yes |
| 3 | GC/MSMS | 25.0 | | -0.77 | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 1.07 | | -6.10 | | Yes | No | No | Yes |
| 5 | GC/MSMS | 26.5 | | -0.43 | | | Yes | Yes | Yes |
| 7 | GC/ECD | 20.7 | 2.59 | -1.73 | -2.57 | No | No | Yes | Yes |
| 10 | GC/MSMS | 34.3 | | 1.3 | | Yes | Yes | No | |
| 15 ¹ | GC/MSMS | 26.5 | 12.2 | -0.44 | -0.28 | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 160 | | 29.21 | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | 22.5 | 2.03 | -1.33 | -2.04 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 29.9 | | 0.32 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 42.8 | 2.07 | 3.20 | 4.90 | Yes | No | No | Yes |
| 36 | GC/MS EI | 5.77 | 0.31 | -5.05 | -8.26 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 36.4 | 2.20 | 1.77 | 2.69 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 8.62 | | -4.42 | | No | No | No | No |
| 47 ¹ | GC/ECD | 18.1 | 7.24 | -2.30 | -2.27 | No | No | No | Yes |
| 48 | GC/ECD | 34.0 | | 1.24 | | Yes | | Yes | Yes |
| 49 ¹ | GC/MS EI | 26.4 | 11 | -0.46 | -0.33 | | No | Yes | Yes |
| 50 | GC/MSMS | 29.8 | 17.9 | 0.3 | 0.14 | Yes | Yes | No | |
| 56 | GC/HRMS | 42.7 | 15.9 | 3.18 | 1.70 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 35.7 | | 1.62 | | Yes | Yes | No | Yes |
| 63 | GC/MS EI | 9.78 | 0.72 | -4.16 | -1.84 | No | No | Yes ⁴ | No |
| 65 | GC/HRMS | 30.1 | 13.2 | 0.36 | 0.22 | Yes | Yes | Yes ⁴ | Yes |
| 66 | GC/MSMS | 20.6 | 2.50 | -1.74 | -2.62 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 165 | | 30.47 | | | | | |
| 72 ¹ | GC/MS EI | 35.4 | 1.10 | 1.55 | 2.49 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 29.2 | | 0.2 | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 31.4 | | 0.7 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 25.7 | 5.15 | -0.60 | -0.72 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 37. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 128

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 14 | GC/MS EI | 6.73 | 3.00 | -1.20 | -0.96 | No | No | No | Yes |
| 25 | GC/MSMS | 8.84 | 0.29 | 0.29 | 0.42 | Yes | | | Yes |
| 36 | GC/MS EI | 1.94 | 0.76 | -4.57 | -6.34 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 8.88 | 3.10 | 0.32 | 0.24 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 9.57 | 5.74 | 0.80 | 0.38 | Yes | Yes | No | |
| 63 | GC/MS EI | 7.48 | 1.09 | -0.7 | -0.13 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 2.63 | 1.16 | -4.09 | -5.21 | Yes | Yes | Yes ³ | Yes |
| 67 ¹ | GC/ECD | 2.99 | | -3.83 | | | | | |
| 72 ¹ | GC/MS EI | 14.6 | 1.10 | 4.35 | 5.61 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 13.8 | 2.76 | 3.8 | 3.21 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 38. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 138

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | 75.4 | 15.1 | 0.67 | 0.72 | No | Yes | No | Yes |
| 3 | GC/MSMS | 65.1 | | -0.29 | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 5.33 | | -5.83 | | Yes | No | No | Yes |
| 5 | GC/MSMS | 85.9 | | 1.65 | | | Yes | Yes | Yes |
| 8 | GC/ECD | 135 | 37.4 | 6.21 | 3.37 | No | | | |
| 10 | GC/MSMS | 64.2 | | -0.4 | | Yes | Yes | No | |
| 14 | GC/MS EI | 57.2 | 25.2 | -1.02 | -0.77 | No | No | No | Yes |
| 15 ¹ | GC/MSMS | 83.4 | 41.7 | 1.42 | 0.70 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 72.7 | 15.3 | 0.42 | 0.45 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 106 | | 3.5 | | | | Yes ⁴ | Yes |
| 24 | GC/MS EI | 51.2 | 3.58 | -1.57 | -2.57 | No | Yes | Yes | Yes |
| 25 | GC/MSMS | 58.8 | 9.9 | -0.86 | -1.13 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 64.2 | | -0.37 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 66.6 | 1.39 | -0.14 | -0.23 | Yes | No | No | Yes |
| 36 | GC/MS EI | 51.7 | 4.65 | -1.53 | -2.35 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 107 | 8 | 3.56 | 4.98 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 3.96 | | -5.96 | | No | No | No | No |
| 47 ¹ | GC/ECD | 36.8 | 14.7 | -2.9 | -3.17 | No | No | No | Yes |
| 49 ¹ | GC/MS EI | 64.7 | 11.2 | -0.32 | -0.40 | | No | Yes | Yes |
| 50 | GC/MSMS | 59.7 | 35.8 | -0.79 | -0.44 | Yes | Yes | No | |
| 53 | GC/MSMS | 82.3 | 41.2 | 1.32 | 0.65 | | Yes | Yes | Yes |
| 56 | GC/HRMS | 90 | 4.56 | 2.03 | 3.14 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 20.8 | | -4.40 | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 83.3 | 15.4 | 1.41 | 1.68 | Yes | No | Yes | No |
| 63 | GC/MS EI | 25.0 | 2.92 | -4.00 | -1.66 | No | No | Yes ⁴ | No |
| 65 | GC/HRMS | 12.2 | 5.37 | -5.2 | -7.87 | Yes | Yes | Yes ⁴ | Yes |
| 66 | GC/MSMS | 40.6 | 3.3 | -2.6 | -4.07 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 243 | | 16.19 | | | | | |
| 71 | GC/ECD ³ | 25.2 | 7.07 | -3.99 | -5.65 | Yes | | No | Yes |
| 72 ¹ | GC/MS EI | 97.8 | 3.20 | 2.75 | 4.38 | Yes | Yes | Yes | Yes |
| 73 | GC/ECD | 73.8 | | 0.52 | | No | No | No | Yes |
| 74 ¹ | GC/MSMS | 63 | | -0.48 | | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 164 | 30 | 8.87 | 5.83 | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | 74 | | 0.54 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 75.2 | 15 | 0.66 | 0.71 | Yes | Yes | Yes | Yes |

¹Laboratory accredited²GC/ECD and peak confirmation with dual column³GC/ECD and peak confirmation with GC/MS⁴Yes of a different matrix of the test sample

TABLE 39. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 146

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|-----------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 10 | GC/MSMS | 19 | | 0.29 | | Yes | Yes | No | |
| 25 | GC/MSMS | 86.7 | 10.25 | 24.58 | 12.77 | Yes | | | Yes |
| 36 | GC/MS EI | 17.3 | 0.48 | -0.34 | -0.59 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 19.0 | 3.3 | 0.29 | 0.45 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 17.6 | 10.53 | -0.24 | -0.12 | Yes | Yes | No | |
| 65 | GC/HRMS | 4.07 | 1.79 | -5.1 | -7.71 | Yes | Yes | Yes ² | Yes |
| 76 | GC/MS EI | 56 | 10.00 | 13.56 | 7.19 | No | Yes | | Yes |

¹ Laboratory accredited² Yes of a different matrix of the test sample

TABLE 40. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 149

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 10 | GC/MSMS | 26.9 | | -0.08 | | Yes | Yes | No | |
| 14 | GC/MS EI | 10.7 | 4.6 | -3.60 | -4.30 | No | No | No | Yes |
| 17 | GC/MSMS | 34.4 | 5.3 | 1.54 | 1.74 | Yes | Yes | Yes | No |
| 25 | GC/MSMS | 23.4 | 0.92 | -0.85 | -1.24 | Yes | | | Yes |
| 27 | GC/ECD | 26.1 | | -0.27 | | Yes | Yes | No | Yes |
| 36 | GC/MS EI | 17.7 | 0.56 | -2.1 | -3.07 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 35.0 | 12.2 | 1.66 | 1.12 | Yes | No | Yes | Yes |
| 48 | GC/ECD | 10.4 | | -3.65 | | Yes | | Yes | Yes |
| 49 ¹ | GC/MS EI | 22.7 | 11 | -0.99 | -0.72 | | No | Yes | Yes |
| 50 | GC/MSMS | 30.1 | 18 | 0.6 | 0.29 | Yes | Yes | No | |
| 63 | GC/MS EI | 11.2 | 1.01 | -3.49 | -1.39 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 6.71 | 2.95 | -4.46 | -5.98 | Yes | Yes | Yes ³ | Yes |
| 66 | GC/MSMS | 23.2 | 2.6 | -0.89 | -1.22 | No | No | No | Yes |
| 67 ¹ | GC/ECD | <0.15 | | | | | | | |
| 72 ¹ | GC/MS EI | 43 | 1.4 | 3.40 | 4.92 | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 88.0 | 16.0 | 13.14 | 7.07 | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | 32.6 | | 1.15 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 27 | 5.4 | -0.07 | -0.08 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 41. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 151

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 9.78 | | 0.28 | | Yes | Yes | No | |
| 25 | GC/MSMS | 7.54 | 1.07 | -1.01 | -1.26 | Yes | | | Yes |
| 36 | GC/MS EI | 6.31 | 0.57 | -1.73 | -2.28 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 10.6 | 3.80 | 0.77 | 0.58 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 11.4 | 6.83 | 1.22 | 0.58 | Yes | Yes | No | |
| 63 | GC/MS EI | 4.62 | 0.73 | -2.7 | -0.97 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 1.97 | 0.87 | -4.25 | -5.44 | Yes | Yes | Yes ³ | Yes |
| 72 ¹ | GC/MS EI | 15 | 0.49 | 3.31 | 4.41 | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 31.7 | 6 | 12.98 | 6.87 | No | Yes | | Yes |
| 79 | GC/ECD ² | 9.52 | 1.90 | 0.1 | 0.14 | Yes | Yes | Yes | Yes |

¹Laboratory accredited²GC/ECD and peak confirmation with GC/MS³Yes of a different matrix of the test sample

TABLE 42. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 153

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | 90.1 | 18 | 0.01 | 0.01 | No | Yes | No | Yes |
| 3 | GC/MSMS | 85.1 | | -0.34 | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 8.5 | | -5.74 | | Yes | No | No | Yes |
| 5 | GC/MSMS | 90.5 | | 0.04 | | | Yes | Yes | Yes |
| 7 | GC/ECD | 82.4 | 12 | -0.54 | -0.73 | No | No | Yes | Yes |
| 8 | GC/ECD | 102 | 27.8 | 0.8 | 0.72 | No | | | |
| 10 | GC/MSMS | 107 | | 1.17 | | Yes | Yes | No | |
| 14 | GC/MS EI | 75.6 | 33.2 | -1.02 | -0.77 | No | No | No | Yes |
| 15 ¹ | GC/MSMS | 112 | 55.8 | 1.52 | 0.74 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 120 | 18.9 | 2.1 | 2.37 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 209 | | 8.39 | | | | Yes ⁴ | Yes |
| 24 | GC/MS EI | 53.7 | 2.90 | -2.55 | -4.18 | No | Yes | Yes | Yes |
| 25 | GC/MSMS | 86.7 | 10.3 | -0.23 | -0.33 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 86.2 | | -0.27 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 94.0 | 2.10 | 0.28 | 0.46 | Yes | No | No | Yes |
| 36 | GC/MS EI | 68.0 | 3.89 | -1.55 | -2.48 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 109 | 7.00 | 1.31 | 1.99 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 8.7 | | -5.7 | | No | No | No | No |
| 47 ¹ | GC/ECD | 53.4 | 21.4 | -2.58 | -2.66 | No | No | No | Yes |
| 50 | GC/MSMS | 104 | 62.4 | 0.98 | 0.43 | Yes | Yes | No | |
| 53 | GC/MSMS | 84.3 | 42.2 | -0.40 | -0.25 | | Yes | Yes | Yes |
| 56 | GC/HRMS | 85.6 | 2.83 | -0.31 | -0.50 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 25.2 | | -4.56 | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 24.6 | 5.01 | -4.61 | -7.35 | Yes | No | Yes | No |
| 63 | GC/MS EI | 32.5 | 3.74 | -4.05 | -1.71 | No | No | Yes ⁴ | No |
| 65 | GC/HRMS | 22.2 | 9.75 | -4.8 | -6.82 | Yes | Yes | Yes ⁴ | Yes |
| 66 | GC/MSMS | 50.3 | 3.80 | -2.8 | -4.48 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 346 | | 18.04 | | | | | |
| 71 | GC/ECD ³ | 26.1 | 7.30 | -4.50 | -6.72 | Yes | | No | Yes |
| 72 ¹ | GC/MS EI | 130 | 4.20 | 2.81 | 4.48 | Yes | Yes | Yes | Yes |
| 73 | GC/ECD | 104 | | 0.97 | | No | No | No | Yes |
| 74 ¹ | GC/MSMS | 105 | | 1.05 | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 103 | | 0.89 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 88.8 | 17.8 | -0.09 | -0.10 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 43. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 156

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 3 | GC/MSMS | 2.63 | | -2.04 | | No | Yes | Yes | Yes |
| 10 | GC/MSMS | 3.98 | | 0.22 | | Yes | Yes | No | |
| 15 ¹ | GC/MSMS | 4.29 | 2.15 | 0.75 | 0.40 | Yes | Yes | Yes | Yes |
| 25 | GC/MSMS | 3.90 | 1.02 | 0.09 | 0.09 | Yes | | | Yes |
| 27 | GC/ECD | 3.54 | | -0.52 | | Yes | Yes | No | Yes |
| 36 | GC/MS EI | 3.44 | 0.74 | -0.7 | -0.79 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 4.27 | 0.30 | 0.71 | 1.15 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 3.97 | 2.38 | 0.21 | 0.10 | Yes | Yes | No | |
| 63 | GC/MS EI | 2.54 | 0.41 | -2.19 | -0.51 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 3.89 | 1.71 | 0.1 | 0.04 | Yes | Yes | Yes ³ | Yes |
| 66 | GC/MSMS | 0.20 | 0.07 | -6.13 | -10.36 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 6.28 | 0.30 | 4.08 | 6.37 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 3.87 | | 0.04 | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 4.23 | | 0.65 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 4.40 | 0.88 | 0.92 | 0.98 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 44. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 170

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 12.6 | | 0.36 | | Yes | Yes | No | |
| 17 | GC/MSMS | 17.7 | 3.38 | 2.98 | 2.76 | Yes | Yes | Yes | No |
| 25 | GC/MSMS | 11.8 | 2.93 | -0.06 | -0.06 | Yes | | | Yes |
| 27 | GC/ECD | 10.5 | | -0.72 | | Yes | Yes | No | Yes |
| 36 | GC/MS EI | 7.88 | 0.74 | -2.09 | -3.13 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 9.92 | 3.50 | -1.0 | -0.96 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 12.5 | 7.53 | 0.31 | 0.15 | Yes | Yes | No | |
| 63 | GC/MS EI | 5.41 | 0.80 | -3.36 | -1.18 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 2.51 | 1.11 | -4.85 | -6.93 | Yes | Yes | Yes ³ | Yes |
| 66 | GC/MSMS | 4.73 | 1.14 | -3.7 | -5.27 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 30.6 | | 9.61 | | | | | |
| 72 ¹ | GC/MS EI | 18.4 | 1.10 | 3.33 | 4.75 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 12.2 | | 0.13 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 13.4 | 2.68 | 0.76 | 0.80 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 45. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 174

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 7.97 | 2.80 | | | Yes | No | Yes | Yes |
| 79 | GC/ECD ² | 7.36 | 1.47 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS

TABLE 46. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 177

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 10.7 | | 1.62 | | Yes | Yes | No | |
| 25 | GC/MSMS | 6.52 | 0.86 | -1.05 | -1.30 | Yes | | | Yes |
| 36 | GC/MS EI | 5.34 | 0.21 | -1.81 | -2.37 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 9.92 | 3.50 | 1.12 | 0.85 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 7.92 | 4.75 | -0.16 | -0.09 | Yes | Yes | No | |
| 63 | GC/MS EI | 8.11 | 0.83 | 0.0 | -0.01 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 2.31 | 1.02 | -3.74 | -4.53 | Yes | Yes | Yes ³ | Yes |
| 72 ¹ | GC/MS EI | 14.2 | 0.89 | 3.86 | 4.76 | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 16.7 | 3 | 5.44 | | No | Yes | | Yes |
| 79 | GC/ECD ² | 8.63 | 1.73 | 0.3 | 0.32 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 47. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 180

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | 32.2 | 6.40 | -0.47 | -0.55 | No | Yes | No | Yes |
| 3 | GC/MSMS | 23.8 | | -2.00 | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 6.54 | | -5.12 | | Yes | No | No | Yes |
| 5 | GC/MSMS | 36.5 | | 0.31 | | | Yes | Yes | Yes |
| 7 | GC/ECD | 30.4 | 4.43 | -0.81 | -1.10 | No | No | Yes | Yes |
| 8 | GC/ECD | 43.3 | 9.06 | 1.5 | 1.50 | No | | | |
| 10 | GC/MSMS | 41.0 | | 1.13 | | Yes | Yes | No | |
| 15 ¹ | GC/MSMS | 37.0 | 18.5 | 0.39 | 0.22 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 51.5 | 8.76 | 3.03 | 3.01 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 56.7 | | 4.0 | | | | Yes ⁴ | Yes |
| 24 | GC/MS EI | 26.8 | 2.27 | -1.46 | -2.37 | No | Yes | Yes | Yes |
| 25 | GC/MSMS | 36.5 | 4.61 | 0.30 | 0.41 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 27.1 | | -1.39 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 51.6 | 2.04 | 3.03 | 4.71 | Yes | No | No | Yes |
| 36 | GC/MS EI | 22.2 | 4.49 | -2.29 | -3.09 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 38.3 | 2.40 | 0.64 | 0.98 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 0.45 | | -6.22 | | No | No | No | No |
| 47 ¹ | GC/ECD | 30.3 | 12.1 | -0.8 | -0.64 | No | No | No | Yes |
| 48 | GC/ECD | 45.5 | | 1.93 | | Yes | | Yes | Yes |
| 49 ¹ | GC/MS EI | 31.3 | 10.3 | -0.64 | -0.57 | | No | Yes | Yes |
| 50 | GC/MSMS | 41.6 | 25.0 | 1.23 | 0.52 | Yes | Yes | No | |
| 53 | GC/MSMS | 31.0 | 15.5 | -0.69 | -0.45 | | Yes | Yes | Yes |
| 56 | GC/HRMS | 41.5 | 5.60 | 1.20 | 1.51 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 8.70 | | -4.73 | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 1.11 | 0.42 | -6.10 | -9.88 | Yes | No | Yes | No |
| 63 | GC/MS EI | 18.0 | 1.39 | -3.0 | -0.92 | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 14.7 | 2.40 | -3.6 | -5.57 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 90 | | 9.99 | | | | | |
| 71 | GC/ECD ³ | 13.9 | 3.90 | -3.78 | -5.23 | Yes | | No | Yes |
| 72 ¹ | GC/MS EI | 46.2 | 2.20 | 2.06 | 3.18 | Yes | Yes | Yes | Yes |
| 73 | GC/ECD | 35.5 | | 0.13 | | No | No | No | Yes |
| 74 ¹ | GC/MSMS | 36.0 | | 0.22 | | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 140.7 | 26.0 | 19.16 | 7.88 | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | 40.1 | | 0.96 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 35.0 | 7 | 0.03 | 0.04 | Yes | Yes | Yes | Yes |

¹Laboratory accredited²GC/ECD and peak confirmation with dual column³GC/ECD and peak confirmation with GC/MS⁴Yes of a different matrix of the test sample

TABLE 48. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 183

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 9.70 | | 0.09 | | Yes | Yes | No | |
| 14 | GC/MS EI | 7.17 | 3.20 | -1.12 | -0.99 | No | No | No | Yes |
| 25 | GC/MSMS | 6.77 | 0.27 | -1.31 | -1.58 | Yes | | | Yes |
| 36 | GC/MS EI | 6.60 | 1.55 | -1.39 | -1.54 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 14.1 | 4.90 | 2.17 | 1.49 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 12.3 | 7.37 | 1.3 | 0.68 | Yes | Yes | No | |
| 63 | GC/MS EI | 4.11 | 0.59 | -2.57 | -1.21 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 2.37 | 1.04 | -3.41 | -3.96 | Yes | Yes | Yes ³ | Yes |
| 72 ¹ | GC/MS EI | 15.8 | 0.89 | 3.00 | 3.53 | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 32.3 | 30 | 10.9 | 1.51 | No | Yes | | Yes |
| 79 | GC/ECD ² | 9.99 | 2 | 0.23 | 0.24 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 49. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 187

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 31.0 | | 0.28 | | Yes | Yes | No | |
| 25 | GC/MSMS | 24.1 | 1.67 | -0.99 | -1.32 | Yes | | | Yes |
| 27 | GC/ECD | 28.7 | | -0.15 | | Yes | Yes | No | Yes |
| 36 | GC/MS EI | 23.0 | 2.90 | -1.19 | -1.53 | No | No | Yes | No |
| 40 ¹ | GC/HRMS | 45.7 | 16 | 2.98 | 1.83 | Yes | No | Yes | Yes |
| 50 | GC/MSMS | 38.1 | 22.9 | 1.6 | 0.71 | Yes | Yes | No | |
| 63 | GC/MS EI | 11.3 | 0.49 | -3.35 | -1.51 | No | No | Yes ³ | No |
| 65 | GC/HRMS | 8.16 | 3.59 | -3.93 | -4.89 | Yes | Yes | Yes ³ | Yes |
| 67 ¹ | GC/ECD | 8.64 | | -3.84 | | | | | |
| 72 ¹ | GC/MS EI | 52.4 | 2.30 | 4.2 | 5.55 | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 122 | 22 | 16.99 | 7.88 | No | Yes | | Yes |
| 79 | GC/ECD ² | 32.1 | 6.42 | 0.48 | 0.51 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 50. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 194

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 17 | GC/MSMS | 4.80 | 1.36 | 0.64 | 0.57 | Yes | Yes | Yes | No |
| 27 | GC/ECD | 3.63 | | -0.86 | | Yes | Yes | No | Yes |
| 40 ¹ | GC/HRMS | 3.36 | 1.20 | -1.20 | -1.13 | Yes | No | Yes | Yes |
| 48 | GC/ECD | 5.41 | | 1.41 | | Yes | | Yes | Yes |
| 49 ¹ | GC/MS EI | 3.49 | 10.8 | -1.0 | -0.15 | | No | Yes | Yes |
| 63 | GC/MS EI | 2.34 | 0.24 | -2.51 | -0.81 | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 1.47 | 0.32 | -3.61 | -4.79 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 5.05 | | 0.96 | | | | | |
| 72 ¹ | GC/MS EI | 5.88 | 0.36 | 2.0 | 2.65 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 4.37 | | 0.09 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 4.33 | 0.87 | 0.04 | 0.05 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 51. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 195

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 1.13 | 0.40 | -1.30 | -1.30 | Yes | No | Yes | Yes |
| 63 | GC/MS EI | 1.72 | 0.07 | 0.28 | 0.06 | No | No | Yes ³ | No |
| 72 ¹ | GC/MS EI | 2.10 | 0.14 | 1.30 | 1.50 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 1.51 | 0.30 | -0.28 | -0.30 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 52. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 201

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 4.03 | 1.40 | | | Yes | No | Yes | Yes |
| 63 | GC/MS EI | 0.95 | 0.15 | | | No | No | Yes ³ | No |
| 79 | GC/ECD ² | 3.14 | 0.63 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 53. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 206

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 3.28 | 1.10 | -0.06 | -0.05 | Yes | No | Yes | Yes |
| 63 | GC/MS EI | 0.85 | 0.15 | -4.81 | -2.74 | No | No | Yes ³ | No |
| 67 ¹ | GC/ECD | <0.15 | | | | | | | |
| 72 ¹ | GC/MS EI | 4.52 | 0.75 | 2.35 | 2.48 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 3.16 | 0.63 | -0.30 | -0.35 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 54. DATA REPORTED AND PERFORMANCE EVALUATION FOR PCB 209

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | 2 | 0.58 | 0.20 | 0.18 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 110 | | 353.36 | | | | Yes ⁴ | Yes |
| 40 ¹ | GC/HRMS | 2.11 | 0.13 | 0.55 | 0.84 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 0.04 | | -6.24 | | No | No | No | No |
| 63 | GC/MS EI | 1.32 | 0.22 | -2.04 | -0.47 | No | No | Yes ³ | No |
| 67 ¹ | GC/ECD | <0.15 | | | | | | | |
| 72 ¹ | GC/MS EI | 3.06 | 0.23 | 3.68 | 5.23 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 1.67 | | -0.89 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 2.01 | 0.40 | 0.22 | 0.25 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample⁴ Yes of a different matrix of the test sample

I.2. RESULTS REPORTED FOR ORGANOCHLORINE PESTICIDES

TABLE 55. DATA REPORTED AND PERFORMANCE EVALUATION FOR HCB

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 1.98 | | -1.91 | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 0.39 | | -5.26 | | Yes | No | No | Yes |
| 7 | GC/ECD | 2.77 | 0.58 | -0.25 | -0.28 | No | No | Yes | Yes |
| 10 | GC/MSMS | 3.72 | | 1.74 | | Yes | Yes | No | |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | 3.35 | 1.68 | 0.97 | 0.51 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 0.37 | 0.13 | -5.30 | -7.97 | Yes | Yes | Yes | No |
| 24 | GC/MS EI | 1.89 | | -2.10 | | No | Yes | Yes | Yes |
| 25 | GC/MSMS | 2.59 | 0.57 | -0.6 | -0.71 | Yes | | | Yes |
| 27 | GC/ECD | 2.13 | | -1.61 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 7.73 | 0.83 | 10.2 | 9.28 | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 3 | 0.19 | 0.22 | 0.33 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 2.08 | | -1.70 | | Yes | Yes | Yes ⁴ | Yes |
| 50 | GC/MSMS | 2.99 | 1.80 | 0.22 | 0.11 | Yes | Yes | No | |
| 53 | GC/MSMS | 3.18 | 1.60 | 0.60 | 0.33 | | Yes | Yes | Yes |
| 54 | GC/MSMS | 3.48 | 1.80 | 1.25 | 0.62 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | 26 | | 48.6 | | No | No | No | No |
| 56 | GC/HRMS | 3.83 | 28 | 2.0 | 0.07 | Yes | No | Yes | Yes |
| 62 | GC/MS EI | 0.19 | 0.06 | -5.68 | -8.71 | Yes | No | Yes | No |
| 66 | GC/MSMS | 2.16 | 0.53 | -1.54 | -1.82 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 105 | | 214 | | | | | |
| 72 ¹ | GC/MS EI | 3.30 | 0.14 | 0.86 | 1.29 | Yes | Yes | Yes | Yes |
| 75 | GC/MS EI | 3.16 | | 0.57 | | | | | |
| 76 | GC/MS EI | 7.33 | | 9.34 | | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | 3.13 | | 0.51 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 1.13 | 0.23 | -3.7 | -5.37 | Yes | Yes | Yes | Yes |

¹Laboratory accredited

²GC/ECD and peak confirmation with dual column

³GC/ECD and peak confirmation with GC/MS

⁴Yes of a different matrix of the test sample

TABLE 56. DATA REPORTED FOR α -HCH

| Lab code | Instrument Type | X_{lab} $\mu\text{g kg}^{-1}$ dw | $U_{lab} (k=2)$ $\mu\text{g kg}^{-1}$ dw | z - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|---------------------------------------|---|----------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 5 | GC/MSMS | 0.02 | | | | | Yes | Yes | Yes |
| 7 | GC/ECD | 1.58 | 0.39 | | | No | No | Yes | Yes |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | <0.01 | | | | Yes | Yes | Yes | No |
| 23 | GC/ECD | 7.28 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 0.19 | | | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 0.20 | 0.02 | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | <0.15 | | | | Yes | No | Yes | Yes |
| 42 | GC/MSMS | <0.08 | | | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 0.32 | | | | No | No | No | No |
| 54 | GC/MSMS | <1.7 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 0.23 | 29.1 | | | Yes | No | Yes | Yes |
| 66 | GC/MSMS | 0.18 | 0.05 | | | No | No | No | Yes |
| 67 ¹ | GC/ECD | 41.1 | | | | | | | |
| 72 ¹ | GC/MS EI | 0.42 | 0.08 | | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | <2 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | <0.01 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 57. DATA REPORTED FOR β -HCH

| Lab code | Instrument Type | X_{lab} $\mu\text{g kg}^{-1}$ dw | $U_{lab} (k=2)$ $\mu\text{g kg}^{-1}$ dw | z - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|---------------------------------------|---|----------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 0.05 | | | | No | Yes | Yes | Yes |
| 5 | GC/MSMS | 0.17 | | | | | Yes | Yes | Yes |
| 7 | GC/ECD | 1.07 | 0.25 | | | No | No | Yes | Yes |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 0.04 | 0.02 | | | Yes | Yes | Yes | No |
| 23 | GC/ECD | 28.3 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 35 | GC/MS EI | 0.43 | 0.06 | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 0.24 | 0.02 | | | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 0.13 | | | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 0.56 | | | | No | No | No | No |
| 54 | GC/MSMS | <1.7 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 0.28 | 16.1 | | | Yes | No | Yes | Yes |
| 66 | GC/MSMS | 1.19 | 0.40 | | | No | No | No | Yes |
| 67 ¹ | GC/ECD | 5.92 | | | | | | | |
| 72 ¹ | GC/MS EI | <0.30 | | | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | <2 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 0.64 | 0.13 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 58. DATA REPORTED FOR γ -HCH (LINDANE)

| Lab code | Instrument Type | X _{lab} $\mu\text{g kg}^{-1}$ dw | U _{lab} ($k=2$) $\mu\text{g kg}^{-1}$ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 0.18 | | | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 0.10 | | | | Yes | No | No | Yes |
| 5 | GC/MSMS | 0.09 | | | | | Yes | Yes | Yes |
| 7 | GC/ECD | 0.69 | 0.29 | | | No | No | Yes | Yes |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 0.06 | 0.06 | | | Yes | Yes | Yes | No |
| 23 | GC/ECD | 4.78 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 0.30 | | | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 1.17 | 0.17 | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 0.12 | 0.01 | | | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 0.12 | | | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 0.79 | | | | No | No | No | No |
| 54 | GC/MSMS | <1.7 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 0.63 | 13.1 | | | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 0.54 | | | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 0.71 | 0.18 | | | Yes | No | Yes | No |
| 63 | GC/MS EI | 2.53 | 0.22 | | | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 2.26 | 0.63 | | | No | No | No | Yes |
| 67 ¹ | GC/ECD | 44.8 | | | | | | | |
| 72 ¹ | GC/MS EI | <0.30 | | | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | <2 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 0.10 | 0.02 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 59. DATA REPORTED FOR δ -HCH

| Lab code | Instrument Type | X _{lab} $\mu\text{g kg}^{-1}$ dw | U _{lab} ($k=2$) $\mu\text{g kg}^{-1}$ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 4 ¹ | GC/MS EI | 0.10 | | | | Yes | No | No | Yes |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | <0.01 | | | | Yes | Yes | Yes | No |
| 23 | GC/ECD | 0.88 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 35 | GC/MS EI | 0.28 | | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | <0.15 | | | | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 0.84 | | | | No | No | No | No |
| 54 | GC/MSMS | <1.7 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 0.52 | | | | Yes | No | Yes | Yes |
| 66 | GC/MSMS | 0.19 | | | | No | No | No | Yes |
| 78 ¹ | GC/MSMS | <2 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 0.10 | 0.02 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 60. DATA REPORTED AND PERFORMANCE EVALUATION FOR *pp'*-DDD

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 1 | GC/MS EI | 0.91 | 0.21 | -5.92 | -9.73 | Yes | No | No | Yes |
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 7.03 | | -2.96 | | No | Yes | Yes | Yes |
| 5 | GC/MSMS | 13.4 | | 0.11 | | | Yes | Yes | Yes |
| 7 | GC/ECD | 7.85 | 2.17 | -2.56 | -3.19 | No | No | Yes | Yes |
| 14 | GC/MS EI | 38.7 | 17 | 12.4 | 2.97 | No | No | No | Yes |
| 17 | GC/MSMS | 13.7 | 3.60 | 0.25 | 0.23 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 42.8 | | 14.35 | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | 17.7 | 1.63 | 2.21 | 3.06 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 11.6 | | -0.73 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 25.3 | 1.16 | 5.90 | 8.82 | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 12.2 | 0.80 | -0.48 | -0.76 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 11.5 | | -0.80 | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 2.92 | | -4.95 | | No | No | No | No |
| 50 | GC/MSMS | 13.2 | 7.9 | 0.01 | 0.01 | Yes | Yes | No | |
| 54 | GC/MSMS | 21.8 | 10.9 | 4.20 | 1.54 | Yes | Yes | No | Yes |
| 56 | GC/HRMS | 24.9 | 8.33 | 5.67 | 2.69 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 12.9 | | -0.1 | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 5.26 | 1.36 | -3.81 | -5.77 | Yes | No | Yes | No |
| 63 | GC/MS EI | 2.08 | 0.42 | -5.36 | -4.56 | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 14.3 | 2.80 | 0.54 | 0.59 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 318 | | 147 | | | | | |
| 71 | GC/ECD ³ | 6.68 | 1.87 | -3.13 | -4.04 | Yes | | No | Yes |
| 72 ¹ | GC/MS EI | 12 | 0.49 | -0.56 | -0.90 | Yes | Yes | Yes | Yes |
| 75 | GC/MS EI | 19.2 | | 2.92 | | | | | |
| 76 | GC/MS EI | 48 | 10 | 16.9 | 6.76 | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | 17 | | 1.9 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 11.6 | 3.21 | -0.74 | -0.75 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 61 DATA REPORTED AND PERFORMANCE EVALUATION FOR *pp'*-DDE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 1 | GC/MS EI | 3.58 | 0.60 | -5.92 | -9.13 | Yes | No | No | Yes |
| 2 ¹ | GC/ECD ² | 108 | 21.6 | -0.68 | -0.78 | No | Yes | No | Yes |
| 3 | GC/MSMS | 120 | | -0.08 | | No | Yes | Yes | Yes |
| 5 | GC/MSMS | 99.8 | | -1.09 | | | Yes | Yes | Yes |
| 7 | GC/ECD | 113 | 25.9 | -0.43 | -0.47 | No | No | Yes | Yes |
| 14 | GC/MS EI | 229 | 91.8 | 5.4 | 2.25 | No | No | No | Yes |
| 15 ¹ | GC/MSMS | 157 | 78.7 | 1.80 | 0.87 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 170 | 33.5 | 2.42 | 2.28 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 5.2 | | -5.83 | | | | Yes ⁴ | Yes |
| 24 | GC/MS EI | 82 | | -2.0 | | No | Yes | Yes | Yes |
| 25 | GC/MSMS | 163 | 7.97 | 2.06 | 3.04 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 101 | | -1.01 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 107 | 5.5 | -0.75 | -1.13 | Yes | No | No | Yes |
| 37 ¹ | GC/MS EI | 98.7 | | -1.14 | | | | | |
| 40 ¹ | GC/HRMS | 145 | 8.9 | 1.17 | 1.71 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 138 | | 0.84 | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 4.9 | | -5.85 | | No | No | No | No |
| 50 | GC/MSMS | 137 | 82.5 | 0.8 | 0.37 | Yes | Yes | No | |
| 53 | GC/MSMS | 111 | 56 | -0.53 | -0.34 | | Yes | Yes | Yes |
| 54 | GC/MSMS | 160 | 79.8 | 1.91 | 0.91 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | 92 | | -1.48 | | No | No | No | No |
| 56 | GC/HRMS | 170 | 16.9 | 2.41 | 3.12 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 78.3 | | -2.17 | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 11.6 | 1.97 | -5.51 | -8.50 | Yes | No | Yes | No |
| 63 | GC/MS EI | 15.3 | 2.72 | -5.33 | -5.30 | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 84.1 | 11.8 | -1.9 | -2.64 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 3251 | | 157.1 | | | | | |
| 71 | GC/ECD ³ | 81.3 | 22.6 | -2.02 | -2.27 | Yes | | No | Yes |
| 72 ¹ | GC/MS EI | 158 | 8.9 | 1.83 | 2.67 | Yes | Yes | Yes | Yes |
| 75 | GC/MS EI | 104 | | -0.89 | | | | | |
| 76 | GC/MS EI | 491 | 46 | 18.54 | 14.01 | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | 143 | | 1.08 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 117 | 23.4 | -0.23 | -0.27 | Yes | Yes | Yes | Yes |

¹Laboratory accredited²GC/ECD and peak confirmation with dual column³GC/ECD and peak confirmation with GC/MS⁴Yes of a different matrix of the test sample

TABLE 62. DATA REPORTED AND PERFORMANCE EVALUATION FOR *pp'*-DDT

| Lab code | Instrument Type | X _{lab} μg kg ⁻¹ dw | U _{lab} (k=2) μg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 1 | GC/MS EI | 2.44 | 0.50 | -5.28 | -8.04 | Yes | No | No | Yes |
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 12.1 | | -2.04 | | No | Yes | Yes | Yes |
| 5 | GC/MSMS | 19.4 | | 0.39 | | | Yes | Yes | Yes |
| 10 | GC/MSMS | 18.2 | | -0.02 | | Yes | Yes | No | |
| 17 | GC/MSMS | 19.8 | 2.98 | 0.5 | 0.63 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 40.8 | | 7.53 | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | 23.8 | 2.94 | 1.87 | 2.30 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 11.0 | | -2.4 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 24.7 | 1.62 | 2.17 | 3.10 | Yes | No | No | Yes |
| 37 ¹ | GC/MS EI | 20.3 | | 0.70 | | | | | |
| 40 ¹ | GC/HRMS | 17.5 | 0.12 | -0.24 | -0.36 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 24.3 | | 2.04 | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 1.38 | | -5.63 | | No | No | No | No |
| 50 | GC/MSMS | 16.9 | 10.2 | -0.43 | -0.24 | Yes | Yes | No | |
| 54 | GC/MSMS | 21.7 | 10.9 | 1.16 | 0.61 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 40.2 | 6.16 | 7.3 | 6.03 | Yes | No | Yes | Yes |
| 62 | GC/MS EI | 9.48 | 2.26 | -2.92 | -4.08 | Yes | No | Yes | No |
| 63 | GC/MS EI | 2.81 | 0.54 | -5.15 | -4.43 | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 20 | 3.20 | 0.60 | 0.71 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 441 | | 141.21 | | | | | |
| 71 | GC/ECD ³ | 13.7 | 3.85 | -1.50 | -1.59 | Yes | | No | Yes |
| 72 ¹ | GC/MS EI | 17.4 | 1.10 | -0.28 | -0.41 | Yes | Yes | Yes | Yes |
| 75 | GC/MS EI | 19.2 | | 0.32 | | | | | |
| 76 | GC/MS EI | 15.0 | 4 | -1.1 | -1.16 | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | 21.7 | | 1.1 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 14.4 | 2.88 | -1.27 | -1.58 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 63. DATA REPORTED AND PERFORMANCE EVALUATION FOR *op'*-DDE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 7 | GC/ECD | 2.52 | 0.37 | 0.82 | 1.00 | No | No | Yes | Yes |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | 2.01 | 1.01 | -0.61 | -0.40 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 2.31 | 0.26 | 0.22 | 0.31 | Yes | Yes | Yes | No |
| 25 | GC/MSMS | 2.19 | 0.37 | -0.1 | -0.14 | Yes | | | Yes |
| 35 | GC/MS EI | 2.60 | 0.22 | 1.04 | 1.50 | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 1.96 | 0.12 | -0.76 | -1.17 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 1.30 | | -2.60 | | Yes | Yes | Yes ⁴ | Yes |
| 50 | GC/MSMS | 2.10 | 1.26 | -0.4 | -0.20 | Yes | Yes | No | |
| 54 | GC/MSMS | 2.32 | 1.20 | 0.25 | 0.14 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 0.10 | 28.5 | -5.97 | -0.15 | Yes | No | Yes | Yes |
| 66 | GC/MSMS | 2.42 | 0.44 | 0.53 | 0.61 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 2.94 | 0.11 | 1.99 | 3.10 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 1.83 | | -1.11 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 5.76 | 1.15 | 9.89 | 5.68 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 64. DATA REPORTED AND PERFORMANCE EVALUATION FOR *op'*-DDD

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 7 | GC/ECD | 11.6 | 3.86 | 16.7 | 4.38 | No | No | Yes | Yes |
| 15 ¹ | GC/MSMS | 2.56 | 1.28 | -0.76 | -0.53 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 4.33 | 0.75 | 2.67 | 2.66 | Yes | Yes | Yes | No |
| 25 | GC/MSMS | 3.34 | 0.67 | 0.74 | 0.79 | Yes | | | Yes |
| 35 | GC/MS EI | 3.43 | 0.22 | 0.9 | 1.28 | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 2.14 | 0.80 | -1.59 | -1.52 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 1.53 | | -2.75 | | Yes | Yes | Yes ⁴ | Yes |
| 50 | GC/MSMS | 1.94 | 1.16 | -1.97 | -1.48 | Yes | Yes | No | |
| 54 | GC/MSMS | 3.65 | 1.90 | 1.4 | 0.72 | Yes | Yes | No | Yes |
| 56 | GC/HRMS | 3.32 | 7.85 | 0.70 | 0.09 | Yes | No | Yes | Yes |
| 66 | GC/MSMS | 5.38 | 0.74 | 4.71 | 4.71 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 3 | 0.14 | 0.09 | 0.13 | Yes | Yes | Yes | Yes |
| 75 | GC/MS EI | 6.34 | | 6.57 | | | | | |
| 78 ¹ | GC/MSMS | 2.77 | | -0.36 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 2.29 | 0.46 | -1.29 | -1.56 | Yes | Yes | Yes | Yes |

¹Laboratory accredited²GC/ECD and peak confirmation with dual column³GC/ECD and peak confirmation with GC/MS⁴Yes of a different matrix of the test sampleTABLE 65. DATA REPORTED AND PERFORMANCE EVALUATION FOR *op'*-DDT

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 7 | GC/ECD | 20.4 | 3.03 | 10.32 | 7.48 | No | No | Yes | Yes |
| 10 | GC/MSMS | 6.97 | | -0.41 | | Yes | Yes | No | |
| 17 | GC/MSMS | 7.53 | 0.83 | 0.04 | 0.05 | Yes | Yes | Yes | No |
| 25 | GC/MSMS | 7.05 | 0.91 | -0.34 | -0.45 | Yes | | | Yes |
| 27 | GC/ECD | 4.02 | | -2.8 | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | 9.73 | 0.60 | 1.81 | 2.56 | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 6.55 | 0.40 | -0.74 | -1.09 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 8.30 | | 0.66 | | Yes | Yes | Yes ⁴ | Yes |
| 50 | GC/MSMS | 6.32 | 3.79 | -0.9 | -0.56 | Yes | Yes | No | |
| 54 | GC/MSMS | 10.2 | 5.10 | 2.19 | 1.00 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 16.3 | 2.49 | 7.03 | 5.85 | Yes | No | Yes | Yes |
| 66 | GC/MSMS | 10.5 | 2.32 | 2.45 | 2.14 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 9.02 | 0.24 | 1.24 | 1.84 | Yes | Yes | Yes | Yes |
| 75 | GC/MS EI | 6.34 | | -0.91 | | | | | |
| 78 ¹ | GC/MSMS | 8.03 | | 0.45 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 4.23 | 0.85 | -2.60 | -3.50 | Yes | Yes | Yes | Yes |

¹Laboratory accredited²GC/ECD and peak confirmation with dual column³GC/ECD and peak confirmation with GC/MS⁴Yes of a different matrix of the test sample

TABLE 66. DATA REPORTED AND PERFORMANCE EVALUATION FOR HEPTACHLOR

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 4 ¹ | GC/MS EI | 0.15 | | | | Yes | No | No | Yes |
| 7 | GC/ECD | 0.97 | 0.37 | | | No | No | Yes | Yes |
| 10 | GC/MSMS | 0.06 | | | | Yes | Yes | No | |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 9.73 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 35 | GC/MS EI | 0.11 | 0.02 | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | <0.16 | | | | Yes | No | Yes | Yes |
| 42 | GC/MSMS | <0.08 | | | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 2.16 | | | | No | No | No | No |
| 54 | GC/MSMS | <1.7 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 0.49 | 4.00 | | | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 0.23 | | | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 0.20 | 0.07 | | | Yes | No | Yes | No |
| 63 | GC/MS EI | 1.62 | 0.27 | | | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 0.40 | 0.13 | | | No | No | No | Yes |
| 67 ¹ | GC/ECD | 93.03 | | | | | | | |
| 72 ¹ | GC/MS EI | <0.30 | | | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | <1 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | <0.01 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 67. DATA REPORTED FOR ALDRIN

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 0.08 | | | | No | Yes | Yes | Yes |
| 7 | GC/ECD | 0.73 | 0.18 | | | No | No | Yes | Yes |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 3.97 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 27 | GC/ECD | 0.28 | | | | Yes | Yes | No | Yes |
| 35 | GC/MS EI | <0.05 | | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | <0.06 | | | | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 2.16 | | | | No | No | No | No |
| 54 | GC/MSMS | <1.7 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 0.33 | 29.9 | | | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 0.87 | | | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 0.61 | 0.11 | | | Yes | No | Yes | No |
| 63 | GC/MS EI | 1.25 | 0.38 | | | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 0.18 | 0.06 | | | No | No | No | Yes |
| 67 ¹ | GC/ECD | 198 | | | | | | | |
| 72 ¹ | GC/MS EI | <0.30 | | | | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 487 | 96 | | | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | <12.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | <0.09 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 68. DATA REPORTED AND PERFORMANCE EVALUATION FOR DIELDRIN

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 1 | GC/MS EI | 8.28 | 1.50 | 2.88 | 2.31 | Yes | No | No | Yes |
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 4.78 | | -0.48 | | No | Yes | Yes | Yes |
| 7 | GC/ECD | 3.91 | 1.32 | -1.31 | -1.31 | No | No | Yes | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 5.80 | | 0.5 | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | 6.93 | 1.65 | 1.59 | 1.44 | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 35 | GC/MS EI | 0.18 | 0.12 | -4.89 | -6.29 | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 6.92 | 0.50 | 1.6 | 1.98 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 4.52 | | -0.72 | | No | No | No | No |
| 47 ¹ | GC/ECD | 49.3 | 16.8 | 42.25 | 5.23 | No | No | No | Yes |
| 50 | GC/MSMS | 5.29 | 3.17 | 0.01 | 0.01 | Yes | Yes | No | |
| 54 | GC/MSMS | <3.3 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 2.19 | 17.50 | -2.96 | -0.35 | Yes | No | Yes | Yes |
| 62 | GC/MS EI | 11.3 | 2.44 | 5.79 | 4.75 | Yes | No | Yes | No |
| 63 | GC/MS EI | 2.11 | 0.89 | -3.04 | -1.40 | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 3.03 | 0.34 | -2.2 | -2.72 | No | No | No | Yes |
| 67 ¹ | GC/ECD | 153 | | 141 | | | | | |
| 72 ¹ | GC/MS EI | 5.38 | 0.40 | 0.10 | 0.12 | Yes | Yes | Yes | Yes |
| 76 | GC/MS EI | 201 | 40 | 187 | 9.76 | No | Yes | | Yes |
| 78 ¹ | GC/MSMS | <12.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 3.28 | 0.66 | -1.92 | -2.29 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 69. DATA REPORTED FOR ENDRIN

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 3 | GC/MSMS | 0.05 | | | | No | Yes | Yes | Yes |
| 4 ¹ | GC/MS EI | 0.58 | | | | Yes | No | No | Yes |
| 7 | GC/ECD | 2.92 | 1.27 | | | No | No | Yes | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 5.42 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | <0.15 | | | | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 3.67 | | | | No | No | No | No |
| 47 ¹ | GC/ECD | 7.88 | 2.68 | | | No | No | No | Yes |
| 54 | GC/MSMS | <3.3 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 1.31 | 30.7 | | | Yes | No | Yes | Yes |
| 62 | GC/MS EI | 2.26 | 0.89 | | | Yes | No | Yes | No |
| 63 | GC/MS EI | 3.86 | 0.39 | | | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 0.86 | 0.26 | | | No | No | No | Yes |
| 78 ¹ | GC/MSMS | <12.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | <0.50 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sampleTABLE 70. DATA REPORTED AND PERFORMANCE EVALUATION FOR *cis*-CHLORDANE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | 6.35 | 3.17 | -0.15 | -0.07 | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 2.83 | | -3.01 | | | | Yes ⁴ | Yes |
| 27 | GC/ECD | 5.24 | | -1.04 | | Yes | Yes | No | Yes |
| 40 ¹ | GC/HRMS | 7.46 | 0.50 | 0.8 | 0.98 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 3.92 | | -2.12 | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 0.67 | | -4.75 | | No | No | No | No |
| 54 | GC/MSMS | 10.6 | 5.30 | 3.27 | 1.41 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | 55 | | 39.3 | | No | No | No | No |
| 56 | GC/HRMS | 8.08 | 30.7 | 1.26 | 0.10 | Yes | No | Yes | Yes |
| 66 | GC/MSMS | 5.43 | 0.58 | -0.89 | -1.14 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 8.38 | 0.42 | 1.50 | 1.95 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 8.40 | | 1.51 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 5.42 | 1.08 | -0.90 | -1.03 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 71. DATA REPORTED AND PERFORMANCE EVALUATION FOR *trans*-CHLORDANE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 14 | GC/MS EI | <10 | | | | No | No | No | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 4.78 | | 71.3 | | | | Yes ⁴ | Yes |
| 27 | GC/ECD | <0.05 | | | | Yes | Yes | No | Yes |
| 40 ¹ | GC/HRMS | 0.33 | 0.03 | -0.9 | -1.30 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 5.43 | | 81.8 | | Yes | Yes | Yes ⁴ | Yes |
| 45 ¹ | GC/ECD | 0.36 | | -0.27 | | No | No | No | No |
| 54 | GC/MSMS | <3.3 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | 17 | | 269 | | No | No | No | No |
| 56 | GC/HRMS | 0.23 | 31.6 | -2.40 | -0.01 | Yes | No | Yes | Yes |
| 63 | GC/MS EI | 0.34 | 0.06 | -0.65 | -0.12 | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 1.51 | 0.36 | 18.3 | 6.14 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 0.39 | 0.02 | 0.22 | 0.33 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | 7.60 | | 117 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 0.42 | 0.08 | 0.70 | 0.77 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sampleTABLE 72. DATA REPORTED AND PERFORMANCE EVALUATION FOR *cis*-NONACHLOR

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 11.3 | 0.7 | 0.01 | 0.03 | Yes | No | Yes | Yes |
| 54 | GC/MSMS | 11.2 | 5.60 | -0.05 | -0.03 | Yes | Yes | No | Yes |
| 72 ¹ | GC/MS EI | 13.8 | 0.9 | 1.50 | 2.48 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 11.3 | 2.25 | -0.02 | -0.03 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MSTABLE 73. DATA REPORTED AND PERFORMANCE EVALUATION FOR *trans*-NONACHLOR

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 40 ¹ | GC/HRMS | 22.1 | 1.40 | 0.45 | 0.62 | Yes | No | Yes | Yes |
| 42 | GC/MSMS | 14.6 | | -1.67 | | Yes | Yes | Yes ³ | Yes |
| 54 | GC/MSMS | 20.2 | 10.1 | -0.10 | -0.06 | Yes | Yes | No | Yes |
| 72 ¹ | GC/MS EI | 24.2 | 0.89 | 1.03 | 1.46 | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | 20 | 4 | -0.15 | -0.17 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 74. DATA REPORTED FOR α -ENDOSULFAN

| Lab code | Instrument Type | X_{lab} $\mu\text{g kg}^{-1}$ dw | $U_{lab} (k=2)$ $\mu\text{g kg}^{-1}$ dw | z - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|---------------------------------------|---|----------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 7 | GC/ECD | 2.19 | 0.50 | | | No | No | Yes | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 7.2 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 35 | GC/MS EI | 0.09 | 0.02 | | | Yes | No | No | Yes |
| 37 ¹ | GC/MS EI | 109 | | | | | | | |
| 40 ¹ | GC/HRMS | <0.5 | | | | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 6.75 | | | | No | No | No | No |
| 54 | GC/MSMS | <3.3 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 1.61 | 29.3 | | | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 2.3 | | | | Yes | Yes | No | Yes |
| 62 | GC/MS EI | 1.23 | 0.47 | | | Yes | No | Yes | No |
| 63 | GC/MS EI | 1.12 | 0.33 | | | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 1.17 | 0.22 | | | No | No | No | Yes |
| 72 ¹ | GC/MS EI | <0.30 | | | | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | <12.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | <0.52 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 75. DATA REPORTED FOR β -ENDOSULFAN

| Lab code | Instrument Type | X_{lab} $\mu\text{g kg}^{-1} \text{ dw}$ | $U_{lab} (k=2)$ $\mu\text{g kg}^{-1} \text{ dw}$ | z - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|---|---|----------------|----------------|-----------|----------------------|-----------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 6.23 | | | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 35 | GC/MS EI | <0.06 | | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | <1 | | | | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 6.26 | | | | No | No | No | No |
| 54 | GC/MSMS | <3.3 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 1.67 | 28 | | | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 1.41 | | | | Yes | Yes | No | Yes |
| 63 | GC/MS EI | 1.49 | 0.21 | | | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 36.2 | 5.51 | | | No | No | No | Yes |
| 78 ¹ | GC/MSMS | <12.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | <0.09 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

TABLE 76. DATA REPORTED FOR ENDOSULFAN SULFATE

| Lab code | Instrument Type | X_{lab} $\mu\text{g kg}^{-1} \text{ dw}$ | $U_{lab} (k=2)$ $\mu\text{g kg}^{-1} \text{ dw}$ | z - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|---|---|----------------|----------------|-----------|----------------------|-----------------------|-------|
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 23 | GC/ECD | 8.6 | | | | | | Yes ³ | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 45 ¹ | GC/ECD | 2.28 | | | | No | No | No | No |
| 54 | GC/MSMS | <3.3 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 57 | GC/MSMS | 0.11 | | | | Yes | Yes | No | Yes |
| 66 | GC/MSMS | 0.74 | 0.21 | | | No | No | No | Yes |
| 78 ¹ | GC/MSMS | <12.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | <0.18 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 77. DATA REPORTED AND PERFORMANCE EVALUATION FOR HEPTACHLOR EPOXIDE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 2 ¹ | GC/ECD ² | <3.0 | | | | No | Yes | No | Yes |
| 7 | GC/ECD | 1.45 | 0.49 | 3.94 | 2.35 | No | No | Yes | Yes |
| 10 | GC/MSMS | 0.48 | | -2.01 | | Yes | Yes | No | |
| 15 ¹ | GC/MSMS | <5.0 | | | | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 0.14 | 0.11 | -4.13 | -4.86 | Yes | Yes | Yes | No |
| 23 | GC/ECD | 8.58 | | 47.9 | | | | Yes ⁴ | Yes |
| 25 | GC/MSMS | <0.10 | | | | Yes | | | Yes |
| 26 | GC/MS EI | <0.0010 | | | | Yes | No | No | Yes |
| 35 | GC/MS EI | <0.05 | | | | Yes | No | No | Yes |
| 40 ¹ | GC/HRMS | 0.84 | 0.05 | 0.2 | 0.25 | Yes | No | Yes | Yes |
| 45 ¹ | GC/ECD | 4.40 | | 22.13 | | No | No | No | No |
| 47 ¹ | GC/ECD | 46.5 | 15.8 | 282 | 5.78 | No | No | No | Yes |
| 53 | GC/MSMS | <3 | | | | | Yes | Yes | Yes |
| 54 | GC/MSMS | <3.3 | | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | No | No |
| 56 | GC/HRMS | 0.72 | 22.4 | -0.55 | -0.01 | Yes | No | Yes | Yes |
| 57 | GC/MSMS | 0.79 | | -0.15 | | Yes | Yes | No | Yes |
| 63 | GC/MS EI | 0.59 | 0.15 | -1.37 | -0.37 | No | No | Yes ⁴ | No |
| 66 | GC/MSMS | 1.60 | 0.29 | 4.9 | 4.04 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 0.97 | 0.05 | 1 | 1.26 | Yes | Yes | Yes | Yes |
| 78 ¹ | GC/MSMS | <12.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ³ | 0.61 | 0.12 | -1.24 | -1.44 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with dual column³ GC/ECD and peak confirmation with GC/MS⁴ Yes of a different matrix of the test sample

I.3. RESULTS REPORTED FOR PBDES

TABLE 78. DATA REPORTED FOR BDE 17

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | <0.10 | | | | Yes | Yes | Yes | No |
| 38 ¹ | GC/HRMS | 0.04 | 0.27 | | | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 0.03 | 0.01 | | | Yes | Yes | Yes | Yes |
| 79 | GC/ECD ² | <0.06 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited

² GC/ECD and peak confirmation with GC/MS

TABLE 79. DATA REPORTED AND PERFORMANCE EVALUATION FOR BDE 28

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MS EI | 1 | | 1.37 | | Yes | Yes | No | Yes |
| 14 | GC/MSMS | 1.45 | 0.64 | 1.19 | 0.70 | Yes | Yes | Yes | Yes |
| 17 | GC/HRMS | 1.65 | 0.40 | 2.15 | 1.80 | Yes | Yes | Yes | No |
| 27 | GC/MSMS | 1.28 | | 0.37 | | Yes | Yes | | Yes |
| 35 | GC/MSMS | 1.36 | 0.02 | 0.74 | 1.08 | Yes | No | No | Yes |
| 38 ¹ | GC/MS NCI | 1.37 | 0.29 | 0.8 | 0.77 | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 1.34 | 0.12 | 0.61 | 0.83 | Yes | Yes | Yes | Yes |
| 46 ¹ | GC/HRMS | 0.004 | 0.32 | -5.86 | -5.68 | | Yes | Yes ² | Yes |
| 49 | GC/MSMS | 1.08 | 30 | -0.62 | -0.01 | No | No | Yes | Yes |
| 50 | GC/MS EI | 0.002 | 0.001 | -5.9 | -8.67 | | | | |
| 56 | GC/HRMS | 0.62 | 6.52 | -2.87 | -0.18 | Yes | No | Yes | Yes |
| 65 | GC/HRMS | 0.38 | 0.17 | -4.04 | -5.12 | Yes | No | No | Yes |
| 66 | GC/MSMS | 0.72 | 0.22 | -2.39 | -2.76 | No | No | No | Yes |
| 72 ¹ | GC/MSMS | 2.32 | 0.09 | 5.41 | 7.59 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MS EI | 1.26 | | 0.25 | | No | Yes | | Yes |
| 75 | GC/MS EI | 0.92 | | -1.41 | | | | | |
| 78 ¹ | GC/MSMS | 1.10 | | -0.53 | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | 0.79 | 0.12 | -2.06 | -2.80 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited

² Yes of a different matrix of the test sample

TABLE 80. DATA REPORTED AND PERFORMANCE EVALUATION FOR BDE 47

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 25.5 | | 0.88 | | Yes | Yes | No | Yes |
| 14 | GC/HRMS | 23.4 | 10.3 | 0.29 | 0.18 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 22.8 | 10.6 | 0.11 | 0.07 | Yes | Yes | Yes | No |
| 27 | GC/MSMS | 22.9 | | 0.15 | | Yes | Yes | | Yes |
| 35 | GC/MS NCI | 24.1 | 0.45 | 0.48 | 0.80 | Yes | No | No | Yes |
| 38 ¹ | GC/HRMS | 24.9 | 3.58 | 0.7 | 0.91 | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 21.3 | 1.60 | -0.31 | -0.48 | Yes | Yes | Yes | Yes |
| 46 ¹ | GC/MSMS | 0.08 | 0.30 | -6.41 | -10.76 | | Yes | Yes ³ | Yes |
| 49 | GC/MS EI | 13.8 | 30 | -2.47 | -0.57 | No | No | Yes | Yes |
| 50 | GC/MSMS | 18.4 | 11.2 | -1.2 | -0.67 | | | | |
| 56 | GC/HRMS | 26.8 | 12.4 | 1.27 | 0.67 | Yes | No | Yes | Yes |
| 65 | GC/MSMS | 7.35 | 3.23 | -4.32 | -5.73 | Yes | No | No | Yes |
| 66 | GC/MSMS | 14.6 | 2 | -2.24 | -3.39 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 28.0 | 1.40 | 1.61 | 2.56 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 21.5 | | -0.26 | | No | Yes | | Yes |
| 75 | GC/MS EI | 11.6 | | -3.10 | | | | | |
| 78 ¹ | GC/MSMS | 20.9 | | -0.43 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 23.0 | 4.60 | 0.16 | 0.19 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 81. DATA REPORTED AND PERFORMANCE EVALUATION FOR BDE 49

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | 1.81 | 0.50 | -0.76 | -0.77 | Yes | Yes | Yes | No |
| 38 ¹ | GC/HRMS | 2.46 | 0.50 | 1.13 | 1.15 | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 2 | 0.70 | -0.21 | -0.18 | Yes | Yes | Yes | Yes |
| 72 ¹ | GC/MS EI | 3.56 | 0.18 | 4.28 | 5.97 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 2.12 | | 0.14 | | No | Yes | | Yes |

¹ Laboratory accredited

TABLE 82. DATA REPORTED AND PERFORMANCE EVALUATION FOR BDE 66

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | 1.97 | 0.90 | -1.22 | -1.01 | Yes | Yes | Yes | No |
| 27 | GC/MSMS | 2.73 | | 0.38 | | Yes | Yes | | Yes |
| 38 ¹ | GC/HRMS | 2.81 | 0.82 | 0.55 | 0.49 | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 2.22 | 0.77 | -0.70 | -0.63 | Yes | Yes | Yes | Yes |
| 72 ¹ | GC/MS EI | 3.56 | 0.18 | 2.12 | 2.77 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 2.02 | | -1.1 | | No | Yes | | Yes |
| 79 | GC/ECD ² | 2.63 | 0.52 | 0.17 | 0.18 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS

TABLE 83. DATA REPORTED FOR BDE 85

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | <0.10 | | | | Yes | Yes | Yes | No |
| 38 ¹ | GC/HRMS | 0.02 | 1.01 | | | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | <0.028 | | | | Yes | Yes | Yes | Yes |
| 72 ¹ | GC/MS EI | <0.30 | | | | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 0.02 | | | | No | Yes | | Yes |
| 79 | GC/ECD ² | 0.10 | 0.03 | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS

TABLE 84. DATA REPORTED AND PERFORMANCE EVALUATION FOR BDE 99

| Lab code | Instrument Type | \bar{X}_{lab} $\mu\text{g kg}^{-1}$ dw | $U_{lab} (k=2)$ $\mu\text{g kg}^{-1}$ dw | z - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|---|---|----------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 1.28 | | -0.54 | | Yes | Yes | No | Yes |
| 14 | GC/HRMS | 1.37 | 0.60 | -0.13 | -0.09 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 1.44 | 0.50 | 0.15 | 0.12 | Yes | Yes | Yes | No |
| 27 | GC/MSMS | 1.47 | | 0.30 | | Yes | Yes | | Yes |
| 35 | GC/MS NCI | 1.57 | 0.06 | 0.73 | 1.12 | Yes | No | No | Yes |
| 38 ¹ | GC/HRMS | 1.90 | 4.82 | 2.2 | 0.20 | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 1.19 | 0.08 | -0.92 | -1.40 | Yes | Yes | Yes | Yes |
| 46 ¹ | GC/MSMS | 0.01 | 0.37 | -6.16 | -5.96 | | Yes | Yes ³ | Yes |
| 49 | GC/MS EI | 4.60 | 30.00 | 14.11 | 0.21 | No | No | Yes | Yes |
| 50 | GC/MSMS | 0.002 | 0.001 | -6.2 | -9.73 | | | | |
| 56 | GC/HRMS | 0.57 | 5.03 | -3.67 | -0.33 | Yes | No | Yes | Yes |
| 65 | GC/MSMS | 0.43 | 0.19 | -4.31 | -5.69 | Yes | No | No | Yes |
| 66 | GC/MSMS | 1.02 | 0.26 | -1.70 | -1.99 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 1.88 | 0.17 | 2.11 | 2.87 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 1.34 | | -0.28 | | No | Yes | | Yes |
| 75 | GC/MS EI | 0.96 | | -1.95 | | | | | |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 1.55 | 0.31 | 0.64 | 0.69 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 85. DATA REPORTED AND PERFORMANCE EVALUATION FOR BDE 100

| Lab code | Instrument Type | \bar{X}_{lab} $\mu\text{g kg}^{-1}$ dw | $U_{lab} (k=2)$ $\mu\text{g kg}^{-1}$ dw | z - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|---|---|----------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 8.29 | | 0.78 | | Yes | Yes | No | Yes |
| 14 | GC/HRMS | 7.72 | 3.40 | 0.33 | 0.22 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 9.33 | 5.97 | 1.60 | 0.68 | Yes | Yes | Yes | No |
| 27 | GC/MSMS | 7.87 | | 0.45 | | Yes | Yes | | Yes |
| 35 | GC/MS NCI | 10.03 | 0.35 | 2.16 | 3.05 | Yes | No | No | Yes |
| 38 ¹ | GC/HRMS | 7.49 | 1.18 | 0.2 | 0.19 | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 7.14 | 0.46 | -0.12 | -0.17 | Yes | Yes | Yes | Yes |
| 46 ¹ | GC/MSMS | 0.03 | 0.29 | -5.74 | -8.17 | | Yes | Yes ³ | Yes |
| 49 | GC/MS EI | 1.55 | 30 | -4.54 | -0.38 | No | No | Yes | Yes |
| 50 | GC/MSMS | 3.59 | 2.24 | -2.9 | -2.60 | | | | |
| 56 | GC/HRMS | 0.59 | 9.91 | -5.30 | -1.33 | Yes | No | Yes | Yes |
| 65 | GC/MSMS | 10065 | 1.33 | 7945.84 | 9137.85 | Yes | No | No | Yes |
| 66 | GC/MSMS | 5.58 | 0.90 | -1.36 | -1.74 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 9.68 | 0.60 | 1.88 | 2.57 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 7.31 | | 0.01 | | No | Yes | | Yes |
| 75 | GC/MS EI | 3.55 | | -2.96 | | | | | |
| 78 ¹ | GC/MSMS | 5.70 | | -1.26 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 6.78 | 1.36 | -0.41 | -0.47 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 86. DATA REPORTED AND PERFORMANCE EVALUATION FOR BDE 153

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 0.69 | | 0.78 | | Yes | Yes | No | Yes |
| 14 | GC/HRMS | 0.77 | 0.34 | 1.51 | 0.85 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 0.65 | 0.25 | 0.36 | 0.26 | Yes | Yes | Yes | No |
| 27 | GC/MSMS | 0.97 | | 3.48 | | Yes | Yes | | Yes |
| 35 | GC/MS NCI | 0.75 | 0.08 | 1.39 | 1.80 | Yes | No | No | Yes |
| 38 ¹ | GC/HRMS | 0.63 | 0.83 | 0.1 | 0.03 | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 0.66 | 0.05 | 0.49 | 0.70 | Yes | Yes | Yes | Yes |
| 46 ¹ | GC/MSMS | 0.004 | 0.28 | -5.92 | -3.89 | | Yes | Yes ³ | Yes |
| 50 | GC/MSMS | 0.002 | 0.001 | -5.95 | -8.92 | | | | |
| 56 | GC/HRMS | 0.63 | 9.20 | 0.2 | 0.00 | Yes | No | Yes | Yes |
| 65 | GC/MSMS | 0.36 | 0.16 | -2.45 | -2.40 | Yes | No | No | Yes |
| 66 | GC/MSMS | 0.42 | 0.14 | -1.91 | -2.00 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 0.98 | 0.05 | 3.63 | 5.16 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 0.68 | | 0.64 | | No | Yes | | Yes |
| 75 | GC/MS EI | 0.38 | | -2.25 | | | | | |
| 78 ¹ | GC/MSMS | <5 | | | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 0.49 | 0.10 | -1.19 | -1.44 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 87. DATA REPORTED AND PERFORMANCE EVALUATION FOR BDE 154

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 10 | GC/MSMS | 5.24 | | 0.69 | | Yes | Yes | No | Yes |
| 14 | GC/HRMS | 5.84 | 2.56 | 1.43 | 0.83 | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | 5.49 | 2.41 | 0.99 | 0.60 | Yes | Yes | Yes | No |
| 27 | GC/MSMS | 7.01 | | 2.88 | | Yes | Yes | | Yes |
| 35 | GC/MS NCI | 6.96 | 0.37 | 2.82 | 3.87 | Yes | No | No | Yes |
| 38 ¹ | GC/HRMS | 5.26 | 0.63 | 0.7 | 0.90 | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 4.18 | 0.27 | -0.63 | -0.89 | Yes | Yes | Yes | Yes |
| 46 ¹ | GC/MSMS | 0.03 | 0.31 | -5.76 | -8.05 | | Yes | Yes ³ | Yes |
| 49 | GC/MS EI | 1.18 | 30 | -4.33 | -0.23 | No | No | Yes | Yes |
| 50 | GC/MSMS | 2.03 | 1.27 | -3.3 | -3.15 | | | | |
| 56 | GC/HRMS | 0.53 | 9.17 | -5.14 | -0.90 | Yes | No | Yes | Yes |
| 65 | GC/MSMS | 1.73 | 0.76 | -3.66 | -4.38 | Yes | No | No | Yes |
| 66 | GC/MSMS | 6.64 | 0.93 | 2.41 | 2.70 | No | No | No | Yes |
| 72 ¹ | GC/MS EI | 6.84 | 0.46 | 2.67 | 3.58 | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 4.88 | | 0.24 | | No | Yes | | Yes |
| 75 | GC/MS EI | 3.24 | | -1.78 | | | | | |
| 78 ¹ | GC/MSMS | 3.73 | | -1.18 | | Yes | Yes | No | Yes |
| 79 | GC/ECD ² | 4.14 | 0.83 | -0.67 | -0.78 | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS³ Yes of a different matrix of the test sample

TABLE 88. DATA REPORTED FOR BDE 183

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 14 | GC/HRMS | <0.008 | | | | Yes | Yes | Yes | Yes |
| 17 | GC/MSMS | <0.20 | | | | Yes | Yes | Yes | No |
| 38 ¹ | GC/HRMS | <0.01 | 1.52 | | | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 0.02 | 0.002 | | | Yes | Yes | Yes | Yes |
| 56 | GC/HRMS | 0.29 | 8.25 | | | Yes | No | Yes | Yes |
| 66 | GC/MSMS | 0.42 | 0.15 | | | No | No | No | Yes |
| 72 ¹ | GC/MS EI | <0.30 | | | | Yes | Yes | Yes | Yes |
| 74 ¹ | GC/MSMS | 0.002 | | | | No | Yes | | Yes |

¹ Laboratory accredited

TABLE 89. DATA REPORTED FOR BDE 209

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|---------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 17 | GC/MSMS | 2.80 | 1.17 | | | Yes | Yes | Yes | No |
| 38 ¹ | GC/HRMS | <2 | 12.33 | | | Yes | Yes | Yes | Yes |
| 40 ¹ | GC/HRMS | 0.05 | 0.01 | | | Yes | Yes | Yes | Yes |
| 56 | GC/HRMS | 80 | 378 | | | Yes | No | Yes | Yes |
| 74 ¹ | GC/MSMS | 0.10 | | | | No | Yes | | Yes |
| 79 | GC/ECD ² | <2.05 | | | | Yes | Yes | Yes | Yes |

¹ Laboratory accredited² GC/ECD and peak confirmation with GC/MS

I.4. RESULTS REPORTED FOR PAHS

TABLE 90. DATA REPORTED AND PERFORMANCE EVALUATION FOR NAPHTHALENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 5 | GC/MSMS | 23.3 | | 1.27 | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 8.91 | | -2.60 | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | 33.7 | 16.7 | 4.09 | 1.70 | Yes | Yes | Yes | Yes |
| 22 | | 0.14 | | -4.96 | | | | Yes | |
| 25 | GC/MSMS | 66.4 | 3.22 | 12.88 | 14.42 | Yes | | | |
| 26 | GC/MS EI | 0.53 | | -4.8 | | No | Yes | | Yes |
| 35 | GC/MS EI | 16.2 | 0.55 | -0.64 | -0.82 | Yes | No | No | Yes |
| 36 | HPLC-Fluo | 2.34 | 0.19 | -4.36 | -5.58 | No | No | Yes | No |
| 37 | GC/MS EI | 127 | | 29.18 | | No | No | | No |
| 42 | HPLC-Fluo | 13.7 | | -1.3 | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.03 | | -4.98 | | No | No | No | No |
| 52 | GC/MSMS | 23.5 | 0.58 | 1.32 | 1.69 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | 339 | | 86.23 | | No | No | | |
| 65 | GC/MS EI | 15.9 | 7 | -0.71 | -0.58 | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 21.5 | 3.20 | 0.78 | 0.88 | No | | No | Yes |
| 67 ¹ | GC/MS EI | 3131 | | 837.42 | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | 11.0 | 0.51 | -2.03 | -2.59 | Yes | Yes | Yes ² | Yes |
| 76 | HPLC-UV | 57.0 | 10 | 10.35 | 6.65 | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | 23.5 | | 1.3 | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | 8.67 | 2.61 | -2.66 | -3.11 | Yes | Yes | Yes ² | Yes |

¹ Laboratory accredited

² Yes of a different matrix of the test sample

TABLE 91. DATA REPORTED AND PERFORMANCE EVALUATION FOR 1-METHYLNAPHTHALENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 12 | GC/MS EI | 9.04 | | -2.13 | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | 18.9 | 9.45 | 1.73 | 0.87 | Yes | Yes | Yes | Yes |
| 25 | GC/MSMS | 21.7 | 2.53 | 2.82 | 3.27 | Yes | | | |
| 49 ¹ | GC/MS EI | 13.7 | 12.2 | -0.31 | -0.13 | No | No | Yes | Yes |
| 55 | GC/MS EI | 15.7 | | 0.47 | | No | No | | |
| 72 ¹ | GC/MS EI | 12.4 | 1.10 | -0.8 | -1.10 | Yes | Yes | Yes ² | Yes |
| 79 | GC/MSMS | 14.1 | 4.22 | -0.17 | -0.15 | Yes | Yes | Yes ² | Yes |

¹ Laboratory accredited

² Yes of a different matrix of the test sample

TABLE 92. DATA REPORTED AND PERFORMANCE EVALUATION FOR 2-METHYLNAPHTHALENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 12 | GC/MS EI | 13.1 | | -1.35 | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | 41.1 | 20.6 | 5.73 | 2.10 | Yes | Yes | Yes | Yes |
| 25 | GC/MSMS | 10.2 | 1.12 | -2.09 | -2.54 | Yes | | | |
| 49 ¹ | GC/MS EI | 22.5 | 12.2 | 1.01 | 0.58 | No | No | Yes | Yes |
| 55 | GC/MS EI | 23.3 | | 1.24 | | No | No | | |
| 72 ¹ | GC/MS EI | 17.2 | 0.89 | -0.3 | -0.38 | Yes | Yes | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <20 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | 21.9 | 6.56 | 0.87 | 0.75 | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 93. DATA REPORTED AND PERFORMANCE EVALUATION FOR ACENAPHTHYLENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 5 | GC/MSMS | 1.98 | | 6.68 | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 2.44 | | 9.64 | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | <5 | | | | Yes | Yes | Yes | Yes |
| 22 | | 0.83 | | -0.70 | | | | Yes | |
| 25 | GC/MSMS | 0.86 | 0.26 | -0.47 | -0.44 | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | 0.25 | 0.02 | -4.39 | -6.62 | Yes | No | No | Yes |
| 45 | GC/MS | 0.003 | | -5.99 | | No | No | No | No |
| 52 | GC/MSMS | 2.19 | 0.32 | 8.07 | 6.61 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.14 | 8.93 | -5.11 | -0.18 | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 5.30 | 2.30 | 28.0 | 3.62 | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 15.1 | 1.81 | 90.7 | 15.6 | No | | No | Yes |
| 67 ¹ | GC/MS EI | <1.75 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | 0.90 | | -0.23 | | Yes | Yes | Yes ² | Yes |
| 76 | HPLC-UV | 261 | 50 | 1668 | 10.4 | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | 0.50 | 0.16 | -2.80 | -3.35 | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 94. DATA REPORTED AND PERFORMANCE EVALUATION FOR ACENAPHTHENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|-----------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 5 | GC/MSMS | 1.8 | | 0.47 | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 5.61 | | 10.7 | | Yes | No | Yes | |
| 14 | HPLC-Fluo | <5 | | | | No | No | No | |
| 15 ¹ | GC/MSMS | <5 | | | | Yes | Yes | Yes | Yes |
| 22 | | 0.81 | | -2.17 | | | | Yes | |
| 25 | GC/MSMS | 1.07 | 0.16 | -1.5 | -1.71 | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | 0.81 | 0.04 | -2.16 | -2.57 | Yes | No | No | Yes |
| 36 | HPLC-Fluo | 1.23 | 1.00 | -1.04 | -0.66 | No | No | Yes | No |
| 42 | HPLC-Fluo | 0.70 | | -2.5 | | Yes | No | Yes | Yes |
| 49 ¹ | GC/MS EI | 3.57 | 11.8 | 5.23 | 0.33 | No | No | Yes | Yes |
| 52 | GC/MSMS | 2.51 | 0.32 | 2.39 | 2.54 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.12 | 25 | -4.02 | -0.12 | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 1.93 | 0.90 | 0.84 | 0.62 | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 3.43 | 0.75 | 4.85 | 3.68 | No | | No | Yes |
| 67 ¹ | GC/MS EI | <1.75 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | 1.70 | 0.14 | 0.22 | 0.25 | Yes | Yes | Yes ² | Yes |
| 76 | HPLC-UV | 267 | 40 | 711 | 13.3 | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | 33.2 | | 85 | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | 1.10 | 0.33 | -1.39 | -1.48 | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 95. DATA REPORTED AND PERFORMANCE EVALUATION FOR FLUORENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|-----------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 5 | GC/MSMS | 4.37 | | 0.44 | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 3.37 | | -0.90 | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | 8.13 | 4.06 | 5.50 | 1.95 | Yes | Yes | Yes | Yes |
| 22 | | 0.13 | | -5.25 | | | | Yes | |
| 25 | GC/MSMS | 4.57 | 0.74 | 0.72 | 0.81 | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | 1.97 | 0.17 | -2.78 | -3.74 | Yes | No | No | Yes |
| 36 | HPLC-Fluo | 3.43 | 0.70 | -0.82 | -0.94 | No | No | Yes | No |
| 42 | HPLC-Fluo | 3.50 | | -0.72 | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.0 | | | | No | No | No | No |
| 47 ¹ | HPLC-Fluo | 33.8 | | 40.1 | 8.70 | No | No | No | Yes |
| 49 ¹ | GC/MS EI | 1.84 | 12.2 | -2.96 | -0.36 | No | No | Yes | Yes |
| 52 | GC/MSMS | 5.89 | 0.32 | 2.50 | 3.26 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | 17 | | 17.4 | | No | No | | |
| 56 | GC/HRMS | 0.18 | 2.47 | -5.19 | -2.85 | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 6.80 | 3 | 3.72 | 1.73 | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 5.34 | 0.96 | 1.75 | 1.79 | No | | No | Yes |
| 67 ¹ | GC/MS EI | <1.75 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | 2.36 | 0.11 | -2.25 | -3.05 | Yes | Yes | Yes ² | Yes |
| 76 | HPLC-UV | 15.3 | 4 | 15.2 | 5.45 | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | 1.90 | 0.58 | -2.87 | -3.46 | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 96. DATA REPORTED AND PERFORMANCE EVALUATION FOR PHENANTHRENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z-score | Zeta-score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|-----------------|--|--|---------|------------|-----------|-------------------|--------------------|-------|
| 5 | GC/MSMS | 15.3 | | 1.58 | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 8.54 | | -1.46 | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | 18.7 | 9.35 | 3.10 | 1.39 | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | 9.36 | | -1.09 | | | Yes | | Yes |
| 22 | | 0.53 | | -5.05 | | | | Yes | |
| 25 | GC/MSMS | 10.6 | 2.05 | -0.6 | -0.63 | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | 11.4 | 0.46 | -0.17 | -0.23 | Yes | No | No | Yes |
| 36 | HPLC-Fluo | 21.2 | 19.7 | 4.24 | 0.94 | No | No | Yes | No |
| 42 | HPLC-Fluo | 16.1 | | 1.9 | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.003 | | -5.29 | | No | No | No | No |
| 47 ¹ | HPLC-Fluo | 691 | | 305 | 9.83 | No | No | No | Yes |
| 49 ¹ | GC/MS EI | 8.57 | 12.2 | -1.44 | -0.51 | No | No | Yes | Yes |
| 52 | GC/MSMS | 18.2 | 0.32 | 2.87 | 3.80 | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.06 | 32.9 | -5.26 | -0.71 | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 23.6 | 10.4 | 5.30 | 2.16 | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 40.5 | 4 | 12.9 | 11 | No | | No | Yes |
| 67 ¹ | GC/MS EI | 401 | | 174 | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | 12.2 | 0.90 | 0.19 | 0.24 | Yes | Yes | Yes ² | Yes |
| 76 | HPLC-UV | 5 | | -3.04 | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | 9.10 | | -1.21 | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | 8.10 | 2.43 | -1.65 | -1.78 | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 97. DATA REPORTED FOR ANTHRACENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 4 ¹ | GC/MS EI | 5.45 | | | | Yes | No | No | |
| 5 | GC/MSMS | 2.12 | | | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 2.36 | | | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | <10 | | | | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | <0.5 | | | | | Yes | | Yes |
| 25 | GC/MSMS | 1.85 | 0.35 | | | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | 0.31 | 0.03 | | | Yes | No | No | Yes |
| 42 | HPLC-Fluo | 0.63 | | | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.0 | | | | No | No | No | No |
| 47 ¹ | HPLC-Fluo | 126 | | | | No | No | No | Yes |
| 52 | GC/MSMS | 1.74 | 0.32 | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.14 | 5.76 | | | Yes | No | Yes | Yes |
| 65 | GC/MS EI | <0.1 | | | | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 0.28 | 0.10 | | | No | | No | Yes |
| 67 ¹ | GC/MS EI | <1.75 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | <0.67 | | | | Yes | Yes | Yes ² | Yes |
| 76 | HPLC-UV | 7.50 | | | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | 1 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.3 | | | | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 98. DATA REPORTED AND PERFORMANCE EVALUATION FOR 1-METHYLPHENANTHRENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 25 | GC/MSMS | 14.3 | 73 | | | Yes | | | |
| 72 ¹ | GC/MS EI | 2.30 | 0.14 | | | Yes | Yes | Yes ² | Yes |
| 79 | GC/MSMS | 3.15 | 0.95 | | | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 99. DATA REPORTED AND PERFORMANCE EVALUATION FOR 2-METHYLPHENANTHRENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 72 ¹ | GC/MS EI | 4.04 | 0.22 | | | Yes | Yes | Yes ² | Yes |
| 79 | GC/MSMS | 5.90 | 1.77 | | | Yes | Yes | Yes ² | Yes |

¹ Laboratory accredited² Yes of a different matrix of the test sample

TABLE 100. DATA REPORTED FOR BENZ[a]ANTHRACENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 4 ¹ | GC/MS EI | 19 | | | | Yes | No | No | |
| 5 | GC/MSMS | 0.49 | | | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 2.63 | | | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | 16.4 | 8.20 | | | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | <0.1 | | | | | Yes | | Yes |
| 22 | | 1.83 | | | | | | Yes | |
| 25 | GC/MSMS | 3.47 | 2.67 | | | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | <0.2 | | | | Yes | No | No | Yes |
| 36 | HPLC-Fluo | 15.2 | 3.79 | | | No | No | Yes | No |
| 42 | HPLC-Fluo | 0.60 | | | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.61 | | | | No | No | No | No |
| 47 ¹ | HPLC-Fluo | 145 | | | | No | No | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.06 | 21.3 | | | Yes | No | Yes | Yes |
| 65 | GC/MS EI | <0.1 | | | | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 0.63 | 0.19 | | | No | | No | Yes |
| 67 ¹ | GC/MS EI | 287 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | <0.53 | | | | Yes | Yes | Yes ² | Yes |
| 74 ¹ | GC/MSMS | <0.3 | | | | No | Yes | No | Yes |
| 76 | HPLC-UV | 40 | 8 | | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.7 | | | | Yes | Yes | Yes ² | Yes |

¹ Laboratory accredited² Yes of a different matrix of the test sample

TABLE 101. DATA REPORTED FOR CHRYSENE (+ TRYPHENYLENE)

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 5 | GC/MSMS | 1.45 | | | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 1.05 | | | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | <10 | | | | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | <0.1 | | | | | Yes | | Yes |
| 22 | | 16.2 | | | | | | Yes | |
| 25 | GC/MSMS | 2.31 | 0.47 | | | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | <0.2 | | | | Yes | No | No | Yes |
| 36 | HPLC-Fluo | 10.8 | 2.34 | | | No | No | Yes | No |
| 42 | HPLC-Fluo | 0.63 | | | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.03 | | | | No | No | No | No |
| 47 ¹ | HPLC-Fluo | 103 | | | | No | No | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.05 | 29.6 | | | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 0.83 | 0.40 | | | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 1.42 | 0.28 | | | No | | No | Yes |
| 67 ¹ | GC/MS EI | <1.75 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | <0.53 | | | | Yes | Yes | Yes ² | Yes |
| 74 ¹ | GC/MSMS | <0.3 | | | | No | Yes | No | Yes |
| 76 | HPLC-UV | 2.33 | | | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.4 | | | | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 102. DATA REPORTED FOR FLUORANTHENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 4 ¹ | GC/MS EI | 14.9 | | | | Yes | No | No | |
| 5 | GC/MSMS | 4.34 | | | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 3.49 | | | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | 24.4 | 12.2 | | | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | <0.5 | | | | | Yes | | Yes |
| 22 | | 0.74 | | | | | | Yes | |
| 25 | GC/MSMS | 4.51 | 4.92 | | | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | 0.42 | 0.17 | | | Yes | No | No | Yes |
| 36 | HPLC-Fluo | 28.48 | 5.43 | | | No | No | Yes | No |
| 42 | HPLC-Fluo | 1.63 | | | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.002 | | | | No | No | No | No |
| 52 | GC/MSMS | 3.58 | 0.32 | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.06 | 21.6 | | | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 8.07 | 3.60 | | | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 5.96 | 1.20 | | | No | | No | Yes |
| 67 ¹ | GC/MS EI | 222 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | 1.02 | 0.10 | | | Yes | Yes | Yes ² | Yes |
| 76 | HPLC-UV | 4 | | | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.25 | | | | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 103. DATA REPORTED FOR PYRENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 5 | GC/MSMS | 3.82 | | | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 0.0 | | | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | <10 | | | | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | <0.5 | | | | | Yes | | Yes |
| 22 | | 0.76 | | | | | | Yes | |
| 25 | GC/MSMS | 4.01 | 3.17 | | | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | 0.44 | 0.12 | | | Yes | No | No | Yes |
| 36 | HPLC-Fluo | 11.1 | 8.61 | | | No | No | Yes | No |
| 42 | HPLC-Fluo | 1.78 | | | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.08 | | | | No | No | No | No |
| 52 | GC/MSMS | 7.17 | 0.28 | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.04 | 26.80 | | | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 4.47 | 2 | | | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 3.39 | 0.82 | | | No | | No | Yes |
| 67 ¹ | GC/MS EI | 144 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | 0.84 | | | | Yes | Yes | Yes ² | Yes |
| 76 | HPLC-UV | 37.7 | 8 | | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.24 | | | | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 104. DATA REPORTED FOR BENZO[a]PYRENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 4 ¹ | GC/MS EI | 18.1 | | | | Yes | No | No | |
| 12 | GC/MS EI | 0.0 | | | | Yes | No | Yes | |
| 5 | GC/MSMS | 0.13 | | | | | Yes | Yes | Yes |
| 14 | HPLC-Fluo | <5 | | | | No | No | No | |
| 15 ¹ | GC/MSMS | <10 | | | | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | <0.1 | | | | | Yes | | Yes |
| 22 | | <0.3 | | | | | | Yes | |
| 25 | GC/MSMS | 2.88 | 1.33 | | | Yes | | | |
| 26 | GC/MS EI | 0.58 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | <0.2 | | | | Yes | No | No | Yes |
| 42 | HPLC-Fluo | <0.5 | | | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.12 | | | | No | No | No | No |
| 47 ¹ | HPLC-Fluo | 14.2 | | | | No | No | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.09 | 13 | | | Yes | No | Yes | Yes |
| 65 | GC/MS EI | <0.1 | | | | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 2.53 | 0.51 | | | No | | No | Yes |
| 67 ¹ | GC/MS EI | 237 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | <0.54 | | | | Yes | Yes | Yes ² | Yes |
| 74 ¹ | GC/MSMS | <0.3 | | | | No | Yes | No | Yes |
| 76 | HPLC-UV | 6.67 | | | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.17 | | | | Yes | Yes | Yes ² | Yes |

¹ Laboratory accredited² Yes of a different matrix of the test sample

TABLE 105. DATA REPORTED FOR BENZO[e]PYRENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (k=2) µg kg ⁻¹ dw | z- score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|--|-------------|----------------|-----------|----------------------|-----------------------|-------|
| 19 ¹ | GC/MSMS | <0.5 | | | | | Yes | | Yes |
| 25 | GC/MSMS | 2.61 | 1.65 | | | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | <0.2 | | | | Yes | No | No | Yes |
| 45 | GC/MS | 0.38 | | | | No | No | No | No |
| 67 ¹ | GC/MS EI | 152 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | <0.53 | | | | Yes | Yes | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.14 | | | | Yes | Yes | Yes ² | Yes |

¹ Laboratory accredited² Yes of a different matrix of the test sample

TABLE 106. DATA REPORTED FOR BENZO[*b+j*]FLUORANTHENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (<i>k</i> =2) µg kg ⁻¹ dw | <i>z</i> - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|---|---------------------|----------------|-----------|----------------------|-----------------------|-------|
| 4 ¹ | GC/MS EI | 36 | | | | Yes | No | No | |
| 5 | GC/MSMS | 0.35 | | | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 0.0 | | | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | <10 | | | | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | <0.1 | | | | | Yes | | Yes |
| 22 | | 3.11 | | | | | | | |
| 25 | GC/MSMS | 2.62 | 0.53 | | | Yes | | | |
| 26 | GC/MS EI | 0.52 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | <0.2 | | | | Yes | No | No | Yes |
| 42 | HPLC-Fluo | 0.67 | | | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.23 | | | | No | No | No | No |
| 47 ¹ | HPLC-Fluo | 29.4 | | | | No | No | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.05 | 27.8 | | | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 0.83 | 0.40 | | | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 3.69 | 0.63 | | | No | | No | Yes |
| 67 ¹ | GC/MS EI | <1.75 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | <0.53 | | | | Yes | Yes | Yes ² | Yes |
| 74 ¹ | GC/MSMS | <0.3 | | | | No | Yes | No | Yes |
| 76 | HPLC-UV | 26.7 | 4 | | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.20 | | | | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

TABLE 107. DATA REPORTED FOR BENZO[*g,h,i*]PERYLENE

| Lab code | Instrument Type | X _{lab} µg kg ⁻¹ dw | U _{lab} (<i>k</i> =2) µg kg ⁻¹ dw | <i>z</i> - score | Zeta- score | Surrogate | Internal Standard | Reference Material | QA/QC |
|-----------------|--------------------|--|---|---------------------|----------------|-----------|----------------------|-----------------------|-------|
| 4 ¹ | GC/MS EI | 19 | | | | Yes | No | No | |
| 5 | GC/MSMS | 0.12 | | | | | Yes | Yes | Yes |
| 12 | GC/MS EI | 0.0 | | | | Yes | No | Yes | |
| 15 ¹ | GC/MSMS | <10 | | | | Yes | Yes | Yes | Yes |
| 19 ¹ | GC/MSMS | <0.5 | | | | | Yes | | Yes |
| 25 | GC/MSMS | 2.74 | 1.26 | | | Yes | | | |
| 26 | GC/MS EI | <0.0010 | | | | No | Yes | | Yes |
| 35 | GC/MS EI | <0.2 | | | | Yes | No | No | Yes |
| 42 | HPLC-Fluo | <0.5 | | | | Yes | No | Yes | Yes |
| 45 | GC/MS | 0.99 | | | | No | No | No | No |
| 47 ¹ | HPLC-Fluo | 7.12 | | | | No | No | No | Yes |
| 52 | GC/MSMS | 10.8 | 1.76 | | | Yes | Yes | No | Yes |
| 55 | GC/MS EI | <10 | | | | No | No | | |
| 56 | GC/HRMS | 0.11 | 6.84 | | | Yes | No | Yes | Yes |
| 65 | GC/MS EI | 0.93 | 0.40 | | | Yes | Yes | Yes ² | Yes |
| 66 | HPLC-Fluo | 2.15 | 0.52 | | | No | | No | Yes |
| 67 ¹ | GC/MS EI | <1.75 | | | | No | No | Yes | No |
| 72 ¹ | GC/MS EI | <0.53 | | | | Yes | Yes | Yes ² | Yes |
| 74 ¹ | GC/MSMS | <0.3 | | | | No | Yes | No | Yes |
| 76 | HPLC-UV | 69.7 | 12 | | | No | No | Yes ² | Yes |
| 78 ¹ | GC/MSMS | <2.5 | | | | Yes | Yes | No | Yes |
| 79 | GC/MSMS | <0.6 | | | | Yes | Yes | Yes ² | Yes |

¹Laboratory accredited²Yes of a different matrix of the test sample

APPENDIX II: REPRESENTATION OF LABORATORY PERFORMANCES

II.1. LABORATORIES PERFORMANCES FOR PCB CONGENERS


Evaluation of Reported data for PCB 18

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 57% | 29% | 14% |
| Zeta-score | 67% | 17% | 17% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 0.71 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.27 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.33 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 9 |
| Number of results accepted | 5 |

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - $X_a \pm U_a (k=2)$
- - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 28

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 39% | 26% | 35% |
| Zeta-score | 33% | 28% | 39% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 1.31 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.37 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.49 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 33 |
| Number of results accepted | 17 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 31

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 69% | 8% | 23% |
| Zeta-score | 67% | 11% | 22% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 0.62 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.17 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.23 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 14 |
| Number of results accepted | 10 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 44

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 75% | 25% | 0% |
| Zeta-score | 83% | 17% | 0% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 1.75 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.53 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.69 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 9 |
| Number of results accepted | 8 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for PCB 49


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 75% | 25% | 0% |
| Zeta-score | 100% | 0% | 0% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 2.15 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.90 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.05 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 4 |
| Number of results accepted | 4 |

Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 52

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 59% | 13% | 28% |
| Zeta-score | 65% | 10% | 25% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 3.90 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.87 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.30 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 34 |
| Number of results accepted | 24 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for PCB 66


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 75% | 0% | 25% |
| Zeta-score | 100% | 0% | 0% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 7.12 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 3.94 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 4.33 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 4 |
| Number of results accepted | 3 |


Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for PCB 70


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | | | |
| Zeta-score | | | |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 5.12 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.83 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.23 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 3 |
| Number of results accepted | 2 |

Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



Evaluation of Reported data for PCB 74


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | | | |
| Zeta-score | | | |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 3.61 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.45 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.71 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 3 |
| Number of results accepted | 3 |

Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



Evaluation of Reported data for PCB 87


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 75% | 25% | 0% |
| Zeta-score | 100% | 0% | 0% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 6.54 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.58 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.27 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 4 |
| Number of results accepted | 4 |

Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 95

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 30% | 20% | 50% |
| Zeta-score | 56% | 11% | 33% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 5.23 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.29 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.84 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 10 |
| Number of results accepted | 5 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for PCB 97


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | | | |
| Zeta-score | | | |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 5.34 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.39 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.93 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 2 |
| Number of results accepted | 2 |

Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error




Evaluation of Reported data for PCB 99

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 55% | 18% | 27% |
| Zeta-score | 50% | 20% | 30% |


| | |
|-------------------------------------|-----------------------------|
| Assigned value (X_a) | 17.40 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 6.68 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 7.97 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 11 |
| Number of results accepted | 8 |

Kernel density plot:




Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 101

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 39% | 26% | 35% |
| Zeta-score | 33% | 28% | 39% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 1.31 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.37 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.49 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 33 |
| Number of results accepted | 17 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 105

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 63% | 19% | 19% |
| Zeta-score | 50% | 20% | 30% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 8 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.64 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.58 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 16 |
| Number of results accepted | 13 |


Kernel density plot:



Error! Not a valid link. Reported


Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 110

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 73% | 9% | 18% |
| Zeta-score | 70% | 20% | 10% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 12.4 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 3.96 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 5.03 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 11 |
| Number of results accepted | 9 |

Kernel density plot:




Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - - $X_a \pm U_a (k=2)$
- - - $X_a \pm 2 \text{ Total error}$



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 118

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 68% | 4% | 29% |
| Zeta-score | 50% | 38% | 13% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 28.4 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 5.48 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 8.98 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 29 |
| Number of results accepted | 20 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a — — $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.




Evaluation of Reported data for PCB 128

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 50% | 0% | 50% |
| Zeta-score | 56% | 0% | 44% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 8.43 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.90 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.84 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 10 |
| Number of results accepted | 5 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 138

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 57% | 11% | 31% |
| Zeta-score | 52% | 9% | 39% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 68.1 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 13.2 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 21.5 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 36 |
| Number of results accepted | 24 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for PCB 146


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 57% | 0% | 43% |
| Zeta-score | 50% | 0% | 50% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 18.2 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 3.21 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 5.57 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 7 |
| Number of results accepted | 4 |


Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - - $X_a \pm U_a (k=2)$
- - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 149

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 59% | 6% | 35% |
| Zeta-score | 62% | 0% | 38% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 27.3 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 6.23 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 9.24 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 18 |
| Number of results accepted | 11 |

Kernel density plot:




Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - - $X_a \pm U_a (k=2)$
- - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 151

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 60% | 10% | 30% |
| Zeta-score | 56% | 11% | 33% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 9.29 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 2.55 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 3.45 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 10 |
| Number of results accepted | 6 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 153

\Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 59% | 15% | 26% |
| Zeta-score | 59% | 14% | 27% |


Kernel density plot:



| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 90 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 17.3 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 28.4 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 35 |
| Number of results accepted | 24 |


Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 156

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 73% | 13% | 13% |
| Zeta-score | 80% | 0% | 20% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 3.85 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.70 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.19 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 15 |
| Number of results accepted | 13 |

Kernel density plot:




Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - - $X_a \pm U_a (k=2)$
- - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 170

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 50% | 14% | 36% |
| Zeta-score | 50% | 10% | 40% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 11.9 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 2.49 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 3.88 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 14 |
| Number of results accepted | 9 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for PCB 174


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | | | |
| Zeta-score | | | |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 7.67 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.50 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.43 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 2 |
| Number of results accepted | 2 |

Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - - $X_a \pm U_a (k=2)$
- - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.


Evaluation of Reported data for PCB 177

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 70% | 0% | 30% |
| Zeta-score | 63% | 13% | 25% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 8.16 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 2.37 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 3.13 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 10 |
| Number of results accepted | 7 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 180

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 60% | 6% | 34% |
| Zeta-score | 59% | 5% | 36% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 34.8 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 6.81 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 11.1 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 36 |
| Number of results accepted | 24 |

Kernel density plot:




Reported results and expanded uncertainties:

$\bullet \pm$ $X_{lab} \pm U_{lab} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 183

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 55% | 18% | 27% |
| Zeta-score | 80% | 0% | 20% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 9.51 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 3.45 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 4.19 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 11 |
| Number of results accepted | 7 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 187

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 50% | 8% | 42% |
| Zeta-score | 67% | 0% | 33% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 29.5 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 7.96 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 10.9 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 12 |
| Number of results accepted | 6 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 194

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 73% | 18% | 9% |
| Zeta-score | 71% | 14% | 14% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 4.30 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.14 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.56 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 12 |
| Number of results accepted | 10 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for PCB 195


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 100% | 0% | 0% |
| Zeta-score | 100% | 0% | 0% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 1.62 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.63 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.75 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 4 |
| Number of results accepted | 4 |


Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for PCB 201


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | | | |
| Zeta-score | | | |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 3.58 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.38 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.65 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 3 |
| Number of results accepted | 2 |

Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Evaluation of Reported data for PCB 206


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 50% | 25% | 25% |
| Zeta-score | 50% | 50% | 0% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 3.31 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.61 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.03 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 5 |
| Number of results accepted | 3 |


Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for PCB 209

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 50% | 13% | 38% |
| Zeta-score | 80% | 0% | 20% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 1.94 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.37 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.61 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 9 |
| Number of results accepted | 5 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



II.2. LABORATORIES PERFORMANCES FOR ORGANOCHLORINE PESTICIDES


Evaluation of Reported data for HCB

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 64% | 4% | 32% |
| Zeta-score | 71% | 0% | 29% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 2.89 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.62 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.95 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 27 |
| Number of results accepted | 17 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for *pp'*-DDD

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 41% | 15% | 44% |
| Zeta-score | 41% | 12% | 47% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 13.1 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 2.51 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 4.13 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 29 |
| Number of results accepted | 14 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for *pp'*-DDE

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 59% | 16% | 25% |
| Zeta-score | 45% | 25% | 30% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 121 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 25.8 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 39.9 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 34 |
| Number of results accepted | 25 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for *pp'*-DDT

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 58% | 19% | 23% |
| Zeta-score | 60% | 7% | 33% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 18.2 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 3.88 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 5.98 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 29 |
| Number of results accepted | 20 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for *op'*-DDE

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 79% | 7% | 14% |
| Zeta-score | 83% | 0% | 17% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 2.23 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.45 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.71 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 17 |
| Number of results accepted | 12 |

Kernel density plot:




Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - $X_a \pm U_a (k=2)$
- - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for *op'*-DDD

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 67% | 13% | 20% |
| Zeta-score | 75% | 8% | 17% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 2.95 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.72 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.03 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 16 |
| Number of results accepted | 11 |

Kernel density plot:




Reported results and expanded uncertainties:

\oplus $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for *op'*-DDT

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 63% | 25% | 13% |
| Zeta-score | 55% | 18% | 27% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 7.48 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.66 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.50 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 18 |
| Number of results accepted | 14 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for Dieldrin

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 50% | 17% | 33% |
| Zeta-score | 50% | 21% | 29% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 5.28 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.62 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.09 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 24 |
| Number of results accepted | 11 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for cis-Chlordane

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 62% | 8% | 31% |
| Zeta-score | 100% | 0% | 0% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 6.53 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.85 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.47 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 15 |
| Number of results accepted | 9 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for trans-Chlordane

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 45% | 9% | 45% |
| Zeta-score | 83% | 0% | 17% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 0.38 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.08 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.12 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 16 |
| Number of results accepted | 4 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for cis-Nonachlor


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 100% | 0% | 0% |
| Zeta-score | 75% | 25% | 0% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 11.3 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.81 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 3.35 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 4 |
| Number of results accepted | 4 |


Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - $X_a \pm U_a (k=2)$
- - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for trans-Nonachlor


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 100% | 0% | 0% |
| Zeta-score | 100% | 0% | 0% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 20.5 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 4.95 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 7.13 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 5 |
| Number of results accepted | 5 |


Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for Heptachlor epoxide

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 46% | 8% | 46% |
| Zeta-score | 56% | 11% | 33% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 0.81 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.25 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.32 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 22 |
| Number of results accepted | 7 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



II.3. LABORATORIES PERFORMANCES FOR PBDEs


Evaluation of Reported data for BDE 28

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 56% | 22% | 22% |
| Zeta-score | 54% | 15% | 31% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 1.21 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.28 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.41 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 18 |
| Number of results accepted | 14 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for BDE 47

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 72% | 11% | 17% |
| Zeta-score | 69% | 8% | 23% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 22.4 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 4.14 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 6.97 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 18 |
| Number of results accepted | 16 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for BDE 49


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 80% | 0% | 20% |
| Zeta-score | 75% | 0% | 25% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 2.07 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.47 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.70 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 5 |
| Number of results accepted | 4 |


Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for BDE 66


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 86% | 14% | 0% |
| Zeta-score | 80% | 20% | 0% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 2.55 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.71 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.95 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 7 |
| Number of results accepted | 7 |


Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for BDE 99

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 59% | 12% | 29% |
| Zeta-score | 69% | 8% | 23% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 1.40 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.29 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.45 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 18 |
| Number of results accepted | 12 |

Kernel density plot:




Reported results and expanded uncertainties:

● $X_{lab} \pm U_{lab} (k=2)$;
 — X_a
 - - $X_a \pm U_a (k=2)$
 - - - $X_a \pm 2 \text{ Total error}$



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for BDE 100

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 61% | 17% | 22% |
| Zeta-score | 62% | 15% | 23% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 7.30 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.76 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 2.53 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 18 |
| Number of results accepted | 14 |

Kernel density plot:




Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - - $X_a \pm U_a (k=2)$
- - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for BDE 153

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 63% | 13% | 25% |
| Zeta-score | 67% | 8% | 25% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 0.61 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.14 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.20 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 17 |
| Number of results accepted | 12 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for BDE 154

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 50% | 22% | 28% |
| Zeta-score | 54% | 8% | 38% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 4.68 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.11 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.62 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 18 |
| Number of results accepted | 9 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



II.4. LABORATORIES PERFORMANCES FOR PAHs


Evaluation of Reported data for Naphthalene

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 37% | 11% | 53% |
| Zeta-score | 50% | 10% | 40% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 18.6 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 5.81 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 7.43 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 20 |
| Number of results accepted | 8 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for 1-Methylnaphthalene


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 71% | 29% | 0% |
| Zeta-score | 80% | 0% | 20% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 14.5 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 3.61 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 5.11 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 7 |
| Number of results accepted | 7 |


Reported results and expanded uncertainties:

$\bullet \pm$ $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for 2-Methylnaphthalene


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 71% | 14% | 14% |
| Zeta-score | 60% | 40% | 0% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 18.4 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 6.43 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 7.91 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 8 |
| Number of results accepted | 6 |


Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for Acenaphthylene

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 23% | 8% | 69% |
| Zeta-score | 25% | 0% | 75% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 0.94 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.21 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.31 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 18 |
| Number of results accepted | 5 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for Acenaphthene

Summary of results:


| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 38% | 25% | 38% |
| Zeta-score | 64% | 18% | 18% |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 1.62 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 0.63 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 0.75 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 21 |
| Number of results accepted | 7 |

Kernel density plot:




Reported results and expanded uncertainties:



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for Fluorene

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 33% | 28% | 39% |
| Zeta-score | 46% | 8% | 46% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 4.04 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.09 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.49 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 22 |
| Number of results accepted | 9 |

Kernel density plot:




Reported results and expanded uncertainties:

● $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$;
 — X_a
- - - $X_a \pm U_a (k=2)$
- - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score




Evaluation of Reported data for Phenanthrene

Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | 48% | 5% | 48% |
| Zeta-score | 67% | 8% | 25% |


| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 11.8 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 3.35 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 4.46 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 23 |
| Number of results accepted | 12 |

Kernel density plot:




Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a -- -- $X_a \pm U_a (k=2)$ -- -- $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Performance evaluation: ■ z-score ■ Zeta-score



Evaluation of Reported data for 1-Methylphenanthrene


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | | | |
| Zeta-score | | | |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 2.73 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 1.28 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 1.45 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 3 |
| Number of results accepted | 2 |

Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

Evaluation of Reported data for 2-Methylphenanthrene


Summary of results:

| | Satisfactory | Questionable | Unsatisfactory |
|------------|--------------|--------------|----------------|
| z-score | | | |
| Zeta-score | | | |

| | |
|-------------------------------------|----------------------------|
| Assigned value (X_a) | 4.97 $\mu\text{g kg}^{-1}$ |
| Assigned uncertainty ($U_a, k=2$) | 2.75 $\mu\text{g kg}^{-1}$ |
| 2 Total error | 3.02 $\mu\text{g kg}^{-1}$ |
| Number of results reported | 2 |
| Number of results accepted | 2 |

Reported results and expanded uncertainties:

\bullet $X_{\text{lab}} \pm U_{\text{lab}} (k=2)$; — X_a - - $X_a \pm U_a (k=2)$ - - $X_a \pm 2$ Total error



*Uncertainty not reported by participant and calculated as: $2 \frac{s}{\sqrt{n}}$ where s is the standard deviation and n is the number of measurements reported.

REFERENCES

- [1] INTERNATIONAL ATOMIC ENERGY AGENCY, Worldwide and regional laboratory Comparison on the Determination of Organochlorine Compounds, Polybrominated Diphenyl Ethers and Petroleum Hydrocarbons in IAEA 451 clam (*glaucum tumidum*) sample, IAEA Analytical Quality in Nuclear Applications Series No. 28 (IAEA/AQ/28), IAEA, Vienna (2013).
- [2] IAEA-MARINE ENVIRONMENT LABORATORY. World-wide and regional intercomparison for the determination of organochlorine compounds and petroleum hydrocarbons in tuna homogenate IAEA-435. Report No.78 (2006).
- [3] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION, ISO Guide 35:2017, Reference Materials – Guidance for characterization and assessment of homogeneity and stability, ISO, Geneva (2017).
- [4] LINSINGER T., PAUWELS J., VAN DER VEEN A., SCHIMMEL H., LAMBERTY A., Homogeneity and stability of reference materials, *Accredit. Qual. Assur.* 6 1 (2001) 20–25.
- [5] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION, ISO 17034: 2016, General requirements for the competence of reference material producers, ISO, Geneva (2016).
- [6] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION, Statistical methods for use in proficiency testing by interlaboratory comparisons, ISO 13528:2005, ISO, Geneva (2005).
- [7] ROYAL SOCIETY OF CHEMISTRY, Statistical Subcommittee of the Analytical Methods Committee (AMC), AMC Technical Brief: Representing data distributions with Kernel density estimates” 2006; [http:// www.rsc.org/amc](http://www.rsc.org/amc).
- [8] THOMPSON, M., ELLISON, M.S., WOOD, R., The international harmonized protocol for the proficiency testing of analytical chemistry laboratories, *Pure Appl. Chem.* **78** (2006) 145-196.

LIST OF PARTICIPANTS

ALBANIA

Vanela PRIFTI

PCBs, OCs Pest.

National Environmental Agency

Rruga "Sami Frasheri" Godina Nr.23

Prane Liqenit Artificial

1001 Tirana

Elda MARKU

PCBs, OCs Pest., PAHs

University of Tirana

Faculty of Natural Sciences

Chemistry Department

Bulevardi Zogu I, 25/1

1000 Tirana

BOSNIA & HERCEGOVINA

Maja STOJANOVIC

PCBs, PAHs

Institut za vode d.o.o. Bijeljina

Ul. Miloša Obilića 51

76300 Bijeljina

BULGARIA

Vesela GENINA

BDEs

Executive Environment Agency (ExEA)

Regional Laboratory Plovdiv

1 Perushtitsa street

4004 Plovdiv

Stela GININ

PCBs, OCs Pest.

Executive Environment Agency

Regional Laboratory

"Stara Planina", Nr.2

6000 Stara Zagora

CANADA

Marivie CEPADA-LEUCEA
Ministry of Environment
Laboratory Services Branch
Conservation and Parks
125 Resources Road
M9P 3V6 Etobicoke

BDEs

COLOMBIA

Robinson Fidel CASANOVA
Centro de Investigaciones Oceanograficas e Hidrograficas del Pacifico (CCCP)
Via "El Morro"
Capitania de Puerto
528501 Tumaco Narino

PAHs

CROATIA

Ivana STRBINIC-ZIDARIC
Institute of Public Health of the County of Istria (ZZJZIZ)
Zavod za javno zdravstvo Istarske županije
Nazorova 23
52100 Pula

PCBs

Marijan AHEL
Ruder Boskovic Institute
Division for Marine and Environmental Research
Bijenicka c. 54
1000 Zagreb

PCBs, OCs Pest., BDEs, PAHs

Paula ZURGA
Teaching Institute of Public Health of PGZ
Kresimirova 52a
51000 Rijeka

PCBs, OCs Pest., BDEs, PAHs

ESTONIA

Urmas MUINASMAA
Estonian Environmental Research Centre
Marja 4d
10617 Tallinn

PCBs, OCs Pest., PAHs

FRANCE

Emmanuel WAFO
Aix-Marseille Université
Faculté de Pharmacie de Marseille
Laboratoire de chimie analytique
27, Bd Jean Moulin
13005 Marseille

PCBs, OCs Pest.

Anne-Gaelle VALADE
La Drôme Laboratoire
37, Avenue de Lautagne
B.P. 118
26904 Valence Cedex 9

PCBs, OCs Pest., BDEs, PAHs

GERMANY

Tanja REEBER
Eurofins WEJ Contaminants GmbH
Neulander Kamp 1
21079 Hamburg

PAHs

GHANA

Anthony Yau KARIKARI
CSIR
Water Research Institute
PO Box 111.32
Accra

PCBs, OCs Pest., PAHs

ISRAEL

Peleg ASTRAHAN
Israel Oceanographic & Limnological Research
Kinneret Limnological Research
PO Box 447
14950 MIGDAL

OCs Pest., PAHs

Lush Helena CERNES
Bactochem Laboratories
Haharash 18
74031 Ness Ziona
0

OCs Pest., PAHs

ITALY

Luca FANTOZZI

ARPA LAZIO RIETI

Via Salaria per l'Aquila 6/8

2100 Rieti

Gianni FORMENTON

ARPAV

Via Lissa 6

30174 Venezia Mestre

PCBs, OCs Pest., BDEs

PCBs, OCs Pest., BDEs, PAHs

Claudio Giovanni ROSCIOLI

CNR

Instituto di Ricerca Sulle Acque

Via del Mulino 19

20861 Brugherio

PCBs, OCs Pest., BDEs, PAHs

Achille PALMA

A.R.P.A.B. C.R.M.

S.S. 106 KM, 448.2

75010 Metaponto

PCBs, OCs Pest., PAHs

Giancarlo CAMPESAN

CNR - ISMAR

Arsenal Tesa 104

Castello 2737/F

30122 Venezia

PCBs, PAHs

Giulio SESTA

ISPRA, CN-LAB

Via di Castel Romano 100

00128 Roma

OCs Pest., PAHs

Paolo ALTEMURA

ARPAT

Laboratorio ARPAT

Via Giovanni Marradi, 114

57126 Livorno

PCBs, OCs Pest., BDEs

Claudia Fornasari
ARPAE Emilia Romagna
Sede Secondaria Laboratorio Multisito
Via Bologna, 534
44124 Ferrara

OCs Pest.

Ivan SCARONI
ARPAE Emilia Romagna
Vial Alberoni 17/19
48121 Ravenna

PCBs, BDEs, PAHs

Rosina ANEDDA
Laboratory ARPA Sardegna - Cagliari
Viale Ciusa, 6
09131 Cagliari

OCs Pest., BDEs

LEBANON

Carol SUKHN
Laboratories for the Environment, Agriculture, and Food (LEAF)
American University of Beirut - Hamra
Bliss St, CCC-SRB Bldg
11072020 Beirut

PCBs

LITHUANIA

Daiva POCKEVIVIUTE
Environmental Protection Agency
Environmental Research Department
Chemical Research Division
Taikos av. 26
91149 Klaipeda

PCBs, OCs Pest.

MAURITANIA

Cherif Ahmed Ahmed ELMAMY
ONISPA
Office National d'Inspection Sanitaire des Produits de la Pêche et d'Aquaculture
222 Nouadhibou

PAHs

MONACO

Tolosa Imma
IAEA-NAEL-MESL
4 Quai Antoine 1er
98000 Monaco

PCBs, OCs Pest., BDEs, PAHs

MONTENEGRO

Zeljka BULATOVIC
Center for Ecotoxicological Research Podgorica
Bulevar Sarla de Gola 2
81000 Podgorica

PCBs, BDEs, PAHs

MOROCCO

Mohammed ELBOUCH
National Laboratory of Studies and Pollution monitoring
LNESP
Avenue Mohamed Ben Abdellah Erreragui
Mandinat Al Irfane
L270 Rabat

PCBs, OCs Pest.

Mostapha BENOMAR
Institut National de Recherche Halieutique (INRH)
Centre de Tanger
Km15, Route de Ksar Sghir, Mnar
90000 Tanger

PAHs

Maria JAYED
Institut National de Recherche Halieutique
INRH
2, Boulevard Sidi Abderrahmane
20180 Casablanca

PCBs, OCs Pest.

Lalla Khadija GHEDDA
Office National de l'Electricité et de l'Eau
ONEE - Branche Eau
Station de traitement
Avenue Mohamed Belhassan El Ouazzani
10220 Rabat

PCBs, OCs Pest.

NORWAY

Bergitte REIRSEN
Institute of Marine Research (IMR)
Havforskningsinstituttet
PO Box 1870 Nordnes
NO-5817 Bergen

PCBs, OCs Pest., BDEs, PAHs

RUSSIA

Andrey SHELEPCHIKOV
(sample must be sent to private address)
Demodevskaja Str., 38-2-411
115852 Moscow

PCBs, OCs Pest., BDEs

Ludmila MALAKHOVA
A.O. Kovalevsky Institute of Biology of the Southern Seas of RAS
Moscow representative office
Room 2, Room XII, Building 3, Ave. Leninsky 38
119991 Moscow

PCBs, OCs Pest.

SLOVAKIA

Ludmila KVASNAKOVA
EL spol. s r.o.
Radlinského 17A
5201 Spišská Nová Ves

PCBs, OCs Pest., PAHs

SLOVENIA

Urška CERVEK
National Laboratory of Health, Environment and Food
Prvomajska 1
2000 Maribor

PCBs, OCs Pest.

SPAIN

Juan Antonio CAMPILLO
Instituto Espanol de Oceanografia (IEO)
Centro Oceanografico de Murcia
c/Varadero, 1
30740 San Pedro del Pinatar

PCBs, OCs Pest.

Lucia VINAS
Instituto Espanol de Oceanografia
Centro Oceanografico de Vigo
Subida a Radio Faro 50
36390 Vigo

PCBs, OCs Pest., BDEs

Pilar Gonzalez VALDES
LABAQUA, S.A.U.
Poligono Industrial Las Atalayas
C/Drama 16-18
03114 Alicante

PAHs

Silvia LACORTE
IDAEA-CSIC
Department of Environmental Chemistry
Jordi Girona 18-26
8034 Barcelona

PCBs, OCs Pest., BDEs, PAHs

SYRIA

Raounak JABOUR
Central Laboratories
Ministry of Local Administration and Environment
Kafer Sosah 17 Nesaan Street
PO Box 3773
00963 Damascus

PCBs, OCs Pest., PAHs

TUNISIA

Salma MASMOUDI
Central Laboratory for Analysis & Testing - L.C.A.E
1 Avenue Mohamed el Melki - El Omrane
1005 Tunis

PCBs

TURKEY

Hakan ATABAY
TÜBİTAK Marmara Research Center
Environment and Cleaner Production Institute
Gebze Kocaeli
PO Box 21
41470 Kocaeli
Filiz KUCUKSEZGIN

PCBs, OCs Pest., PAHs

PCBs, OCs Pest.

Dokuz Eylul University
Institute of Marine Sciences and Technology
Deniz Bilimleri ve Teknolojisi Enstitüsü
Haydar Aliyev Bulvari N°32 Inciralti
35430 Izmir

UKRAINE

Yuriy DENGA
Ukrainian Scientific Center of Ecology of the Sea
89 Frantsuzki Blvd
65009 Odessa

PCBs, OCs Pest., PAHs

USA

Gina YLITALO
Northwest Fisheries Science Center
Environmental and Fisheries Sciences Division
2725 Montlake Boulevard E
98112 Seattle, WA

PCBs, OCs Pest., BDEs, PAHs

CONTRIBUTORS TO DRAFTING AND REVIEW

| | |
|----------------------|------------------------------------|
| Cassi, R. | International Atomic Energy Agency |
| Tolosa, I. | International Atomic Energy Agency |
| Huertas, D. | International Atomic Energy Agency |
| Barilaro-Hamonic, L. | International Atomic Energy Agency |

ANNEX: DOCUMENTS SENT TO PARTICIPANTS



Atoms for Peace and Development

الوكالة الدولية للطاقة الذرية

国际原子能机构

International Atomic Energy Agency

Agence internationale de l'énergie atomique

Международное агентство по атомной энергии

Organismo Internacional de Energía Atómica

IAEA Environment Laboratories

4 Quai Antoine 1er, MC 98000 Monaco, Principality of Monaco

Phone: (+377) 97.97.72.72 • Fax: (+377) 97.97.72.73

Email: el@iaea.org • Internet: <http://www.iaea.org/environmentlaboratories>

In reply please refer to: Proficiency Test ILC-IAEA-MESL-2019-01-OC

INFORMATION SHEET: ILC-IAEA-MESL-2019-01-OC PCBs, organochlorine pesticides, PBDEs and PAHs in a fish sample.

PLEASE READ THIS INFORMATION CAREFULLY BEFORE STARTING THE ANALYSES AND FILLING IN THE ONLINE DATA REPORTING FORM!

The present exercise is specifically organized for the determination of **PCBs, organochlorine pesticides, PBDEs and PAHs in a fish sample: ILC-IAEA-MESL-2019-01-OC.**

Description of the material:

A marine fish sample was collected for interlaboratory comparison purpose. This sample was deep-frozen, freeze-dried, ground and sieved through a 250 µm stainless steel sieve.

This powder, with a particle size of less than 250 µm was homogenized by mixing in a stainless-steel rotating drum for two weeks. Then aliquots of about 30 grams were packaged into glass bottles with aluminium screw caps and sealed with Teflon tape.

Homogeneity test:

The homogeneity of the material for PCBs, organochlorine pesticides, PBDEs and PAHs was checked by determining the concentration of these compounds in several samples taken randomly in the bulk of the powder. A one-way variance analysis indicated that the material can be considered as homogenous.

Moisture content:

Since the moisture content may change with the ambient humidity and temperature, it is recommended that the water content of this material always be determined in a separate sub-sample (not taken for analysis) by drying to **constant weight at 105 °C for 8 hrs.**

All results are to be reported on a dry mass basis using the unit specified on the reporting form.

Compounds to be determined:

The participants are requested to determine as many compounds as possible from the attached list using their established analytical methods.

Analytical quality control:

Procedures of quality control and laboratory quality assurance are recommended to be applied. The results of the analyses of the quality control (QC) sample should be reported together with the results from PT sample on the same reporting forms.

Reporting of results:

1. The participants are requested to make at least **three independent determinations** and to report the results together with a short **description of the method used**. Participants are also required to provide **proof of traceability** of obtained results to the International System of Units (SI), via standard calibration solutions and CRM applied, as a part of their analytical procedures.
2. The concentrations reported must be calculated on a dry mass basis and given as ng g^{-1} , leaving as many significant figures as justified by the precision of the method used.
3. If a compound is not detected by the method used, the corresponding limit of detection should be given rather than the statement “not detected”.
4. For each class of compounds, the participants **MUST** report information about:
 - Water content (in %);
 - % Lipid;
 - Extraction technique;
 - Quantity extracted;
 - Solvent used;
 - Samples clean up;
 - Samples fractionation;
 - Detection method;
 - Injector;
 - GC or HPLC column;
 - Results quantification;
 - Statement on traceability of the obtained measurement results (standards and reference materials used);
 - QA/QC;
 - Surrogates spiked before extraction;
 - Internal standards spiked before injection;
 - Use of CRMs;
 - Use of validated method;
 - Accreditation;

5. For each compound the participants **MUST** report:

- Detection limit;
- Quantification limit;
- At least three determinations;
- Combined uncertainty;
- Expanded uncertainty;
- Coverage factor;
- Surrogate used;
- Result of CRM used for QC;
- CRM value (on certificate);

OTHER NOTES

1. About two weeks before the deadline, the organizers of the Proficiency Test will send to all participating laboratories deadline reminder and further instructions for the on-line submission of results by Email.

The deadline for returning the results is **3 March 2020**.

IAEA Environment Laboratories
Marine Environment Studies Laboratory
4 Quai Antoine 1er, MC 98000 Monaco,
Principality of Monaco
Phone: (+377) 97.97.72.72
Fax: (+377) 97.97.72.73
Email: NAEL.MESL-ORGANIC@iaea.org
Internet: <http://www.iaea.org/monaco>

List of Chlorinated pesticides and PCBs compounds to be measured:

| | |
|-----------------------|----------------|
| HCB | PCB 70 |
| α HCH | PCB 74 |
| β HCH | PCB 95 |
| γ HCH- Lindane | PCB 87 |
| <i>pp'</i> DDE | PCB 97 |
| <i>pp'</i> DDD | PCB 99 |
| <i>pp'</i> DDT | PCB 101 |
| δ HCH | PCB 105 |
| <i>op'</i> DDE | PCB 110 |
| <i>op'</i> DDD | PCB 118 |
| <i>op'</i> DDT | PCB 128 |
| Heptachlor | PCB 138 |
| Aldrin | PCB 138 (+169) |
| Dieldrin | PCB 149 |
| Endrin | PCB 146 |
| cis-Chlordane | PCB 151 |
| trans-Chlordane | PCB 153 |
| cis-Nonachlor | PCB 156 |
| trans-Nonachlor | PCB 170 |
| α Endosulfan | PCB 174 |
| β Endosulfan | PCB 177 |
| Endosulfan sulfate | PCB 180 |
| Heptachlor epoxide | PCB 180 (+193) |
| PCB 8 | PCB 183 |
| PCB 18 | PCB 187 |
| PCB 28 | PCB 194 |
| PCB 31 | PCB 195 |
| PCB 44 | PCB 201 |
| PCB 49 | PCB 206 |
| PCB 52 | PCB 209 |
| PCB 66 | |

List of Polybrominated Diphenylethers (PBDEs) to be measured:

BDE 17

BDE 28

BDE 28 + 33

BDE 47

BDE 49

BDE 66

BDE 85

BDE 99

BDE 100

BDE 153

BDE 154

BDE 183

BDE 209

List of Polycyclic Aromatic Hydrocarbons (PAHs) and aliphatic hydrocarbons to be measured:

Naphthalene

2-Methylnaphthalene

1-Methylnaphthalene

1,6-Dimethylnaphthalene + 2,7 -Dimethylnaphthalene

Biphenyl

Acenaphthene

Acenaphthylene

Fluorene

Phenanthrene

Anthracene

1-Methylphenanthrene

2-Methylphenanthrene

3,6-Dimethylphenanthrene

Fluoranthene

Pyrene

Benz(a)anthracene

Chrysene

Benzo(b)fluoranthene

Benzo(e)pyrene

Benzo(a)pyrene

Benzo(g,h,i)perylene

List of aliphatic hydrocarbons to be measured:

n-C17

Pristane

n-C18

Phytane



ORDERING LOCALLY

IAEA priced publications may be purchased from the sources listed below or from major local booksellers.

Orders for unpriced publications should be made directly to the IAEA. The contact details are given at the end of this list.

NORTH AMERICA

Bernan / Rowman & Littlefield

15250 NBN Way, Blue Ridge Summit, PA 17214, USA

Telephone: +1 800 462 6420 • Fax: +1 800 338 4550

Email: orders@rowman.com • Web site: www.rowman.com/bernan

REST OF WORLD

Please contact your preferred local supplier, or our lead distributor:

Eurospan Group

Gray's Inn House
127 Clerkenwell Road
London EC1R 5DB
United Kingdom

Trade orders and enquiries:

Telephone: +44 (0)176 760 4972 • Fax: +44 (0)176 760 1640

Email: eurospan@turpin-distribution.com

Individual orders:

www.eurospanbookstore.com/iaea

For further information:

Telephone: +44 (0)207 240 0856 • Fax: +44 (0)207 379 0609

Email: info@eurospangroup.com • Web site: www.eurospangroup.com

Orders for both priced and unpriced publications may be addressed directly to:

Marketing and Sales Unit

International Atomic Energy Agency

Vienna International Centre, PO Box 100, 1400 Vienna, Austria

Telephone: +43 1 2600 22529 or 22530 • Fax: +43 1 26007 22529

Email: sales.publications@iaea.org • Web site: www.iaea.org/publications

