IAEA/AQ/74

IAEA Analytical Quality in Nuclear Applications Series No. 74

Interlaboratory Comparison on the Determination of Trace Elements and Methyl Mercury in Sediment Sample IAEA-158A

INTERLABORATORY COMPARISON ON THE DETERMINATION OF TRACE ELEMENTS AND METHYL MERCURY IN SEDIMENT SAMPLE IAEA-158A

AFGHANISTAN ALBANIA ALGERIA ANGOLA ANTIGUA AND BARBUDA ARGENTINA ARMENIA AUSTRALIA AUSTRIA AZERBAIJAN BAHAMAS BAHRAIN BANGLADESH BARBADOS BELARUS **BELGIUM BELIZE** BENIN BOLIVIA, PLURINATIONAL STATE OF BOSNIA AND HERZEGOVINA BOTSWANA BRAZIL BRUNEI DARUSSALAM BULGARIA BURKINA FASO BURUNDI CABO VERDE CAMBODIA CAMEROON CANADA CENTRAL AFRICAN REPUBLIC **CHAD** CHILE CHINA COLOMBIA COMOROS CONGO COSTA RICA CÔTE D'IVOIRE CROATIA CUBA **CYPRUS** CZECH REPUBLIC DEMOCRATIC REPUBLIC OF THE CONGO DENMARK DJIBOUTI DOMINICA DOMINICAN REPUBLIC ECUADOR EGYPT EL SALVADOR ERITREA ESTONIA ESWATINI ETHIOPIA FIJI FINLAND FRANCE GABON GAMBIA GEORGIA

GERMANY GHANA GREECE GRENADA GUATEMALA GUINEA GUYANA HAITI HOLY SEE HONDURAS HUNGARY ICELAND INDIA INDONESIA IRAN, ISLAMIC REPUBLIC OF $IRAO$ IRELAND ISRAEL ITALY JAMAICA JAPAN JORDAN KAZAKHSTAN KENYA KOREA, REPUBLIC OF KUWAIT KYRGYZSTAN LAO PEOPLE'S DEMOCRATIC REPUBLIC LATVIA LEBANON LESOTHO LIBERIA LIBYA LIECHTENSTEIN LITHUANIA LUXEMBOURG MADAGASCAR MALAWI MALAYSIA MALI MALTA MARSHALL ISLANDS MAURITANIA MAURITIUS MEXICO MONACO MONGOLIA MONTENEGRO MOROCCO MOZAMBIQUE MYANMAR NAMIBIA NEPAL NETHERLANDS, KINGDOM OF THE NEW ZEALAND NICARAGUA NIGER NIGERIA NORTH MACEDONIA NORWAY OMAN PAKISTAN

PALAU PANAMA PAPUA NEW GUINEA PARAGUAY PERU PHILIPPINES POLAND PORTUGAL QATAR REPUBLIC OF MOLDOVA ROMANIA RUSSIAN FEDERATION RWANDA SAINT KITTS AND NEVIS SAINT LUCIA SAINT VINCENT AND THE GRENADINES SAMOA SAN MARINO SAUDI ARABIA SENEGAL SERBIA SEYCHELLES SIERRA LEONE SINGAPORE SLOVAKIA SLOVENIA SOUTH AFRICA SPAIN SRI LANKA SUDAN SWEDEN SWITZERLAND SYRIAN ARAB REPUBLIC TAJIKISTAN THAILAND TOGO TONGA TRINIDAD AND TOBAGO TUNISIA TÜRKİYE TURKMENISTAN UGANDA UKRAINE UNITED ARAB EMIRATES UNITED KINGDOM OF GREAT BRITAIN AND NORTHERN IRELAND UNITED REPUBLIC OF TANZANIA UNITED STATES OF AMERICA URUGUAY UZBEKISTAN VANUATU VENEZUELA, BOLIVARIAN REPUBLIC OF VIET NAM YEMEN ZAMBIA ZIMBABWE

The Agency's Statute was approved on 23 October 1956 by the Conference on the Statute of the IAEA held at United Nations Headquarters, New York; it entered into force on 29 July 1957. The Headquarters of the Agency are situated in Vienna. Its principal objective is "to accelerate and enlarge the contribution of atomic energy to peace, health and prosperity throughout the world''.

IAEA Analytical Quality in Nuclear Applications Series No. 74

INTERLABORATORY COMPARISON ON THE DETERMINATION OF TRACE ELEMENTS AND METHYL MERCURY IN SEDIMENT SAMPLE IAEA-158A

INTERNATIONAL ATOMIC ENERGY AGENCY VIENNA, 2024

COPYRIGHT NOTICE

All IAEA scientific and technical publications are protected by the terms of the Universal Copyright Convention as adopted in 1952 (Geneva) and as revised in 1971 (Paris). The copyright has since been extended by the World Intellectual Property Organization (Geneva) to include electronic and virtual intellectual property. Permission may be required to use whole or parts of texts contained in IAEA publications in printed or electronic form. Please see www.iaea.org/publications/rights-and-permissions for more details. Enquiries may be addressed to:

Publishing Section International Atomic Energy Agency Vienna International Centre PO Box 100 1400 Vienna, Austria tel.: +43 1 2600 22529 or 22530 email: sales.publications@iaea.org www.iaea.org/publications

For further information on this publication, please contact:

Marine Environmental Studies Laboratory Marine Environment Laboratories International Atomic Energy Agency 4a Quai Antoine 1er, 98000 Principality of Monaco

INTERLABORATORY COMPARISON ON THE DETERMINATION OF TRACE ELEMENTS AND METHYL MERCURY IN SEDIMENT SAMPLE IAEA-158A IAEA, VIENNA, 2024 IAEA/AQ/74 ISSN 2074-7659

© IAEA, 2024

Printed by the IAEA in Austria July 2024

FOREWORD

The identification of environmental pollution is based on monitoring campaigns that periodically assess the quality of seawater, marine sediments and biota samples. The reliability and comparability of analytical results produced in this context are crucial for the management of the marine environment in general, for example when taking decisions and meaningful actions in relation to remediation policies.

The IAEA provides support to Member States in the field of data quality and quality assurance by organizing interlaboratory comparisons and producing marine certified reference materials (biota and sediments) which are characterized for trace elements and the methylmercury mass fractions.

To ensure compliance with the international standard ISO/IEC 17034:2016, certified reference materials produced by the IAEA are characterized with the participation of analytical laboratories with demonstrated measurement competence. This is ensured by regularly organizing targeted interlaboratory comparisons involving these laboratories. Interlaboratory comparisons involve the comparison of participants' respective results to an assigned value, which is usually derived as a consensus value from the overall population of obtained results.

This publication summarizes the results of the interlaboratory comparison on the determination of trace elements and methylmercury in a sediment sample organized in 2022.

The IAEA is grateful to the Government of Monaco for its support and wishes to thank the participants and laboratories involved in this comparison exercise. The IAEA officer responsible for this publication was S. Azemard of the Marine Environment Laboratories.

EDITORIAL NOTE

This publication has been prepared from the original material as submitted by the contributors and has not been edited by the editorial staff of the IAEA. The views expressed remain the responsibility of the contributors and do not necessarily reflect those of the IAEA or the governments of its Member States.

Guidance and recommendations provided here in relation to identified good practices represent expert opinion but are not made on the basis of a consensus of all Member States.

Neither the IAEA nor its Member States assume any responsibility for consequences which may arise from the use of this publication. This publication does not address questions of responsibility, legal or otherwise, for acts or omissions on the part of any person.

The use of particular designations of countries or territories does not imply any judgement by the publisher, the IAEA, as to the legal status of such countries or territories, of their authorities and institutions or of the delimitation of their boundaries.

The mention of names of specific companies or products (whether or not indicated as registered) does not imply any intention to infringe proprietary rights, nor should it be construed as an endorsement or recommendation on the part of the IAEA.

The IAEA has no responsibility for the persistence or accuracy of URLs for external or third party Internet web sites referred to in this publication and does not guarantee that any content on such web sites is, or will remain, accurate or appropriate.

CONTENTS

1 INTRODUCTION

1.1 BACKGROUND

The Marine Environmental Studies Laboratory (MESL) of the International Atomic Energy Agency's Marine Environment Laboratories (NAML) has the programmatic responsibility to provide assistance to Member States' laboratories in maintaining and improving the reliability of analytical measurement results, both for trace elements and organic contaminants. This is accomplished through the provision of certified reference materials (CRM's) of marine origin, validated analytical procedures, training on the implementation of internal quality control, and through the evaluation of measurement performance via interlaboratory comparisons (ILC).

The production process of CRM's followed by MESL implies to perform a characterization exercise with laboratories with demonstrated analytical capabilities so called "experts". The demonstration of competence of collaborating laboratories is a way to ensure compliance with ISO/IEC 17034:2016 [1]. The results of ILC or Proficiency Tests (PT) provide clear information on measurement capabilities of the participating laboratories.

1.2 OBJECTIVE

The ILC presented in this report was designed to evaluate the measurement performance and analytical capabilities of laboratories already identified as "experts", new identified as potential 'experts" based on worldwide ILC's organized by MESL and laboratories that have analytical capabilities for rare earth elements (REE's). REEs are considered as emerging contaminants and are often used as tracers in geological and hydrothermal systems, but only very few marine sediments CRM's have been certified for REEs. The characterization of REEs in a candidate CRM's implies to develop a network of laboratories with demonstrated capabilities in the determination of REEs.

1.3 SCOPE

The present ILC study was designed to evaluate the measurement performance of selected laboratories for trace and rare earth elements and methyl mercury (MeHg) in sediment. The scope of this publication is to describe obtained results of mentioned ILC.

1.4 STRUCTURE

This publication is structured in five sections, section 1 being the introduction. Section 2 describes the test sample. The individual performance assessment with z and Zeta-scores is explained in section 3 and obtained results are reviewed in section 4. Section 5 provides some conclusions on results obtained in this ILC.

2 STUDY SET UP

In February 2022, invitation letters were sent to 43 laboratories from 24 Member States, which previously participated in the IAEA characterization exercises or have been selected as potential collaborators. Positive responses with intent to participate were received from 24 laboratories in 17 Member States.

In April 2022 each laboratory received one bottle of the test sample, accompanied by an information sheet. Participants were requested to determine as many elements as possible from the following list: Ag, Al, As, Ba, Ca, Cd, Ce, Co, Cr, Cu, Cs, Dy, Er, Eu, Fe, Gd, Hg, Hf, Ho, K, La, Li, Lu, MeHg, Mg, Mn, Mo, Na, Nd, Ni, Pb, Pr, Rb, Sb, Se, Sc, Sm, Sn, Sr, Ta, Tb, Th,

Ti, Tm, U, V, Y, Yb and Zn) using the analytical procedures routinely applied in their laboratories. The deadline for reporting results was set at the end of June 2022.

Participating laboratories were requested to report their results for the ILC sample together with standard and expanded uncertainties, description and results of internal quality control samples (e.g CRMs or other reference materials (RMs)), analyte recovery, detection and quantification limits, digestion and instrumental technique used.

In addition, participating laboratories were requested to answer some questions on their analytical procedure, calibration, recovery correction, uncertainties estimation, moisture determination, validation of analytical method, CRM or RM used, quality procedures and accreditation.

In total, 16 laboratories from 12 Member States, reported results back to MESL. The data submitted by the laboratories, together with the technical and statistical evaluations of the results for the requested trace elements, are included in this report. All results were treated confidentially, and each laboratory was identified with a unique confidential code number.

3 DESCRIPTION OF ILC TEST MATERIAL

The ILC test sample (IAEA-158A) is an estuarine sediment sample from the North Sea, which was characterized in a separate exercise. All details on sample preparation, homogeneity, stability, and assignment of values can be found in the certification report [2].

The assigned values used in this report to evaluate the reported results are presented in Table 1. The assigned value has a direct impact on the conclusions about the 'measurement capability' of the participating laboratories, and therefore, the most metrologically credible value should be sought. Therefore, the IAEA-158A certified values determined according to the requirements of the ISO/IEC 17034:2016 [1] and ISO/IEC guide 35:2017 [3] for Al, As, Ba, Ca, Cd, Ce, Co, Cr, Eu, Fe, K, La, Li, Mg, Mn, Na, Nd, Ni, Pb, Rb, Sm, Sr, U, V, Y, Yb have been used as assigned values in this ILC. For other analytes, mass fractions provided as information in the certificate of IAEA-158A were considered sufficiently credible if arising from at least 5 results and with associated expanded uncertainty being less than 20%. As a result, for Dy, Ho, Lu, MeHg, Sc, Ta, Tb, Th and Ti information values of IAEA-158A are considered as assigned values for this exercise. Results reported by the participating laboratories for other analytes (i.e. Ag, Cs, Cu, Er, Gd, Hg, Hf, Pr, Sb, Se, Sn and Tm) will not be further discussed in this report.

| Element | Assigned value ¹ | U $(k=2)^2$ | |
|---------------------------|-----------------------------|-----------------------|--|
| | $(mg kg-1)$ | $(mg kg-1)$ | |
| \mathbf{Al} | 52.2×10^{3} | 2.4×10^{3} | |
| As | 12.0 | 1.0 | |
| Ba | 1.031×10^{3} | 0.083×10^{3} | |
| Ca | 65.3×10^{3} | 2.8×10^{3} | |
| Cd | 0.361 | 0.043 | |
| Ce | 54.1 | 5.6 | |
| Co | 9.19 | 0.90 | |
| Cr | 77.0 | 8.2 | |
| Dy | 3.35 | 0.41 | |
| Eu | 1.098 | 0.085 | |
| $\rm Fe$ | 26.6×10^{3} | 1.2×10^{3} | |
| Ho | 0.692 | 0.074 | |
| K | 20.5×10^{3} | 1.5×10^{3} | |
| La | 28.5 | 3.1 | |
| Li | 33.5 | 3.3 | |
| Lu | 0.307 | 0.026 | |
| MeHg ³ | 1.80×10^{-3} | 0.26×10^{-3} | |
| Mg | 10.86×10^3 | 0.58×10^{3} | |
| Mn | 367 | 19 | |
| Na | 23.91×10^{3} | 0.88×10^{3} | |
| Nd | 25.1 | 2.6 | |
| Ni | 31.1 | 2.1 | |
| Pb | 41.0 | 5.2 | |
| Rb | 87.4 | 7.2 | |
| Sc | 8.20 | 0.84 | |
| Sm | 4.64 | 0.52 | |
| Sr | 478 | 54 | |
| ${\it Ta}$ | 0.984 | 0.14 | |
| Tb | 0.633 | 0.061 | |
| Th | 8.39 | 0.64 | |
| Ti | 3.29×10^{3} | 0.59×10^3 | |
| $\mathbf U$ | 2.40 | 0.33 | |
| $\ensuremath{\mathbf{V}}$ | 74.1 | 5.2 | |
| Y | 17.1 | 1.1 | |
| Yb | 2.02 | 0.24 | |
| Zn | 141 | 18 | |

TABLE 1. ASSIGNED VALUES AND UNCERTAINTY FOR THE ILC TEST SAMPLE

Analyte in italic are information values.

¹ The value is the mean of the means of sets of data, each set being obtained by a different laboratory. The information values are reported on dry mass basis and are traceable to the SI.

² The uncertainty is expressed as an expanded uncertainty with a coverage factor $k=2$, corresponding to a level of confidence of about 95%, estimated in accordance with the JCGM 100:2008 [4], and ISO/IEC Guide 35:2017 [3]. 3 as Hg.

4 EVALUATION OF ANALYTICAL PERFORMANCE

The individual laboratory performance was expressed in terms of z-scores and Zeta-scores, in accordance with ISO/IEC 17043:2010 [5].

The determination of target standard deviation σ_p , for the proficiency assessment was based on the outcome of previous ILCs organized by MESL for the same population of laboratories and similar sample matrices and was set as 12.5% of the assigned values. The appropriateness of this level of tolerated variability of results was confirmed by calculation of the robust standard deviation of the participant's results and the uncertainty of the assigned values for the respective measurands.

The z-score, which is calculated as shown in Eq. (1), defines the difference between the mean value provided by the laboratory and the reference value, expressed in the units of the target standard deviation.

$$
z = \frac{x_{lab} - x_{ass}}{\sigma_p} \tag{1}
$$

where:

xlab is the result reported by the participating laboratory

xass is the assigned value

 σ_p is the target standard deviation

The Zeta-score, which is calculated as shown in Eq. (2), demonstrates the agreement of the results reported by participating laboratories with the reference value within the respective uncertainties. The denominator in the Eq. (2) is calculated from the combined uncertainty of the assigned value and the combined uncertainty reported by the respective participant $(k=1)$.

$$
\text{Zeta} = \frac{x_{lab} - x_{ass}}{\sqrt{u_{x_{lab}}^2 + u_{x_{ass}}^2}} \tag{2}
$$

where:

xlab is the result reported by the participating laboratory

 x_{ass} is the assigned value

uxlab is the combined uncertainty reported by the participating laboratory

uxass is the combined uncertainty of the assigned value.

The interpretation of a laboratory's performance was evaluated according to the following internationally accepted limits [5]:

5 RESULTS AND DISCUSSION

Sixteen sets of data for 49 analytes were submitted by participating laboratories, comprising 312 numerical results. As explained above, z-scores and Zeta-scores were only calculated for 36 analytes (258 numerical results). Eight participating laboratories reported results for both trace element and REEs, 4 only for trace elements and 4 only for mercury and MeHg.

The reported results (mean \pm expanded uncertainty) per analyte are shown in the Appendix using a scatter plot along assigned value, assigned expanded uncertainty and 2 times standard deviation. In addition, the reported results have been displayed using Pomplot [6]. The PomPlot graphical method displays the relative deviation of individual results (x_{lab}) from the assigned value (xass) on the horizontal axis and relative uncertainties on the vertical axis. Distances (D) and uncertainties (u) are calculated as described in Eq. (5) and Eq. (6) and are expressed as a multiple of median absolute deviation (MAD) calculated with Eq. (7).

$$
D_i = x_{lab,i} - x_{ass}, \quad (i=1,...n)
$$
 (5)

$$
u = \sqrt{u_{xlab}^2 + u_{xass}^2} \tag{6}
$$

 $\text{MAD} = \text{Median} |D_i|, \quad (i = 1, ... n)$ (7) Where:

n is the number of reported values per analyte

xlab is the result reported by the participating laboratory

xass is the assigned value

 u_{xlab} is the combined uncertainty reported by the participating laboratory

uxass is the combined uncertainty of the assigned value.

As shown in Figure 1, the points on the right and left side of the graph correspond to biased results, results reported with small uncertainties are shown on the top of the graph, while points at the bottom of the graph represent results reported with large uncertainties. The assigned value (named Ref value in the graphs) and the value(s) from the organizer are also shown on graphs presented in the Appendix for comparison purpose.

FIG. 1. Interpretation of a PomPlot (adapted from [6]).

Figures 2 and 4 summarize the overall performance as defined by z-scores, by participating laboratories and by analyte respectively. Figure 3 and 5 summarize the overall performance as defined by Zeta-scores by participating laboratories and by analytes respectively.

Table 2 and Table 3 show the overall performance as defined by z-scores and Zeta-scores respectively by trace element.

FIG. 2. z-scores calculated from the results reported by the participating laboratories per laboratory. Numbers provided in the bars are the number of analytes reported.

FIG. 3. Zeta-scores calculated from the results reported by the participating laboratories per laboratory. Numbers provided in the bars are the number of analytes reported.

FIG. 4. z-scores calculated from the results reported by the participating laboratories for each analyte. Numbers provided in the bars are the number of participating laboratories reporting results.

FIG. 5. Zeta-scores calculated from the results reported by the participating laboratories for each analyte. Numbers provided in the bars arethe number of participating laboratories reporting results..

| | | Laboratory Codes | | | | | | | | | | | | | | |
|----------------|--------------|------------------|---------|---------|---------|-----|--------|--------|---------|--------|--------|---------|---------|--------|---------|--------|
| Analyte | $\mathbf{1}$ | $\overline{4}$ | τ | $8\,$ | 9 | 10 | 13 | 14 | 15 | 16 | 18 | 19 | 20 | 21 | 23 | 24 |
| \mathbf{A} l | | -0.1 | | -0.2 | | | -1.3 | | 0.1 | -0.3 | -0.6 | -0.1 | | | | |
| As | -0.4 | -0.1 | 0.2 | -0.1 | 0.4 | | | | -0.3 | | | -0.3 | 0.1 | | | |
| Ba | -1.5 | 0.1 | | -0.5 | -0.3 | | 0.0 | | 0.1 | 0.1 | | 0.2 | 0.5 | | | |
| Ca | -0.3 | | | -2.8 | $0.0\,$ | | -0.4 | | -0.3 | 0.1 | | -0.2 | 0.6 | | | |
| Cd | | | 0.5 | | 0.4 | | | -1.2 | | | -0.2 | -0.1 | | | | |
| Ce | 0.2 | 0.5 | | 3.1 | -0.4 | | | | 0.3 | | 2.5 | -1.4 | 1.3 | | | |
| Co | -0.6 | -0.5 | $0.4\,$ | 5.4 | $0.2\,$ | | -0.3 | | $0.2\,$ | 0.3 | -0.2 | $0.2\,$ | $0.7\,$ | | | |
| Cr | -0.2 | 0.1 | 0.7 | 1.5 | 0.5 | | -0.7 | -1.4 | -0.2 | -0.1 | -0.1 | -0.9 | | | | |
| Dy | | -1.3 | | -0.1 | 0.4 | | | | -0.3 | | 1.7 | | | | | |
| Eu | -0.5 | -1.8 | | 0.6 | -0.1 | | | | $0.0\,$ | | 0.9 | | -0.1 | | | |
| Fe | -0.6 | -0.2 | | 0.7 | 0.0 | | -0.6 | -0.4 | 0.3 | -0.6 | -0.4 | 0.3 | $0.7\,$ | | | |
| Ho | | | | | 0.0 | | | | -0.8 | | 1.4 | -0.5 | | | | |
| K | | | | -0.3 | -0.3 | | -1.8 | | -0.1 | -0.1 | | 0.7 | $1.0\,$ | | | |
| La | $0.0\,$ | $0.0\,$ | | $0.2\,$ | -1.0 | | | | $0.2\,$ | | 1.8 | | $0.7\,$ | | | |
| Li | | | 0.2 | | | | | | -0.9 | 0.2 | 0.3 | 0.9 | | | | |
| Lu | | $0.0\,$ | | | -0.1 | | | | -0.1 | | 0.3 | | $0.4\,$ | | | |
| MeHg | | | | | | 2.0 | | | -0.1 | | | | | -0.9 | $0.8\,$ | -0.9 |
| Mg | | 0.2 | | | -0.7 | | -0.8 | | 0.3 | -0.2 | | -0.3 | 10.5 | | | |
| Mn | | -0.2 | 1.1 | -0.4 | 0.1 | | -0.1 | -0.7 | -0.2 | -0.2 | -0.1 | 0.6 | 0.1 | | | |
| Na | -0.1 | $0.0\,$ | | 0.2 | -0.4 | | -0.8 | | -0.2 | 0.9 | | 0.4 | -0.4 | | | |
| Nd | 0.3 | 0.1 | | | -0.8 | | | | -0.2 | | 1.9 | | 1.3 | | | |
| Ni | | | $0.0\,$ | | 0.1 | | 0.7 | | -0.8 | 1.0 | -0.2 | 0.3 | | | | |
| Pb | | 2.8 | -0.1 | | 0.5 | | | -0.1 | -0.1 | -0.4 | 2.4 | $0.2\,$ | | | | |
| Rb | -0.7 | 0.0 | | $0.5\,$ | | | -0.6 | | -0.2 | | | 0.4 | -1.1 | | | |
| Sc | -0.6 | -0.5 | | 0.3 | | | | | | | | | 0.6 | | | |
| Sm | -0.1 | -0.6 | | -0.2 | -0.7 | | | | 0.0 | | | | 1.1 | | | |

TABLE 2. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (Z-SCORE) BY TRACE ELEMENT (<mark>|Z|>3, 2< | Z | <3)</mark>

| | Laboratory Codes | | | | | | | | | | | | | | | |
|----------------|-------------------------|----------------|--------|--------|--------|----|--------|--------|--------|--------|-----|--------|---------|----|----|----|
| Analyte | | $\overline{4}$ | | 8 | 9 | 10 | 13 | 14 | 15 | 16 | 18 | 19 | 20 | 21 | 23 | 24 |
| Sr | -0.6 | -0.1 | | -1.1 | -0.4 | | 0.0 | | -0.1 | 0.1 | | 0.4 | | | | |
| Ta | -1.1 | -0.8 | | 2.1 | | | | | | | | 0.1 | 0.5 | | | |
| Tb | -0.9 | -0.7 | | 1.3 | 0.0 | | | | 0.6 | | 0.9 | | -1.2 | | | |
| Th | -0.3 | -0.3 | | 0.2 | | | | | | | | | 0.3 | | | |
| T _i | | 1.0 | | 0.3 | 0.3 | | -1.0 | | 0.6 | 0.0 | | -0.4 | | | | |
| $\mathbf U$ | -0.6 | -0.1 | -0.4 | 1.4 | -0.1 | | | | 0.3 | | | -0.7 | | | | |
| V | | 0.4 | 0.5 | -0.1 | 0.5 | | -0.5 | | 0.4 | -0.1 | 0.6 | -0.2 | 2.3 | | | |
| Y | | 0.4 | | | 0.3 | | | | -0.3 | | | -0.1 | | | | |
| Yb | -0.4 | 0.1 | | 0.5 | -0.2 | | | | 0.1 | | 0.5 | | 0.8 | | | |
| Zn | 6.5 | 0.4 | -1.1 | | 0.0 | | 0.2 | -0.6 | -0.1 | -0.6 | 0.1 | 0.1 | $0.0\,$ | | | |

–
TABLE 3. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (Z-SCORE) BY TRACE ELEMENT (<mark>|Z|>3, 2< | Z | <3</mark>) cont.

TABLE 4. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (ZETA-SCORE) BY TRACE ELEMENT (|ZETA|>3, 2< | ZETA | <3)

| | | Laboratory Codes | | | | | | | | | | | | | | |
|---------|--------|-------------------------|-----|---------|--------|----|---------|--------|--------|--------|--------|--------|--------|----|----|----|
| Analyte | | 4 | | 8 | 9 | 10 | 13 | 14 | 15 | 16 | 18 | 19 | 20 | 21 | 23 | 24 |
| Al | | -0.3 | | -0.4 | | | -4.8 | | 0.2 | -1.6 | -2.9 | -0.4 | | | | |
| As | -0.9 | -0.1 | 0.1 | -0.1 | 1.0 | | | | -0.6 | | | -0.5 | 0.0 | | | |
| Ba | -3.6 | 0.1 | | -0.4 | -0.8 | | $0.0\,$ | | 0.3 | 0.3 | | 0.4 | 1.5 | | | |
| Ca | -0.9 | | | -12.8 | 0.2 | | -1.4 | | -0.9 | 0.3 | | -0.8 | 1.8 | | | |
| Cd | | | 0.4 | | 0.8 | | | -1.4 | | | -0.4 | -0.1 | | | | |
| Ce | 0.4 | 0.8 | | 1.3 | -0.8 | | | | 0.6 | | 3.7 | -2.7 | 3.1 | | | |
| Co | -1.2 | -0.9 | 0.5 | 5.2 | 0.5 | | -0.4 | | 0.4 | 0.7 | -0.4 | 0.5 | 1.7 | | | |
| Cr | -0.5 | 0.2 | 0.9 | 0.8 | 1.0 | | -1.2 | -1.3 | -0.3 | -0.3 | -0.1 | -1.9 | | | | |
| Dy | | -1.1 | | $0.0\,$ | 0.6 | | | | -0.6 | | 2.5 | | | | | |
| Eu | -1.1 | -4.1 | | 0.4 | -0.1 | | | | 0.1 | | 1.6 | | -0.2 | | | |
| Fe | -1.7 | -0.5 | | 1.1 | 0.1 | | -2.1 | -1.8 | 0.6 | -2.7 | -2.0 | 0.8 | 3.9 | | | |

| | | Laboratory Codes | | | | | | | | | | | | | | |
|----------------|--------------|------------------|---------|--------|---------|-----|---------|--------|---------|---------|---------|---------|---------|--------|---------|--------|
| Analyte | $\mathbf{1}$ | 4 | τ | $8\,$ | 9 | 10 | 13 | 14 | 15 | 16 | 18 | 19 | 20 | 21 | 23 | 24 |
| Ho | | | | | $0.0\,$ | | | | -1.7 | | 2.1 | -0.8 | | | | |
| $\rm K$ | | | | -0.8 | -1.1 | | -5.5 | | -0.2 | -0.2 | | 1.9 | 0.6 | | | |
| La | 0.1 | -0.1 | | 0.3 | -1.9 | | | | $0.4\,$ | | 2.8 | | 1.6 | | | |
| Li | | | 0.0 | | | | | | -1.7 | 0.4 | 0.5 | 1.6 | | | | |
| Lu | | 0.0 | | | -0.3 | | | | -0.1 | | 0.6 | | $1.0\,$ | | | |
| MeHg | | | | | | 1.4 | | | -0.1 | | | | | -1.5 | $0.5\,$ | -1.4 |
| Mg | | $0.3\,$ | | | -1.1 | | -2.7 | | $0.6\,$ | -0.9 | | -1.0 | 6.3 | | | |
| Mn | | -0.4 | 3.0 | -0.8 | 0.3 | | -0.2 | -1.7 | -0.7 | -1.2 | -0.2 | 1.6 | $0.5\,$ | | | |
| Na | -0.4 | 0.1 | | 0.3 | -1.6 | | -3.5 | | -0.6 | 1.9 | | 1.5 | -1.8 | | | |
| Nd | $0.5\,$ | 0.2 | | | -1.5 | | | | -0.4 | | 3.0 | | 1.8 | | | |
| Ni | | | $0.0\,$ | | 0.3 | | 2.2 | | -2.0 | 3.6 | -0.3 | 0.6 | | | | |
| Pb | | 1.9 | -0.2 | | 0.9 | | | -0.1 | -0.2 | -0.8 | 3.3 | 0.3 | | | | |
| Rb | -1.6 | -0.1 | | 0.6 | | | -1.6 | | -0.5 | | | $0.9\,$ | -3.1 | | | |
| Sc | -1.1 | -1.0 | | 0.3 | | | | | | | | | 1.4 | | | |
| \mbox{Sm} | -0.2 | -1.1 | | -0.2 | -1.3 | | | | $0.0\,$ | | | | 0.4 | | | |
| Sr | -1.2 | -0.2 | | | -0.8 | | $0.0\,$ | | -0.1 | 0.3 | | $0.7\,$ | | | | |
| Ta | -1.7 | -0.5 | | 0.4 | | | | | | | | $0.1\,$ | $0.8\,$ | | | |
| Tb | -1.9 | -1.0 | | 0.3 | -0.1 | | | | $1.4\,$ | | 1.5 | | -0.7 | | | |
| Th | -0.6 | -0.6 | | 0.3 | | | | | | | | | 1.0 | | | |
| T _i | | 1.1 | | 0.3 | $0.5\,$ | | -1.3 | | 0.7 | $0.0\,$ | | -0.5 | | | | |
| ${\bf U}$ | -0.9 | -0.1 | -0.1 | 0.9 | -0.1 | | | | $0.4\,$ | | | -1.0 | | | | |
| $\mathbf V$ | | $0.8\,$ | 0.3 | -0.2 | 1.3 | | -0.9 | | 1.1 | -0.4 | 1.1 | -0.4 | 2.0 | | | |
| $\mathbf Y$ | | $0.8\,$ | | | $0.6\,$ | | | | -0.8 | | | -0.1 | | | | |
| Yb | -0.7 | 0.1 | | 0.4 | -0.4 | | | | $0.2\,$ | | $0.7\,$ | | 1.6 | | | |
| Zn | 11.0 | 0.6 | -1.1 | | 0.0 | | 0.3 | -1.1 | -0.1 | -1.2 | 0.1 | 0.1 | -0.1 | | | |

TABLE 5. OVERALL ASSESSMENT OF LABORATORIES PERFORMANCE (ZETA-SCORE) BY TRACE ELEMENT ([|]ZETA|>3, 2< | ZETA | <3) cont.

5.1 z-SCORES

The z-scores compare the participating laboratories deviation from the assigned value with the target standard deviation σ_p for proficiency assessment. σ_p was set by the ILC organizer to 12.5 %, so the maximum acceptable deviation ($|z| \le 2$) was 25% of the assigned value.

As indicated in section 2, z-scores were only calculated for 36 analytes. As a result, out of the 16 datasets received from participating laboratories, 258 z-scores were calculated. From these 258 calculated z-scores, 95.7% were satisfactory with $|z| \le 2$, and 1.6% were considered to be unsatisfactory with $|z| > 3$. Among the 16 participating laboratories, 10 (62.5% of participating laboratories) achieved satisfactory z- scores $|z| \leq 2$ for all their reported values.

Participating laboratories with results assessed as questionable and/or unsatisfactory are encouraged to carefully check laboratory procedures and applied working instructions.

5.2 ZETA-SCORES

The Zeta-score shows the agreement of the laboratory result with the reference value considering the respective uncertainties. The denominator in Eq. (2) includes the combined uncertainties of the reference values and the reported values by the participating laboratories.

As it can be seen on Figures 1 to 5, the comparison of measurement performances evaluated with z-score and Zeta-score indicates that the number of unsatisfactory Zeta-scores is slightly higher than the number of unsatisfactory z-score (1.6% of calculated z-scores and 6% of calculated Zeta-scores). Only 6 participating laboratories (37.5% of participating laboratories) reported values which were evaluated as 100% satisfactory with both $|z|$ and $|Zeta| \leq 2$, out of which 3 are participating laboratories reporting only MeHg.

Zeta-scores include the estimation of uncertainties, so values receiving $|Zeta| > 3$ while $|z|$ < 3 could indicate an underestimation of uncertainties. In Figure 5, absolute Zeta-scores are plotted against expanded uncertainties of results reported by participants. Almost 40% of the results receiving questionable or unsatisfactory Zeta-scores have been reported with expended uncertainties lower than 5%, many being estimated as standard deviation of replicate analyses. In general, laboratories should keep in mind that uncertainties based only on the precision of measurement results (measurement standard deviation) are frequently underestimated. In many cases, they just reflect variations coming from the measurement step and usually do not include the contribution of uncertainty coming from other major contributors such as recovery, procedural blank, moisture content etc.

On the other hand (as shown in Figure 5), some results were reported with high uncertainties (i.e., above 40%) which does not appear consistent with the uncertainties reported for the same analytes measured by the same instrumental techniques.

Participating laboratories reporting values receiving $|z|$ and $|Zeta| > 3$ are encouraged to review their analytical procedures, as already mentioned in 4.1. Indeed 3 out of 4 reported results receiving an unsatisfactory z-score ($|z| > 3$) also received an unsatisfactory Zeta-score.

FIG. 6. Zeta-scores versus reported expanded uncertainties.

5.3 ANALYTICAL METHODS

Table 4 shows the distribution of values reported by different techniques as well as the number of participating laboratories being equipped with each instrumentation. Analytical methods used by participating laboratories in this ILC can be divided to three groups: nondestructive techniques (neutron activation analysis, X-ray fluorescence spectroscopy); plasma spectrometric methods (inductively coupled plasma mass spectrometry and inductively coupled plasma optical emission spectrometry) and atomic absorption spectroscopy methods. The most used instrumental methods were inductively coupled plasma mass spectrometry and neutron activation, which accounted for almost 75% of reported values and were used by 69% of the participating laboratories. This is a slightly different picture than for previous ILC's on the same matrix where typically neutron activation did not represent more than 10% of the reported results. This is related to the fact that a large part of the analytical laboratories determining REEs are using neutron activation.

5.4 REVIEW OF QUESTIONNAIRES

The regular use of a CRM as part of the internal quality control process is one way to ensure the quality of results produced in a laboratory as recommended under ISO/IEC 17025:2017 [7]. All participating laboratories reported results for a CRM along their results of the ILC sample for at least part of the reported results.

It should be noted that CRM's used by the participating laboratories are generally characterized for only part of the analytes of this exercise. As a result, 3 participating laboratories declared as not having applied fully validated methodologies to perform the measurements of the ILC sample. This underlines the importance of this exercise to help participating laboratories validate their methods for a wide range of analytes including trace elements, methylmercury and REEs.

All participating laboratories have quality control procedure in place, but only 75% of them have a quality system. Only 2 participating laboratories declare to be accredited for the determination of trace elements in marine sediments.

ILC participating laboratories were requested to report the detection limit of their analytical procedures used in this ILC. All results were reported with the associated detection and quantification limit of the applied analytical procedure.

All information reported by participating laboratories underlay their proper application of the quality control procedure and traceability concept.

6 CONCLUSIONS

The current ILC was designed to evaluate the analytical capabilities of selected laboratories. The obtained results in the ILC demonstrate the measurement capabilities of invited laboratories with very few exceptions.

More than 95% of reported values were assessed as satisfactory based on z-scores which demonstrates the accuracy of results produced by the selected laboratories. On the other hand, some results were not considered as satisfactory based on Zeta-scores (6%), indicating a tendency of under-estimating uncertainties associated with the reported results.

Participating laboratories are encouraged to carefully investigate the cause of any unsatisfactory scores (i.e., $|z|$ or $|Zeta| > 3$) and put in place the necessary corrective actions to prevent reoccurrence of the problem. This is a requirement for accreditation to ISO/IEC 17025:2017 [7].

As a post action of their participation in this ILC, participating laboratories are encouraged to contact the organizers to get more information on the above discussed points, if necessary.

APPENDIX: REPORTED RESULTS BY ELEMENTS

A.1. EVALUATION OF REPORTED DATA FOR Al

FIG. 7. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.2. EVALUATION OF REPORTED DATA FOR As

FIG. 9. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.3. EVALUATION OF REPORTED DATA FOR Ba

Summary of results:

FIG. 11. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 12. Reported results and expanded uncertainties.

A.4. EVALUATION OF REPORTED DATA FOR Ca

FIG. 13. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.5. EVALUATION OF REPORTED DATA FOR Cd

FIG. 15. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.6. EVALUATION OF REPORTED DATA FOR Ce

FIG. 17. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.7. EVALUATION OF REPORTED DATA FOR Co

FIG. 19. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.8. EVALUATION OF REPORTED DATA FOR Cr

FIG. 21. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.9. EVALUATION OF REPORTED DATA FOR Dy

FIG. 23. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.10. EVALUATION OF REPORTED DATA FOR Eu

FIG. 25. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.11. EVALUATION OF REPORTED DATA FOR Fe

FIG. 27. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.12. EVALUATION OF REPORTED DATA FOR Ho

FIG. 29. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.13. EVALUATION OF REPORTED DATA FOR K

FIG. 31. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.14. EVALUATION OF REPORTED DATA FOR La

FIG. 33. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.15. EVALUATION OF REPORTED DATA FOR Li

FIG. 35. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.16. EVALUATION OF REPORTED DATA FOR Lu

FIG. 37. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.17. EVALUATION OF REPORTED DATA FOR MeHg

FIG. 39. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\mathbf{\Phi}}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 40. Reported results and expanded uncertainties.

A.18. EVALUATION OF REPORTED DATA FOR Mg

FIG. 41. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 42. Reported results and expanded uncertainties.

A.19. EVALUATION OF REPORTED DATA FOR Mn

FIG. 43. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 44. Reported results and expanded uncertainties.

A.20. EVALUATION OF REPORTED DATA FOR Na

FIG. 45. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 46. Reported results and expanded uncertainties.

A.21. EVALUATION OF REPORTED DATA FOR Nd Summary of results:

FIG. 47. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.22. EVALUATION OF REPORTED DATA FOR Ni Summary of results:

FIG. 49. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 50. Reported results and expanded uncertainties.

A.23. EVALUATION OF REPORTED DATA FOR Pb Summary of results:

FIG. 51. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 52. Reported results and expanded uncertainties.

A.24. EVALUATION OF REPORTED DATA FOR Rb Summary of results:

FIG. 53. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 54. Reported results and expanded uncertainties.

A.25. EVALUATION OF REPORTED DATA FOR Sc Summary of results:

FIG. 55. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 56. Reported results and expanded uncertainties.

A.26. EVALUATION OF REPORTED DATA FOR Sm Summary of results:

FIG. 57. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

 X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$
FIG. 58. Reported results and expanded uncertainties.

A.27. EVALUATION OF REPORTED DATA FOR Sr Summary of results:

FIG. 59. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 60. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}}$ ± U_{lab}; ---- X_{ass} ± 2 σ_{p} ; ---- X_{ass} ± U_{ass}(k=2)

A.28. EVALUATION OF REPORTED DATA FOR Ta Summary of results:

FIG. 61. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 62. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$

A.29. EVALUATION OF REPORTED DATA FOR Tb Summary of results:

FIG. 63. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 64. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$

A.30. EVALUATION OF REPORTED DATA FOR Th Summary of results:

FIG. 65. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 66. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$

A.31. EVALUATION OF REPORTED DATA FOR Ti Summary of results:

FIG. 67. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 68. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$

A.32. EVALUATION OF REPORTED DATA FOR U Summary of results:

FIG. 69. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 70. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$

A.33. EVALUATION OF REPORTED DATA FOR V Summary of results:

FIG. 71. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

A.34. EVALUATION OF REPORTED DATA FOR Y Summary of results:

FIG. 73. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 74. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$

A.35. EVALUATION OF REPORTED DATA FOR Yb Summary of results:

FIG. 75. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 76. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}} \pm U_{\text{lab}}$; ---- $X_{\text{ass}} \pm 2\sigma_{\text{p}}$; ---- $X_{\text{ass}} \pm U_{\text{ass}}(k=2)$

A.36. EVALUATION OF REPORTED DATA FOR Zn Summary of results:

FIG. 77. PomPlot: Numbers are laboratory codes. (See Section 4. for more details).

FIG. 78. Reported results and expanded uncertainties. X_{ass} ; $\overline{\Phi}X_{\text{lab}}$ ± U_{lab}; ---- X_{ass} ± 2 σ_{p} ; ---- X_{ass} ± U_{ass}(k=2)

REFERENCES

- [1] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION, ISO 17034:2016, General requirements for the competence of reference material producers, ISO, Geneva (2016).
- [2] INTERNATIONAL ATOMICENERGY AGENCY, Certification of trace and rare earth elements mass fractions in estuarine sediment IAEA-158A, IAEA Analytical Quality in Nuclear Application Series N°73 (IAEA/AQ/73), IAEA, Vienna (2024).
- [3] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION, ISO Guide 35 2017(E)-Reference materials -- Guidance for characterization and assessment of homogeneity and stability, ISO, Geneva (2017).
- [4] JOINT COMMITTEE FOR GUIDES IN METROLOGY (JCGM), Evaluation of measurement data – Guide to the expression of uncertainty in measurement, JCGM 100: 2008 (GUM 1995 with minor corrections), (2008).
- [5] INTERNATIONAL ORGANIZATION FOR STANDARDIZATION, ISO 17043, Conformity assessment — General requirements for proficiency testing, ISO, Geneva (2010).
- [6] POMMÉ, S., An intuitive visualisation of intercomparison results applied to the KCDB, Appl. Radiat. Isot.64 (2006) 10–11.
- [7] INTERNATIONAL ORGANIZATION FORSTANDARDIZATION, ISO 17025:2017. General requirements for the competence of testing and calibration laboratories, ISO, Geneva (2017).

LIST OF PARTICIPANTS

CONTRIBUTORS TO DRAFTING AND REVIEW

ORDERING LOCALLY

IAEA priced publications may be purchased from the sources listed below or from major local booksellers.

Orders for unpriced publications should be made directly to the IAEA. The contact details are given at the end of this list.

NORTH AMERICA

Bernan / Rowman & Littlefield

15250 NBN Way, Blue Ridge Summit, PA 17214, USA Telephone: +1 800 462 6420 • Fax: +1 800 338 4550 Email: orders@rowman.com • Web site: www.rowman.com/bernan

REST OF WORLD

Please contact your preferred local supplier, or our lead distributor:

Eurospan

1 Bedford Row London WC1R 4BU United Kingdom

Trade Orders and Enquiries:

Tel: +44 (0)1235 465576 Email: trade.orders@marston.co.uk

Individual Customers:

Tel: +44 (0)1235 465577 Email: direct.orders@marston.co.uk www.eurospanbookstore.com/iaea

For further information:

Tel. +44 (0) 207 240 0856 Email: info@eurospan.co.uk www.eurospan.co.uk

Orders for both priced and unpriced publications may be addressed directly to:

Marketing and Sales Unit International Atomic Energy Agency Vienna International Centre, PO Box 100, 1400 Vienna, Austria Telephone: +43 1 2600 22529 or 22530 • Fax: +43 1 26007 22529 Email: sales.publications@iaea.org • Web site: www.iaea.org/publications

INTERNATIONAL ATOMIC ENERGY AGENCY VIENNA