

IAEA Analytical Quality in Nuclear Applications Series No. 42

IAEA-RML-2013-01

Proficiency Test for Determination of Radionuclides in Sea Water



IAEA

International Atomic Energy Agency

**IAEA-RML-2013-01 PROFICIENCY TEST
FOR DETERMINATION OF
RADIONUCLIDES IN SEA WATER**

The following States are Members of the International Atomic Energy Agency:

AFGHANISTAN	GERMANY	OMAN
ALBANIA	GHANA	PAKISTAN
ALGERIA	GREECE	PALAU
ANGOLA	GUATEMALA	PANAMA
ARGENTINA	GUYANA	PAPUA NEW GUINEA
ARMENIA	HAITI	PARAGUAY
AUSTRALIA	HOLY SEE	PERU
AUSTRIA	HONDURAS	PHILIPPINES
AZERBAIJAN	HUNGARY	POLAND
BAHAMAS	ICELAND	PORTUGAL
BAHRAIN	INDIA	QATAR
BANGLADESH	INDONESIA	REPUBLIC OF MOLDOVA
BELARUS	IRAN, ISLAMIC REPUBLIC OF	ROMANIA
BELGIUM	IRAQ	RUSSIAN FEDERATION
BELIZE	IRELAND	RWANDA
BENIN	ISRAEL	SAN MARINO
BOLIVIA, PLURINATIONAL STATE OF	ITALY	SAUDI ARABIA
BOSNIA AND HERZEGOVINA	JAMAICA	SENEGAL
BOTSWANA	JAPAN	SERBIA
BRAZIL	JORDAN	SEYCHELLES
BRUNEI DARUSSALAM	KAZAKHSTAN	SIERRA LEONE
BULGARIA	KENYA	SINGAPORE
BURKINA FASO	KOREA, REPUBLIC OF	SLOVAKIA
BURUNDI	KUWAIT	SLOVENIA
CAMBODIA	KYRGYZSTAN	SOUTH AFRICA
CAMEROON	LAO PEOPLE'S DEMOCRATIC REPUBLIC	SPAIN
CANADA	LATVIA	SRI LANKA
CENTRAL AFRICAN REPUBLIC	LEBANON	SUDAN
CHAD	LESOTHO	SWAZILAND
CHILE	LIBERIA	SWEDEN
CHINA	LIBYA	SWITZERLAND
COLOMBIA	LIECHTENSTEIN	SYRIAN ARAB REPUBLIC
CONGO	LITHUANIA	TAJIKISTAN
COSTA RICA	LUXEMBOURG	THAILAND
CÔTE D'IVOIRE	MADAGASCAR	THE FORMER YUGOSLAV REPUBLIC OF MACEDONIA
CROATIA	MALAWI	TOGO
CUBA	MALAYSIA	TRINIDAD AND TOBAGO
CYPRUS	MALI	TUNISIA
CZECH REPUBLIC	MALTA	TURKEY
DEMOCRATIC REPUBLIC OF THE CONGO	MARSHALL ISLANDS	UGANDA
DENMARK	MAURITANIA	UKRAINE
DJIBOUTI	MAURITIUS	UNITED ARAB EMIRATES
DOMINICA	MEXICO	UNITED KINGDOM OF GREAT BRITAIN AND NORTHERN IRELAND
DOMINICAN REPUBLIC	MONACO	UNITED REPUBLIC OF TANZANIA
ECUADOR	MONGOLIA	UNITED STATES OF AMERICA
EGYPT	MONTENEGRO	URUGUAY
EL SALVADOR	MOROCCO	UZBEKISTAN
ERITREA	MOZAMBIQUE	VANUATU
ESTONIA	MYANMAR	VENEZUELA, BOLIVARIAN REPUBLIC OF
ETHIOPIA	NAMIBIA	VIET NAM
FIJI	NEPAL	YEMEN
FINLAND	NETHERLANDS	ZAMBIA
FRANCE	NEW ZEALAND	ZIMBABWE
GABON	NICARAGUA	
GEORGIA	NIGER	
	NIGERIA	
	NORWAY	

The Agency's Statute was approved on 23 October 1956 by the Conference on the Statute of the IAEA held at United Nations Headquarters, New York; it entered into force on 29 July 1957. The Headquarters of the Agency are situated in Vienna. Its principal objective is "to accelerate and enlarge the contribution of atomic energy to peace, health and prosperity throughout the world".

IAEA-RML-2013-01 PROFICIENCY TEST
FOR DETERMINATION OF
RADIONUCLIDES IN SEA WATER

COPYRIGHT NOTICE

All IAEA scientific and technical publications are protected by the terms of the Universal Copyright Convention as adopted in 1952 (Berne) and as revised in 1972 (Paris). The copyright has since been extended by the World Intellectual Property Organization (Geneva) to include electronic and virtual intellectual property. Permission to use whole or parts of texts contained in IAEA publications in printed or electronic form must be obtained and is usually subject to royalty agreements. Proposals for non-commercial reproductions and translations are welcomed and considered on a case-by-case basis. Enquiries should be addressed to the IAEA Publishing Section at:

Marketing and Sales Unit, Publishing Section
International Atomic Energy Agency
Vienna International Centre
PO Box 100
1400 Vienna, Austria
fax: +43 1 2600 29302
tel.: +43 1 2600 22417
email: sales.publications@iaea.org
<http://www.iaea.org/books>

For further information on this publication, please contact:

IAEA Environment Laboratories, Monaco
Radiometrics Laboratory
International Atomic Energy Agency
4a Quai Antoine 1er, MC 98000
Principality of Monaco

FOREWORD

The Radiometrics Laboratory of the IAEA Environment Laboratories (NAEL) has been providing quality support products and services for the past 50 years. These include the organization of proficiency tests and laboratory comparisons, and the production of certified reference materials, including a wide range of marine sample matrices and radionuclide levels.

As part of these activities, a second proficiency test was organized in the framework of the technical cooperation project entitled Marine Benchmark Study on the Possible Impact of the Fukushima Radioactive Releases in the Asia-Pacific Region, to test the performance of participating laboratories in an analysis of radionuclides in a seawater sample. This exercise was initiated to support Member States in seawater analyses of strontium-90 and caesium isotopes in relation to the accident at the Fukushima Daiichi nuclear power plant, in March 2011, and subsequent contamination of the marine environment.

The IAEA wishes to thank all the participants and laboratories who took part in this proficiency test. The IAEA is also grateful to the Government of Monaco for its support, and to G. Bilancia for providing technical assistance.

The IAEA officers responsible for this publication were A.V. Harms, M.K. Pham and H. Nies of the IAEA Environment Laboratories.

EDITORIAL NOTE

This publication has been prepared from the original material as submitted by the contributors and has not been edited by the editorial staff of the IAEA. The views expressed remain the responsibility of the contributors and do not necessarily reflect those of the IAEA or the governments of its Member States.

This publication has not been edited by the editorial staff of the IAEA. It does not address questions of responsibility, legal or otherwise, for acts or omissions on the part of any person.

The use of particular designations of countries or territories does not imply any judgement by the publisher, the IAEA, as to the legal status of such countries or territories, of their authorities and institutions or of the delimitation of their boundaries.

The mention of names of specific companies or products (whether or not indicated as registered) does not imply any intention to infringe proprietary rights, nor should it be construed as an endorsement or recommendation on the part of the IAEA.

The contributors are responsible for having obtained the necessary permission for the IAEA to reproduce, translate or use material from sources already protected by copyrights.

The IAEA has no responsibility for the persistence or accuracy of URLs for external or third party Internet web sites referred to in this publication and does not guarantee that any content on such web sites is, or will remain, accurate or appropriate.

CONTENTS

1.	INTRODUCTION	1
2.	MATERIAL AND METHODS	1
2.1.	Proficiency test objectives	1
2.2.	Participants.....	1
2.3.	Material distribution and reporting requirements.....	1
3.	PERFORMANCE CRITERIA.....	3
3.1.	Accuracy	3
3.2.	Precision and trueness.....	3
3.3.	Final evaluation.....	3
3.4.	Comparison of participants' values with the IAEA values	4
4.	RESULTS AND DISCUSSION	5
	APPENDIX I. PERFORMANCE EVALUATION TABLES SORTED BY RADIONUCLIDE	9
	APPENDIX II. PERFORMANCE EVALUATION TABLES SORTED BY LABORATORY CODE	21
	APPENDIX III. LIST OF PARTICIPATING RCA RAS/7/021 LABORATORIES.....	33
	APPENDIX IV. LIST OF OTHER PARTICIPATING LABORATORIES	35
	CONTRIBUTORS TO DRAFTING AND REVIEW.....	39

1. INTRODUCTION

The IAEA Environment Laboratories (IAEA-EL) in Monaco and Seibersdorf regularly organize proficiency tests (PT) for radionuclides in environmental samples to support laboratories in IAEA Member States. The organization and the results of a second PT organised in the frame of the IAEA Technical Cooperation project RAS/7/021 entitled Marine Benchmark Study on the Possible Impact of the Fukushima Radioactive Releases in the Asia-Pacific Region are described in this report.

2. MATERIAL AND METHODS

2.1. PROFICIENCY TEST OBJECTIVES

This proficiency test was organized in the frame of the IAEA Technical Cooperation project RAS/7/021. This second exercise was initiated to support Member States in seawater analyses of strontium-90 and caesium isotopes, in relation to the Fukushima Daiichi nuclear power station accident in Japan (March 2011) and subsequent contamination to the marine environment. The results of this exercise allowed the participating laboratories to evaluate their performance in the analysis of the radionuclides for this sample type.

2.2. PARTICIPANTS

A total of 31 laboratories from 24 countries participated in the exercise. In addition to the 11 laboratories taking part in the IAEA Technical Cooperation project RAS/7/021 (from Australia, Bangladesh, China, Indonesia, Malaysia, New Zealand, Pakistan, Philippines, Sri Lanka, Thailand and Viet Nam), 20 other laboratories from Thailand, USA, Viet Nam and OSPAR and HELCOM contracting parties also took part in the exercise. The 11 laboratories associated with the RAS/7/021 project are indicated with the symbol * following their participation code (i.e. Participants 1*, 4*, 5*, 10*, 11*, 13*, 16*, 19*, 20*, 22* and 26*). The full list of participants is given on pages 33–37.

2.3. MATERIAL DISTRIBUTION AND REPORTING REQUIREMENTS

A sample containing 5 L of filtered and acidified Mediterranean seawater spiked by the IAEA with the radionuclides ^{90}Sr , ^{134}Cs and ^{137}Cs was distributed to the participants, with the massic activities only known to the IAEA. The massic activities were traceable to a standard provided by Czech Metrology Institute ČMI. The combined massic activities in the exercise samples were lower than the natural activity level of ^{40}K in seawater (which is approximately 12 Bq kg $^{-1}$). The ^{90}Sr , ^{134}Cs and ^{137}Cs massic activities for the samples were approximately 0.17, 0.33 and 0.42 Bq kg $^{-1}$, respectively. The sample also contained a non-active Cs-carrier at 10 mg kg $^{-1}$ in order to stabilise the solution and to avoid adsorption of Cs to the container walls. No Sr was added to the sample as non-active Sr present naturally in seawater will act as a carrier for ^{90}Sr .

The participants were required to report to the IAEA the ^{90}Sr , ^{134}Cs and ^{137}Cs massic activities (in Bq kg $^{-1}$) of the proficiency test sample combined with the associated uncertainties (also in Bq kg $^{-1}$). Additionally, the participants were asked to submit the following:

- A short description of the analytical method used for the sample analysis. The Information Sheet sent to the participants suggested ^{90}Sr precipitation from seawater as mixed Ca/Sr oxalate or carbonate follow by a standard ^{90}Sr procedure (e.g. a radiochemical procedure such as precipitation with fuming nitric acid, liquid-liquid extraction or extraction chromatography followed by a measurement technique such as gas-flow proportional counting or liquid scintillation counting). For ^{134}Cs and ^{137}Cs three methods were suggested: (i) direct gamma spectrometry, (ii) adsorption on AMP (ammonium molybdochophosphate, $(\text{NH}_4)_3\text{PO}_4\text{Mo}_{12}\text{O}_{36}$) and subsequent gamma spectrometry or (iii) adsorption on copper hexacyanoferrate ($\text{Cu}_2[\text{Fe}(\text{CN})_6]$) and subsequent gamma spectrometry;
- Type of calibration and software used for gamma ray spectrometry;
- Nuclear data used;
- An uncertainty budget for the measurement results.

The reference date for reporting massic activities was set at 1 July 2013. At this date, the ranges for the traceable massic activities in the proficiency test exercise samples sent to the participants were 0.1721–0.1788 Bq kg⁻¹⁹⁰Sr, 0.3303–0.3433 Bq kg⁻¹¹³⁴Cs and 0.416–0.432 Bq kg⁻¹¹³⁷Cs, respectively (see Table 1).

TABLE 1. IAEA ASSIGNED VALUES

Participant	⁹⁰ Sr massic activity (Bq kg ⁻¹)	¹³⁴ Cs massic activity (Bq kg ⁻¹)	¹³⁷ Cs massic activity (Bq kg ⁻¹)
1*	0.1788 ± 0.0011	0.3433 ± 0.0014	0.432 ± 0.003
2	0.1740 ± 0.0011	0.3339 ± 0.0013	0.421 ± 0.003
3	0.1768 ± 0.0011	0.3393 ± 0.0014	0.428 ± 0.003
4*	0.1743 ± 0.0011	0.3344 ± 0.0013	0.421 ± 0.003
5*	0.1759 ± 0.0011	0.3375 ± 0.0014	0.425 ± 0.003
6	0.1724 ± 0.0010	0.3308 ± 0.0013	0.417 ± 0.003
7	0.1766 ± 0.0011	0.3389 ± 0.0014	0.427 ± 0.003
8	0.1767 ± 0.0011	0.3391 ± 0.0014	0.427 ± 0.003
9	0.1758 ± 0.0011	0.3373 ± 0.0014	0.425 ± 0.003
10*	0.1753 ± 0.0011	0.3365 ± 0.0014	0.424 ± 0.003
11*	0.1770 ± 0.0011	0.3397 ± 0.0014	0.428 ± 0.003
12	0.1782 ± 0.0011	0.3420 ± 0.0014	0.431 ± 0.003
13*	0.1767 ± 0.0011	0.3391 ± 0.0014	0.427 ± 0.003
14	0.1752 ± 0.0011	0.3362 ± 0.0014	0.424 ± 0.003
15	0.1780 ± 0.0011	0.3416 ± 0.0014	0.430 ± 0.003
16*	0.1737 ± 0.0011	0.3333 ± 0.0013	0.420 ± 0.003
17	0.1721 ± 0.0010	0.3303 ± 0.0013	0.416 ± 0.003
18	0.1725 ± 0.0011	0.3311 ± 0.0013	0.417 ± 0.003
19*	0.1729 ± 0.0011	0.3318 ± 0.0013	0.418 ± 0.003
20*	0.1745 ± 0.0011	0.3348 ± 0.0013	0.422 ± 0.003
21	0.1761 ± 0.0011	0.3379 ± 0.0014	0.426 ± 0.003
22*	0.1757 ± 0.0011	0.3372 ± 0.0014	0.425 ± 0.003
23	0.1786 ± 0.0011	0.3427 ± 0.0014	0.432 ± 0.003
24	0.1748 ± 0.0011	0.3355 ± 0.0013	0.423 ± 0.003
25	0.1788 ± 0.0011	0.3431 ± 0.0014	0.432 ± 0.003
26*	0.1733 ± 0.0011	0.3326 ± 0.0013	0.419 ± 0.003
27	0.1744 ± 0.0011	0.3346 ± 0.0013	0.422 ± 0.003
28	0.1737 ± 0.0011	0.3334 ± 0.0013	0.420 ± 0.003
29	0.1754 ± 0.0011	0.3366 ± 0.0014	0.424 ± 0.003
30	0.1770 ± 0.0011	0.3397 ± 0.0014	0.428 ± 0.003
31	0.1776 ± 0.0011	0.3408 ± 0.0014	0.429 ± 0.003
32	0.1761 ± 0.0011	0.3379 ± 0.0014	0.426 ± 0.003

3. PERFORMANCE CRITERIA

The methodology adopted for this exercise was slightly updated from the existing IAEA methodology. The scoring system takes into account the accuracy, precision and trueness of the reported data and includes in the evaluation both the combined standard uncertainty of the target value and the combined standard uncertainty reported by the participating laboratories. The IAEA target values, which will be used for data evaluation, were the certified values of radionuclides at the reference date. A result must pass three tests to be assigned the status “Accepted”, otherwise it will be assigned the status “Warning” or “Not accepted”.

3.1. ACCURACY

The first step in producing a score for a result Value_{Analyst} is the estimation of the bias. The relative bias between the Analyst’s value and the IAEA target value is calculated as follows and expressed as a percentage:

$$Bias_{relative} = \frac{Value_{Analyst} - Value_{IAEA}}{Value_{IAEA}} \times 100\% \quad (1)$$

The absolute value of the relative bias was compared to the Maximal Accepted Relative Bias (MARB). Participants’ results are scored as “Pass” for accuracy when:

$$|Bias_{relative}| \leq MARB \quad (2)$$

The MARB values used in this evaluation were 15% for both ¹³⁴Cs and ¹³⁷Cs and 25% for ⁹⁰Sr.

3.2. PRECISION AND TRUENESS

The precision P for each result was calculated according to the following equation:

$$P = \sqrt{\left(\frac{unc_{IAEA}}{Value_{IAEA}}\right)^2 + \left(\frac{unc_{Analyst}}{Value_{Analyst}}\right)^2} \times 100\% \quad (3)$$

The precision P was compared to the Limit of Accepted Precision (LAP). The participants’ results are scored as “Pass” for precision when:

$$P \leq LAP \quad (4)$$

The Limit of Accepted Precision (LAP) values used in this evaluation were 15% for both ¹³⁴Cs and ¹³⁷Cs and 25% for ⁹⁰Sr.

The participants’ results for trueness were scored as “Pass” when:

$$|Bias_{relative}| \leq \frac{Value_{Analyst}}{Value_{IAEA}} 2.58 P \quad (5)$$

3.3. FINAL EVALUATION

For the final evaluation, all three scores were combined (see Table 2). The result was considered as “Accepted” if it passed all three tests. If the accuracy test was failed, the result was considered as “Not accepted”. If the accuracy test was passed but one of the other two tests was failed, the result was assigned the “Warning” status. The “Warning” status will reflect cases in which the reported result is close enough to the assigned property value, but its associated uncertainty is deemed to be either too small or too large.

TABLE 2. PERFORMANCE EVALUATION CRITERIA

Accuracy	Precision	Trueness	Final evaluation
Pass	Pass	Pass	Accepted
Pass	Fail	Pass	Warning
Pass	Pass	Fail	Warning
Fail	Pass/Fail	Pass/Fail	Not accepted

The evaluation criteria can also be illustrated by plotting the relative bias against the relative uncertainty of the participants' result (see Figure 1). In the illustrated case, a relative uncertainty of 1% for the IAEA target value is assumed with MARB and LAP limits of 15%.




FIG. 1. Visualisation of performance evaluation criteria.

The plot consists of six zones (Zone 1 “Accepted”; Zones 2, 3 and 4 “Warning”; Zones 5 and 6 “Not accepted”), whose areas are defined by the three tests used above to evaluate the data. The areas of Zones 1, 3 and 4 are finite, while the areas of Zones 2, 5 and 6 are infinite. A result located in Zone 1 passes all three tests (evaluation “Accepted”). A result located in Zone 2 fails only the precision test as its associated uncertainty is deemed to be too large (evaluation: “Warning”). A result located in Zones 3 and 4 fails only the trueness test as its associated uncertainty is deemed to be too small (evaluation: “Warning”). A result located in Zones 5 and 6 fails (at least) the accuracy test as its relative bias is larger than the MARB (evaluation: “Not accepted”).

3.4. COMPARISON OF PARTICIPANTS' VALUES WITH THE IAEA VALUES

The means and the uncertainties for the combined participants' results were calculated according to a method developed by Cox (i.e. the weighted mean of the largest consistent subset containing p results)¹ and subsequently compared with the IAEA values. The mean of the combined participants' results, $\text{Value}_{\text{Combined}}$, was tested against the IAEA value, $\text{Value}_{\text{IAEA}}$, using this equation:

$$t = \frac{\text{Value}_{\text{Combined}} - \text{Value}_{\text{IAEA}}}{\sqrt{(\text{unc}_{\text{Combined}})^2 + (\text{unc}_{\text{IAEA}})^2}} \quad (6)$$

¹ Cox, M.G., 2007. The evaluation of key comparison data: determining the largest consistent subset. Metrologia 44 187-200.

The effective degrees of freedom v_{eff} were determined with the Welch-Satterthwaite equation. The effective degrees of freedom v_{eff} were rounded and the critical value t_{crit} for this value was identified. The criterion for passing the t test was:

$$|t| < t_{\text{crit}} \quad (7)$$

If the absolute value of t was greater than the critical value t_{crit} , this indicates there was a significant difference between the combined participants' results and the IAEA value.

4. RESULTS AND DISCUSSION

Measurement results were reported by 32 laboratories from 24 countries for ^{90}Sr , ^{134}Cs and ^{137}Cs . The overall evaluation of these results showed that 69% of all reported measurement results fulfilled the PT criteria of acceptability, while 20% of the individual measurement results were not accepted with the remaining 11% having the warning status. The performance evaluation for ^{90}Sr , ^{134}Cs and ^{137}Cs is summarised in Table 3. For the subset of subset of RCA RAS/7/021 results, the values were: 52% 'Accepted', 32% 'Not accepted' and 16% 'Warning'.

TABLE 3. SUMMARY EVALUATION OF THE RADIONUCLIDES REPORTED

Radionuclide	Number of submitted results	Accepted	Warning	Not accepted
^{90}Sr	13 (5)	8 (1)	1 (0)	4 (4)
^{134}Cs	31 (10)	21 (6)	3 (2)	7 (2)
^{137}Cs	30 (10)	22 (6)	4 (2)	4 (2)

NOTE: The numbers in brackets represent the subset of RCA RAS/7/021 results

For ^{134}Cs there was a significant negative bias of the combined participants' results (-3.8%), while there was no significant bias for both ^{90}Sr (-0.5%) and ^{137}Cs (-2.0%) (see Table 4). For ^{90}Sr and ^{137}Cs about half of the participants' results showed a negative bias (62% and 46% of the results, respectively), while for ^{134}Cs the majority of participants' results (81% of the results) showed a negative bias. Coincidence summing is a problem for ^{134}Cs as it leads to signal loss and hence underestimation of the activity levels for this radionuclide. It is clear from the results that some participants did not make a sufficient correction for coincidence summing.

TABLE 4. COMPARISON OF THE COMBINED PARTICIPANTS' RESULTS WITH THE IAEA VALUE

Nuclide	Combined participants' result (Bq kg^{-1})	IAEA value (Bq kg^{-1})	Size of the LCS ^{\$} (%)	t-value	Critical t-value	Bias (%)
^{90}Sr	0.174 ± 0.006	0.175 ± 0.004	9 (69%)	-0.24	2.16	-0.5
^{134}Cs	0.321 ± 0.005	0.337 ± 0.007	26 (84%)	-3.78	2.03	-4.9
^{137}Cs	0.416 ± 0.006	0.425 ± 0.009	25 (83%)	-1.51	2.02	-2.0

^{\$}Size of the LCS (largest consistent subset) denotes the percentage of the returned results contributing to combined participants' results

Table 5 summarises the data evaluation sorted by laboratory code. Six laboratories (Participants 2, 6, 8, 11*, 17 and 30) scored 'Accepted' ratings for all three radionuclides. Eleven laboratories (Participants 1*, 7, 13*, 14, 15, 16*, 18, 21, 27, 28 and 29) scored two 'Accepted' ratings (when no ^{90}Sr results were submitted). A full overview of all ratings is given in Table 6.

TABLE 5. SUMMARY EVALUATION

Lab code	⁹⁰ Sr	¹³⁴ Cs	¹³⁷ Cs
1*		A	A
2	A	A	A
3	W	W	W
4*	NA	W	NA
5*	NA	NA	A
6	A	A	A
7		A	A
8	A	A	A
9	A	A	W
10*	NA	A	A
11*	A	A	A
12		NA	NA
13*		A	A
14		A	A
15		A	A
16*		A	A
17	A	A	A
18		A	A
19*	NA	NA	NA
20*		W	W
21		A	A
22*		A	W
23		NA	A
24		NA	
25		NA	A
26*			
27		A	A
28		A	A
29		A	A
30	A	A	A
31	A	A	NA
32		NA	A

TABLE 6. OVERVIEW OF PERFORMANCE

Rating	Number of participants	Lab code
3 A	6	2, 6, 8, 11*, 17 and 30
2 A	11	1*, 7, 13*, 14, 15, 16*, 18, 21, 27, 28 and 29
2 A and 1 W	1	9
2 A and 1 NA	2	10* and 31
1 A and 1 W	1	22*
1 A and 1 NA	3	23, 25 and 32
1 A and 2 NA	1	5*
2 W	1	20*
3 W	1	3
1 NA	1	24
2 NA	1	12
1 W and 2 NA	1	4*
3 NA	1	19*

The performance evaluation sorted by radionuclide and the bias plots are presented in Appendix I (see Tables 7–9 and Figures 2–7). In the bias plots, the ‘Accepted’ results are represented by dark blue points. ‘Warning’ and ‘Not accepted’ results are represented by the yellow and red points, respectively. The error bars represent the standard uncertainties of the bias (with a coverage factor of $k = 1$). The dotted lines represent a relative bias of $\pm 25\%$ or $\pm 15\%$ (Maximum Accepted Bias for ^{90}Sr and $^{134}\text{Cs}/^{137}\text{Cs}$, respectively). The performance evaluation sorted by laboratory code is presented in Appendix II. All laboratories reported their values with expanded uncertainties ($k = 2$). However, a large spread in the submitted uncertainties was observed with relative uncertainties (at $k = 2$) ranging from 3.4% to 37% for ^{90}Sr , from 1.1% to 25% for ^{134}Cs and from 4.0% to 34% for ^{137}Cs . However, most reported relative uncertainties were in the range 7% to 24% for ^{90}Sr and 4% to 14% for both ^{134}Cs and ^{137}Cs .

For Sr-90, the majority of the participants used gas-flow proportional counting (GPC) as the analysis technique, while three laboratories (Participants 8, 11* and 30) used liquid scintillation counting (LSC), in all cases after chemical separation of Sr from the seawater matrix. Chemical separation techniques used included precipitations of Sr-oxalate or Sr-carbonate, cation-exchange chromatography, Sr-extraction chromatography and direct liquid-liquid extraction of Y-90. The results based on GPC show a larger variability (the range of the relative bias was between –37% and 81%, although two GPC results [Participants 17 and 31, who both used direct liquid-liquid extraction of Y-90] had a relative bias of less than $\pm 1.5\%$) than the three results based on LSC (these three results all had a relative bias of less than –4%).

For ^{134}Cs and ^{137}Cs , the majority of the participants measured the seawater samples directly with gamma spectrometry, while a minority of laboratories either used a pre-concentration technique (e.g. adsorption on either ammonium molybdophosphate or hexacyanoferrates containing either copper, nickel or cobalt) to separate the caesium radionuclides from the seawater matrix followed by gamma spectrometry (Participants 4*, 10*, 18, 20*, 28, 29 and 31) or they used a combination of the two techniques (Participants 3, 11* and 13*). No significant difference in the performance between the two techniques is apparent although in general the reported relative uncertainties (at $k = 2$) for the direct measurements were somewhat larger, typically around 12% for the direct measurements and around 9% for the pre-concentration methods.

APPENDIX I. PERFORMANCE EVALUATION TABLES SORTED BY RADIONUCLIDE

TABLE 7. EVALUATION RESULTS FOR ^{90}Sr

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
1*	—	—	0.1788	0.0011	—	—	—	—	—	—	Not reported
2	0.18	0.02	0.1740	0.0011	3.5	11	30	Pass	Pass	Pass	Accepted
3	0.208	0.008	0.1768	0.0011	18	3.7	11	Pass	Pass	Fail	Warning
4*	0.130	0.024	0.1743	0.0011	-25	18	36	Fail	Pass	Pass	Not accepted
5*	0.110	0.005	0.1759	0.0011	-37	4.6	7.5	Fail	Pass	Fail	Not accepted
6	0.162	0.020	0.1724	0.0010	-6.0	12	29	Pass	Pass	Pass	Accepted
7	—	—	0.1766	0.0011	—	—	—	—	—	—	Not reported
8	0.17	0.02	0.1767	0.0011	-3.8	12	29	Pass	Pass	Pass	Accepted
9	0.205	0.016	0.1758	0.0011	17	7.8	23	Pass	Pass	Pass	Accepted
10*	0.317	0.016	0.1753	0.0011	81	4.9	23	Fail	Pass	Fail	Not accepted
11*	0.174	0.010	0.1770	0.0011	-1.7	5.8	15	Pass	Pass	Pass	Accepted
12	—	—	0.1782	0.0011	—	—	—	—	—	—	Not reported
13*	—	—	0.1767	0.0011	—	—	—	—	—	—	Not reported
14	—	—	0.1752	0.0011	—	—	—	—	—	—	Not reported
15	—	—	0.1780	0.0011	—	—	—	—	—	—	Not reported
16*	—	—	0.1737	0.0011	—	—	—	—	—	—	Not reported

continues

continued

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
17	0.174	0.003	0.1721	0.0010	1.1	1.8	4.8	Pass	Pass	Pass	Accepted
18	—	—	0.1725	0.0011	—	—	—	—	—	—	Not reported
19*	0.129	0.015	0.1729	0.0011	-25	11	22	Fail	Pass	Fail	Not accepted
20*	—	—	0.1745	0.0011	—	—	—	—	—	—	Not reported
21	—	—	0.1761	0.0011	—	—	—	—	—	—	Not reported
22*	—	—	0.1757	0.0011	—	—	—	—	—	—	Not reported
23	—	—	0.1786	0.0011	—	—	—	—	—	—	Not reported
24	—	—	0.1748	0.0011	—	—	—	—	—	—	Not reported
25	—	—	0.1788	0.0011	—	—	—	—	—	—	Not reported
26*	—	—	0.1733	0.0011	—	—	—	—	—	—	Not reported
27	—	—	0.1744	0.0011	—	—	—	—	—	—	Not reported
28	—	—	0.1737	0.0011	—	—	—	—	—	—	Not reported
29	—	—	0.1754	0.0011	—	—	—	—	—	—	Not reported
30	0.174	0.010	0.1770	0.0011	-1.7	5.5	14	Pass	Pass	Pass	Accepted
31	0.175	0.006	0.1776	0.0011	-1.4	3.6	9.0	Pass	Pass	Pass	Accepted
32	—	—	0.1761	0.0011	—	—	—	—	—	—	Not reported

* RCA RAS/7/021 participant

TABLE 8. EVALUATION RESULTS FOR ^{134}Cs

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
1*	0.342	0.024	0.3433	0.0014	-0.4	6.9	18	Pass	Pass	Pass	Accepted
2	0.332	0.027	0.3339	0.0013	-0.6	8.1	21	Pass	Pass	Pass	Accepted
3	0.304	0.007	0.3393	0.0014	-10	2.2	5.0	Pass	Pass	Fail	Warning
4*	0.314	0.006	0.3344	0.0013	-6.2	1.9	4.6	Pass	Pass	Fail	Warning
5*	0.28	0.02	0.3375	0.0014	-17	7.2	15	Fail	Pass	Fail	Not accepted
6	0.330	0.013	0.3308	0.0013	-0.2	3.8	9.8	Pass	Pass	Pass	Accepted
7	0.335	0.019	0.3389	0.0014	-1.2	5.5	14	Pass	Pass	Pass	Accepted
8	0.320	0.025	0.3391	0.0014	-5.6	7.8	19	Pass	Pass	Pass	Accepted
9	0.32	0.02	0.3373	0.0014	-5.1	6.3	15	Pass	Pass	Pass	Accepted
10*	0.304	0.014	0.3365	0.0014	-9.7	4.5	10	Pass	Pass	Pass	Accepted
11*	0.333	0.022	0.3397	0.0014	-2.0	6.5	16	Pass	Pass	Pass	Accepted
12	0.285	0.015	0.3420	0.0014	-17	5.1	11	Fail	Pass	Fail	Not accepted
13*	0.32	0.02	0.3391	0.0014	-5.6	6.3	15	Pass	Pass	Pass	Accepted
14	0.33	0.01	0.3362	0.0014	-1.9	3.1	7.7	Pass	Pass	Pass	Accepted
15	0.324	0.011	0.3416	0.0014	-5.2	3.3	8.0	Pass	Pass	Pass	Accepted
16*	0.327	0.021	0.3333	0.0013	-1.9	6.4	16	Pass	Pass	Pass	Accepted
17	0.331	0.007	0.3303	0.0013	0.2	2.2	5.6	Pass	Pass	Pass	Accepted
18	0.320	0.015	0.3311	0.0013	-3.4	4.7	12	Pass	Pass	Pass	Accepted

continues

continued

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
19*	0.242	0.016	0.3318	0.0013	-27	6.6	12	Fail	Pass	Fail	Not accepted
20*	0.286	0.012	0.3348	0.0013	-15	4.2	9.3	Pass	Pass	Fail	Warning
21	0.331	0.022	0.3379	0.0014	-2.0	6.5	16	Pass	Pass	Pass	Accepted
22*	0.310	0.031	0.3372	0.0014	-8.1	9.8	23	Pass	Pass	Pass	Accepted
23	0.56	0.03	0.3427	0.0014	63	5.4	23	Fail	Pass	Fail	Not accepted
24	0.450	0.003	0.3355	0.0013	34	0.7	2.2	Fail	Pass	Fail	Not accepted
25	0.41	0.03	0.3431	0.0014	19	7.3	23	Fail	Pass	Pass	Not accepted
26*	< 1.6	-	0.3326	0.0013	-	-	-	-	-	-	Not evaluated
27	0.358	0.022	0.3346	0.0013	6.9	6.1	17	Pass	Pass	Pass	Accepted
28	0.324	0.007	0.3334	0.0013	-2.8	2.1	5.4	Pass	Pass	Pass	Accepted
29	0.330	0.007	0.3366	0.0014	-2.0	2.1	5.4	Pass	Pass	Pass	Accepted
30	0.370	0.025	0.3397	0.0014	8.9	6.8	19	Pass	Pass	Pass	Accepted
31	0.332	0.020	0.3408	0.0014	-2.6	6.0	15	Pass	Pass	Pass	Accepted
32	0.280	0.035	0.3379	0.0014	-17	13	27	Fail	Pass	Pass	Not accepted

* RCA RAS/7/021 participant

TABLE 9. EVALUATION RESULTS FOR ^{137}Cs

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
1*	0.403	0.022	0.432	0.003	-6.8	5.4	13	Pass	Pass	Pass	Accepted
2	0.385	0.031	0.421	0.003	-8.5	8.1	19	Pass	Pass	Pass	Accepted
3	0.400	0.008	0.428	0.003	-6.4	2.1	5.1	Pass	Pass	Fail	Warning
4*	0.491	0.010	0.421	0.003	17	2.2	6.4	Fail	Pass	Fail	Not accepted
5*	0.42	0.03	0.425	0.003	-1.2	7.2	18	Pass	Pass	Pass	Accepted
6	0.417	0.019	0.417	0.003	0.1	4.6	12	Pass	Pass	Pass	Accepted
7	0.409	0.024	0.427	0.003	-4.2	5.9	15	Pass	Pass	Pass	Accepted
8	0.430	0.025	0.427	0.003	0.6	5.8	15	Pass	Pass	Pass	Accepted
9	0.41	0.07	0.425	0.003	-3.5	17	42	Pass	Fail	Pass	Warning
10*	0.407	0.012	0.424	0.003	-4.0	3.0	7.5	Pass	Pass	Pass	Accepted
11*	0.433	0.026	0.428	0.003	1.2	6.0	16	Pass	Pass	Pass	Accepted
12	0.349	0.016	0.431	0.003	-19	4.6	9.7	Fail	Pass	Fail	Not accepted
13*	0.380	0.035	0.427	0.003	-11	9.2	21	Pass	Pass	Pass	Accepted
14	0.44	0.02	0.424	0.003	3.9	4.6	12	Pass	Pass	Pass	Accepted
15	0.44	0.04	0.430	0.003	1.5	9.2	24	Pass	Pass	Pass	Accepted
16*	0.424	0.020	0.420	0.003	1.0	4.8	12	Pass	Pass	Pass	Accepted
17	0.428	0.010	0.416	0.003	2.8	2.4	6.4	Pass	Pass	Pass	Accepted
18	0.40	0.02	0.417	0.003	-4.1	5.0	12	Pass	Pass	Pass	Accepted

continues

continued

Lab Code	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
19*	0.352	0.018	0.418	0.003	-16	5.0	11	Fail	Pass	Fail	Not accepted
20*	0.479	0.018	0.422	0.003	14	3.8	11	Pass	Pass	Fail	Warning
21	0.433	0.030	0.426	0.003	1.7	7.0	18	Pass	Pass	Pass	Accepted
22*	0.392	0.009	0.425	0.003	-7.7	2.3	5.4	Pass	Pass	Fail	Warning
23	0.41	0.02	0.432	0.003	-5.0	4.9	12	Pass	Pass	Pass	Accepted
24	-	-	0.423	0.003	-	-	-	-	-	-	Not reported
25	0.370	0.045	0.432	0.003	-14	12	27	Pass	Pass	Pass	Accepted
26*	< 1.3	-	0.419	0.003	-	-	-	-	-	-	Not evaluated
27	0.437	0.024	0.422	0.003	3.6	5.6	15	Pass	Pass	Pass	Accepted
28	0.430	0.009	0.420	0.003	2.4	2.2	5.8	Pass	Pass	Pass	Accepted
29	0.440	0.009	0.424	0.003	3.8	2.2	5.8	Pass	Pass	Pass	Accepted
30	0.450	0.045	0.428	0.003	5.1	10	27	Pass	Pass	Pass	Accepted
31	0.561	0.034	0.429	0.003	31	6.0	20	Fail	Pass	Fail	Not accepted
32	0.44	0.02	0.426	0.003	3.3	4.6	12	Pass	Pass	Pass	Accepted

* RCA RAS/7/021 participant




FIG. 2. *Sr-90 bias.*




FIG. 3. Sr-90 bias RCA RAS/07/021 participants.




FIG. 4. Cs-134 bias.




FIG. 5. Cs-134 bias RCA RAS/07/021 participants.




FIG. 6. Cs-137 bias.




FIG. 7. Cs-137 bias RCA RAS/07/021 participants.

APPENDIX II. PERFORMANCE EVALUATION TABLES SORTED BY LABORATORY CODE

LABORATORY CODE 1.*

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1788	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.342	0.024	0.3433	0.0014	-0.4	6.9	18	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.403	0.022	0.432	0.003	-6.8	5.4	13	Pass	Pass	Pass	Accepted

* RCA RAS/7/021 participant

LABORATORY CODE 2.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.18	0.02	0.1740	0.0011	3.5	11	30	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.332	0.027	0.3339	0.0013	-0.6	8.1	21	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.385	0.031	0.421	0.003	-8.5	8.1	19	Pass	Pass	Pass	Accepted

LABORATORY CODE 3.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.208	0.008	0.1768	0.0011	18	3.7	11	Pass	Pass	Fail	Warning
¹³⁴ Cs	0.304	0.007	0.3393	0.0014	-10	2.2	5.0	Pass	Pass	Fail	Warning
¹³⁷ Cs	0.400	0.008	0.428	0.003	-6.4	2.1	5.1	Pass	Pass	Fail	Warning

LABORATORY CODE 4*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.130	0.024	0.1743	0.0011	-25	18	36	Fail	Pass	Pass	Not accepted
¹³⁴ Cs	0.314	0.006	0.3344	0.0013	-6.2	1.9	4.6	Pass	Pass	Fail	Warning
¹³⁷ Cs	0.491	0.010	0.421	0.003	17	2.2	6.4	Fail	Pass	Fail	Not accepted

* RCA RAS/7/021 participant

LABORATORY CODE 5*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.110	0.005	0.1759	0.0011	-37	4.6	7.5	Fail	Pass	Fail	Not accepted
¹³⁴ Cs	0.28	0.02	0.3375	0.0014	-17	7.2	15	Fail	Pass	Fail	Not accepted
¹³⁷ Cs	0.42	0.03	0.425	0.003	-1.2	7.2	18	Pass	Pass	Pass	Accepted

* RCA RAS/7/021 participant

LABORATORY CODE 6.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.162	0.020	0.1724	0.0010	-6.0	12	29	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.330	0.013	0.3308	0.0013	-0.2	3.8	9.8	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.417	0.019	0.417	0.003	0.1	4.6	12	Pass	Pass	Pass	Accepted

LABORATORY CODE 7.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1766	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.335	0.019	0.3389	0.0014	-1.2	5.5	14	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.409	0.024	0.427	0.003	-4.2	5.9	15	Pass	Pass	Pass	Accepted

LABORATORY CODE 8.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.17	0.02	0.1767	0.0011	-3.8	12	29	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.320	0.025	0.3391	0.0014	-5.6	7.8	19	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.430	0.025	0.427	0.003	0.6	5.8	15	Pass	Pass	Pass	Accepted

LABORATORY CODE 9.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.205	0.016	0.1758	0.0011	17	7.8	23	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.32	0.02	0.3373	0.0014	-5.1	6.3	15	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.41	0.07	0.425	0.003	-3.5	17	42	Pass	Fail	Pass	Warning

LABORATORY CODE 10*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.317	0.016	0.1753	0.0011	81	4.9	23	Fail	Pass	Fail	Not accepted
¹³⁴ Cs	0.304	0.014	0.3365	0.0014	-9.7	4.5	10	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.407	0.012	0.424	0.003	-4.0	3.0	7.5	Pass	Pass	Pass	Accepted

* RCA RAS7/021 participant

LABORATORY CODE 11*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.174	0.010	0.1770	0.0011	-1.7	5.8	15	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.333	0.022	0.3397	0.0014	-2.0	6.5	16	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.433	0.026	0.428	0.003	1.2	6.0	16	Pass	Pass	Pass	Accepted

* RCA RAS7/021 participant

LABORATORY CODE 12.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1782	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.285	0.015	0.3420	0.0014	-17	5.1	11	Fail	Pass	Fail	Not accepted
¹³⁷ Cs	0.349	0.016	0.431	0.003	-19	4.6	9.7	Fail	Pass	Fail	Not accepted

LABORATORY CODE 13*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1767	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.32	0.02	0.3391	0.0014	-5.6	6.3	15	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.380	0.035	0.427	0.003	-11	9.2	21	Pass	Pass	Pass	Accepted

* RCA RAS/7/021 participant

LABORATORY CODE 14.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1752	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.33	0.01	0.3362	0.0014	-1.9	3.1	7.7	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.44	0.02	0.424	0.003	3.9	4.6	12	Pass	Pass	Pass	Accepted

LABORATORY CODE 15.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1780	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.324	0.011	0.3416	0.0014	-5.2	3.3	8.0	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.44	0.04	0.430	0.003	1.5	9.2	24	Pass	Pass	Pass	Accepted

LABORATORY CODE 16*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1737	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.327	0.021	0.3333	0.0013	-1.9	6.4	16	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.424	0.020	0.420	0.003	1.0	4.8	12	Pass	Pass	Pass	Accepted

* RCA RAS7/021 participant

LABORATORY CODE 17.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.174	0.003	0.1721	0.0010	1.1	1.8	4.8	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.331	0.007	0.3303	0.0013	0.2	2.2	5.6	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.428	0.010	0.416	0.003	2.8	2.4	6.4	Pass	Pass	Pass	Accepted

LABORATORY CODE 18.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1725	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.320	0.015	0.3311	0.0013	-3.4	4.7	12	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.40	0.02	0.417	0.003	-4.1	5.0	12	Pass	Pass	Pass	Accepted

LABORATORY CODE 19*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.129	0.015	0.1729	0.0011	-25	11	22	Fail	Pass	Fail	Not accepted
¹³⁴ Cs	0.242	0.016	0.3318	0.0013	-27	6.6	12	Fail	Pass	Fail	Not accepted
¹³⁷ Cs	0.352	0.018	0.418	0.003	-16	5.0	11	Fail	Pass	Fail	Not accepted

* RCA RAS/7/021 participant

LABORATORY CODE 20*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1745	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.286	0.012	0.3348	0.0013	-15	4.2	9.3	Pass	Pass	Fail	Warning
¹³⁷ Cs	0.479	0.018	0.422	0.003	14	3.8	11	Pass	Pass	Fail	Warning

* RCA RAS/7/021 participant

LABORATORY CODE 21.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1761	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.331	0.022	0.3379	0.0014	-2.0	6.5	16	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.433	0.030	0.426	0.003	1.7	7.0	18	Pass	Pass	Pass	Accepted

LABORATORY CODE 22*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1757	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.310	0.031	0.3372	0.0014	-8.1	9.8	23	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.392	0.009	0.425	0.003	-7.7	2.3	5.4	Pass	Pass	Fail	Warning

* RCA RAS/7/021 participant

LABORATORY CODE 23.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1786	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.56	0.03	0.3427	0.0014	63	5.4	23	Fail	Pass	Fail	Not accepted
¹³⁷ Cs	0.41	0.02	0.432	0.003	-5.0	4.9	12	Pass	Pass	Pass	Accepted

LABORATORY CODE 24.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1748	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.450	0.003	0.3355	0.0013	34	0.7	2.2	Fail	Pass	Fail	Not accepted
¹³⁷ Cs	—	—	0.423	0.003	—	—	—	—	—	—	Not reported

LABORATORY CODE 25.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1788	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.41	0.03	0.3431	0.0014	1.9	7.3	23	Fail	Pass	Pass	Not accepted
¹³⁷ Cs	0.370	0.045	0.432	0.003	-14	12	27	Pass	Pass	Pass	Accepted

LABORATORY CODE 26*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1733	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	< 1.6	—	0.3326	0.0013	—	—	—	—	—	—	Not evaluated
¹³⁷ Cs	< 1.3	—	0.419	0.003	—	—	—	—	—	—	Not evaluated

* RCA RAS/7/021 participant

LABORATORY CODE 27.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1744	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.358	0.022	0.3346	0.0013	6.9	6.1	17	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.437	0.024	0.422	0.003	3.6	5.6	15	Pass	Pass	Pass	Accepted

LABORATORY CODE 28.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1737	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.324	0.007	0.3334	0.0013	-2.8	2.1	5.4	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.430	0.009	0.420	0.003	2.4	2.2	5.8	Pass	Pass	Pass	Accepted

LABORATORY CODE 29.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1754	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.330	0.007	0.3366	0.0014	-2.0	2.1	5.4	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.440	0.009	0.424	0.003	3.8	2.2	5.8	Pass	Pass	Pass	Accepted

LABORATORY CODE 30.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.174	0.010	0.1770	0.0011	-1.7	5.5	14	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.370	0.025	0.3397	0.0014	8.9	6.8	19	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.450	0.045	0.428	0.003	5.1	10	27	Pass	Pass	Pass	Accepted

LABORATORY CODE 31.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	0.175	0.006	0.1776	0.0011	-1.4	3.6	9.0	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.332	0.020	0.3408	0.0014	-2.6	6.0	15	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.561	0.034	0.429	0.003	31	6.0	20	Fail	Pass	Fail	Not accepted

LABORATORY CODE 32.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
⁹⁰ Sr	—	—	0.1761	0.0011	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.280	0.035	0.33379	0.0014	-17	13	27	Fail	Pass	Pass	Not accepted
¹³⁷ Cs	0.44	0.02	0.426	0.003	3.3	4.6	12	Pass	Pass	Pass	Accepted

APPENDIX III. LIST OF PARTICIPATING RCA RAS/7/021 LABORATORIES

AUSTRALIA

Sdraulig, S.
Australian Radiation Protection and Nuclear Safety Agency (ARPANSA)
619 Lower Plenty Road
Yallambie 3085

BANGLADESH

Zafrul Kabir, M.
Atomic Energy Research Establishment (AERE)
Ganakbari, Savar
Dhaka

CHINA

Li, Y.
Laboratory of Marine Isotopic Technology and Environmental Risk Assessment
Third Institute of Oceanography
State Oceanic Administration
184# Daxue Rd, Xiamen 361005

INDONESIA

Suseno, H.
Radioactive Waste Technology Center (PTLR)
National Nuclear Energy Agency (BATAN)
Gedung 71 lantai 3 Kawasan
Puspiptek Serpong
Tangerang 15311

MALAYSIA

Kadir bin Ishak, A.
Radiochemistry and Environment Laboratory
Building 23
Malaysian Nuclear Agency
Bangi
43000 Kajang

NEW ZEALAND

Hermannspahn, N.
Institute of Environmental Science and Research
ESR Christchurch Science Centre
27 Creyke Road
Ilam Christchurch 8041

PAKISTAN

Mashiatullah, A.
Radiation and Isotope Applications Division Pinstech
P.O. Box 1482
Nilore
Islamabad

PHILIPPINES

Enriquez, E. B.
Health Physics Research Section
Philippine Nuclear Research Institute
Commonwealth Avenue, Diliman
1101 Quezon City

SRI LANKA

Waduge, V.A.
Life Sciences Division
60/460, Baseline Road
Orugodawatte
Wellampitiya (Colombo)

THAILAND

Udomsomporn, S.
Office of Atoms for Peace
16 Vibhavadi Rungsit Road
Chatuchak
Ladyao
Bangkok 10900

VIET NAM

Nhu Sieu, L.
Centre for Analytical Techniques and Environmental Monitoring
Nuclear Research Institute (NRI)
Vietnam Atomic Energy Institute
1 Nguyen Tu Luc Street
Dalat

APPENDIX IV. LIST OF OTHER PARTICIPATING LABORATORIES

BELGIUM

Sneyers, L.
Unit Low-level Radioactivity Measurements Environment Health & Safety
Belgian Nuclear Research Centre SCK CEN
Boeretang 200
2400 Mol

DENMARK

Nielsen, S.P.
Technical University of Denmark
Center for Nuclear Technologies
DTU Risø Campus Building 204
Frederiksborgvej 399
4000 Roskilde

ESTONIA

Jakobson, E.
Environmental Board Radiation Safety Department
Kopli 76
10416 Tallinn

FINLAND

Vartti, V.-P.
STUK (Radiation and Nuclear Safety Authority)
Laippatie 4
00880 Helsinki

GERMANY

Herrmann, J.
Bundesamt für Seeschifffahrt und Hydrographie (BSH)
Wüstland 2
22589 Hamburg

Kanisch, G.
Thünen-Institut für Fischereiökologie
Marckmannstraße 129b
20539 Hamburg

IRELAND

McKittrick, L.
Radiological Protection Institute of Ireland
3 Clonskeagh Square
Clonskeagh Road
Dublin 14

NETHERLANDS

Engeler, C.
Rijkswaterstaat Centre for Water Management
Zuiderwagenplein 2
8224 AD Lelystad

NORWAY

Lind, B.
Norwegian Radiation Protection Authority (NRPA)
1361 Østerås

POLAND

Zalewska, T.
Institute of Meteorology and WaterManagement
National Research Institute Maritime Branch
Waszyngtona 42
81-342 Gdynia

PORTUGAL

Madruga, M.J.
Instituto Superior Técnico
Laboratório de Proteção e Segurança Radiológica
Campus Tecnológico e Nuclear
Estrada Nacional 10 (ao km 139,7)
2695-066 Bobadela LRS

SPAIN

de Pablo Sanmartín, Á.
CEDEX (Mº FOMENTO)
C/ Alfonso XII 3 y 5
28014 Madrid

Masque, P.
Universidad Autónoma de Barcelona
Bellaterra
08193 Barcelona

SPAIN (continued)

García Tenorio, R.
Escuela Técnica Superior de Arquitectura
Avenida Reina Mercedes, 2
41012 Sevilla

SWEDEN

Eriksson, M.
Swedish Radiation Safety Authority
Solna strandväg 96
SE-17116 Stockholm

THAILAND

Chanyotha, S.
Department of Nuclear Engineering
Faculty of Engineering
Chulalongkorn University
Bangkok 10330

Pakkong, P.
Department of Applied Radiation and Isotopes
Kasetsart University
50 Phahon Yothin Rd.
Chatuchak, Ladyao
Bangkok 10900

Srisuksawad, K.
Nuclear Research and Development Group
Thailand Institute of Nuclear Technology
9/9 Saimoon, Ongkharak
Nakorn Nayok 26120

VIET NAM

Quang Long, N.
Institute for Nuclear Science and Technology
179 Hoang Quoc Viet Street
Cau Giay
Hanoi

UNITED STATES OF AMERICA

Pike, S.
Woods Hole Oceanographic Institution
266 Woods Hole Road
Woods Hole MA
02543-1050

CONTRIBUTORS TO DRAFTING AND REVIEW

Harms, A.V.	International Atomic Energy Agency
Pham, M.K.	International Atomic Energy Agency
Nies, H.	International Atomic Energy Agency



ORDERING LOCALLY

In the following countries, IAEA priced publications may be purchased from the sources listed below or from major local booksellers.

Orders for unpriced publications should be made directly to the IAEA. The contact details are given at the end of this list.

AUSTRALIA

DA Information Services

648 Whitehorse Road, Mitcham, VIC 3132, AUSTRALIA
Telephone: +61 3 9210 7777 • Fax: +61 3 9210 7788
Email: books@dadirect.com.au • Web site: <http://www.dadirect.com.au>

BELGIUM

Jean de Lannoy

Avenue du Roi 202, 1190 Brussels, BELGIUM
Telephone: +32 2 5384 308 • Fax: +32 2 5380 841
Email: jean.de.lannoy@euronet.be • Web site: <http://www.jean-de-lannoy.be>

CANADA

Renouf Publishing Co. Ltd.

5369 Canotek Road, Ottawa, ON K1J 9J3, CANADA
Telephone: +1 613 745 2665 • Fax: +1 643 745 7660
Email: order@renoufbooks.com • Web site: <http://www.renoufbooks.com>

Bernan Associates

4501 Forbes Blvd., Suite 200, Lanham, MD 20706-4391, USA
Telephone: +1 800 865 3457 • Fax: +1 800 865 3450
Email: orders@bernan.com • Web site: <http://www.bernan.com>

CZECH REPUBLIC

Suweco CZ, spol. S.r.o.

Klecakova 347, 180 21 Prague 9, CZECH REPUBLIC
Telephone: +420 242 459 202 • Fax: +420 242 459 203
Email: nakup@suweco.cz • Web site: <http://www.suweco.cz>

FINLAND

Akateeminen Kirjakauppa

PO Box 128 (Keskuskatu 1), 00101 Helsinki, FINLAND
Telephone: +358 9 121 41 • Fax: +358 9 121 4450
Email: akatilaus@akateeminen.com • Web site: <http://www.akateeminen.com>

FRANCE

Form-Edit

5 rue Janssen, PO Box 25, 75921 Paris CEDEX, FRANCE
Telephone: +33 1 42 01 49 49 • Fax: +33 1 42 01 90 90
Email: fabien.boucard@formedit.fr • Web site: <http://www.formedit.fr>

Lavoisier SAS

14 rue de Provigny, 94236 Cachan CEDEX, FRANCE
Telephone: +33 1 47 40 67 00 • Fax: +33 1 47 40 67 02
Email: livres@lavoisier.fr • Web site: <http://www.lavoisier.fr>

L'Appel du livre

99 rue de Charonne, 75011 Paris, FRANCE
Telephone: +33 1 43 07 50 80 • Fax: +33 1 43 07 50 80
Email: livres@appeldulivre.fr • Web site: <http://www.appeldulivre.fr>

GERMANY

Goethe Buchhandlung Teubig GmbH

Schweitzer Fachinformationen
Willstätterstrasse 15, 40549 Düsseldorf, GERMANY
Telephone: +49 (0) 211 49 8740 • Fax: +49 (0) 211 49 87428
Email: s.dehaan@schweitzer-online.de • Web site: <http://www.goethebuch.de>

HUNGARY

Librotade Ltd., Book Import

PF 126, 1656 Budapest, HUNGARY
Telephone: +36 1 257 7777 • Fax: +36 1 257 7472
Email: books@librotade.hu • Web site: <http://www.librotade.hu>

INDIA**Allied Publishers**

1st Floor, Dubash House, 15, J.N. Heredi Marg, Ballard Estate, Mumbai 400001, INDIA
Telephone: +91 22 2261 7926/27 • Fax: +91 22 2261 7928
Email: alliedpl@vsnl.com • Web site: <http://www.alliedpublishers.com>

Bookwell

3/79 Nirankari, Delhi 110009, INDIA
Telephone: +91 11 2760 1283/4536
Email: bkwell@nde.vsnl.net.in • Web site: <http://www.bookwellindia.com>

ITALY**Libreria Scientifica "AEIOU"**

Via Vincenzo Maria Coronelli 6, 20146 Milan, ITALY
Telephone: +39 02 48 95 45 52 • Fax: +39 02 48 95 45 48
Email: info@libreriaaeiou.eu • Web site: <http://www.libreriaaeiou.eu>

JAPAN**Maruzen Co., Ltd.**

1-9-18 Kaigan, Minato-ku, Tokyo 105-0022, JAPAN
Telephone: +81 3 6367 6047 • Fax: +81 3 6367 6160
Email: journal@maruzen.co.jp • Web site: <http://maruzen.co.jp>

NETHERLANDS**Martinus Nijhoff International**

Koraalrood 50, Postbus 1853, 2700 CZ Zoetermeer, NETHERLANDS
Telephone: +31 793 684 400 • Fax: +31 793 615 698
Email: info@nijhoff.nl • Web site: <http://www.nijhoff.nl>

Swets Information Services Ltd.

PO Box 26, 2300 AA Leiden
Dellaertweg 9b, 2316 WZ Leiden, NETHERLANDS
Telephone: +31 88 4679 387 • Fax: +31 88 4679 388
Email: tbeyens@nl.swets.com • Web site: <http://www.swets.com>

SLOVENIA**Cankarjeva Zalozba dd**

Kopitarjeva 2, 1515 Ljubljana, SLOVENIA
Telephone: +386 1 432 31 44 • Fax: +386 1 230 14 35
Email: import.books@cankarjeva-z.si • Web site: http://www.mladinska.com/cankarjeva_zalozba

SPAIN**Diaz de Santos, S.A.**

Librerias Bookshop • Departamento de pedidos
Calle Albasanz 2, esquina Hermanos Garcia Noblejas 21, 28037 Madrid, SPAIN
Telephone: +34 917 43 48 90 • Fax: +34 917 43 4023
Email: compras@diazdesantos.es • Web site: <http://www.diazdesantos.es>

UNITED KINGDOM**The Stationery Office Ltd. (TSO)**

PO Box 29, Norwich, Norfolk, NR3 1PD, UNITED KINGDOM
Telephone: +44 870 600 5552
Email (orders): books.orders@tso.co.uk • (enquiries): book.enquiries@tso.co.uk • Web site: <http://www.tso.co.uk>

UNITED STATES OF AMERICA**Bernan Associates**

4501 Forbes Blvd., Suite 200, Lanham, MD 20706-4391, USA
Telephone: +1 800 865 3457 • Fax: +1 800 865 3450
Email: orders@bernan.com • Web site: <http://www.bernan.com>

Renouf Publishing Co. Ltd.

812 Proctor Avenue, Ogdensburg, NY 13669, USA
Telephone: +1 888 551 7470 • Fax: +1 888 551 7471
Email: orders@renoufbooks.com • Web site: <http://www.renoufbooks.com>

United Nations

300 East 42nd Street, IN-919J, New York, NY 1001, USA
Telephone: +1 212 963 8302 • Fax: 1 212 963 3489
Email: publications@un.org • Web site: <http://www.unp.un.org>

Orders for both priced and unpriced publications may be addressed directly to:

IAEA Publishing Section, Marketing and Sales Unit, International Atomic Energy Agency
Vienna International Centre, PO Box 100, 1400 Vienna, Austria
Telephone: +43 1 2600 22529 or 22488 • Fax: +43 1 2600 29302
Email: sales.publications@iaea.org • Web site: <http://www.iaea.org/books>

INTERNATIONAL ATOMIC ENERGY AGENCY
VIENNA
ISSN 2074-7659