

IAEA Analytical Quality in Nuclear Applications Series No. 43

IAEA-RML-2014-01

Proficiency Test for Determination of Radionuclides in Sea Water



IAEA

International Atomic Energy Agency

**IAEA-RML-2014-01 PROFICIENCY TEST
FOR DETERMINATION OF
RADIONUCLIDES IN SEA WATER**

The following States are Members of the International Atomic Energy Agency:

AFGHANISTAN	GERMANY	OMAN
ALBANIA	GHANA	PAKISTAN
ALGERIA	GREECE	PALAU
ANGOLA	GUATEMALA	PANAMA
ARGENTINA	GUYANA	PAPUA NEW GUINEA
ARMENIA	HAITI	PARAGUAY
AUSTRALIA	HOLY SEE	PERU
AUSTRIA	HONDURAS	PHILIPPINES
AZERBAIJAN	HUNGARY	POLAND
BAHAMAS	ICELAND	PORTUGAL
BAHRAIN	INDIA	QATAR
BANGLADESH	INDONESIA	REPUBLIC OF MOLDOVA
BELARUS	IRAN, ISLAMIC REPUBLIC OF	ROMANIA
BELGIUM	IRAQ	RUSSIAN FEDERATION
BELIZE	IRELAND	RWANDA
BENIN	ISRAEL	SAN MARINO
BOLIVIA, PLURINATIONAL STATE OF	ITALY	SAUDI ARABIA
BOSNIA AND HERZEGOVINA	JAMAICA	SENEGAL
BOTSWANA	JAPAN	SERBIA
BRAZIL	JORDAN	SEYCHELLES
BRUNEI DARUSSALAM	KAZAKHSTAN	SIERRA LEONE
BULGARIA	KENYA	SINGAPORE
BURKINA FASO	KOREA, REPUBLIC OF	SLOVAKIA
BURUNDI	KUWAIT	SLOVENIA
CAMBODIA	KYRGYZSTAN	SOUTH AFRICA
CAMEROON	LAO PEOPLE'S DEMOCRATIC REPUBLIC	SPAIN
CANADA	LATVIA	SRI LANKA
CENTRAL AFRICAN REPUBLIC	LEBANON	SUDAN
CHAD	LESOTHO	SWAZILAND
CHILE	LIBERIA	SWEDEN
CHINA	LIBYA	SWITZERLAND
COLOMBIA	LIECHTENSTEIN	SYRIAN ARAB REPUBLIC
CONGO	LITHUANIA	TAJIKISTAN
COSTA RICA	LUXEMBOURG	THAILAND
CÔTE D'IVOIRE	MADAGASCAR	THE FORMER YUGOSLAV REPUBLIC OF MACEDONIA
CROATIA	MALAWI	TOGO
CUBA	MALAYSIA	TRINIDAD AND TOBAGO
CYPRUS	MALI	TUNISIA
CZECH REPUBLIC	MALTA	TURKEY
DEMOCRATIC REPUBLIC OF THE CONGO	MARSHALL ISLANDS	UGANDA
DENMARK	MAURITANIA	UKRAINE
DJIBOUTI	MAURITIUS	UNITED ARAB EMIRATES
DOMINICA	MEXICO	UNITED KINGDOM OF GREAT BRITAIN AND NORTHERN IRELAND
DOMINICAN REPUBLIC	MONACO	UNITED REPUBLIC OF TANZANIA
ECUADOR	MONGOLIA	UNITED STATES OF AMERICA
EGYPT	MONTENEGRO	URUGUAY
EL SALVADOR	MOROCCO	UZBEKISTAN
ERITREA	MOZAMBIQUE	VANUATU
ESTONIA	MYANMAR	VENEZUELA, BOLIVARIAN REPUBLIC OF
ETHIOPIA	NAMIBIA	VIET NAM
FIJI	NEPAL	YEMEN
FINLAND	NETHERLANDS	ZAMBIA
FRANCE	NEW ZEALAND	ZIMBABWE
GABON	NICARAGUA	
GEORGIA	NIGER	
	NIGERIA	
	NORWAY	

The Agency's Statute was approved on 23 October 1956 by the Conference on the Statute of the IAEA held at United Nations Headquarters, New York; it entered into force on 29 July 1957. The Headquarters of the Agency are situated in Vienna. Its principal objective is "to accelerate and enlarge the contribution of atomic energy to peace, health and prosperity throughout the world".

IAEA Analytical Quality in Nuclear Applications No. 43

**IAEA-RML-2014-01 PROFICIENCY TEST
FOR DETERMINATION OF
RADIONUCLIDES IN SEA WATER**

INTERNATIONAL ATOMIC ENERGY AGENCY
VIENNA, 2015

COPYRIGHT NOTICE

All IAEA scientific and technical publications are protected by the terms of the Universal Copyright Convention as adopted in 1952 (Berne) and as revised in 1972 (Paris). The copyright has since been extended by the World Intellectual Property Organization (Geneva) to include electronic and virtual intellectual property. Permission to use whole or parts of texts contained in IAEA publications in printed or electronic form must be obtained and is usually subject to royalty agreements. Proposals for non-commercial reproductions and translations are welcomed and considered on a case-by-case basis. Enquiries should be addressed to the IAEA Publishing Section at:

Marketing and Sales Unit, Publishing Section
International Atomic Energy Agency
Vienna International Centre
PO Box 100
1400 Vienna, Austria
fax: +43 1 2600 29302
tel.: +43 1 2600 22417
email: sales.publications@iaea.org
<http://www.iaea.org/books>

For further information on this publication, please contact:

IAEA Environment Laboratories, Monaco
Radiometrics Laboratory
International Atomic Energy Agency
4a Quai Antoine 1er, MC 98000
Principality of Monaco

FOREWORD

The Radiometrics Laboratory of the IAEA Environment Laboratories (NAEL) has been providing quality support products and services for the past 50 years. These include the organization of proficiency tests and laboratory comparisons, and the production of certified reference materials, including a wide range of marine sample matrices and radionuclide levels.

As part of these activities, a proficiency test was organized at the request of the Nuclear Regulation Authority (NRA) of Japan to test the performance of participating laboratories in an analysis of radionuclides in a seawater sample. This exercise was initiated to support laboratories in seawater analyses of tritium, strontium-90 and caesium isotopes in relation to the accident at the Fukushima Daiichi nuclear power plant, in March 2011, and subsequent contamination of the marine environment.

The IAEA wishes to thank all the participants and laboratories who took part in this proficiency test. The IAEA is also grateful to the Government of Monaco for its support.

The IAEA officers responsible for this publication were A.V. Harms, I. Osvath and D. Osborn of the IAEA Environment Laboratories.

EDITORIAL NOTE

This publication has been prepared from the original material as submitted by the contributors and has not been edited by the editorial staff of the IAEA. The views expressed remain the responsibility of the contributors and do not necessarily reflect those of the IAEA or the governments of its Member States.

This publication has not been edited by the editorial staff of the IAEA. It does not address questions of responsibility, legal or otherwise, for acts or omissions on the part of any person.

The use of particular designations of countries or territories does not imply any judgement by the publisher, the IAEA, as to the legal status of such countries or territories, of their authorities and institutions or of the delimitation of their boundaries.

The mention of names of specific companies or products (whether or not indicated as registered) does not imply any intention to infringe proprietary rights, nor should it be construed as an endorsement or recommendation on the part of the IAEA.

The contributors are responsible for having obtained the necessary permission for the IAEA to reproduce, translate or use material from sources already protected by copyrights.

The IAEA has no responsibility for the persistence or accuracy of URLs for external or third party Internet web sites referred to in this publication and does not guarantee that any content on such web sites is, or will remain, accurate or appropriate.

CONTENTS

1.	INTRODUCTION	1
2.	MATERIAL AND METHODS	1
2.1.	Proficiency test objectives	1
2.2.	Participants.....	1
2.3.	Material distribution and reporting requirements.....	1
3.	PERFORMANCE CRITERIA.....	3
3.1.	Accuracy	3
3.2.	Precision and trueness.....	3
3.3.	Final evaluation.....	3
3.4.	Comparison of participants' values with the IAEA values	4
4.	RESULTS AND DISCUSSION	5
	APPENDIX I. PERFORMANCE EVALUATION TABLES SORTED BY RADIONUCLIDE	9
	APPENDIX II. PERFORMANCE EVALUATION TABLES SORTED BY LABORATORY CODE	25
	APPENDIX III. LIST OF PARTICIPATING LABORATORIES.....	41
	CONTRIBUTORS TO DRAFTING AND REVIEW.....	45

1. INTRODUCTION

The IAEA Environment Laboratories (IAEA-EL) in Monaco and Seibersdorf regularly organize proficiency tests (PT) for radionuclides in environmental samples to support laboratories in IAEA Member States. The organization and the results of a PT organised in cooperation with the Nuclear Regulation Authority (NRA) in Japan are described in this report.

2. MATERIAL AND METHODS

2.1. PROFICIENCY TEST OBJECTIVES

This proficiency test was initiated to support laboratories in seawater analyses of tritium, strontium-90 and caesium isotopes. The results of this exercise allowed the participating laboratories to evaluate their performance in the analysis of these radionuclides for this sample type.

2.2. PARTICIPANTS

A total of 30 laboratories from 15 countries participated in this exercise during the period June – August 2014. This included 12 laboratories from Japan and 18 laboratories from IAEA Member States in Europe and the USA. The full list of participants is given on pages 41–44.

2.3. MATERIAL DISTRIBUTION AND REPORTING REQUIREMENTS

A sample containing 5 L of filtered and acidified Mediterranean seawater spiked by the IAEA with the radionuclides ^3H , ^{90}Sr , ^{134}Cs and ^{137}Cs was distributed to the participants, with the massic activities only known to the IAEA. The massic activities were traceable to a standard provided by Czech Metrology Institute ČMI. The combined massic activities in the exercise samples were lower than the natural activity level of ^{40}K in seawater (which is approximately 12 Bq kg^{-1}). The ^3H , ^{90}Sr , ^{134}Cs and ^{137}Cs massic activities for the samples were approximately 2.8, 0.35, 0.12 and 0.31 Bq kg^{-1} , respectively. The sample also contained a non-active Cs-carrier at 10 mg kg^{-1} in order to stabilise the solution and to avoid adsorption of Cs to the container walls. No Sr was added to the sample as non-active Sr present naturally in seawater will act as a carrier for ^{90}Sr .

The participants were required to report to the IAEA the ^3H , ^{90}Sr , ^{134}Cs and ^{137}Cs massic activities (in Bq kg^{-1}) of the proficiency test sample combined with the associated uncertainties (also in Bq kg^{-1}). Additionally, the participants were asked to submit the following:

- A short description of the analytical method used for the sample analysis. The Information Sheet sent to the participants suggested for ^3H distillation followed by liquid scintillation counting. For ^{90}Sr precipitation ^{90}Sr precipitation from seawater as mixed Ca/Sr oxalate or carbonate follow by a standard ^{90}Sr procedure (e.g. a radiochemical procedure such as precipitation with fuming nitric acid, liquid-liquid extraction or extraction chromatography followed by a measurement technique such as gas-flow proportional counting or liquid scintillation counting). For ^{134}Cs and ^{137}Cs three methods were suggested: (i) direct gamma spectrometry, (ii) adsorption on AMP (ammonium molybdate phosphate, $(\text{NH}_4)_3\text{PO}_4\text{Mo}_{12}\text{O}_{36}$) and subsequent gamma spectrometry or (iii) adsorption on copper hexacyanoferrate ($\text{Cu}_2[\text{Fe}(\text{CN})_6]$) and subsequent gamma spectrometry;
- Type of calibration and software used for gamma ray spectrometry;
- Nuclear data used;
- An uncertainty budget for the measurement results.

The reference date for reporting massic activities was set at 1 August 2014. At this date, the ranges for the traceable massic activities in the proficiency test exercise samples sent to the participants were $2.78\text{--}2.83 \text{ Bq kg}^{-1}$ ^3H , $0.3453\text{--}0.3515 \text{ Bq kg}^{-1}$ ^{90}Sr , $0.1154\text{--}0.1175 \text{ Bq kg}^{-1}$ ^{134}Cs and $0.3064\text{--}0.3118 \text{ Bq kg}^{-1}$ ^{137}Cs , respectively (see Table 1).

TABLE 1. IAEA ASSIGNED VALUES

Participant	^3H massic activity (Bq kg $^{-1}$)	^{90}Sr massic activity (Bq kg $^{-1}$)	^{134}Cs massic activity (Bq kg $^{-1}$)	^{137}Cs massic activity (Bq kg $^{-1}$)
1*	2.82 ± 0.06	0.3502 ± 0.0021	0.1171 ± 0.0005	0.3107 ± 0.0019
2*	2.79 ± 0.06	0.3472 ± 0.0021	0.1161 ± 0.0005	0.3080 ± 0.0019
3*	2.80 ± 0.06	0.3484 ± 0.0021	0.1165 ± 0.0005	0.3091 ± 0.0019
4*	2.81 ± 0.06	0.3492 ± 0.0021	0.1167 ± 0.0005	0.3098 ± 0.0019
5*	2.79 ± 0.06	0.3463 ± 0.0021	0.1158 ± 0.0005	0.3072 ± 0.0019
6*	2.80 ± 0.06	0.3475 ± 0.0021	0.1162 ± 0.0005	0.3083 ± 0.0019
7*	2.80 ± 0.06	0.3481 ± 0.0021	0.1164 ± 0.0005	0.3088 ± 0.0019
8*	2.81 ± 0.06	0.3489 ± 0.0021	0.1166 ± 0.0005	0.3095 ± 0.0019
9*	2.80 ± 0.06	0.3480 ± 0.0021	0.1163 ± 0.0005	0.3087 ± 0.0019
10*	2.79 ± 0.06	0.3470 ± 0.0021	0.1160 ± 0.0005	0.3079 ± 0.0019
11*	2.82 ± 0.06	0.3500 ± 0.0021	0.1170 ± 0.0005	0.3105 ± 0.0019
12*	2.81 ± 0.06	0.3492 ± 0.0021	0.1168 ± 0.0005	0.3098 ± 0.0019
13	2.82 ± 0.06	0.3502 ± 0.0021	0.1171 ± 0.0005	0.3106 ± 0.0019
14	2.82 ± 0.06	0.3509 ± 0.0021	0.1173 ± 0.0005	0.3113 ± 0.0019
15	2.79 ± 0.06	0.3465 ± 0.0021	0.1159 ± 0.0005	0.3074 ± 0.0019
16	2.80 ± 0.06	0.3486 ± 0.0021	0.1165 ± 0.0005	0.3092 ± 0.0019
17	2.81 ± 0.06	0.3486 ± 0.0021	0.1166 ± 0.0005	0.3093 ± 0.0019
18	2.78 ± 0.06	0.3453 ± 0.0021	0.1154 ± 0.0005	0.3064 ± 0.0019
19	2.79 ± 0.06	0.3468 ± 0.0021	0.1160 ± 0.0005	0.3077 ± 0.0019
20	2.79 ± 0.06	0.3468 ± 0.0021	0.1159 ± 0.0005	0.3077 ± 0.0019
21	2.82 ± 0.06	0.3500 ± 0.0021	0.1170 ± 0.0005	0.3105 ± 0.0019
22	2.81 ± 0.06	0.3490 ± 0.0021	0.1167 ± 0.0005	0.3097 ± 0.0019
23	2.81 ± 0.06	0.3498 ± 0.0021	0.1169 ± 0.0005	0.3103 ± 0.0019
24	2.80 ± 0.06	0.3485 ± 0.0021	0.1165 ± 0.0005	0.3091 ± 0.0019
25	2.80 ± 0.06	0.3475 ± 0.0021	0.1162 ± 0.0005	0.3083 ± 0.0019
26	2.81 ± 0.06	0.3492 ± 0.0021	0.1167 ± 0.0005	0.3098 ± 0.0019
27	2.83 ± 0.06	0.3515 ± 0.0021	0.1175 ± 0.0005	0.3118 ± 0.0019
28	2.79 ± 0.06	0.3470 ± 0.0021	0.1160 ± 0.0005	0.3078 ± 0.0019
29	2.81 ± 0.06	0.3499 ± 0.0021	0.1170 ± 0.0005	0.3104 ± 0.0019
30	2.81 ± 0.06	0.3493 ± 0.0021	0.1168 ± 0.0005	0.3099 ± 0.0019

* Participant from Japan

3. PERFORMANCE CRITERIA

The methodology adopted for this exercise was slightly updated from the existing IAEA methodology. The scoring system took into account the accuracy, precision and trueness of the reported data and included in the evaluation both the combined standard uncertainty of the IAEA value and the combined standard uncertainty reported by the participating laboratories. The IAEA values, which were used for the data evaluation, were the certified values of radionuclides at the reference date. A result must pass three tests to be assigned the status “Accepted”, otherwise it was assigned the status “Warning” or “Not accepted”.

3.1. ACCURACY

The first step in producing a score for a result $\text{Value}_{\text{Analyst}}$ was the estimation of the bias. The relative bias between the Analyst’s value and the IAEA target value was calculated as follows and expressed as a percentage:

$$\text{Bias}_{\text{relative}} = \frac{\text{Value}_{\text{Analyst}} - \text{Value}_{\text{IAEA}}}{\text{Value}_{\text{IAEA}}} \times 100\% \quad (1)$$

The absolute value of the relative bias was compared to the Maximal Accepted Relative Bias (MARB). Participants’ results were scored as “Pass” for accuracy when:

$$|\text{Bias}_{\text{relative}}| \leq \text{MARB} \quad (2)$$

The MARB values used in this evaluation were 20% for both ^{134}Cs and ^{137}Cs and 25% for both ^3H and ^{90}Sr .

3.2. PRECISION AND TRUENESS

The precision P for each result was calculated according to the following equation:

$$P = \sqrt{\left(\frac{\text{unc}_{\text{IAEA}}}{\text{Value}_{\text{IAEA}}}\right)^2 + \left(\frac{\text{unc}_{\text{Analyst}}}{\text{Value}_{\text{Analyst}}}\right)^2} \times 100\% \quad (3)$$

The precision P was compared to the Limit of Accepted Precision (LAP). The participants’ results were scored as “Pass” for precision when:

$$P \leq \text{LAP} \quad (4)$$

The Limit of Accepted Precision (LAP) values used in this evaluation were 20% for both ^{134}Cs and ^{137}Cs and 25% for both ^3H and ^{90}Sr .

The participants’ results for trueness were scored as “Pass” when:

$$|\text{Bias}_{\text{relative}}| \leq \frac{\text{Value}_{\text{Analyst}}}{\text{Value}_{\text{IAEA}}} 2.58 P \quad (5)$$

3.3. FINAL EVALUATION

For the final evaluation, all three scores were combined (see Table 2). The result was considered as “Accepted” if it passed all three tests. If the accuracy test was failed, the result was considered as “Not accepted”. If the accuracy test was passed but one of the other two tests was failed, the result was assigned the “Warning” status. The “Warning” status will reflect cases in which the reported result was close enough to the assigned property value, but its associated uncertainty was deemed to be either too small or too large.

TABLE 2. PERFORMANCE EVALUATION CRITERIA

Accuracy	Precision	Trueness	Final evaluation
Pass	Pass	Pass	Accepted
Pass	Fail	Pass	Warning
Pass	Pass	Fail	Warning
Fail	Pass/Fail	Pass/Fail	Not accepted

These evaluation criteria can also be illustrated by plotting the relative bias against the relative uncertainty of the participants' result (see Figure 1). In the illustrated case, a relative uncertainty of 1% for the IAEA value is assumed with MARB and LAP limits of 20%.




FIG. 1. Visualisation of performance evaluation criteria.

The plot consists of six zones (Zone 1 “Accepted”; Zones 2, 3 and 4 “Warning”; Zones 5 and 6 “Not accepted”), whose areas are defined by the three tests used above to evaluate the data. The areas of Zones 1, 3 and 4 are finite, while the areas of Zones 2, 5 and 6 are infinite. A result located in Zone 1 passes all three tests (evaluation “Accepted”). A result located in Zone 2 fails only the precision test as its associated uncertainty is deemed to be too large (evaluation: “Warning”). A result located in Zones 3 and 4 fails only the trueness test as its associated uncertainty is deemed to be too small (evaluation: “Warning”). A result located in Zones 5 and 6 fails (at least) the accuracy test as its relative bias is larger than the MARB (evaluation: “Not accepted”).

3.4. COMPARISON OF PARTICIPANTS' VALUES WITH THE IAEA VALUES

The means and the uncertainties for the combined participants' results were calculated according to a method developed by Cox (i.e. the weighted mean of the largest consistent subset containing p results)¹ and subsequently compared with the IAEA values. The mean of the combined participants' results, $\text{Value}_{\text{Combined}}$, was tested against the IAEA value, $\text{Value}_{\text{IAEA}}$, using this equation:

$$t = \frac{\text{Value}_{\text{Combined}} - \text{Value}_{\text{IAEA}}}{\sqrt{(\text{unc}_{\text{Combined}})^2 + (\text{unc}_{\text{IAEA}})^2}} \quad (6)$$

The effective degrees of freedom v_{eff} were determined with the Welch-Satterthwaite equation. The effective degrees of freedom v_{eff} were rounded and the critical value t_{crit} for this value was identified.

¹ Cox, M.G., 2007. The evaluation of key comparison data: determining the largest consistent subset. Metrologia 44 187-200

The criterion for passing the t test was:

$$|t| < t_{crit} \quad (7)$$

If the absolute value of t was greater than the critical value t_{crit} , this indicates there was a significant difference between the combined participants' results and the IAEA value.

4. RESULTS AND DISCUSSION

In total 90 measurement results were reported by 30 laboratories from 15 countries for ^3H , ^{90}Sr , ^{134}Cs and ^{137}Cs . The evaluation of these results showed that 74% of all reported results were "Accepted", while 13% of the individual measurement results were "Not accepted" with the remaining 13% having the "Warning" status. The performance evaluation for ^3H , ^{90}Sr , ^{134}Cs and ^{137}Cs is summarised in Table 3. The subset of the 12 Japanese participants' results showed that 86% of all reported results were "Accepted", while 3% of the results were "Not accepted", with the remaining 11% having the "Warning" status.

TABLE 3. SUMMARY EVALUATION OF THE RADIONUCLIDES REPORTED

Radionuclide	Number of submitted results	Accepted	Warning	Not accepted
^3H	17 (7)	12 (7)	1 (0)	4 (0)
^{90}Sr	17 (6)	12 (5)	4 (1)	1 (0)
^{134}Cs	27 (11)	20 (9)	3 (2)	4 (0)
^{137}Cs	29 (11)	22 (9)	4 (1)	3 (1)

NOTE: The numbers in brackets represent the subset of Japanese results

For ^{134}Cs there was a significant negative bias of the combined participants' results (-9.4%), while there was no significant bias for ^3H (-0.3%), ^{90}Sr (-0.4%) and ^{137}Cs (1.2%) (see Table 4). For ^3H , ^{90}Sr and ^{137}Cs about half of the participants' results showed a negative bias (53%, 47% and 48% of the results, respectively), while for ^{134}Cs the 19 of participants' results (70% of the results) showed a negative bias. Coincidence summing is a problem for ^{134}Cs as it leads to signal loss and hence underestimation of the activity levels for this radionuclide. It is clear from the results that some participants did not make a sufficient correction for coincidence summing. For the combined Japanese results, there were no significant relative differences with the IAEA values for all four radionuclides. Table 5 summarises the data evaluation sorted by laboratory code.

TABLE 4. COMPARISON OF THE COMBINED PARTICIPANTS' RESULTS WITH THE IAEA VALUE

Nuclide	Combined participants' result (Bq kg^{-1})	IAEA value (Bq kg^{-1})	Size of the LCS* (%)	t-value	Critical t-value	Bias (%)
^3H	2.80 ± 0.03	2.807 ± 0.010	14 (82%)	-0.25	2.12	-0.3
^{90}Sr	0.347 ± 0.006	0.3486 ± 0.0017	13 (76%)	-0.20	2.16	-0.4
^{134}Cs	0.1055 ± 0.0013	0.1165 ± 0.0005	26 (96%)	-7.81	2.04	-9.4
^{137}Cs	0.3128 ± 0.0024	0.3092 ± 0.0013	26 (90%)	1.30	2.02	1.2

*Size of the LCS (largest consistent subset) denotes the percentage of the returned results contributing to combined participants' results

TABLE 5. SUMMARY EVALUATION

Lab Code	^3H	^{90}Sr	^{134}Cs	^{137}Cs
1*	A	A	A	A
2*		A	A	A
3*	A			
4*			A	A
5*	A	A	A	W
6*			A	NA
7*	A	A	A	A
8*			A	A
9*			A	A
10*	A	A	A	A
11*	A	W	W	A
12*	A		W	A
13	NA		NA	A
14		NA		W
15		A	A	A
16	A		A	A
17	NA		A	A
18			W	A
19		A	A	A
20				NA
21			A	A
22	NA	A	NA	W
23	A	A	NA	NA
24			A	A
25		W	NA	W
26	NA	W	A	A
27	W	A	A	A
28	A	A	A	A
29	A	A	A	A
30	A	W	A	A

* Participant from Japan

The performance evaluation sorted by radionuclide and the bias plots are presented in Appendix I (see Tables 6–9 and Figures 2–9). In the bias plots, the ‘Accepted’ results are represented by dark blue points. ‘Warning’ and ‘Not accepted’ results are represented by the yellow and red points, respectively. The error bars represent the standard uncertainties of the bias (with a coverage factor of $k = 1$). The dotted lines represent a relative bias of $\pm 25\%$ or $\pm 20\%$ (Maximum Accepted Bias for ${}^3\text{H}/{}^{90}\text{Sr}$ and ${}^{134}\text{Cs}/{}^{137}\text{Cs}$, respectively). The performance evaluation sorted by laboratory code is presented in Appendix II. All laboratories reported their values with standard uncertainties ($k = 1$). However, a large spread in the submitted uncertainties was observed with relative uncertainties (at $k = 1$) ranging from 1.1% to 31% for ${}^3\text{H}$, from 1.2% to 21% for ${}^{90}\text{Sr}$, from 1.1% to 38% for ${}^{134}\text{Cs}$ and from 0.6% to 38% for ${}^{137}\text{Cs}$. However, most reported relative uncertainties were in the range of 4% – 10% for ${}^{90}\text{Sr}$ and ${}^{137}\text{Cs}$ and 4% – 15% for ${}^3\text{H}$ and ${}^{134}\text{Cs}$.

For ${}^3\text{H}$, practically all participants performed a distillation of the seawater samples, followed by liquid scintillation counting (LSC). Two participants (3 and 16) performed electrolytic enrichment after distillation (and obtained excellent results with small uncertainties of less than 2% at $k = 1$). The direct measurements after distillation yielded slightly larger uncertainties between 4% – 15%.

For ${}^{90}\text{Sr}$, all participants used gas-flow proportional counting (GPC) of chemically separated ${}^{90}\text{Y}$ as the analysis technique, except Participants 26 and 27 who used LSC. Chemical separation techniques used included precipitations of Sr-oxalate or Sr-carbonate, barium chromate precipitations, nitric acid precipitations, cation-exchange chromatography, Sr-extraction chromatography and direct liquid-liquid extraction of ${}^{90}\text{Y}$ with HDEHP. No significant difference between the performances of the chemical separation techniques is apparent.

For ${}^{134}\text{Cs}$ and ${}^{137}\text{Cs}$, a large majority of the participants measured the seawater samples directly with gamma spectrometry, while a small minority of laboratories either used a pre-concentration technique (e.g. adsorption on either ammonium molybdate or hexacyanoferates containing either copper, nickel or cobalt) to separate the caesium radionuclides from the seawater matrix followed by gamma spectrometry (Participants 1, 5, 18 and 23) or they used a combination of the two techniques (Participant 19). No significant difference in the performance between the two techniques is apparent although in general the reported relative uncertainties (at $k = 1$) for the direct measurements were somewhat larger, typically between 5% – 15% for the direct measurements and between 2% – 8% for the pre-concentration methods.

APPENDIX I. PERFORMANCE EVALUATION TABLES SORTED BY RADIONUCLIDE

TABLE 6. EVALUATION RESULTS FOR ^{3}H

Lab Code	Lab Value (Bq kg $^{-1}$)	Lab uncert. (Bq kg $^{-1}$)	IAEA Value (Bq kg $^{-1}$)	IAEA uncert. (Bq kg $^{-1}$)	Rel. bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
1*	2.83	0.16	2.82	0.06	0.5	6.0	16	Pass	Pass	Pass	Accepted
2*	< 1.6	—	2.79	0.06	—	—	—	—	—	—	Not evaluated
3*	2.780	0.031	2.80	0.06	-0.9	2.3	6.0	Pass	Pass	Pass	Accepted
4*	—	—	2.81	0.06	—	—	—	—	—	—	Not reported
5*	3.04	0.45	2.79	0.06	9.0	15	41	Pass	Pass	Pass	Accepted
6*	—	—	2.80	0.06	—	—	—	—	—	—	Not reported
7*	2.70	0.20	2.80	0.06	-3.6	7.7	19	Pass	Pass	Pass	Accepted
8*	—	—	2.81	0.06	—	—	—	—	—	—	Not reported
9*	—	—	2.80	0.06	—	—	—	—	—	—	Not reported
10*	2.43	0.16	2.79	0.06	-13	6.9	16	Pass	Pass	Pass	Accepted
11*	2.90	0.17	2.82	0.06	3.0	6.2	16	Pass	Pass	Pass	Accepted
12*	2.90	0.29	2.81	0.06	3.2	10	27	Pass	Pass	Pass	Accepted
13	6.32	0.56	2.82	0.06	124	9.1	51	Fail	Pass	Fail	Not accepted
14	—	—	2.82	0.06	—	—	—	—	—	—	Not reported
15	< 5.35	—	2.79	0.06	—	—	—	—	—	—	Not evaluated
16	2.910	0.055	2.80	0.06	3.8	2.8	7.3	Pass	Pass	Pass	Accepted

continues

continued

Lab Code	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Rel. bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
17	1.75	0.45	2.81	0.06	-38	26	42	Fail	Fail	Pass	Not accepted
18	-	-	2.78	0.06	-	-	-	-	-	-	Not reported
19	-	-	2.79	0.06	-	-	-	-	-	-	Not reported
20	-	-	2.79	0.06	-	-	-	-	-	-	Not reported
21	-	-	2.82	0.06	-	-	-	-	-	-	Not reported
22	2.03	0.31	2.81	0.06	-28	15	29	Fail	Pass	Pass	Not accepted
23	2.9	0.4	2.81	0.06	3.1	14	37	Pass	Pass	Pass	Accepted
24	-	-	2.80	0.06	-	-	-	-	-	-	Not reported
25	-	-	2.80	0.06	-	-	-	-	-	-	Not reported
26	1.71	0.15	2.81	0.06	-39	9.0	15	Fail	Pass	Fail	Not accepted
27	2.6	0.8	2.83	0.06	-8.0	31	73	Pass	Fail	Pass	Warning
28	2.85	0.12	2.79	0.06	2.0	4.7	12	Pass	Pass	Pass	Accepted
29	2.2	0.3	2.81	0.06	-22	14	28	Pass	Pass	Pass	Accepted
30	2.72	0.11	2.81	0.06	-3.2	4.5	11	Pass	Pass	Pass	Accepted

TABLE 7. EVALUATION RESULTS FOR ^{90}Sr

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
1*	0.305	0.025	0.3502	0.0021	-13	8.2	18	Pass	Pass	Pass	Accepted
2*	0.37	0.02	0.3472	0.0021	6.6	5.4	15	Pass	Pass	Pass	Accepted
3*	-	-	0.3484	0.0021	-	-	-	-	-	-	Not reported
4*	-	-	0.3492	0.0021	-	-	-	-	-	-	Not reported
5*	0.358	0.020	0.3463	0.0021	3.4	5.6	15	Pass	Pass	Pass	Accepted
6*	-	-	0.3475	0.0021	-	-	-	-	-	-	Not reported
7*	0.343	0.019	0.3481	0.0021	-1.5	5.6	14	Pass	Pass	Pass	Accepted
8*	-	-	0.3489	0.0021	-	-	-	-	-	-	Not reported
9*	-	-	0.3480	0.0021	-	-	-	-	-	-	Not reported
10*	0.345	0.029	0.3470	0.0021	-0.6	8.4	22	Pass	Pass	Pass	Accepted
11*	0.3000	0.0037	0.3500	0.0021	-14	1.4	3.1	Pass	Pass	Fail	Warning
12*	-	-	0.3492	0.0021	-	-	-	-	-	-	Not reported
13	-	-	0.3502	0.0021	-	-	-	-	-	-	Not reported
14	0.52	0.08	0.3509	0.0021	48	15	59	Fail	Pass	Pass	Not accepted
15	0.34	0.02	0.3465	0.0021	-1.9	5.9	15	Pass	Pass	Pass	Accepted
16	-	-	0.3486	0.0021	-	-	-	-	-	-	Not reported
17	-	-	0.3486	0.0021	-	-	-	-	-	-	Not reported
18	-	-	0.3453	0.0021	-	-	-	-	-	-	Not reported

continues

continued

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
19	0.296	0.023	0.3468	0.0021	-15	7.8	17	Pass	Pass	Pass	Accepted
20	-	-	0.3468	0.0021	-	-	-	-	-	-	Not reported
21	-	-	0.3500	0.0021	-	-	-	-	-	-	Not reported
22	0.38	0.08	0.3490	0.0021	8.9	21	59	Pass	Pass	Pass	Accepted
23	0.38	0.04	0.3498	0.0021	9.8	10	28	Pass	Pass	Pass	Accepted
24	-	-	0.3485	0.0021	-	-	-	-	-	-	Not reported
25	0.390	0.011	0.3475	0.0021	12	2.9	8.3	Pass	Pass	Fail	Warning
26	0.386	0.010	0.3492	0.0021	11	2.5	7.2	Pass	Pass	Fail	Warning
27	0.36	0.02	0.3515	0.0021	2.4	5.6	15	Pass	Pass	Pass	Accepted
28	0.359	0.016	0.3470	0.0021	3.5	4.5	12	Pass	Pass	Pass	Accepted
29	0.342	0.016	0.3499	0.0021	-2.3	4.7	12	Pass	Pass	Pass	Accepted
30	0.310	0.012	0.3493	0.0021	-11	3.9	9.0	Pass	Pass	Fail	Warning

TABLE 8. EVALUATION RESULTS FOR ^{134}Cs

Lab Code	Lab Value (Bq kg $^{-1}$)	Lab uncert. (Bq kg $^{-1}$)	IAEA Value (Bq kg $^{-1}$)	IAEA uncert. (Bq kg $^{-1}$)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
1*	0.113	0.009	0.1171	0.0005	-3.5	8.0	20	Pass	Pass	Pass	Accepted
2*	0.114	0.013	0.1161	0.0005	-1.8	11	29	Pass	Pass	Pass	Accepted
3*	-	-	0.1165	0.0005	-	-	-	-	-	-	Not reported
4*	0.127	0.013	0.1167	0.0005	8.8	10	29	Pass	Pass	Pass	Accepted
5*	0.104	0.005	0.1158	0.0005	-10	5.2	12	Pass	Pass	Pass	Accepted
6*	0.135	0.019	0.1162	0.0005	16	14	42	Pass	Pass	Pass	Accepted
7*	0.131	0.026	0.1164	0.0005	13	20	58	Pass	Pass	Pass	Accepted
8*	0.137	0.022	0.1166	0.0005	17	16	49	Pass	Pass	Pass	Accepted
9*	0.106	0.018	0.1163	0.0005	-8.9	17	40	Pass	Pass	Pass	Accepted
10*	0.116	0.016	0.1160	0.0005	0.0	14	36	Pass	Pass	Pass	Accepted
11*	0.110	0.029	0.1170	0.0005	-6.0	26	64	Pass	Fail	Pass	Warning
12*	0.10	0.03	0.1168	0.0005	-14	30	66	Pass	Fail	Pass	Warning
13	0.093	0.004	0.1171	0.0005	-21	4.3	8.9	Fail	Pass	Fail	Not accepted
14	<0.3	-	0.1173	0.0005	-	-	-	-	-	-	Not evaluated
15	0.107	0.006	0.1159	0.0005	-7.6	5.2	12	Pass	Pass	Pass	Accepted
16	0.118	0.008	0.1165	0.0005	1.3	6.8	18	Pass	Pass	Pass	Accepted
17	0.110	0.016	0.1166	0.0005	-6.1	15	36	Pass	Pass	Pass	Accepted
18	0.105	0.002	0.1154	0.0005	-9.1	1.9	4.6	Pass	Pass	Fail	Warning

continues

continued

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
19	0.106	0.005	0.1160	0.0005	-8.6	4.7	11	Pass	Pass	Pass	Accepted
20	-	-	0.1159	0.0005	-	-	-	-	-	-	Not reported
21	0.117	0.013	0.1170	0.0005	0.0	11	29	Pass	Pass	Pass	Accepted
22	0.071	0.027	0.1167	0.0005	-39	38	60	Fail	Fail	Pass	Not accepted
23	0.093	0.007	0.1169	0.0005	-20	7.5	15	Fail	Pass	Fail	Not accepted
24	0.120	0.009	0.1165	0.0005	3.0	7.5	20	Pass	Pass	Pass	Accepted
25	0.1590	0.0017	0.1162	0.0005	37	1.1	3.9	Fail	Pass	Fail	Not accepted
26	0.098	0.010	0.1167	0.0005	-16	10	22	Pass	Pass	Pass	Accepted
27	0.107	0.010	0.1175	0.0005	-9.0	9.4	22	Pass	Pass	Pass	Accepted
28	0.119	0.009	0.1160	0.0005	2.6	7.6	20	Pass	Pass	Pass	Accepted
29	0.116	0.011	0.1170	0.0005	-0.8	9.5	24	Pass	Pass	Pass	Accepted
30	0.110	0.008	0.1168	0.0005	-5.8	7.3	18	Pass	Pass	Pass	Accepted

TABLE 9. EVALUATION RESULTS FOR ^{137}Cs

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
1*	0.326	0.015	0.3107	0.0019	4.9	4.6	13	Pass	Pass	Pass	Accepted
2*	0.304	0.017	0.3080	0.0019	-1.3	5.6	14	Pass	Pass	Pass	Accepted
3*	-	-	0.3091	0.0019	-	-	-	-	-	-	Not reported
4*	0.312	0.016	0.3098	0.0019	0.7	5.1	13	Pass	Pass	Pass	Accepted
5*	0.279	0.010	0.3072	0.0019	-9.2	3.6	8.4	Pass	Pass	Fail	Warning
6*	0.372	0.030	0.3083	0.0019	21	8.1	25	Fail	Pass	Pass	Not accepted
7*	0.322	0.033	0.3088	0.0019	4.3	10	28	Pass	Pass	Pass	Accepted
8*	0.286	0.024	0.3095	0.0019	-7.6	8.3	20	Pass	Pass	Pass	Accepted
9*	0.321	0.024	0.3087	0.0019	4.0	7.5	20	Pass	Pass	Pass	Accepted
10*	0.297	0.024	0.3079	0.0019	-3.5	8.1	20	Pass	Pass	Pass	Accepted
11*	0.290	0.037	0.3105	0.0019	-6.6	13	31	Pass	Pass	Pass	Accepted
12*	0.350	0.024	0.3098	0.0019	13	6.9	20	Pass	Pass	Pass	Accepted
13	0.281	0.022	0.3106	0.0019	-9.5	7.9	18	Pass	Pass	Pass	Accepted
14	0.29	0.11	0.3113	0.0019	-6.8	38	91	Pass	Fail	Pass	Warning
15	0.310	0.025	0.3074	0.0019	0.8	8.1	21	Pass	Pass	Pass	Accepted
16	0.275	0.021	0.3092	0.0019	-11	7.7	18	Pass	Pass	Pass	Accepted
17	0.327	0.017	0.3093	0.0019	5.6	5.3	14	Pass	Pass	Pass	Accepted
18	0.295	0.006	0.3064	0.0019	-3.7	2.1	5.3	Pass	Pass	Pass	Accepted

continues

continued

Lab Code	Lab Value (Bq kg^{-1})	Lab uncert. (Bq kg^{-1})	IAEA Value (Bq kg^{-1})	IAEA uncert. (Bq kg^{-1})	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
19	0.309	0.014	0.3077	0.0019	0.4	4.6	12	Pass	Pass	Pass	Accepted
20	0.23	0.02	0.3077	0.0019	-25	8.7	17	Fail	Pass	Fail	Not accepted
21	0.309	0.014	0.3105	0.0019	-0.5	4.5	11	Pass	Pass	Pass	Accepted
22	0.27	0.06	0.3097	0.0019	-13	22	50	Pass	Fail	Pass	Warning
23	0.247	0.018	0.3103	0.0019	-20	7.3	15	Fail	Pass	Fail	Not accepted
24	0.330	0.024	0.3091	0.0019	6.7	7.3	20	Pass	Pass	Pass	Accepted
25	0.318	0.002	0.3083	0.0019	3.1	0.9	2.3	Pass	Pass	Fail	Warning
26	0.322	0.018	0.3098	0.0019	3.8	5.5	15	Pass	Pass	Pass	Accepted
27	0.298	0.014	0.3118	0.0019	-4.4	4.7	12	Pass	Pass	Pass	Accepted
28	0.330	0.021	0.3078	0.0019	7.2	6.4	18	Pass	Pass	Pass	Accepted
29	0.333	0.025	0.3104	0.0019	7.3	7.5	21	Pass	Pass	Pass	Accepted
30	0.340	0.017	0.3099	0.0019	9.7	5.0	14	Pass	Pass	Pass	Accepted




FIG. 2. $H-3$ bias.




FIG. 3. Sr-90 bias.




FIG. 4. Cs-134 bias.




FIG. 5. Cs-137 bias.




FIG. 6. H-3 bias Japanese participants.




FIG. 7. Sr-90 bias Japanese participants.




FIG. 8. Cs-134 bias Japanese participants.




FIG. 9. Cs-137 bias Japanese participants.

APPENDIX II. PERFORMANCE EVALUATION TABLES SORTED BY LABORATORY CODE

LABORATORY CODE 1*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.83	0.16	2.82	0.06	0.5	6.0	16	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.305	0.025	0.3502	0.0021	-13	8.2	18	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.113	0.009	0.1171	0.0005	-3.5	8.0	20	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.326	0.015	0.3107	0.0019	4.9	4.6	13	Pass	Pass	Pass	Accepted

LABORATORY CODE 2*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	< 1.6	—	2.79	0.06	—	—	—	—	—	—	Not evaluated
⁹⁰ Sr	0.37	0.02	0.3472	0.0021	6.6	5.4	15	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.114	0.013	0.1161	0.0005	-1.8	11	29	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.304	0.017	0.3080	0.0019	-1.3	5.6	14	Pass	Pass	Pass	Accepted

LABORATORY CODE 3*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.780	0.031	2.80	0.06	-0.9	2.3	6.0	Pass	Pass	Pass	Accepted
⁹⁰ Sr	—	—	0.3484	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	—	—	0.1165	0.0005	—	—	—	—	—	—	Not reported
¹³⁷ Cs	—	—	0.3091	0.0019	—	—	—	—	—	—	Not reported

LABORATORY CODE 4*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.81	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	—	—	0.3492	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.127	0.013	0.1167	0.0005	8.8	10	29	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.312	0.016	0.3098	0.0019	0.7	5.1	13	Pass	Pass	Pass	Accepted

LABORATORY CODE 5*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	3.04	0.45	2.79	0.06	9.0	15	41	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.358	0.020	0.3463	0.0021	3.4	5.6	15	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.104	0.005	0.1158	0.0005	-10	5.2	12	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.279	0.010	0.3072	0.0019	-9.2	3.6	8.4	Pass	Pass	Fail	Warning

LABORATORY CODE 6*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.80	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	—	—	0.3475	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.135	0.019	0.1162	0.0005	16	14	42	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.372	0.030	0.3083	0.0019	21	8.1	25	Fail	Pass	Pass	Not accepted

LABORATORY CODE 7*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.70	0.20	2.80	0.06	-3.6	7.7	19	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.343	0.019	0.3481	0.0021	-1.5	5.6	14	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.131	0.026	0.1164	0.0005	13	20	58	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.322	0.033	0.3088	0.0019	4.3	10	28	Pass	Pass	Pass	Accepted

LABORATORY CODE 8*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.81	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	—	—	0.3489	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.137	0.022	0.1166	0.0005	17	16	49	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.286	0.024	0.3095	0.0019	-7.6	8.3	20	Pass	Pass	Pass	Accepted

LABORATORY CODE 9*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.80	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	—	—	0.3480	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.106	0.018	0.1163	0.0005	-8.9	17	40	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.321	0.024	0.3087	0.0019	4.0	7.5	20	Pass	Pass	Pass	Accepted

LABORATORY CODE 10*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.43	0.16	2.79	0.06	-13	6.9	16	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.345	0.029	0.3470	0.0021	-0.6	8.4	22	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.116	0.016	0.1160	0.0005	0.0	14	36	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.297	0.024	0.3079	0.0019	-3.5	8.1	20	Pass	Pass	Pass	Accepted

LABORATORY CODE 11*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.90	0.17	2.82	0.06	3.0	6.2	16	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.3000	0.0037	0.3500	0.0021	-14	1.4	3.1	Pass	Pass	Fail	Warning
¹³⁴ Cs	0.110	0.029	0.1170	0.0005	-6.0	26	64	Pass	Fail	Pass	Warning
¹³⁷ Cs	0.290	0.037	0.3105	0.0019	-6.6	13	31	Pass	Pass	Pass	Accepted

LABORATORY CODE 12*.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.90	0.29	2.81	0.06	3.2	10	27	Pass	Pass	Pass	Accepted
⁹⁰ Sr	—	—	0.3492	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.10	0.03	0.1168	0.0005	-14	30	66	Pass	Fail	Pass	Warning
¹³⁷ Cs	0.350	0.024	0.3098	0.0019	13	6.9	20	Pass	Pass	Pass	Accepted

LABORATORY CODE 13.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	6.32	0.56	2.82	0.06	124	9.1	51	Fail	Pass	Fail	Not accepted
⁹⁰ Sr	—	—	0.3502	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.093	0.004	0.1171	0.0005	-21	4.3	8.9	Fail	Pass	Fail	Not accepted
¹³⁷ Cs	0.281	0.022	0.3106	0.0019	-9.5	7.9	18	Pass	Pass	Pass	Accepted

LABORATORY CODE 14.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.82	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	0.52	0.08	0.3509	0.0021	48	15	59	Fail	Pass	Pass	Not accepted
¹³⁴ Cs	<0.3	—	0.1173	0.0005	—	—	—	—	—	—	Not evaluated
¹³⁷ Cs	0.29	0.11	0.3113	0.0019	-6.8	38	91	Pass	Fail	Pass	Warning

LABORATORY CODE 15.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	< 5.35	—	2.79	0.06	—	—	—	—	—	—	Not evaluated
⁹⁰ Sr	0.34	0.02	0.3465	0.0021	-1.9	5.9	15	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.107	0.006	0.1159	0.0005	-7.6	5.2	12	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.310	0.025	0.3074	0.0019	0.8	8.1	21	Pass	Pass	Pass	Accepted

LABORATORY CODE 16.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.910	0.055	2.80	0.06	3.8	2.8	7.3	Pass	Pass	Pass	Accepted
⁹⁰ Sr	—	—	0.3484	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.118	0.008	0.1165	0.0005	1.3	6.8	18	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.275	0.021	0.3092	0.0019	-11	7.7	18	Pass	Pass	Pass	Accepted

LABORATORY CODE 17.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	1.75	0.45	2.81	0.06	-38	26	42	Fail	Fail	Pass	Not accepted
⁹⁰ Sr	—	—	0.3486	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.110	0.016	0.1166	0.0005	-6.1	15	36	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.327	0.017	0.3093	0.0019	5.6	5.3	14	Pass	Pass	Pass	Accepted

LABORATORY CODE 18.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.78	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	—	—	0.3453	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.105	0.002	0.1154	0.0005	-9.1	1.9	4.6	Pass	Pass	Fail	Warning
¹³⁷ Cs	0.295	0.006	0.3064	0.0019	-3.7	2.1	5.3	Pass	Pass	Pass	Accepted

LABORATORY CODE 19.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.79	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	0.296	0.023	0.3468	0.0021	-15	7.8	17	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.106	0.005	0.1160	0.0005	-8.6	4.7	11	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.309	0.014	0.3077	0.0019	0.4	4.6	12	Pass	Pass	Pass	Accepted

LABORATORY CODE 20.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.79	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	—	—	0.3468	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	—	—	0.1159	0.0005	—	—	—	—	—	—	Not reported
¹³⁷ Cs	0.23	0.02	0.3077	0.0019	-25	8.7	17	Fail	Pass	Fail	Not accepted

LABORATORY CODE 21.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.82	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	—	—	0.3500	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.117	0.013	0.1170	0.0005	0.0	11	29	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.309	0.014	0.3105	0.0019	-0.5	4.5	11	Pass	Pass	Pass	Accepted

LABORATORY CODE 22.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.03	0.31	2.81	0.06	-28	15	29	Fail	Pass	Pass	Not accepted
⁹⁰ Sr	0.38	0.08	0.3490	0.0021	8.9	21	59	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.071	0.027	0.1167	0.0005	-39	38	60	Fail	Fail	Pass	Not accepted
¹³⁷ Cs	0.27	0.06	0.3097	0.0019	-13	22	50	Pass	Fail	Pass	Warning

LABORATORY CODE 23.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.9	0.4	2.81	0.06	3.1	14	37	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.38	0.04	0.3498	0.0021	9.8	10	28	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.093	0.007	0.1169	0.0005	-20	7.5	15	Fail	Pass	Fail	Not accepted
¹³⁷ Cs	0.247	0.018	0.3103	0.0019	-20	7.3	15	Fail	Pass	Fail	Not accepted

LABORATORY CODE 24.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.80	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	—	—	0.3485	0.0021	—	—	—	—	—	—	Not reported
¹³⁴ Cs	0.120	0.009	0.1165	0.0005	3.0	7.5	20	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.330	0.024	0.3091	0.0019	6.7	7.3	20	Pass	Pass	Pass	Accepted

LABORATORY CODE 25.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	—	—	2.80	0.06	—	—	—	—	—	—	Not reported
⁹⁰ Sr	0.390	0.011	0.3475	0.0021	12	2.9	8.3	Pass	Pass	Fail	Warning
¹³⁴ Cs	0.1590	0.0017	0.1162	0.0005	37	1.1	3.9	Fail	Pass	Fail	Not accepted
¹³⁷ Cs	0.318	0.002	0.3083	0.0019	3.1	0.9	2.3	Pass	Pass	Fail	Warning

LABORATORY CODE 26.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	1.71	0.15	2.81	0.06	-39	9.0	15	Fail	Pass	Fail	Not accepted
⁹⁰ Sr	0.386	0.010	0.3492	0.0021	11	2.5	7.2	Pass	Pass	Fail	Warning
¹³⁴ Cs	0.098	0.010	0.1167	0.0005	-16	10	22	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.322	0.018	0.3098	0.0019	3.8	5.5	15	Pass	Pass	Pass	Accepted

LABORATORY CODE 27.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.6	0.8	2.83	0.06	-8.0	31	73	Pass	Fail	Pass	Warning
⁹⁰ Sr	0.36	0.02	0.3515	0.0021	2.4	5.6	15	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.107	0.010	0.1175	0.0005	-9.0	9.4	22	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.298	0.014	0.3118	0.0019	-4.4	4.7	12	Pass	Pass	Pass	Accepted

LABORATORY CODE 28.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.85	0.12	2.79	0.06	2.0	4.7	12	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.359	0.016	0.3470	0.0021	3.5	4.5	12	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.119	0.009	0.1160	0.0005	2.6	7.6	20	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.330	0.021	0.3078	0.0019	7.2	6.4	18	Pass	Pass	Pass	Accepted

LABORATORY CODE 29.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.2	0.3	2.81	0.06	-22	14	28	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.342	0.016	0.3499	0.0021	-2.3	4.7	12	Pass	Pass	Pass	Accepted
¹³⁴ Cs	0.116	0.011	0.1170	0.0005	-0.8	9.5	24	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.333	0.025	0.3104	0.0019	7.3	7.5	21	Pass	Pass	Pass	Accepted

LABORATORY CODE 30.

Radio-nuclide	Lab Value (Bq kg ⁻¹)	Lab uncert. (Bq kg ⁻¹)	IAEA Value (Bq kg ⁻¹)	IAEA uncert. (Bq kg ⁻¹)	Relative bias (%)	P (%)	Trueness limit (%)	Accuracy	Precision	Trueness	Final Score
³ H	2.72	0.11	2.81	0.06	-3.2	4.5	11	Pass	Pass	Pass	Accepted
⁹⁰ Sr	0.310	0.012	0.3493	0.0021	-11	3.9	9.0	Pass	Pass	Fail	Warning
¹³⁴ Cs	0.110	0.008	0.1168	0.0005	-5.8	7.3	18	Pass	Pass	Pass	Accepted
¹³⁷ Cs	0.340	0.017	0.3099	0.0019	9.7	5.0	14	Pass	Pass	Pass	Accepted

APPENDIX III. LIST OF PARTICIPATING LABORATORIES

JAPAN

Akimoto, S.
Environmental Radioactivity Monitoring Center of Fukushima
8-1 Minaminakayaji, Sasakino
Fukushima-shi 960-8057

Fujikawa, T.
General Environmental Technos Co Ltd
3-1-1 Higashikuraji
Katano
Osaka 576-0061

Fujita, H.
Japan Atomic Energy Agency
4-33 Muramatsu Naka-gun
Tokai-mura, Ibaraki 319-1194

Hayakawa, K.
Fukushima Nuclear Power Branch Office
Tokyo Power Technology Ltd.
232-11 Maekawahara
Kegaya
Tomioka-cho, Futaba County 979-0622

Kato, Y.
Geo-Science Laboratory Co Ltd
1-608 Uedahonmachi
Tenpaku-ku
Nagoya-city 468-0007

Ohta, T.
Japan Chemical Analysis Center
295-3 Sanno-cho, Inage-ku
Chiba 263-0002

Seimiya, K.
Foundation for Promotion of Material Science and Technology of Japan
1-18-6 Kitami, Setagaya-ku
Tokyo 157-0067

Takasawa, K.
Environmental Affairs Division
Tokyo Power Technology Ltd.
2-3-6 Onodai
Midori-ku 267-0056

Takayama, Y.
IDEA Consultants Inc
3-15-1 Komazawa
Setagaya-ku
Tokyo 154-8585

JAPAN (continued)

Tamari, T.
Kyushu Environmental Evaluation Association
1-10-1 Matsukadai, Higashi-ku
Fukuoka 813-0004

Tomatsuri, S.
Kaken Inc
Hori 1044
Mito, Ibaraki 310-0903

Yasunami, I.
Chugai Technos Cooperation
2-2-16 Ohnodox, Midori-Ku
Chiba-city 267-0056

BELGIUM

Verheyen, L.
Unit Low-level Radioactivity Measurements
Environment Health & Safety
Belgian Nuclear Research Centre SCK CEN
Boeretang 200
2400 Mol

DENMARK

Nielsen, S.P.
Technical University of Denmark
Center for Nuclear Technologies
DTU Risø Campus Building 204
Frederiksborgvej 399
4000 Roskilde

ESTONIA

Jakobson, E.
Environmental Board
Radiation Safety Department
Kopli 76
10416 Tallinn

FRANCE

Gurriaran, R.
Laboratoire de Mesure de la Radioactivité de l'Environnement
IRSN/PRP-ENV/STEME
Bât. 501, Bois des Rames
91400 Orsay

FRANCE (continued)

Loyer, J.
Laboratoire de Traitement des Echantillons
IRSN/PRP-ENV/STEME
31 rue de l'Ecluse
78110 Le Vesinet

GERMANY

Aust, M.-O.
Thünen-Institut für Fischereiökologie
Marckmannstraße 129b
20539 Hamburg

Schmied, S.
Bundesamt für Seeschifffahrt und Hydrographie (BSH)
Wüstland 2
22589 Hamburg

IRELAND

O'Colmáin, M.
Environmental Protection Agency
Office of Radiological Protection
3 Clonskeagh Square
Clonskeagh Road
Dublin 14

NETHERLANDS

Engeler, C.
Rijkswaterstaat Centre for Water Management
Zuiderwagenplein 2
8224 AD Lelystad

NORWAY

Sværen, I.
Institute of Marine Research
P. B. 1870 Nordnes
5817 Bergen

POLAND

Zalewska, T.
Institute of Meteorology and Water Management
National Research Institute Maritime Branch
Waszyngtona 42
81-342 Gdynia

PORUGAL

Madruga, M.J.
Instituto Superior Técnico
Laboratório de Proteção e Segurança Radiológica
Campus Tecnológico e Nuclear
Estrada Nacional 10 (ao km 139,7)
2695-066 Bobadela LRS

SPAIN

Garcia Orellana, J.
Universidad Autónoma de Barcelona
Bellaterra
08193 Barcelona

García Tenorio, R.
Escuela Técnica Superior de Arquitectura
Avenida Reina Mercedes, 2
41012 Sevilla

de Pablo Sanmartín, A.
CEDEX (Mº FOMENTO)
C/ Alfonso XII 3 y 5
28014 Madrid

SWEDEN

Eriksson, S.
Swedish Radiation Safety Authority
Solna strandväg 96
17116 Stockholm

UNITED KINGDOM OF GREAT BRITAIN AND NORTHERN IRELAND

Blowers, P.
CEFAS
Pakefield Road, Pakefield
Lowestoft, Suffolk
NR33 0HT

UNITED STATES OF AMERICA

Pike, S.
Woods Hole Oceanographic Institution
266 Woods Hole Road
Woods Hole MA
02543-1050

CONTRIBUTORS TO DRAFTING AND REVIEW

Harms, A.	International Atomic Energy Agency
Osvath, I.	International Atomic Energy Agency
Osborn, D.	International Atomic Energy Agency



ORDERING LOCALLY

In the following countries, IAEA priced publications may be purchased from the sources listed below or from major local booksellers.

Orders for unpriced publications should be made directly to the IAEA. The contact details are given at the end of this list.

AUSTRALIA

DA Information Services

648 Whitehorse Road, Mitcham, VIC 3132, AUSTRALIA
Telephone: +61 3 9210 7777 • Fax: +61 3 9210 7788
Email: books@dadirect.com.au • Web site: <http://www.dadirect.com.au>

BELGIUM

Jean de Lannoy

Avenue du Roi 202, 1190 Brussels, BELGIUM
Telephone: +32 2 5384 308 • Fax: +32 2 5380 841
Email: jean.de.lannoy@euronet.be • Web site: <http://www.jean-de-lannoy.be>

CANADA

Renouf Publishing Co. Ltd.

5369 Canotek Road, Ottawa, ON K1J 9J3, CANADA
Telephone: +1 613 745 2665 • Fax: +1 643 745 7660
Email: order@renoufbooks.com • Web site: <http://www.renoufbooks.com>

Bernan Associates

4501 Forbes Blvd., Suite 200, Lanham, MD 20706-4391, USA
Telephone: +1 800 865 3457 • Fax: +1 800 865 3450
Email: orders@bernan.com • Web site: <http://www.bernan.com>

CZECH REPUBLIC

Suweco CZ, spol. S.r.o.

Klecakova 347, 180 21 Prague 9, CZECH REPUBLIC
Telephone: +420 242 459 202 • Fax: +420 242 459 203
Email: nakup@suweco.cz • Web site: <http://www.suweco.cz>

FINLAND

Akateeminen Kirjakauppa

PO Box 128 (Keskuskatu 1), 00101 Helsinki, FINLAND
Telephone: +358 9 121 41 • Fax: +358 9 121 4450
Email: akatilaus@akateeminen.com • Web site: <http://www.akateeminen.com>

FRANCE

Form-Edit

5 rue Janssen, PO Box 25, 75921 Paris CEDEX, FRANCE
Telephone: +33 1 42 01 49 49 • Fax: +33 1 42 01 90 90
Email: fabien.boucard@formedit.fr • Web site: <http://www.formedit.fr>

Lavoisier SAS

14 rue de Provigny, 94236 Cachan CEDEX, FRANCE
Telephone: +33 1 47 40 67 00 • Fax: +33 1 47 40 67 02
Email: livres@lavoisier.fr • Web site: <http://www.lavoisier.fr>

L'Appel du livre

99 rue de Charonne, 75011 Paris, FRANCE
Telephone: +33 1 43 07 50 80 • Fax: +33 1 43 07 50 80
Email: livres@appeldulivre.fr • Web site: <http://www.appeldulivre.fr>

GERMANY

Goethe Buchhandlung Teubig GmbH

Schweitzer Fachinformationen
Willstätterstrasse 15, 40549 Düsseldorf, GERMANY
Telephone: +49 (0) 211 49 8740 • Fax: +49 (0) 211 49 87428
Email: s.dehaan@schweitzer-online.de • Web site: <http://www.goethebuch.de>

HUNGARY

Librotade Ltd., Book Import

PF 126, 1656 Budapest, HUNGARY
Telephone: +36 1 257 7777 • Fax: +36 1 257 7472
Email: books@librotade.hu • Web site: <http://www.librotade.hu>

INDIA

Allied Publishers

1st Floor, Dubash House, 15, J.N. Heredi Marg, Ballard Estate, Mumbai 400001, INDIA
Telephone: +91 22 2261 7926/27 • Fax: +91 22 2261 7928
Email: alliedpl@vsnl.com • Web site: <http://www.alliedpublishers.com>

Bookwell

3/79 Nirankari, Delhi 110009, INDIA
Telephone: +91 11 2760 1283/4536
Email: bkwell@nde.vsnl.net.in • Web site: <http://www.bookwellindia.com>

ITALY

Libreria Scientifica "AEIOU"

Via Vincenzo Maria Coronelli 6, 20146 Milan, ITALY
Telephone: +39 02 48 95 45 52 • Fax: +39 02 48 95 45 48
Email: info@libreriaaeiou.eu • Web site: <http://www.libreriaaeiou.eu>

JAPAN

Maruzen Co., Ltd.

1-9-18 Kaigan, Minato-ku, Tokyo 105-0022, JAPAN
Telephone: +81 3 6367 6047 • Fax: +81 3 6367 6160
Email: journal@maruzen.co.jp • Web site: <http://maruzen.co.jp>

NETHERLANDS

Martinus Nijhoff International

Koraalrood 50, Postbus 1853, 2700 CZ Zoetermeer, NETHERLANDS
Telephone: +31 793 684 400 • Fax: +31 793 615 698
Email: info@nijhoff.nl • Web site: <http://www.nijhoff.nl>

Swets Information Services Ltd.

PO Box 26, 2300 AA Leiden
Dellaertweg 9b, 2316 WZ Leiden, NETHERLANDS
Telephone: +31 88 4679 387 • Fax: +31 88 4679 388
Email: tbeyens@nl.swets.com • Web site: <http://www.swets.com>

SLOVENIA

Cankarjeva Zalozba dd

Kopitarjeva 2, 1515 Ljubljana, SLOVENIA
Telephone: +386 1 432 31 44 • Fax: +386 1 230 14 35
Email: import.books@cankarjeva-z.si • Web site: http://www.mladinska.com/cankarjeva_zalozba

SPAIN

Diaz de Santos, S.A.

Librerías Bookshop • Departamento de pedidos
Calle Albasanz 2, esquina Hermanos García Noblejas 21, 28037 Madrid, SPAIN
Telephone: +34 917 43 48 90 • Fax: +34 917 43 4023
Email: compras@diazdesantos.es • Web site: <http://www.diazdesantos.es>

UNITED KINGDOM

The Stationery Office Ltd. (TSO)

PO Box 29, Norwich, Norfolk, NR3 1PD, UNITED KINGDOM
Telephone: +44 870 600 5552
Email (orders): books.orders@tso.co.uk • (enquiries): book.enquiries@tso.co.uk • Web site: <http://www.tso.co.uk>

UNITED STATES OF AMERICA

Bernan Associates

4501 Forbes Blvd., Suite 200, Lanham, MD 20706-4391, USA
Telephone: +1 800 865 3457 • Fax: +1 800 865 3450
Email: orders@bernan.com • Web site: <http://www.bernan.com>

Renouf Publishing Co. Ltd.

812 Proctor Avenue, Ogdensburg, NY 13669, USA
Telephone: +1 888 551 7470 • Fax: +1 888 551 7471
Email: orders@renoufbooks.com • Web site: <http://www.renoufbooks.com>

United Nations

300 East 42nd Street, IN-919J, New York, NY 1001, USA
Telephone: +1 212 963 8302 • Fax: 1 212 963 3489
Email: publications@un.org • Web site: <http://www.unp.un.org>

Orders for both priced and unpriced publications may be addressed directly to:

IAEA Publishing Section, Marketing and Sales Unit, International Atomic Energy Agency
Vienna International Centre, PO Box 100, 1400 Vienna, Austria
Telephone: +43 1 2600 22529 or 22488 • Fax: +43 1 2600 29302
Email: sales.publications@iaea.org • Web site: <http://www.iaea.org/books>

INTERNATIONAL ATOMIC ENERGY AGENCY
VIENNA
ISSN 2074-7659